



# Acceptance Manual

- Bruker NMR Product Test (NMRPT) for Avance NMR Systems with “Topspin Version 4.1.4” Software

Version 042





Copyright © by Bruker Corporation

All rights reserved. No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form, or by any means without the prior consent of the publisher. Product names used are trademarks or registered trademarks of their respective holders.

This manual was modified by

Bruker BioSpin - Docu Group

© October 01 2021: Bruker Corporation

Rheinstetten, Germany

P/N: ZUEP0102

DWG-Nr.:





<b>1 Introduction.....</b>	<b>18</b>
<b>2 Installation Qualification (IQ).....</b>	<b>22</b>
<b>3 Operational qualification (OQ) and customer training.....</b>	<b>26</b>
<b>4 NMR Product Test (NMRPT).....</b>	<b>40</b>
<b>5 NMRPT Experiments.....</b>	<b>42</b>
<b>6 Appendix.....</b>	<b>540</b>
<b>7 Contact.....</b>	<b>548</b>

1 Introduction.....	18
1.1 Purpose .....	18
1.2 Area of application .....	18
1.3 Warning / Notice Signs .....	18
1.4 Referenced documents .....	19
1.4.1 Test procedures, Test protocols and Test report (forms) .....	19
1.4.2 Installation Manuals and Service Manuals .....	19
1.4.3 User Manuals .....	20
2 Installation Qualification (IQ).....	22
2.1 General setup of the console .....	22
2.2 Firmware check .....	22
2.3 Cortab .....	22
2.4 Sample Lift and Spin calibration .....	23
2.5 Software license installed .....	23
2.6 Helium and Nitrogen level log files active .....	23
2.7 MICS installed .....	24
2.8 Special precautions if a Prodigy Cryoprobe will be installed .....	24
2.9 Customer Information .....	24
3 Operational qualification (OQ) and customer training.....	26
3.1 Basic safety precautions .....	26
3.1.1 Magnet safety .....	27
3.1.2 Handling of cryogenic liquids and magnet refilling .....	27
3.1.3 Control of nitrogen and helium level .....	29
3.1.4 Homogeneity and stability disturbance .....	29
3.1.5 Earthquake precautions .....	31
3.1.6 Refilling of the nitrogen vessel for Prodigy CryoProbes .....	31
3.2 Hardware overview .....	31
3.2.1 Console on/off operation .....	31
3.2.2 Basic operation .....	32
3.2.3 Troubleshooting .....	34
3.2.4 Backup (nmr_save, Images) .....	34
3.2.5 Introduction to IconNMR .....	34
3.2.6 AssureSST / Performance check .....	34

3.3 Optional Components .....	34
3.3.1 Sample Changer .....	35
3.3.2 MAS Controller .....	35
3.3.3 High power equipment .....	35
3.3.4 LC-NMR .....	36
3.3.5 Liquid Handler SamplePro Tube .....	36
3.3.6 Micro-Imaging .....	36
3.3.7 Diffusion .....	36
3.3.8 CryoProbe / Cryoplatform .....	36
3.3.9 BSNL (Bruker Smart Nitrogen Liquefier) .....	37
3.3.10 Additional cooling/heating units (like BCU1 / BCU2) .....	37
3.3.11 LT-MAS (Low Temperature MAS equipment) .....	37
3.3.12 Gyrotron magnet and DNP console .....	37
3.4 Acceptance and warranty .....	38
3.4.1 Explanation of warranty .....	38
3.4.2 System documentation .....	38
3.4.3 Customer support hotlines .....	38
<b>4 NMR Product Test (NMRPT).....</b>	<b>40</b>
<b>5 NMRPT Experiments.....</b>	<b>42</b>
5.1 Introduction .....	42
5.2 Experiments for High Resolution Probes (HR) .....	42
5.2.1 13C background with sample (NPT_13C_backgr_withsample).....	43
5.2.2 DEPT135 (NPT_13C_dept135).....	45
5.2.3 DEPT90 (NPT_13C_dept90).....	47
5.2.4 13C test for artifacts with 1H decoupling (NPT_13C_fullsw_dec1h).....	49
5.2.5 13C lineshape without sample rotation (NPT_13C_lineshape_nrot).....	51
5.2.6 13C lineshape with sample rotation (NPT_13C_lineshape_wrot).....	53
5.2.7 P90 13C pulse calibration (NPT_13C_p90det_astm_13c).....	55
5.2.8 13C ringing test with 1H decoupling (NPT_13C_ringing_dec1h).....	57
5.2.9 13C sensitivity (NPT_13C_sensitivity).....	59
5.2.10 13C sensitivity with 1H decoupling (NPT_13C_sensitivity_dec1h).....	61
5.2.11 15N test for artifacts (NPT_15N_fullsw_inept).....	63
5.2.12 P90 15N pulse calibration (NPT_15N_p90det_formamide_15n).....	65

5.2.13 15N sensitivity with 1H decoupling (NPT_15N_sensitivity_dec1h).....	67
5.2.14 15N sensitivity (INEPT) with 1H decoupling (NPT_15N_sensitivity_inept).....	69
5.2.15 19F B1 homogeneity integral (NPT_19F_b1homogeneityInt_19f).....	71
5.2.16 19F B1 homogeneity integral on H-coil (NPT_19F_b1homogeneityInt_hcoil).....	73
5.2.17 19F background with sample and 1H decoupling (NPT_19F_backgr_withsample).....	75
5.2.18 19F background with sample on H-coil (NPT_19F_backgr_withsample_hcoil).....	77
5.2.19 CPD 19F pulse calibration (NPT_19F_cpddeterminationf1_19f).....	79
5.2.20 19F test for artifacts (NPT_19F_fullsw_dec1h).....	81
5.2.21 P90 19F pulse calibration (NPT_19F_p90determinationf1_19f).....	83
5.2.22 19F P90 pulse calibration on H-coil (NPT_19F_p90determinationf1_hcoil).....	85
5.2.23 19F sensitivity (NPT_19F_sensitivity).....	87
5.2.24 19F sensitivity on H-coil (NPT_19F_sensitivity_hcoil).....	89
5.2.25 19F sensitivity with 1H decoupling and LB=0.5 (NPT_19F_sensitivity_lb05_dec1h).....	91
5.2.26 13 degree pulse stability test (NPT_1H_13degtest).....	93
5.2.27 30 degree pulse stability test (NPT_1H_30degtest).....	95
5.2.28 90 degree pulse stability test (NPT_1H_90degtest).....	97
5.2.29 Amplitude stability after gradient echo with strong gradient pulses (NPT_1H_ampStabGradientEchoStrong).....	99
5.2.30 Amplitude stability after gradient echo with weak gradient pulses (NPT_1H_ampStabGradientEchoWeak).....	101
5.2.31 B0 magnet drift experiment (NPT_1H_b0drifttest).....	103
5.2.32 13C B1 homogeneity integral (NPT_1H_b1homogeneityInt_13c).....	105
5.2.33 15N B1 homogeneity integral (NPT_1H_b1homogeneityInt_15n).....	107
5.2.34 1H B1 homogeneity integral (NPT_1H_b1homogeneityInt_1h).....	109
5.2.35 1H background without sample (NPT_1H_backgr_nosample).....	111
5.2.36 1H background with sample (NPT_1H_backgr_withsample).....	113
5.2.37 2D COSY (NPT_1H_cosydfphpr).....	115
5.2.38 CPD 1H pulse calibration (NPT_1H_cpddeterminationf1_1h).....	117
5.2.39 Indirect CPD 13C pulse calibration (NPT_1H_cpddeterminationf2_13c).....	119
5.2.40 Indirect CPD 15N pulse calibration (NPT_1H_cpddeterminationf2_15n).....	121
5.2.41 13C CPMG test (NPT_1H_cpmgtestf2_13c).....	123
5.2.42 15N CPMG test (NPT_1H_cpmgtestf2_15n).....	125
5.2.43 13C decoupler profile Chirp (NPT_1H_decProfile_chirp_13c).....	127
5.2.44 13C decoupler profile Garp (NPT_1H_decProfile_garp_13c).....	129

5.2.45 Low Current Diffusion Test for Z-direction (NPT_1H_diffusionLowCurrentZ).....	131
5.2.46 High Current Diffusion Test for Z-direction (NPT_1H_diffusionHighCurrentZ).....	133
5.2.47 13C decoupler profile Waltz (NPT_1H_decProfile_waltz_13c).....	135
5.2.48 1H detection with 13C garp decoupling (NPT_1H_garpdectestf2_13c).....	137
5.2.49 1H detection with 13C hard pulse and 13C garp decoupling (NPT_1H_garp_pulse13c_dec13c).....	139
5.2.50 1H detection with 13C and 15N hard pulses and 13C garp decoupling (NPT_1H_garp_simpul13c15n_dec13c).....	141
5.2.51 13C detection with 13C and 15N hard pulses and 15N and 1H garp decoupling (NPT_13C_garp_pulses13c15n_dec15n1h).....	143
5.2.52 15N detection with 15N and 13C hard pulses and 13C and 1H garp decoupling (NPT_15N_garp_pulses15n13c_dec13c1h).....	145
5.2.53 1H Z-gradient profile [-] (NPT_1H_gradientprofile_neg).....	147
5.2.54 1H Z-gradient profile [+] (NPT_1H_gradientprofile_pos).....	149
5.2.55 1H X-gradient profile [-] (NPT_1H_gradprofX_neg).....	151
5.2.56 1H X-gradient profile [+] (NPT_1H_gradprofX_pos).....	153
5.2.57 1H Y-gradient profile [-] (NPT_1H_gradprofY_neg).....	155
5.2.58 1H Y-gradient profile [+] (NPT_1H_gradprofY_pos).....	157
5.2.59 1H Z-gradient profile [-] (NPT_1H_gradprofZ_neg).....	159
5.2.60 1H Z-gradient profile [+] (NPT_1H_gradprofZ_pos).....	161
5.2.61 Gradient recovery stability test (NPT_1H_gradrec_stest_1h).....	163
5.2.62 Gradient recovery test for X-direction [-] (NPT_1H_gradrecX_sqn_1h).....	165
5.2.63 Gradient recovery test for X-direction [+] (NPT_1H_gradrecX_sqp_1h).....	167
5.2.64 Gradient recovery test for Y-direction [-] (NPT_1H_gradrecY_sqn_1h).....	169
5.2.65 Gradient recovery test for Y-direction [+] (NPT_1H_gradrecY_sqp_1h).....	171
5.2.66 Gradient recovery test for Z-direction [-] (NPT_1H_gradrecZ_sqn_1h).....	173
5.2.67 Gradient recovery test for Z-direction [+] (NPT_1H_gradrecZ_sqp_1h).....	175
5.2.68 Gradient recovery test for X-direction [-] with trapezoid pulses (NPT_1H_gradrecX_trapezoid_neg_1h).....	177
5.2.69 Gradient recovery test for X-direction [+] with trapezoid pulses (NPT_1H_gradrecX_trapezoid_pos_1h).....	179
5.2.70 Gradient recovery test for Y-direction [-] with trapezoid pulses (NPT_1H_gradrecY_trapezoid_neg_1h).....	181
5.2.71 Gradient recovery test for Y-direction [+] with trapezoid pulses (NPT_1H_gradrecY_trapezoid_pos_1h).....	183

5.2.72 Gradient recovery test for Z-direction [-] with trapezoid pulses (NPT_1H_gradrecZ_trapezoid_neg_1h).....	185
5.2.73 Gradient recovery test for Z-direction [+] with trapezoid pulses (NPT_1H_gradrecZ_trapezoid_pos_1h).....	187
5.2.74 Inverse spin echo difference (NPT_1H_hmqc1df2_13c).....	189
5.2.75 Inverse spin echo difference experiment [2D] (NPT_1H_hmqc2df2_13c).....	191
5.2.76 1H of pseudo honey with noesy for solvent suppression (NPT_1H_honeyNoesy).....	193
5.2.77 1H homodecoupling (NPT_1H_homodec).....	195
5.2.78 2D 1H-13C HSQC (NPT_1H_hsqc_10EB_13c_2d).....	197
5.2.79 2D 1H-13C HSQC (NPT_1H_hsqc_etsisp_13c_2d).....	199
5.2.80 2D 1H-13C HSQC with adiabatic 13C decoupling (NPT_1H_hsqc_etsisp_adia13c_2d).	201
5.2.81 1H integral sensitivity (NPT_1H_inno).....	203
5.2.82 Linearity test with constant flip angle (NPT_1H_linearityConstFlipAngle).....	205
5.2.83 Linearity test with decreasing power (NPT_1H_linearityDecreasingPower).....	207
5.2.84 1H lineshape without sample rotation (NPT_1H_lineshape_nrot).....	209
5.2.85 1H lineshape with sample rotation and NS = 4 (NPT_1H_lineshape_wrot).....	211
5.2.86 1H lineshape with sample rotation and NS = 1 (NPT_1H_lineshape_wrot_ns1).....	213
5.2.87 1H lineshape stability test (NPT_1H_lineshapeStability).....	215
5.2.88 2D NOESY (NPT_1H_noesyphpr).....	217
5.2.89 P90 1H pulse calibration 0.5M NaCl (NPT_1H_p90_05M_NaCl_1h).....	219
5.2.90 P90 1H pulse calibration acetone (NPT_1H_p90_acetone_1h).....	221
5.2.91 P90 1H pulse calibration (NPT_1H_p90det_astm_1h).....	223
5.2.92 P90 1H pulse calibration (NPT_1H_p90determinationf1_1h).....	225
5.2.93 Indirect P90 13C pulse calibration (NPT_1H_p90determinationf2_13c).....	227
5.2.94 Indirect P90 15N pulse calibration (NPT_1H_p90determinationf2_15n).....	229
5.2.95 Phase propagation test (NPT_1H_phase_propagation).....	231
5.2.96 Phase shifting test (NPT_1H_phase_shifting).....	233
5.2.97 Phase cycle cancelation (NPT_1H_phaseCycleCancelation).....	235
5.2.98 Phase cycle cancelation after gradient pulse (NPT_1H_phaseCycleCancelationGrad)...	237
5.2.99 1H quantification reference (NPT_1H_quant_ref).....	239
5.2.100 1H low flipangle single scan experiment (NPT_1H_rd).....	241
5.2.101 1H selective excitation (NPT_1H_selex).....	243
5.2.102 1H sensitivity (NPT_1H_sensitivity).....	245
5.2.103 Triple resonance (NPT_1H_sensitivity_dec13c15n).....	247

5.2.104 1H sensitivity with 19F GARP decoupling (NPT_1H_sensitivity_dec19f).....	249
5.2.105 1H sensitivity with HSQC selection and 13C garp decoupling (NPT_1H_sensitivity_hsqc13c).....	251
5.2.106 Simultaneous hard pulses on 13C and 15N (NPT_1H_simpul_13c15n).....	253
5.2.107 Simultaneous hard pulses on 15N and 13C (NPT_1H_simpul_15n13c).....	255
5.2.108 1H temperature calibration with 99.8% MeOD (NPT_1H_tempcalib_998meod).....	257
5.2.109 Vibration Test using Doped Water Sample (NPT_1H_vibration_doped_water).....	259
5.2.110 Vibration Test using Lineshape Sample (NPT_1H_vibration_lineshape).....	261
5.2.111 Watersuppression NaCl with recommended gas flow (NPT_1H_watersupp_NaCl_recflo).....	263
5.2.112 Watersuppression with 270 l/h gas flow (NPT_1H_watersuppression_270l).....	265
5.2.113 Watersuppression with 400 l/h gas flow (NPT_1H_watersuppression_400l).....	267
5.2.114 Watersuppression with 535 l/h gas flow (NPT_1H_watersuppression_535l).....	269
5.2.115 Watersuppression with 670 l/h gas flow (NPT_1H_watersuppression_670l).....	271
5.2.116 Watersuppression with recommended gas flow (NPT_1H_watersuppression_recflo).....	273
5.2.117 29Si background without sample (NPT_29Si_backgr_nosample).....	275
5.2.118 P90 29Si pulse calibration (NPT_29Si_p90determination_29si).....	277
5.2.119 29Si sensitivity (NPT_29Si_sensitivity).....	279
5.2.120 31P B1 homogeneity integral (NPT_31P_b1homogeneityInt_31p).....	281
5.2.121 31P background with sample (NPT_31P_backgr_withsample).....	283
5.2.122 CPD 31P pulse calibration (NPT_31P_cpddeterminationf1_31p).....	285
5.2.123 31P test for artifacts (NPT_31P_fullsw_dec1h).....	287
5.2.124 P90 31P pulse calibration (NPT_31P_p90determinationf1_31p).....	289
5.2.125 31P sensitivity (NPT_31P_sensitivity).....	291
5.2.126 31P sensitivity with 1H decoupling (NPT_31P_sensitivity_dec1h).....	293
5.2.127 P90 39K pulse calibration (NPT_39K_p90determination_39k).....	295
5.2.128 39K sensitivity (NPT_39K_sensitivity).....	297
5.2.129 Atma test (NPT_prep_atma_test).....	299
5.2.130 2H B1 homogeneity integral (NPT_prep_b1homogeneityInt_d).....	301
5.2.131 CPD 2H pulse calibration (NPT_prep_cpddeterminationf1_d).....	303
5.2.132 Optimization of 2H frequency (NPT_prep_fieldsetting_d).....	305
5.2.133 2H lineshape with sample rotation (NPT_prep_lineshape_wrot).....	307
5.2.134 Optimization of 19F locksetting (NPT_prep_locksettings_19f).....	309
5.2.135 Optimization of 2H locksetting (NPT_prep_locksettings_d).....	311

5.2.136 P90 2H pulse calibration (NPT_prep_p90det_astm_d).....	313
5.2.137 P90 2H pulse calibration (NPT_prep_p90det_d).....	315
5.2.138 2H sensitivity, 1% D2O (NPT_prep_sensitivity_1_d).....	317
5.2.139 2H sensitivity, 10% D2O (NPT_prep_sensitivity_10_d).....	319
5.2.140 Automated shim optimization (NPT_prep_tsopt).....	321
5.2.141 1H inno (NPT_1H_sensitivity_inno).....	323
5.2.142 13C inno (NPT_13C_sensitivity_inno).....	325
5.2.143 19F inno (NPT_19F_sensitivity_inno).....	327
5.2.144 19F inno on 1H/19F-coil (NPT_19F_sensitivity_inno_hcoil).....	329
5.2.145 31P inno (NPT_31P_sensitivity_inno).....	331
5.2.146 1H Z-gradient profile [+] (NPT_1H_CMR_gradientprofile_pos).....	333
5.2.147 1H Z-gradient profile [-] (NPT_1H_CMR_gradientprofile_neg).....	335
5.2.148 Gradient recovery test for Z-direction [+] (NPT_1H_CMR_gradientrecovery_pos).....	337
5.2.149 Gradient recovery test for Z-direction [-] (NPT_1H_CMR_gradientrecovery_neg).....	339
5.3 Experiments for HR-Probes with Flow Inserts (LC) .....	341
5.3.1 1H lineshape (NPT_1H_LC_lineshape).....	341
5.3.2 loop transfer time determination starting with empty cell (NPT_1H_LC_loopTransferTimeEmptyCell).....	343
5.3.3 loop transfer time determination starting with filled cell (NPT_1H_LC_loopTransferTimeFilledCell).....	345
5.3.4 1H integrated system performance peak multi-trapping/transfe (NPT_1H_LC_multiTrappingTransfer_spe).....	347
5.3.5 Indirect P90 13C pulse calibration, LC (NPT_1H_LC_p90det_13c).....	349
5.3.6 P90 1H pulse calibration, LC (NPT_1H_LC_p90det_1h).....	351
5.3.7 1H system performance, LC (NPT_1H_LC_performance).....	353
5.3.8 1H integrated system performance stop-flow peak A (NPT_1H_LC_performanceStopFlow_d2o_peak_A).....	355
5.3.9 1H integrated system performance stop-flow peak B (NPT_1H_LC_performanceStopFlow_d2o_peak_B).....	357
5.3.10 1H integrated system performance loop-sapmling/transfer peak A (NPT_1H_LC_performanceTransfer_d2o_peak_A).....	359
5.3.11 1H integrated system performance loop-sapmling/transfer peak B (NPT_1H_LC_performanceTransfer_d2o_peak_B).....	361
5.3.12 1H integrated system performance peak-trapping/transfer peak A (NPT_1H_LC_performanceTransfer_spe_peak_A).....	363



5.3.13 1H integrated system performance peak-trapping/transfer peak B (NPT_1H_LC_performanceTransfer_spe_peak_B).....	365
5.3.14 1H sensitivity, LC (NPT_1H_LC_sensitivity).....	367
5.3.15 stop-flow time determination (NPT_1H_LC_stopFlowTime).....	369
5.3.16 P90 2H pulse calibration, LC (NPT_prep_LC_p90det_d).....	371
5.3.17 peak transfer volume determination starting with empty cell (NPT_prep_LC_peakTransferVolume).....	373
5.4 Experiments for Magic Angle Spinning Probes (Solids) .....	375
5.4.1 13C B1 homogeneity, MAS (NPT_13C_MAS_b1homogeneity_13c).....	375
5.4.2 Double CP 1H-15N-13C, MAS (NPT_13C_MAS_double_cp1h15n_13c).....	377
5.4.3 Optimization of 13C frequency (NPT_13C_MAS_fieldsetting_dec1h).....	380
5.4.4 P90 13C pulse calibration, MAS (NPT_13C_MAS_p90det_13c).....	383
5.4.5 P90 13C 19F-13C CP pulse calibration, MAS (NPT_13C_MAS_p90det_cp19f_13c).....	385
5.4.6 P90 13C 19F-13C CP pulse calibration using H-coil, MAS (NPT_13C_MAS_p90det_cp19f_13c_hcoil).....	388
5.4.7 P90 13C 1H-13C CP pulse calibration, MAS (NPT_13C_MAS_p90det_cp1h_13c).....	392
5.4.8 CP 1H-13C parameter optimization, MAS (NPT_13C_MAS_paropt_cp1h_13c).....	396
5.4.9 P90 13C 19F-13C CP shortest pulse calibration, MAS (NPT_13C_MAS_shortestPulse_cp19f_13c).....	399
5.4.10 P90 13C 19F-13C CP shortest pulse calibration using H-coil, MAS (NPT_13C_MAS_shortestPulse_cp19f13c_hcoil).....	402
5.4.11 P90 13C 1H-13C CP shortest pulse calibration, MAS (NPT_13C_MAS_shortestPulse_cp1h_13c).....	406
5.4.12 13C sensitivity, MAS (NPT_13C_MAS_sino_13c).....	410
5.4.13 CP 19F-13C sensitivity, MAS (NPT_13C_MAS_sino_cp19f_13c).....	413
5.4.14 CP 19F-13C sensitivity using H-coil, MAS (NPT_13C_MAS_sino_cp19f_13c_hcoil).....	416
5.4.15 CP 1H-13C sensitivity, MAS (NPT_13C_MAS_sino_cp1h_13c).....	420
5.4.16 CP 19F-13C power stability MAS (NPT_13C_MAS_stab_cp19f_13c).....	424
5.4.17 CP 19F-13C power stability using H-coil, MAS (NPT_13C_MAS_stab_cp19f_13c_hcoil).....	428
5.4.18 CP 1H-13C power stability MAS (NPT_13C_MAS_stab_cp1h_13c).....	432
5.4.19 P90 15N pulse calibration, MAS (NPT_15N_MAS_p90det_15n).....	436
5.4.20 P90 15N 1H-15N CP pulse calibration, MAS (NPT_15N_MAS_p90det_cp1h_15n).....	439
5.4.21 CP 1H-15N parameter optimization, MAS (NPT_15N_MAS_paropt_cp1h_15n).....	442
5.4.22 P90 15N 1H-15N CP shortest pulse calibration, MAS (NPT_15N_MAS_shortestPulse_cp1h_15n).....	445

5.4.23 CP 1H-15N sensitivity, MAS (NPT_15N_MAS_sino_cp1h_15n).....	448
5.4.24 CP 1H-15N power stability MAS (NPT_15N_MAS_stab_cp1h_15n).....	452
5.4.25 19F B1 homogeneity, MAS (NPT_19F_MAS_b1homogeneity_19f).....	456
5.4.26 19F B1 homogeneity on H-coil, MAS (NPT_19F_MAS_b1homogeneity_19f_hcoil).....	459
5.4.27 P90 19F pulse calibration, MAS (NPT_19F_MAS_p90det_19f).....	461
5.4.28 P90 19F pulse calibration on h-coil, MAS (NPT_19F_MAS_p90det_19f_hcoil).....	463
5.4.29 P90 19F shortest pulse calibration, MAS (NPT_19F_MAS_shortestPulse_19f).....	465
5.4.30 P90 19F shortest pulse calibration on h-coil, MAS (NPT_19F_MAS_shortestPulse_19f_hcoil).....	467
5.4.31 19F sensitivity, MAS (NPT_19F_MAS_sino_19f).....	469
5.4.32 19F sensitivity on h-coil, MAS (NPT_19F_MAS_sino_19f_hcoil).....	471
5.4.33 1H B1 homogeneity, MAS (NPT_1H_MAS_b1homogeneity_1h).....	473
5.4.34 P90 1H pulse calibration, MAS (NPT_1H_MAS_p90det_1h).....	475
5.4.35 P90 1H shortest pulse calibration, MAS (NPT_1H_MAS_shortestPulse_1h).....	477
5.4.36 1H sensitivity, MAS (NPT_1H_MAS_sino_1h).....	479
5.4.37 P90 31P pulse calibration, MAS (NPT_31P_MAS_p90det_31p).....	481
5.4.38 P90 31P 1H-31P CP pulse calibration, MAS (NPT_31P_MAS_p90det_cp1h_31p).....	483
5.4.39 P90 31P 1H-31P CP shortest pulse calibration, MAS (NPT_31P_MAS_shortestPulse_cp1h_31p).....	486
5.4.40 CP 1H-31P sensitivity, MAS (NPT_31P_MAS_sino_cp1h_31p).....	490
5.4.41 CP 1H-31P power stability MAS (NPT_31P_MAS_stab_cp1h_31p).....	494
5.4.42 Optimization of 79Br frequency (NPT_79Br_MAS_fieldsetting).....	498
5.4.43 Magic Angle setting, MAS (NPT_79Br_MAS_magicAngle).....	501
5.4.44 Maximum spin rate testing, MAS (NPT_79Br_MAS_maxSpinRate).....	503
5.4.45 P90 79Br pulse calibration, MAS (NPT_79Br_MAS_p90det_79br).....	505
5.4.46 Temperature controll test on KBr, MAS (NPT_79Br_MAS_temperatureTestKBr).....	507
5.5 Experiments for High Resolution Magic Angle Spinning Probes (HRMAS) .....	509
5.5.1 1H lineshape with magic angle spinning (NPT_1H_HRMAS_lineshape).....	510
5.5.2 Watersuppression (NPT_1H_HRMAS_watersuppression).....	513
5.5.3 Magic Angle setting, HRMAS (NPT_79Br_HRMAS_magicAngle).....	515
5.5.4 Maximum spin rate testing, HRMAS (NPT_79Br_HRMAS_maxSpinRate).....	517
5.5.5 P90 79Br pulse calibration, HRMAS (NPT_79Br_HRMAS_p90det_79br).....	519
5.5.6 Temperature controll test on KBr, HRMAS (NPT_79Br_HRMAS_temperatureTestKBr)....	521
5.6 Experiments for Fourier Spectrometer (CMR) .....	523

5.6.1 P90 1H pulse calibration (NPT_1H_CMR_p90determination_1h).....	523
5.6.2 P90 13C pulse calibration (NPT_13C_CMR_p90determination_13c).....	525
5.6.3 1H lineshape (NPT_1H_CMR_lineshape).....	527
5.6.4 1H sensitivity (NPT_1H_CMR_sensitivity).....	529
5.6.5 1H background with sample (NPT_1H_CMR_background).....	531
5.6.6 13C sensitivity with 1H decoupling (NPT_13C_CMR_sensitivity_dec1h).....	533
5.6.7 P90 31P pulse calibration (NPT_31P_CMR_p90determination_31p).....	535
5.6.8 31P sensitivity with 1H decoupling (NPT_31P_CMR_sensitivity_dec1h).....	537
<b>6 Appendix.....</b>	<b>540</b>
6.1 Sample List .....	540
<b>7 Contact.....</b>	<b>548</b>
7.1 Manufacturer .....	548
7.2 NMR Hotlines .....	548





## 1 Introduction

### 1.1 Purpose

---

This manual describes the NMR tests, which have been scheduled logically to efficiently demonstrate the full **installation and operational qualification (IQ/OQ)** of a Bruker AVANCE-series instrument. It includes basic specification tests such as line shape, resolution and signal-to-noise ratio for <sup>1</sup>H and <sup>13</sup>C as well as signal-to-noise ratio tests for other nuclei. Additional tests will show the high stability and precision of the electronics as well as that of the magnet and the probes.

#### NOTICE

##### Test procedures and Tests:

Please note that Bruker Service Engineers are not responsible for, nor are they trained to run, any extra tests or test spectra in addition to those described in the NMR Product Test manual ZUEP0102.

The aforementioned NMR Product Test procedure constitutes the agreed tests for demonstrating the correct function and performance of the instrument.

Upon successful completion of the tests, both the customer and the Bruker service engineer shall sign the acceptance report ZFPT0008 and the instrument shall be deemed to have been accepted in full according to the terms of the sales agreement.

=> All test procedures and typical performance values are subject to change without notice.

##### Additional Tests:

Any tests additional to those described in this document must be agreed upon at the time of sale and must be documented in the sales agreement. They will be performed by the Bruker application expert subsequent to the installation and they do not form part of the technical acceptance (IQ/OQ) of the instrument.

Some experiments of the NMR Product Test procedure are also used when installing a probe or accessories on an existing spectrometer. For the IQ/OQ procedure of such components, the acceptance reports ZFPT0009 (for probes) and ZFPT0014 (for accessories) will be generated automatically from the NMRPT software.

### 1.2 Area of application

---

With any installation of an NMR system, the customer will receive an introduction and basic training from the Bruker service engineer. The goal of this demonstration is to familiarize the user with safety requirements, system features, and the handling of the OEM products. In addition to the installation procedure, a list of technical demonstrations is outlined in order to provide an effective introduction to the NMR system.

### 1.3 Warning / Notice Signs

---

Safety instructions in this manual and labels of devices are marked with symbols. The safety instructions are introduced using indicative words which express the extent of the hazard.

In order to avoid accidents, personal injury or damage to property, always observe safety instructions and proceed with care.

### **WARNING**



**WARNING** indicates a hazardous situation, which, if not avoided, could result in death or serious injury.

This is the consequence of not following the warning.

1. This is the safety condition.
- => This is the safety instruction.

### **NOTICE**

**NOTICE** indicates a property damage message.

This is the consequence of not following the notice.

1. This is the safety condition.
- => This is the safety instruction.



This symbol highlights useful tips and recommendations as well as information designed to ensure efficient and smooth operation.

## 1.4 Referenced documents

### 1.4.1 Test procedures, Test protocols and Test report (forms)

ZUEP0103	Test procedures for final test of Avance NMR systems
ZFPT0008	Acceptance test protocol for systems
ZFPT0009	Acceptance test protocol for probes (add-on installation)
ZFPT0014	Acceptance test protocol for accessories (add-on installation)
ANAPH011	LC-NMR Acceptance test description
FBAPH009	LC-NMR Acceptance test report

### 1.4.2 Installation Manuals and Service Manuals

Z31555	CryoProbe System Installation Manual
Z31984	CryoProbe Prodigy Installation Manual
H9153	Diffusion Installation and User Manual
Z31750	SampleJet Installation Manual
Z31973	SampleMail / SampleCase Installation Manual
Z31901	SampleXpress Service Manual
Z31942	DNP-NMR Control System Console Service Manual

## 1.4.3 User Manuals

---

Z31836	General Safety Considerations User Manual (English version). Note: This manual is also available in other languages.
Z31633e	Avance SGU Based Frequency Generation Beginners Guide (English version). Note: This manual is also available in other languages.
H9775SA3	Acquisition Reference Manual – TopSpin online help
Z31326	User manual for NMR magnet systems and refilling procedures
Z33092	Liquid Handler SamplePro Tube
Z33075	Micro Imaging for AVANCE III Systems User Manual
Z33117	BSNL (Bruker Smart Nitrogen Liquefier)
Z33119	BCU I User Manual
Z33120	BCU II User Manual
W153895	LT-MAS (Low temperature MAS equipment)
Z31943	DNP-NMR Control System Console User Manual
H146901	MAS III Pneumatic Unit User Manual
Z33045	CryoProbe Prodigy Order Information





## 2 Installation Qualification (IQ)

The system must first be installed and certain initial tests must be performed to ensure general functionality before the operational qualification can be started. The installation qualification includes the following steps.

### 2.1 General setup of the console

---

The engineer checks that all cables in the console and also all cables leading to peripherals are firmly connected and not bent in an unfavourable way that may compromise the functionality of the connected device or otherwise negatively influence the performance of the system. Next, the computer controlling the system, the console, and all peripherals are connected to power and turned on. The engineer makes sure that all power supplies work correctly and that all hardware components including the peripherals show no error indicators.

In particular, the following connections to site-specific supply lines or installations must be checked. Together with the system owner and local authorities it must be made sure that these connections comply with local regulations. The following list serves as a guideline, but the check is not limited to these devices. Instead, all site-specific connections need to be validated.



- => Check the correct **grounding** of the console, preamplifier and magnet in accordance with the descriptions given in the "General Safety Considerations" User Manual Z31836 (available in several languages) on the BASH-CD-ROM. This applies also to the following checks.
- => Check all **electrical connections** together with a local electrician. If a UPS is installed, make sure that it's connected properly to the wall outlet and that proper grounding of the UPS is established.
- => Check all **gas connectors** (to nitrogen, air, or helium gas supply or air-cooled compressors) and make sure that the required pressure and flow is delivered. This applies to the magnet stands, automation devices, the console, the cryo platform, and possibly additional devices.
- => Check all connections to **water supplies** (e.g. for water-cooled compressors).
- => Check any other connections to local infrastructural devices.

Upon signing the acceptance documents, the future responsibility for these connections now lies with the system owner.

---

Topspin is started and some basic software checks are performed as outlined in the next chapters.

### 2.2 Firmware check

---

The engineer checks the firmware version on all devices and makes sure that the appropriate version is installed.



Engineers are provided with a Bruker internal document describing the download procedures:

- => **Firmware\_Updates.pptx**  
Customers can be provided with this document upon request.
- 

### 2.3 Cortab

---

The engineer checks together with the customer that all necessary Cortab tables (linearization tables of the NMR spectrometer) for the required nuclei and routings are present on the system.



The cortab handling is explained in the acquisition manual H9775SA3 which is part of the TopSpin online documentation.

## 2.4 Sample Lift and Spin calibration

---

The engineer sets up the sample lift and checks the sample spinning.

## 2.5 Software license installed

---

Bruker supports different types of software licenses for all computers such as:

### a) Floating licenses

Floating licenses allow the user to run the licensed program on any host in the licensed network. The number of programs that can run simultaneously on any host in this network is limited by the number of available floating licenses only.

### b) Spectrometer routing license (up to TopSpin3.5)

TopSpin offers one additional license type: the spectrometer routing license. On a host computer that controls a spectrometer, TopSpin will always work even if no dedicated TopSpin license is available. This spectrometer routing license allows basic operation of the spectrometer, but some advanced features may not be available.



Please note that the spectrometer routing license is just a security feature that will be used only if for any reason (e.g. hardware problems) a regular license is temporarily unavailable. If a regular license (floating, node-locked, demo) is available, TopSpin will always use the regular license and not the spectrometer routing license.

### c) Demo licenses

Demo licenses are node-locked licenses that expire after three months. They are available for all TopSpin versions. In addition, they can be used as an emergency license for TopSpin4 as a replacement of the spectrometer routing license (see point b above).

The Bruker engineer should see to it that an appropriate license has been installed.

To request any type of license, please use the Bruker license request form.



<https://www.bruker.com/service/support-upgrades/license-requests/nmr-license-requests.html>

For further information please refer to the Software Release Letter of your TopSpin version or see [www.bruker.com](http://www.bruker.com) for more information on software packages and license types.

## 2.6 Helium and Nitrogen level log files active

---



The procedures that write log files for the Helium level measurement and, if installed, the Nitrogen level measurement must be set up and checked.

## 2.7 MICS installed

The Magnet Information and Control System software (MICS) must be installed and the correct magnet BIS file must be in place. The alarm settings must be explained to the customer. All details are described in the MICS manual, which is part of the MICS software online documentation.

You can configure your system to allow remote monitoring of your system health by Bruker. Remote monitoring allows us to take preventive action prior to system failure and help you obtain maximum uptime. This remote monitoring service is free of charge.

If you have subscribed on a **LabScape** Agreement, or if you are within 12 months of system acceptance, Bruker will, in the event that we detect any potential issue with your system, analyze the situation after which we will contact you from the local office to plan remedial action. Additional responsibilities covered by Bruker will depend on the type of **LabScape** Agreement.

In addition, you will also receive emails at approximately three month intervals reporting upon the status of your magnet and summarizing events that have taken place (e.g. warnings, number of refilling operations etc.) during this period.



For further information, visit our website at [www.bruker.com/LabScape](http://www.bruker.com/LabScape) or contact your local office.

## 2.8 Special precautions if a Prodigy Cryoprobe will be installed

If a Prodigy CryoProbe will be installed with the system, special precautions are required. The Prodigy CryoProbe will be cooled through evaporation of liquid nitrogen. Approximately 10 liters of liquid nitrogen will be evaporated during 24 hours of operation. This results in approximately 7000 liters (7 m<sup>3</sup>) of nitrogen gas in the atmosphere of the spectrometer room. The ventilation system must be capable to replace this additional nitrogen gas with air.

### NOTICE

#### Use of Oxygen Sensors:

Bruker strongly recommends installing oxygen sensors in the spectrometer room to detect a possible drop in the oxygen level.

=> Consult the CryoProbe Prodigy Order Information Z33045 for details.

## 2.9 Customer Information

The Customer Information must be entered in TopSpin with the edcstm command.

It is recommended to print the resulting file and to keep it with the installation documentation.

The system is now ready and the operational qualification (OQ) can be started.



# 3 Operational qualification (OQ) and customer training

The operational qualification of the system is performed through the execution of NMR measurements outlined in subsequent chapters of this manual. The type and number of experiments is determined through the type and number of probes that will be used with the NMR system. A protocol with the results from these NMR measurements will be presented to and discussed with the customer.

The goal of this demonstration is to familiarize the user with the safety requirements, with the system features, and with the handling of the peripherals and, if present, additional OEM products.



At the end of the demonstration, the acceptance protocol will be signed.

## 3.1 Basic safety precautions

### NOTICE

#### Prevention of hazards:

The installation and operation of a superconducting magnet system presents a number of hazards that all laboratory staff must be aware of.

- => The Service Engineer will explain the effects of the magnetic field on magnetic materials such as tools, cell phones, watches, magnetic tapes, credit cards, and surgical implants.
- => **The customer is responsible after this training to forward these instructions to all laboratory staff, also to new employees in the future.**
- => Be aware that during normal operation only 3-5m<sup>3</sup>/day of nitrogen is evaporated, but during a quench 50-100m<sup>3</sup> of helium gas is produced within a short time. Because of that a magnet system must not be installed in an airtight room.
- => Normal-sized windows and doors are usually sufficient for ventilation even during and after a quench.



For more details consult the magnet manual specific for your magnet system. Also refer to the "General Safety Considerations" User Manual Z31836 (available in several languages) on the BASH-CD-ROM.

## 3.1.1 Magnet safety

### NOTICE

#### Before charging of the magnet:

The appropriate warning signs must be posted by the customer at each entrance to any room containing an NMR-System.

=> If the magnet is placed in an open setting, warning signs have to be posted at least 5 m from the magnetic center such that people cannot approach without first seeing the signs.

=> Also post the warning signs at rooms above and below the magnet, if the lines of 5G (0.5 mT) and 30 G (3.0 mT) penetrate into these areas.

=> Ensure that all loose ferromagnetic objects are removed from within the 5G (0.5 mT) area.

=> The customer should be aware of the fact that the operation of other equipment may be affected by the presence of large magnetic fields above 30 G (3.0 mT).

=> Items such as watches, cell phones, tape recorders and cameras may be magnetized and irreparably damaged if they are exposed to fields above 100 G (10.0 mT).

=> **The customer must provide on-site training about magnetic field hazards to people who may be exposed to 5G (0.5mT) or higher magnetic fields.**

=> The vertical and horizontal stray field plots are available in the Bruker site planning manuals and in the magnet manual specific for your magnet system.

## 3.1.2 Handling of cryogenic liquids and magnet refilling

The safe handling of cryogenic liquids requires some knowledge of the physical properties of these liquids, common sense, and sufficient understanding to predict the reactions of such liquids under certain physical conditions.

### NOTICE

#### Prevention of hazards:

The very large increase in volume accompanying the vaporization of the liquid into gas and the subsequent process of warming up is approx. 700:1 for helium and nitrogen.

=> Therefore, we recommend that containers for cryogenic liquids must not be closed completely to avoid a large pressure build-up.

This will present an explosion hazard and may lead to large product losses.



## WARNING

### Risk of injury during refilling procedure

Risk of blindness if cryogenic liquids come into contact with the eyes!

Risk of severe cold-burns if skin comes into contact with cryogenic fluids!

Risk of skin adhesion with cooled metal parts.



1. Always wear protective goggles when carrying out the refilling procedure.
2. Always wear protective gloves and closed clothing when carrying out the refilling procedure.
3. Read the user manual for NMR magnet systems Z31326 for more details on the refilling procedures.
4. Refer to the Bruker movies explaining the correct helium and nitrogen refilling procedures. Movies to these topics are available for registered customers on [www.bruker.com](http://www.bruker.com) here: <https://www.bruker.com/service/information-communication/bruker-academy-videos.html>.

## NOTICE

### Refilling of liquid Nitrogen:

The nitrogen tank of the magnet (if present) should be refilled once a week or as often as described in the magnet manual.

=> Never apply a transfer pressure of more than 350 mbar (5.0 psi) to the nitrogen vessel.

=> Make sure that during the transfer all nitrogen neck tubes are fully open.

=> Before refilling the helium tank of the magnet, the level in the refill vessel must be checked with the dip-stick.

### Liquid Nitrogen:



For the refilling procedure of **liquid nitrogen** follow the description in the user manual for NMR magnet systems Z31326 or refer to the **nitrogen refill movie** on [www.bruker.com](http://www.bruker.com) or on the BASH-CD-ROM.



## NOTICE

### Refilling of liquid Helium:

- => Always refill nitrogen (see above) before refilling helium.
- => Never insert a "warm helium transfer line" into the cryostat, the warm helium gas could lead to a quench of the magnet.
- => Never use any extensions on the helium transfer line because they might reach the siphon.
- => Never apply a transfer pressure of more than 100 mbar (1.45 psi) to the helium vessel.
- => Always make sure that the outlet of the helium manifold is fully open to the atmosphere or to a helium recovery system.
- => After refilling of helium, check that all nitrogen neck tubes are free of any ice blockages.

### Liquid Helium:



For the refilling procedure of **liquid helium** follow the description in the user manual for NMR magnet systems Z31326 or refer to the **helium refill movie** on [www.bruker.com](http://www.bruker.com) or on the BASH-CD-ROM.

### 3.1.3 Control of nitrogen and helium level

The NMR instrument measures the helium level once a day at 04.15 am (in TopSpin2.x and older versions it was measured at 03:00 am). This level is recorded by software...

**Windows:** c:\bruker\diskless\prog\logfiles\heliumlog

or

**Linux:** /usr/diskless/prog/logfiles/heliumlog)

...and can be checked manually on the BSMS Keyboard (He-Level) or in the BSMS display.

This level must be recorded at least once a week. You can find prepared sheets at the end of the magnet manual.

The MICS software (see above) can be configured such that it will read the levels recorded in the heliumlog file.

With this information, MICS can provide graphical presentations of helium and nitrogen boil-off.



We recommend that the helium vessel is refilled within the specified hold time period and certainly before the level falls below the allowed minimum level (check the user manual for NMR magnet systems Z31326).

## 3.1.4 Homogeneity and stability disturbance

---

### Ferromagnetic Material



The presence of any ferromagnetic materials in the immediate vicinity of the magnet will decrease the magnet homogeneity. The site-planning manuals on the BASH-CD-ROM provide guidelines for minimum requirements.

---

### Electromagnetic Interferences



Large electromagnetic interference can affect the long term stability of an NMR system. Possible sources of electromagnetic interference may be produced, for instance, by elevators, trams, subways, mass spectrometers, centrifuges, large electric motors, air conditioning systems, heavy duty transformers etc.

The site planning manuals contain detailed information on the tolerated limits.

---

## 3.1.5 Earthquake precautions

---



If the system is placed in a territory with higher seismic activities, we recommend securing the magnet system either with ropes from a ceiling hook or with non-magnetic brackets around the magnet stand.

---

## 3.1.6 Refilling of the nitrogen vessel for Prodigy CryoProbes

---

The refilling of the nitrogen vessel for Prodigy CryoProbes and necessary precautions will be explained in cases where such a probe will be installed on the system.

## 3.2 Hardware overview

---

The Service Engineer will outline the function of the NMR system. The individual hardware units are explained to familiarize the customer with all system features.



Detailed circuit descriptions will not be covered. Special customer training courses are offered by Bruker here:

<https://www.bruker.com/service/education-training/training-courses/magnetic-resonance.html>

---

### NOTICE

#### Damage to the probe because of excessive power:

Applying too much power may damage the probe and will most likely lead to an expensive probe repair.

=> Each probe has a maximum power limit for every possible nucleus.

=> For probes that are delivered with a PICS module (probe information control system), these values are directly imported with the command "edprobe".

=> The maximum RF power values for the corresponding nuclei are also provided in the "Customer Certificate" that is delivered with the probe.

=> Make sure this maximum power is never exceeded in any experiment.

=> If possible always turn powercheck on in Topspin.

### 3.2.1 Console on/off operation

---

The customer will be trained on the different procedures of how to power down a console carefully (like RF power down, computer power down, and main power down) as well as how to start it up again. Also the reset functions will be shown on all units.

The customer will be trained to do some basic system tests like DRU and ELCB tests through the respective web interfaces (TopSpin command "ha").



Bruker offers periodically scheduled training courses for operation and troubleshooting of spectrometers especially for technicians and engineers. See here:

<https://www.bruker.com/service/education-training/training-courses/magnetic-resonance.html>

### NOTICE

#### **Safety Demonstration of power down:**

For safety reasons a fast power down will be demonstrated.

**The customer is warned that only trained service staff should open the units because of hazardous voltages.**

### 3.2.2 Basic operation

An introduction to standard measurements on 1H- and X-nuclei is given by the Bruker service engineer. A demonstration of peripheral units like temperature control, gradient control, gradient amplifiers, or solids accessories will be performed (including mechanical operations like probe change).

### NOTICE

#### **Read Manual before Demonstration:**

We recommend that the customer reviews the operation manuals before attending the demonstration in order to benefit most.

A general guide to shimming as well as tuning and matching will be part of the introduction. Engineers are provided with a presentation explaining the basic features of TopSpin.

=> **TopSpin3.5\_introduction\_3h.pptx**

Customers can be provided with this document upon request.



A basic introduction is provided with the Beginners' Guide Z31633e, which is available in several languages on the BASH-CD-ROM as well as on the Bruker website at

<https://www.bruker.com/service/information-communication/user-manuals/nmr.html>

Bruker offers periodically scheduled training courses for operation and usage of the TopSpin and IconNMR software packages for users of the spectrometers. See here:

<https://www.bruker.com/service/education-training/training-courses/magnetic-resonance.html>

### NOTICE

**Installation and Operation of products from other sources:**

Bruker service engineers are not responsible for, nor are they trained to install, locally ordered OEM-products from other sources but Bruker.

## 3.2.3 Troubleshooting

---

The complete range of software service tools such as web tools, BSMS, HPPR, and IPSO software tools will be demonstrated.

## 3.2.4 Backup (nmr\_save, Images)

---

The TopSpin software provides functionality to periodically archive the current spectrometer configuration to a file. In case of computer hardware problems, this archive can be used to restore a computer to the last known spectrometer status.



Bruker also provides a software tool to create a complete computer system disk image. The customer will be informed about these possibilities.

---

## 3.2.5 Introduction to IconNMR

---

The basic functionality of IconNMR should be presented and explained to the customer, especially if a sample changer is installed on the system. An example run should be started and the quality of the results should be verified.



Engineers are provided with a presentation explaining the basic features of IconNMR.

=> **IconNMR-acquisition-overview.pptx**

Customers can be provided with this document upon request.

---

## 3.2.6 AssureSST / Performance check

---

AssureSST (System Suitability Test) is a part of IconNMR to check system performance on a regular basis (e.g. for instruments that are installed in a GxP environment). AssureSST should be setup and demonstrated.



Engineers are provided with a presentation explaining the basic features of AssureSST.

=> **AssureSST\_overview.pptx**

Customers can be provided with this document upon request.

---

## 3.3 Optional Components

---

The customer will be informed about the handling and documentation of optional or additional components like...:

- => Sample Changer
- => MAS controller
- => High power equipment
- => LC-NMR
- => Liquid Handler SamplePro Tube
- => Micro-Imaging

- => Diffusion
- => CryoProbe / Cryoplatfrom
- => BSNL
- => Additional cooling/heating units (like BCU1 / BCU2)
- => LT-MAS (Low Temperature MAS equipment)
- => Gyrotron magnet and DNP console

## 3.3.1 Sample Changer

---

The Sample Changer will be installed in accordance with its corresponding service or user manual. If the device may be installed by the customer, then the instructions are contained in the user manual. Otherwise, the instructions are contained in the service manual.



An acceptance protocol for the device is part of the installation instructions. The engineer will complete this acceptance protocol together with the customer and both will sign this acceptance document.

---

### 3.3.1.1 SampleJet

---

The installation manual Z31750 contains the installation instructions and the acceptance protocol.

### 3.3.1.2 SampleMail / SampleCase

---

The installation manual Z31973 contains the installation instructions and the acceptance protocol.

### 3.3.1.3 SampleXpress / SampleXpress Lite

---

The service manual Z31901 contains the installation instructions and the acceptance protocol.

## 3.3.2 MAS Controller

---

The MAS Controller will be installed in accordance with its corresponding user manual (H146901), which also covers the installation.



The successful installation will be shown through the measurement of relevant samples that show the proper functionality in combination with the spectroscopic test results obtained with a suitable MAS probe.

---

## 3.3.3 High power equipment

---

High power equipment refers to special amplifiers that provide high power output typically for solid-state applications. The basic functionality of these amplifiers is tested during the Cortab procedure (see above 2.3).



The proper functionality will be shown through the measurement of relevant samples in combination with the spectroscopic test results obtained with a suitable MAS probe.

---

## 3.3.4 LC-NMR

---

The LC-NMR equipment will be installed in accordance with the document ANAPH011 and the acceptance test report will be documented in FBAPH009.



The proper functionality will be shown through the measurement of relevant samples in combination with the spectroscopic test results obtained with a suitable probe.

---

## 3.3.5 Liquid Handler SamplePro Tube

---

The basic installation of the Liquid Handler SamplePro Tube will be done in accordance with the document Z33092.

### NOTICE

#### Application-specific Tests

Additional application-specific tests and adjustments are not part of the basic installation. They are part of, and defined for, an application specific installation.

## 3.3.6 Micro-Imaging

---

The Micro-Imaging equipment will be installed in accordance with the document Z33075.



The proper functionality will be shown through the measurement of relevant samples in combination with the spectroscopic test results obtained with a suitable Micro-Imaging probe.

---

## 3.3.7 Diffusion

---

The Diffusion equipment will be installed in accordance with the document H9153.



The proper functionality will be shown through the measurement of relevant samples in combination with the spectroscopic test results obtained with a suitable Diffusion probe.

---

## 3.3.8 CryoProbe / Cryoplatform

---

The Cryoprobe equipment will be installed in accordance with the documents listed be-low.



The proper functionality will be shown through the measurement of relevant samples in combination with the spectroscopic test results obtained through the NMRPT procedure.

---



## 3.3.8.1 Helium-cooled cryoprobes

---

The CryoProbe / Cryoplatform will be installed in accordance with the document Z31555.

## 3.3.8.2 Nitrogen-cooled CryoProbes (Prodigy probes)

---

The CryoProbe / Cryoplatform will be installed in accordance with the document Z31984.

## 3.3.9 BSNL (Bruker Smart Nitrogen Liquefier)

---

The BSNL unit will be installed in accordance with the document Z33117.

## 3.3.10 Additional cooling/heating units (like BCU1 / BCU2)

---

The additional cooling units will be installed in accordance with the documents Z33119 and Z33120, respectively.

## 3.3.11 LT-MAS (Low Temperature MAS equipment)

---

The LT-MAS (Low Temperature MAS equipment) will be installed in accordance with the document W153895.



The proper functionality will be shown through the measurement of relevant samples in combination with the spectroscopic test results obtained with a suitable MAS probe.

---

## 3.3.12 Gyrotron magnet and DNP console

---

The Gyrotron equipment will be installed in accordance with the document Z31942. The operation of the system is described in the user manual Z31943.



The proper functionality will be shown through the measurement of relevant samples in combination with the spectroscopic test results obtained with a suitable probe.

---

## 3.4 Acceptance and warranty

---

### 3.4.1 Explanation of warranty

---

#### NOTICE

The customer will be informed that the warranty will begin with the acceptance of the instrument. The warranty period will be in accordance with the terms of sale and must be inserted in the acceptance report ZFPT0008.

The items covered under warranty are also described in the sales agreement and normally include all defects in material and workmanship. Glassware and parts, subject to wear and tear, are not included.

In the case of parts delivered late, the warranty for these parts will begin on the acceptance of these parts.

### 3.4.2 System documentation

---

For each Bruker NMR-System a whole set of operation and reference manuals is provided either in printed form, or stored on CD/DVD, or as online help in the respective software packages.



The service engineer will give a short introduction on how to use this documentation.

---

### 3.4.3 Customer support hotlines

---



Customer support hotlines will be entered in the corresponding software interfaces.


General contact information is provided in the Contact chapter of this manual.

---



## 4 NMR Product Test (NMRPT)

*NMRPT* software provides an automated Acceptance Test and Service Test procedure of spectrometers and probes.



This software is generating all necessary acquisition and processing parameters for Acceptance and Service Testing of all kind of probes and consoles (AVIII and newer). Evaluation of the results is provided. Customer information, spectrometer and probe information is collected by the *NMRPT* software for documentation and acquisition purposes. All processing and plotting parameters are created and applied automatically. Measurement results can be explored individually and selected measurements of each experiment are collected into one spectra PDF. The collected spectra and a Test Certificate are the two parts of the Test Report. The type of Test Report depends on the selected test mode. This Test Report is created by the Service Engineer after all measurements are finished. If the results meet the specifications according to the Customer Certificate (ZFUT0015), the Service Engineer can finalize this Test Report to obtain an official Bruker Test Document.

*NMRPT* automatically stores a shimfile for the Lineshape Nonspinning and the Water Suppression experiments. These shim file names are composed of the part number and serial number as well as an extension containing the experiment type, e.g. the file Z104275\_0118\_1H\_Is\_nrot is a shimfile of the 5 mm BBO probe Z104275\_0118 shimmed on 1H for Line Shape Nonspinning experiment.

Note: All probe specifications (Line Shape, Resolution, Sensitivity, Hump and Splitting) will automatically be evaluated from measurement by *NMRPT*. The *NMRPT* software therefore always calculates the 1H Sensitivity for each, a noise region of 2 ppm and a noise region of 200 Hertz. See the probe specifications (Customer Certificate ZFUT0015) for the corresponding values ("Rated Specification", "Rated Sample" and "Conditions") of the respective nuclei.



# 5 NMRPT Experiments

## 5.1 Introduction

---

All standardized experiments provided by *NMRPT* are described in this chapter. But the user can select only from the subset feasible to the current hardware (combination of probe, console, and routing). The datasets included with the *NMRPT* are automatically adjusted and optimized to the current hardware. Therefore the parameters shown in this chapter are only example values and may differ depending on the hardware to be tested. If user interaction is necessary, the software will guide the user (e.g. through the parameter optimization).

A PDF file documents the results and spectra for each successful measurement. *NMRPT* collates the results and spectra of all acquired experiments (one measurement per experiment) in a report. This report is the test document for the acceptance or a service of a spectrometer, parts of a spectrometer, a probe, or accessories.

### Receiver Gain (RG)

The value for RG is predefined in each experiment. It is adjusted for each probe during setup. The usual standard values are  $RG_{\min}$  or  $RG_{\max}$ . In first case, *NMRPT* executes RGA to set the optimum value. In second case, *NMRPT* adopts the value. For a user controlled RG, set RG in the Preparation Panel of *NMRPT* Acquisition page.  $RG_{\min}$  and  $RG_{\max}$  are derived from TopSpin configuration information (after execution of cf, see TopSpin manual). In *NMRPT* the maximum value of  $RG_{\max}$  is limited to 203. If the  $RG_{\max}$  value of the TopSpin configuration is lower, this value will be used.

### Carrier Frequencies (O1 and O2)

The carrier frequency offset can be changed in all experiments where *NMRPT* does not optimize these settings. It is not recommended to change this except for the water suppression experiments when using specific options of L23 (see subsequent section of watersuppression experiment).

### Loopcounter Parameter L23

The loopcounter L23 is used to apply special conditions in *NMRPT* experiments. The allowed values for each experiment are provided in the experiment description in this chapter. For some experiments (e.g. L23=27 in watersuppression experiment) changing the L23 value results in a non-standard experiment without a regular experiment output.

### Gradient Strength (GPZx)

The used gradient strength (given in %) of all experiments is related to a maximum gradient current of the probe. If the maximum current of a gradient amplifier differs from the one of the probe, the gradient strength is recalculated accordingly.

### Pre-Scan Delay (DE)

The value of DE is set during execution of getprosol. For a user controlled DE, set DE and the option 'Skip Getprosol' in the Preparation Panel of *NMRPT* Acquisition Page or change the DE entry in the Prosol table. For the following experiments this handling is overruled by setting DE explicitly after getprosol: NPT\_1H\_cosydfphpr, NPT\_1H\_noesyphpr, NPT\_13C\_ringing\_dec1h, NPT\_prep\_fieldsetting\_d, NPT\_prep\_lineshape\_wrot

## 5.2 Experiments for High Resolution Probes (HR)

---

### Sample Rotation Frequency (RO)

The sample rotation frequency is set and interpreted by *NMRPT* as provided in sample state in the experiment descriptions of this chapter. Selected experiments (e.g. 2D experiments) are measured always with  $RO = 0$ . For all the other experiments RO is set according the following rules:

Sample diameter  $\leq 1.7\text{mm}$   $\Rightarrow$   $RO = 0$   
Sample diameter =  $3.0\text{mm}$   $\Rightarrow$   $RO = 0$  (RT probes)

Sample diameter = 3.0mm => RO = 20 (CRP probes)

Sample diameter = 5mm => RO = 20

Sample diameter = 8mm => RO = 18

Sample diameter = 10mm => RO = 14

### **Probes with the same nucleus on multiple channels (13C, 15N, 31P, 19F)**

When a nucleus occurs twice on a probe, different experiments are used for each probe channel. The standard experiment (e.g. NPT\_13C\_sensitivity) is measured on the selective probe channel (excluding 1H/19F channel). The second experiment, labeled with “\_2nd” (e.g. NPT\_13C\_sensitivity\_2nd) is measured on the broadband channel. The “\_2nd” experiments are not documented in this chapter, since they are similar to the corresponding standard experiments.


### **Probes with 19F tuneable on 1H channel**

In case of 19F measurements on a 1H/19F channel the experiments need special handling for certain hardware. Therefore the experiment names for 19F measurements on the 1H/19F channel are indicated with “\_hcoil”.

## 5.2.1 <sup>13</sup>C background with sample (NPT\_13C\_backgr\_withsample)

---

**Test Sample:** 10% Ethylbenzene (EB) in Chloroform-D  
Z10153, Z10154, Z10034, Z10253, Z10292, Z10723  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

<sup>13</sup>C Background signal spectrum with sample. Sharp signals arise from sample and solvent, broad signal could arise from solid compound in the probe.

### Control Option for Acquisition (L23)

1 default



## Parameters

<b>F1 ACQU</b>			Parameters	<b>CNST 10</b>	20.000		Flip angle for P90
NUC1	13C			<b>F1 PROC</b>			Parameters
NUC2	1H			SI	65536		
PULPROG	npt_zg0dc			WDW	1		
NS	1000			LB	5.000	Hz	
DS	4			PC	1.400		
RG	101.000		no optim.	F1P	180.000	ppm	
O1P	199.988	ppm		F2P	-20.000	ppm	
O2P	5.000	ppm					
CPDPRG2	waltz64		decoupl. sequence				
SW	496.855	ppm					
TD	65536						
AQ	0.655	s	field dependent				
FIDRES	1.526	Hz	field dependent				
D 1	1.430	s	AQ+D1=const				
P 0		us	P 1 * CNST 10 / 90				
P 1	9.0	us	90deg NUC1				
PLW 1	40.2	W	Pow@90deg(Specs) NUC1				
PLW 12	0.1	W	Pow@CPD NUC2				
DIGMOD	3		baseopt				
TE	298.000	K	default				


## Experiment Description

Background signal measurements are executed using a small flip angle due to two reasons. First, to get more efficient excitation bandwidth, secondly to compensate for usually long T1 relaxation time of solid signals.

Full spectrum is normally printed without any baseline correction.

## 5.2.2 DEPT135 (NPT\_13C\_dept135)

**Test Sample:** 10% Ethylbenzene (EB) in Chloroform-D  
Z10153, Z10154, Z10034, Z10253, Z10292, Z10723  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

DEPT ('Distortionless Enhancement of Polarization Transfer') using a 135 degree decoupler pulse. The experiment shows carbon-13 signals where one (CH) or three protons (CH<sub>3</sub>) are attached with positive sign, whereas the negative signals in the spectrum arise from carbon with two protons (CH<sub>2</sub>). Carbons with no protons attached are virtually suppressed, as well as carbon signals from solvent molecules.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F1 ACQU</b>			Parameters	P 13	2000.0	us	180deg shaped pulse
NUC1	13C			PLW 1	19.3	W	Pow@90deg(Specs) NUC1
NUC2	1H			PLW 2	5.6	W	Pow@90deg(Specs) NUC2
PULPROG	deptsp135			PLW 12	0.08	W	Pow@90deg(CPD)
NS	128		RT: 128, Cryo: 8	SPNAM5	Gaus1.1000		shape name
DS	8			SPOAL 5	0.500		phase align.
RG	101.000		no optim.	SPOFFS 5	0.000	Hz	offset freq.
O1P	89.981	ppm		SPW 5	3.6	W	Pow@Shape
O2P	5.000	ppm		TE	298.000	K	default
SW	216.048	ppm		<b>F1 PROC</b>			Parameters
TD	130434		field dependent	SI	131072		
AQ	3.000	s		WDW	0		
FIDRES	0.333	Hz	field dependent	LB	1.000	Hz	
D 1	4.000	s		PC	1.400		
D 2	0.003	s	1s/(CNST2*2)	F1P	175.000	ppm	
CPDPRG2	waltz65		decoupl. sequence	F2P	-5.000	ppm	
CNST 2	145.000	Hz	J[XH] coupling	CY	11.000	cm	
P 1	11.0	us	90deg NUC1	<b>NMRPT</b>			Parameters
P 3	14.0	us	90deg NUC2	CNST 48	1.000		SN_WSUP(res.)
P 4	28.0	us	180deg NUC2				
PCPD 2	115.0	us	90deg CPD				

## Experiment Description


DEPT experiments are executed as integral and functional test of the system. The outcome of this experiment is strongly dependent on correct decoupler pulses (1H). The pulse sequence version used in NMRPT is using shaped pulses for 180 degree carbon refocussing pulse to suppress off-resonance effects which are often responsible for phase distortions at higher field strengths.

The evaluation of the experiment is based on the comparison of the signal integral at ~126 ppm (reference) and ~29 ppm (check). For the DEPT135 the integral ratio should be close to 100%.

## 5.2.3 DEPT90 (NPT\_13C\_dept90)

---

**Test Sample:** 10% Ethylbenzene (EB) in Chloroform-D  
Z10153, Z10154, Z10034, Z10253, Z10292, Z10723  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

DEPT ('Distortionless Enhancement of Polarization Transfer') using a 90 degree decoupler pulse. The experiment shows only carbon-13 signals where one proton (CH) is attached. All other carbon signals including signals from solvent molecules are virtually suppressed.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F1 ACQU</b>			Parameters	P 13	2000.0	us	180deg shaped pulse
NUC1	13C			PLW 1	19.3	W	Pow@90deg(Specs) NUC1
NUC2	1H			PLW 2	5.6	W	Pow@90deg(Specs) NUC2
PULPROG	deptsp90			PLW 12	0.08	W	Pow@90deg(CPD)
NS	128		RT: 128, Cryo: 8	SPNAM5	Gaus1.1000		shape name
DS	8			SPOAL 5	0.500		phase align.
RG	101.000		no optim.	SPOFFS 5	0.000	Hz	offset freq.
O1P	89.981	ppm		SPW 5	3.6	W	Pow@Shape
O2P	5.000	ppm		TE	298.000	K	default
SW	216.048	ppm		<b>F1 PROC</b>			Parameters
TD	130434		field dependent	SI	131072		
AQ	3.000	s		WDW	0		
FIDRES	0.333	Hz	field dependent	LB	1.000	Hz	
D 1	4.000	s		PC	1.400		
D 2	0.003	s	1s/(CNST2*2)	F1P	175.000	ppm	
CPDPRG2	waltz65		decoupl. sequence	F2P	-5.000	ppm	
CNST 2	145.000	Hz	J[XH] coupling	CY	11.000	cm	
P 1	11.0	us	90deg NUC1	<b>NMRPT</b>			Parameters
P 3	14.0	us	90deg NUC2	CNST 48	1.000		SN_WSUP(res.)
P 4	28.0	us	180deg NUC2				
PCPD 2	115.0	us	90deg CPD				

## Experiment Description


DEPT experiments are executed as integral and functional test of the system. The outcome of this experiment is strongly dependent on correct decoupler pulses (1H). The pulse sequence version used in NMRPT is using shaped pulses for 180 degree carbon refocussing pulse to suppress off-resonance effects which are often responsible for phase distortions at higher field strengths.

The evaluation of the experiment is based on the comparison of the signal integral at ~126 ppm (reference) and ~29 ppm (check). For the DEPT90 the integral ratio should be close to 0%.

## 5.2.4 <sup>13</sup>C test for artifacts with 1H decoupling (NPT\_13C\_fullsw\_dec1h)

---

**Test Sample:** 10% Ethylbenzene (EB) in Chloroform-D  
Z10153, Z10154, Z10034, Z10253, Z10292, Z10723  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Full range <sup>13</sup>C spectrum with 1H decoupling.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU		Parameters		F1 PROC		Parameters	
NUC1	13C			SI	524288		
NUC2	1H			WDW	1		
PULPROG	npt_zg0dc			LB	0.300	Hz	
NS	64			PC	1.400		
DS	4			F1P	180.000	ppm	
RG	101.000		no optim.	F2P	-20.000	ppm	
O1P	164.987	ppm		CY	11.000	cm	
O2P	5.000	ppm					
SW	354.909	ppm					
TD	262144						
AQ	3.670	s	field dependent				
FIDRES	0.272	Hz	field dependent				
D 1	1.230	s	AQ+D1=const				
P 1	9.0	us	90deg NUC1				
PLW 1	40.2	W	Pow@90deg(Specs) NUC1				
PLW 12	0.1	W	Pow@CPD NUC2				
CPDPRG2	waltz64		decoupl. sequence				
DIGMOD	3		baseopt				
TE	298.000	K	default				

## Experiment Description

Full range spectra (spectroscopically useful range) are acquired as test for RF artifacts (spikes). Currently no evaluation is provided and visual inspection of the data is required. Data are treated with baseline correction.

## 5.2.5 <sup>13</sup>C lineshape without sample rotation (NPT\_13C\_lineshape\_nrot)

**Test Sample:** 40% Dioxane in Benzene-D6 (ASTM)  
Z10163, Z100929, Z10164, Z10035, Z10255, Z10242, Z10724  
**Solvent:** C6D6  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Carbon-13 line shape spectrum from +10.0 Hz to -10.0 Hz as overview spectrum. The expansion plot scaled to full intensity on the left shows the dioxane (1,4-dioxane) signal with higher resolution (+2.0 Hz, -2.0 Hz).

### Control Option for Acquisition (L23)

- 1 Broad band decoupling using waltz65, PULPROG=zgpg
- 11 CW decoupling based on PLW12, automatic O2 optimization, PULPROG=zgcv
- 21 CW decoupling based on PLW12, O2 from parameter set, no optimization, PULPROG=zgcv



## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	13C				SI	32768			
NUC2	1H				WDW	0			
PULPROG	zgpg				LB	0.000	Hz		
NS	4				PC	1.400			
DS	0				F1P	0.099	ppm		
RG	101.000			optim. by RGA	F2P	-0.099	ppm		
O1P	66.488	ppm			CY	1000.000	cm		
O2P	3.474	ppm							
SWH	396.825	Hz							
TD	32768								
AQ	41.288	s							
FIDRES	0.024	Hz							
D 1	33.712	s		AQ+D1=const					
P 1	9.0	us		90deg NUC1					
PCPD 2	120.0	us		CPD NUC2					
PLW 1	39.3	W		Pow@90deg(Specs) NUC1					
PLW 12	0.1	W		Pow@CPD NUC2					
PLW 13	0.05	W		Pow@CPD NOE NUC2					
CPDPRG2	waltz65			decoupl. sequence					
TE	298.000	K		default					

## Experiment Description

The Carbon-13 line shape experiment with 1H decoupling using as default waltz65 and decoupler pulse length (PCPD2) of 120 us. The corresponding decoupling powers (PLW12 and PLW13) are calculated from prosol entries for P12 and PLW12.


The exact signal position is determined with NS=1 prior to the line shape measurement. The carrier position is afterwards set to this position

O1 optimized = peakFreqHz - (SWH/4)

This procedure allows the application of a strip transformation with the parameters STSR=0, STSI=SI/2 resulting in the signal position on-resonance and in most cases resulting in an optimal automatic phase correction.

## 5.2.6 <sup>13</sup>C lineshape with sample rotation (NPT\_13C\_lineshape\_wrot)

**Test Sample:** 40% Dioxane in Benzene-D6 (ASTM)  
Z10163, Z100929, Z10164, Z10035, Z10255, Z10242, Z10724  
**Solvent:** C6D6  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Carbon-13 line shape spectrum from +10.0 Hz to -10.0 Hz as overview spectrum. The expansion plot scaled to full intensity on the left shows the dioxane (1,4-dioxane) signal with higher resolution (+2.0 Hz, -2.0 Hz).

The PDF comprises a second page (not shown in documentation) which shows carbon-13 line shape spectrum with a printing range from +45.0 Hz to -45.0 Hz (including spinning side bands).

### Control Option for Acquisition (L23)

- 1 Broad band decoupling using waltz65, PULPROG=zgpg
- 11 CW decoupling based on PLW12, automatic O2 optimization, PULPROG=zgcw
- 21 CW decoupling based on PLW12, O2 from parameter set, no optimization, PULPROG=zgcw

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	13C				SI	32768			
NUC2	1H				WDW	0			
PULPROG	zgpg				LB	0.000	Hz		
NS	4				PC	1.400			
DS	0				F1P	0.099	ppm		
RG	101.000			optim. by RGA	F2P	-0.099	ppm		
O1P	66.488	ppm			CY	1000.000	cm		
O2P	3.474	ppm							
SWH	396.825	Hz							
TD	32768								
AQ	41.288	s							
FIDRES	0.024	Hz							
D 1	33.712	s		AQ+D1=const					
P 1	9.0	us		90deg NUC1					
PCPD 2	120.0	us		CPD NUC2					
PLW 1	39.3	W		Pow@90deg(Specs) NUC1					
PLW 12	0.1	W		Pow@CPD NUC2					
PLW 13	0.05	W		Pow@CPD NOE NUC2					
CPDPRG2	waltz65			decoupl. sequence					
TE	298.000	K		default					

## Experiment Description

The Carbon-13 line shape experiment with 1H decoupling using as default waltz65 and decoupler pulse length (PCPD2) of 120 us. The corresponding decoupling powers (PLW12 and PLW13) are calculated from prosol entries for P12 and PLW12.


The exact signal position is determined with NS=1 prior to the line shape measurement. The carrier position is afterwards set to this position

O1 optimized = peakFreqHz - (SWH/4)

This procedure allows the application of a strip transformation with the parameters STSR=0, STSI=SI/2 resulting in the signal position on-resonance and in most cases resulting in an optimal automatic phase correction.

## 5.2.7 P90 13C pulse calibration (NPT\_13C\_p90det\_astm\_13c)

**Test Sample:** 40% Dioxane in Benzene-D6 (ASTM)  
 Z10163, Z100929, Z10164, Z10035, Z10255, Z10242, Z10724  
**Solvent:** C6D6  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments around 360 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 1024	
PARMODE 0	Data Dimension	LB 3.500	Hz
PULPROG zg		F1P 130.975	ppm
NS 1		F2P 123.025	ppm
DS 0		CY 5.500	cm
RG 0.250	optim. by RGA		
SWH 2000.000	Hz		
TD 1048			
AQ 0.262	s		
FIDRES 3.817	Hz		
O1P 127.000	ppm		
P 1 14.0	us		
PLW 1 6.6	W		
DIGMOD 3	90deg Pulse		
DSPFIRM 4	Pow@90deg(Specs)		
TE 298.000	baseopt		
	rectangle		
	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.


The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

## 5.2.8 <sup>13</sup>C ringing test with <sup>1</sup>H decoupling (NPT\_13C\_ringing\_dec1h)

**Test Sample:** 10% Ethylbenzene (EB) in Chloroform-D  
Z10153, Z10154, Z10034, Z10253, Z10292, Z10723  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

<sup>13</sup>C spectrum of 10% ethyl benzene in CDCl<sub>3</sub> scaled to show baseline distortions.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F1 ACQU</b>			Parameters	<b>CNST 10</b>	45.000		Flip angle for P90
NUC1	13C			<b>F1 PROC</b>			Parameters
NUC2	1H			SI	524288		
PULPROG	npt_zg0dc			WDW	1		
NS	500			LB	2.000	Hz	
DS	4			PC	1.400		
RG	101.000		no optim.	F1P	180.000	ppm	
O1P	130.000	ppm		F2P	-20.000	ppm	
O2P	5.000	ppm		CY	11.000	cm	
SW	354.922	ppm					
TD	262144						
AQ	3.670	s	field dependent				
FIDRES	0.272	Hz	field dependent				
D 1	1.230	s	AQ+D1=const				
P 0		us	P 1 * CNST 10 / 90				
P 1	9.0	us	90deg NUC1				
PLW 1	40.2	W	Pow@90deg(Specs) NUC1				
PLW 12	0.1	W	Pow@CPD NUC2				
CPDPRG2	waltz64		decoupl. sequence				
TE	298.000	K	default				
DIGMOD	3		baseopt				
DE	42.000	us	set after getprosol				


## Experiment Description

13C-ringing test full range spectrum. The fullrange spectrum is first baseline corrected before being fitted, without the peaks, with a polynomial function. The ratio distortion/noise=(max(fit)-min(fit))/noise(250 to 210 ppm) gives a measure of the baseline distortion due to ringing. Procno 2 is used for noise determination, procno 3 contains the baseline where all the peaks have been replaced with piece-wise linear noisy sections, whereas procno 4 contains the fit itself.

## 5.2.9 <sup>13</sup>C sensitivity (NPT\_13C\_sensitivity)

---

**Test Sample:** 40% Dioxane in Benzene-D6 (ASTM)  
Z10163, Z100929, Z10164, Z10035, Z10255, Z10242, Z10724, Z142224  
**Solvent:** C6D6  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Carbon-13 sensitivity test.

### Control Option for Acquisition (L23)

1 default



## Parameters

F1 ACQU		Parameters		F1 PROC		Parameters	
NUC1	13C			SI	131072		
PULPROG	zg			WDW	1		
NS	1			LB	3.500	Hz	
DS	0			PC	1.400		
RG	101.000		no optim.	F1P	140.000	ppm	
O1P	99.987	ppm		F2P	60.000	ppm	
SW	198.762	ppm		CY	11.000	cm	
TD	65536						
AQ	1.638	s	field dependent				
FIDRES	0.610	Hz	field dependent				
D 1	828.362	s	AQ+D1=const				
P 1	9.0	us	90deg NUC1				
PLW 1	39.6	W	Pow@90deg(Specs) NUC1				
DIGMOD	3		baseopt				
TE	298.000	K	default				


## Experiment Description

Carbon-13 sensitivity test (no 1H decoupling). Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from 140.0 to 124.0 ppm, while the best 40 ppm noise region is determined over the range from 124.0 to 80.0 ppm.

## 5.2.10 <sup>13</sup>C sensitivity with <sup>1</sup>H decoupling (NPT\_13C\_sensitivity\_dec1h)

---

**Test Sample:** 10% Ethylbenzene (EB) in Chloroform-D  
Z10153, Z100929, Z10154, Z10034, Z10253, Z10292, Z10723  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Carbon-13 sensitivity test with <sup>1</sup>H decoupling.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU		Parameters		F1 PROC		Parameters	
NUC1	13C			SI	524288		
NUC2	1H			WDW	1		
PULPROG	zgpg			LB	0.300	Hz	
NS	1			PC	1.400		
DS	0			F1P	180.000	ppm	
RG	101.000		no optim.	F2P	-20.000	ppm	
O1P	79.987	ppm		CY	11.000	cm	
O2P	4.000	ppm					
SW	198.766	ppm					
TD	262144						
AQ	6.554	s	field dependent				
FIDRES	0.153	Hz	field dependent				
D 1	132.446	s	AQ+D1=const				
P 1	9.0	us	90deg NUC1				
PLW 1	39.6	W	Pow@90deg(Specs) NUC1				
PLW 12	0.1	W	Pow@CPD NUC2				
PLW 13	0.05	W	Pow@CPD NOE NUC2				
DIGMOD	3		baseopt				
CPDPRG2	waltz64		decoupl. sequence				
TE	298.000	K	default				


## Experiment Description

Carbon-13 sensitivity test with 1H decoupling. Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from 140.0 to 124.0 ppm, while the best 40 ppm noise region is determined over the range from 124.0 to 80.0 ppm.

## 5.2.11 15N test for artifacts (NPT\_15N\_fullsw\_inept)

---

**Test Sample:** 90% Formamide (HCONH<sub>2</sub>) in Dimethyl Sulfoxide-D<sub>6</sub>  
Z10187, Z10188, Z10039, Z10256, Z10136, Z142227  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Full range 15N spectrum with 1H decoupling.

### Control Option for Acquisition (L23)

1 default

## Parameters


<b>F1 ACQU</b>			Parameters	<b>CNST 10</b>	45.000		Flip angle for P90
NUC1	15N			<b>F1 PROC</b>			Parameters
NUC2	1H			SI	131072		
PULPROG	ineptd			WDW	1		
NS	32			LB	1.000	Hz	
DS	2			PC	1.000		
RG	101.000		no optim.	F1P	122.000	ppm	
O1P	112.421	ppm		F2P	104.000	ppm	
O2P	7.300	ppm		CY	11.000	cm	
SW	493.226	ppm					
TD	65536						
AQ	1.638	s	field dependent				
FIDRES	0.610	Hz	field dependent				
D 1	2.862	s	AQ+D1=const				
P 0		us	P 1 * CNST 10 / 90				
P 1	14.0	us	90deg NUC1				
PLW 1	82.034	W	Pow@90deg(Specs) NUC1				
PLW 12	0.1	W	Pow@CPD NUC2				
CPDPRG2	waltz65		decoupl. sequence				
TE	298.000	K	default				

## Experiment Description

Full range spectra (spectroscopically useful range) are acquired as test for RF artifacts (spikes). Currently no evaluation is provided and visual inspection of the data is required. Data are treated with baseline correction.

## 5.2.12 P90 15N pulse calibration (NPT\_15N\_p90det\_formamide\_15n)

**Test Sample:** 90% Formamide (HCONH<sub>2</sub>) in Dimethyl Sulfoxide-D<sub>6</sub>  
 Z10187, Z10188, Z10039, Z10256, Z10136, Z142227  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows three experiments around 360 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 15N		SI 4096	
NUC2 1H		LB 0.500	Hz
PARMODE 0	Data Dimension	F1P 113.233	ppm
PULPROG zgig		F2P 110.767	ppm
NS 1		CY 5.500	cm
DS 0			
RG 101.000	optim. by RGA		
SWH 526.316	Hz		
TD 4192			
AQ 3.982	s		
FIDRES 0.251	Hz		
O1P 112.000	ppm		
P 1 14.0	us		90deg Pulse
PLW 1 6.6	W		Pow@90deg(Specs)
O2P 7.200	ppm		
PCPD 2 90.0	us		PCPD2 NUC2
PLW 12 0.13	W		Pow@90deg(Specs)
DIGMOD 3			baseopt
DSPFIRM 4			rectangle
TE 298.000	K		default

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 8 is not achieved. SINO check can be skipped with L23=2, 12, or 22


If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

## 5.2.13 <sup>15</sup>N sensitivity with <sup>1</sup>H decoupling (NPT\_15N\_sensitivity\_dec1h)

---

**Test Sample:** 90% Formamide (HCONH<sub>2</sub>) in Dimethyl Sulfoxide-D<sub>6</sub>  
Z10187, Z10188, Z10039, Z10256, Z10136, Z142227  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Nitrogen-15 sensitivity test with <sup>1</sup>H decoupling.

### Control Option for Acquisition (L23)

1 default



## Parameters

F1 ACQU		Parameters		F1 PROC		Parameters	
NUC1	15N			SI	16384		
NUC2	1H			WDW	1		
PULPROG	zgig			LB	0.300	Hz	
NS	1			PC	1.000		
DS	0			F1P	122.000	ppm	
RG	101.000		no optim.	F2P	104.000	ppm	
O1P	112.421	ppm		CY	11.000	cm	
O2P	7.300	ppm					
SW	20.214	ppm					
TD	32768						
AQ	19.988	s	field dependent				
FIDRES	0.050	Hz	field dependent				
D 1	110.011	s	AQ+D1=const				
P 1	14.0	us	90deg NUC1				
PLW 1	84.4	W	Pow@90deg(Specs) NUC1				
PLW 12	0.2	W	Pow@CPD NUC2				
CPDPRG2	waltz64		decoupl. sequence				
TE	298.000	K	default				

## Experiment Description


Nitrogen-15 sensitivity test with 1H decoupling. Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from 113.0 to 111.0 ppm, while the best 2 ppm noise region is determined over the range from 122.0 to 113.0 ppm.

Before the acquisition, the proton 90 degrees pulse is calibrated using pulscal, which result is stored in the corresponding derived dataset. The pulse determination is skipped if experiment is measured with option 'Skip Getprosol'.

## 5.2.14 <sup>15</sup>N sensitivity (INEPT) with <sup>1</sup>H decoupling (NPT\_15N\_sensitivity\_inept)

---

**Test Sample:** 90% Formamide (HCONH<sub>2</sub>) in Dimethyl Sulfoxide-D<sub>6</sub>  
Z10187, Z10188, Z10039, Z10256, Z10136, Z142227  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Nitrogen-15 sensitivity test based on INEPT (with <sup>1</sup>H decoupling).

### Control Option for Acquisition (L23)

- 1 default
- 2 no decoupling during acquisition

## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	15N			SI	16384		field dependent
NUC2	1H			WDW	1		
PULPROG	ineptd			LB	0.300	Hz	
NS	8			PC	1.000		
DS	2			F1P	122.000	ppm	
RG	101.000		no optim.	F2P	104.000	ppm	
O1P	112.421	ppm		CY	11.000	cm	
O2P	7.300	ppm					
SW	20.214	ppm					
TD	4918		field dependent				
AQ	3.000	s					
FIDRES	0.333	Hz	field dependent				
D 1	10.500	s					
CPDPRG2	waltz64		decoupl. sequence				
CNST 2	88.000	Hz	J(NH) coupling				
CNST 11	6.000		6=all NH[n] pos.				
P 1	14.0	us	90deg NUC1				
P 3	11.0	us	90deg NUC2				
PLW 1	84.4	W	Pow@90deg(Specs) NUC1				
PLW 2	19.3	W	Pow@90deg(Specs) NUC2				
PLW 12	0.08	W	Pow@CPD NUC2				
TE	298.000	K	default				

## Experiment Description

Nitrogen-15 sensitivity test based on INEPT pulse sequence with 1H decoupling during acquisition. Signal enhancement compared to 15N{1H} experiment should be  $\sim BF1[1H]/BF1[15N]=10*\sqrt{NS=8}$ .


Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from 113.0 to 111.0 ppm, while the best 2 ppm noise region is determined over the range from 122.0 to 113.0 ppm.

Before the acquisition, the proton 90 degrees pulse is calibrated using pulscal, which result is stored in the corresponding derived dataset. The pulse determination is skipped if experiment is measured with option 'Skip Getprosol'.

Option L23=1 is standard whereas option L23=2 is non-standard and will not be considered as regular test from the 'NMRPT Control Structure'.

## 5.2.15 <sup>19</sup>F B1 homogeneity integral (NPT\_19F\_b1homogeneityInt\_19f)

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
 Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a fraction around signal maximum (-60 deg, +60 deg) is acquired and shown in the figure. The variable pulse list which is used for acquisition is shown to the right.

The B1-homogeneities based on amplitudes and integrals are computed from the peak-picking list amplitudes and the magnitude of the second complex FID data point respectively.

### Control Option for Acquisition (L23)

- 1 default PLW 1 is used for determination of B1 homogeneity.
- 2 PLW 1 will be set to CNST 10 and resultant pulses are used for determination of B1 homogeneity. Experiment will be set to irregular.
- 3 PLW 1 will be set to power corresponding to specified customer pulses and resultant pulses are used for determination of B1 homogeneity.

## Parameters

<b>F2 ACQU</b>		Parameters F2	<b>F2 PROC</b>		Parameters F2
NUC1	19F		SI	4096	
PARMODE	1	Data Dimension	WDW	1	
PULPROG	npt_p1b1hom2d		LB	1.000	Hz
NS	1		PH_mod	1	pk
DS	0		ME_mod	2	LPfc
RG	101.000	optim. by RGA	NCOEF	20	
O1P	-62.766		ABSF1	1000.000	ppm
SWH	396.825		ABSF2	-1000.000	ppm
TD	1024		F1P	-62.574	ppm
AQ	1.290		F2P	-62.874	ppm
FIDRES	0.775		<b>F1 ACQU</b>		Parameters F1
D 1	23.527	AQ+D1=const	NUC1	19F	
P 1	14.0	90deg NUC1	TD	43	No of incr.
PLW 1	6.6	Pow@90deg(Specs) NUC1	<b>F1 PROC</b>		Parameters F1
TE	298.000	default	SI	64	
			<b>NMRPT</b>		Parameters
			L 4	6	integ. fraction of 90deg
			L 5	8	# of step per maxima

## Experiment Description

The B1 homogeneity measurement is normally acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range is from 90 to 810 degree. In order to speed up the measurements NMRPT is acquiring only the partial ranges around the positive signal maxima for the pulse angles 90, 450, and 810 degree.


By this approach it is possible to get a better incremental resolution and a shorter overall measuring time. Prior to the main acquisition O1P (PROCNO=2) and phase correction (PROCNO=3) will be optimized. Intermediate results of the CONVTO1D routine is stored in PROCNO=998, result of CONVTO1D in PROCNO=999.

As control of stability, reproducibility and 90 degree calibration a spectrum with 180 degree pulse is acquired before and after each maxima. If relaxation conditions and instrument stability is sufficient, the result should not differ among them.

For the prediction of the signal maxima at 450 and 810 degree the propagation delay is included in the calculation of the VP list. (PDelay is calculated as difference of 360 and 180 degree pulse lengths during the corresponding 90 degree pulse determination experiment.) In this measurement O1 is determined before the main acquisition is started. Evaluation is achieved after processing the data in F2 only (XF2) by peak picking and integral determination of the most intense signal per maxima.

## 5.2.16 <sup>19</sup>F B1 homogeneity integral on H-coil (NPT\_19F\_b1homogeneityInt\_hcoil)

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a fraction around signal maximum (-60 deg, +60 deg) is acquired and shown in the figure. The variable pulse list which is used for acquisition is shown to the right.

The B1-homogeneities based on amplitudes and integrals are computed from the peak-picking list amplitudes and the magnitude of the second complex FID data point respectively.

### Control Option for Acquisition (L23)

- 1 default PLW 1 is used for determination of B1 homogeneity.
- 2 PLW 1 will be set to CNST 10 and resultant pulses are used for determination of B1 homogeneity. Experiment will be set to irregular.
- 3 PLW 1 will be set to power corresponding to specified customer pulses and resultant pulses are used for determination of B1 homogeneity.

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	19F			SI	4096		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_p1b1hom2d			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	0			ME_mod	2		LPfc
RG	101.000		optim. by RGA	NCOEF	20		
O1P	-62.766	ppm		ABSF1	1000.000	ppm	
SWH	396.825	Hz		ABSF2	-1000.000	ppm	
TD	1024			F1P	-62.574	ppm	
AQ	1.290	s		F2P	-62.874	ppm	
FIDRES	0.775	Hz		<b>F1 ACQU</b>			Parameters F1
D 1	23.527	s	AQ+D1=const	NUC1	19F		
P 1	14.0	us	90deg NUC1	TD	43		No of incr.
PLW 1	6.6	W	Pow@90deg(Specs) NUC1	<b>F1 PROC</b>			Parameters F1
TE	298.000	K	default	SI	64		
				<b>NMRPT</b>			Parameters
				L 4	6		integ. fraction of 90deg
				L 5	8		# of step per maxima

## Experiment Description

The B1 homogeneity measurement is normally acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range is from 90 to 810 degree. In order to speed up the measurements NMRPT is acquiring only the partial ranges around the positive signal maxima for the pulse angles 90, 450, and 810 degree.


By this approach it is possible to get a better incremental resolution and a shorter overall measuring time. Prior to the main acquisition O1P (PROCNO=2) and phase correction (PROCNO=3) will be optimized. Intermediate results of the CONVTO1D routine is stored in PROCNO=998, result of CONVTO1D in PROCNO=999.

As control of stability, reproducibility and 90 degree calibration a spectrum with 180 degree pulse is acquired before and after each maxima. If relaxation conditions and instrument stability is sufficient, the result should not differ among them.

For the prediction of the signal maxima at 450 and 810 degree the propagation delay is included in the calculation of the VP list. (PDelay is calculated as difference of 360 and 180 degree pulse lengths during the corresponding 90 degree pulse determination experiment.) In this measurement O1 is determined before the main acquisition is started. Evaluation is achieved after processing the data in F2 only (XF2) by peak picking and integral determination of the most intense signal per maxima.

## 5.2.17 <sup>19</sup>F background with sample and <sup>1</sup>H decoupling (NPT\_19F\_backgr\_withsample)

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

<sup>19</sup>F Background signal spectrum with sample. Sharp signal arises from sample, broad signal around -175 ppm arises from solid compound in the probe.

### Control Option for Acquisition (L23)

1 default



## Parameters

<b>F1 ACQU</b>		Parameters	<b>CNST 10</b>	45.000	Flip angle for P90
NUC1	19F		<b>F1 PROC</b>		Parameters
NUC2	1H		SI	131072	
PULPROG	npt_zg0ig		WDW	1	
NS	1000		LB	2.000	Hz
DS	0		PC	1.000	
RG	32.000	no optim.	F1P	24.747	ppm
O1P	-139.998	ppm	F2P	-150.759	ppm
O2P	5.000	ppm			
CPDPRG2	waltz64	decoupl. sequence			
SWH	147058.828	Hz			
TD	131072				
AQ	0.446	s			
FIDRES	2.244	Hz			
D 1	1.000	s			
P 0		us			
P 1	9.0	us			
PLW 1	25.1	W			
PLW 12	0.23	W			
TE	298.000	K			
		AQ+D1=const			
		P 1 * CNST 10 / 90			
		90deg NUC1			
		Pow@90deg(Specs) NUC1			
		Pow@CPD NUC2			
		default			

## Experiment Description


Background signal measurements are executed using a small flip angle due to two reasons. First, to get more efficient excitation bandwidth, secondly to compensate for usually long T1 relaxation time of solid signals.

Full spectrum is normally printed without any baseline correction.

Spectrum is processed with MC always.

## 5.2.18 <sup>19</sup>F background with sample on H-coil (NPT\_19F\_backgr\_withsample\_hcoil)

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

<sup>19</sup>F Background signal spectrum with sample. Sharp signal arises from sample, broad signal around -175 ppm arises from solid compound in the probe.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F1 ACQU</b>			Parameters	<b>CNST 10</b>	45.000		Flip angle for P90
NUC1	19F			<b>F1 PROC</b>			Parameters
NUC2	off			SI	131072		
PULPROG	npt_zg0			WDW	1		
NS	1000			LB	2.000	Hz	
DS	0			PC	1.000		
RG	32.000		no optim.	F1P	24.747	ppm	
O1P	-139.998	ppm		F2P	-150.759	ppm	
O2P	-139.998	ppm					
CPDPRG2	waltz64		decoupl. sequence				
SWH	147058.828	Hz					
TD	131072						
AQ	0.446	s					
FIDRES	2.244	Hz					
D 1	1.000	s	AQ+D1=const				
P 0		us	P 1 * CNST 10 / 90				
P 1	15.0	us	90deg NUC1				
PLW 1	10.0	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				

## Experiment Description


Background signal measurements are executed using a small flip angle due to two reasons. First, to get more efficient excitation bandwidth, secondly to compensate for usually long T1 relaxation time of solid signals.

Full spectrum is normally printed without any baseline correction.

Spectrum is processed with MC always.

## 5.2.19 CPD 19F pulse calibration (NPT\_19F\_cpddeterminationf1\_19f)

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
 Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 19F		SI 16384	
PARMODE 0	Data Dimension	WDW 3	
PULPROG zg		LB 2.000 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 0.250	optim. by RGA	ME_mod 2	LPfc
O1P -62.800 ppm		NCOEF 20	
SWH 396.825 Hz		ABSF1 -57.199 ppm	
TD 1000		ABSF2 -68.200 ppm	
AQ 1.260 s		F1P -57.700 ppm	
FIDRES 0.794 Hz		F2P -67.700 ppm	
D 1 4.300 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.


A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.20 <sup>19</sup>F test for artifacts (NPT\_19F\_fullsw\_dec1h)

---

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Full range <sup>19</sup>F spectrum with <sup>1</sup>H decoupling.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	19F			SI	131072		
NUC2	1H			WDW	1		
PULPROG	npt_zg0ig			LB	2.000	Hz	
NS	64			PC	1.000		
DS	4			F1P	24.747	ppm	
RG	101.000		no optim.	F2P	-150.759	ppm	
O1P	-62.766	ppm		CY	11.000	cm	
O2P	5.000	ppm					
SW	132.811	ppm					
TD	65536						
AQ	0.655	s	field dependent				
FIDRES	1.526	Hz	field dependent				
D 1	0.545	s	AQ+D1=const				
P 1	9.0	us	90deg NUC1				
PLW 1	22.473	W	Pow@90deg(Specs) NUC1				
PLW 12	0.1	W	Pow@CPD NUC2				
CPDPRG2	waltz64		decoupl. sequence				
TE	298.000	K	default				

## Experiment Description

Full range spectra (spectroscopically useful range) are acquired as test for RF artifacts (spikes). Currently no evaluation is provided and visual inspection of the data is required. Data are treated with baseline correction.


## 5.2.21 P90 19F pulse calibration (NPT\_19F\_p90determinationf1\_19f)

**Test Sample:** (a) 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
 (b) 45% Chloroform-D (CDCl<sub>3</sub>) and 45% Chloroform (CHCl<sub>3</sub>) in 10% Hexafluorobenzene (C<sub>6</sub>F<sub>6</sub>).  
 Z10234, Z100937, Z10235, Z10040, Z10728, Z10079, Z142228

**Solvent:** (a) CDCl<sub>3</sub>  
 (b) C<sub>6</sub>F<sub>6</sub>

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 19F		SI 16384	
PARMODE 0	Data Dimension	WDW 3	(TFT Sample)
PULPROG zg		WDW 1	(Fluor Lock Sample)
NS 1		LB 2.000 Hz	(TFT Sample)
DS 0		LB 1.0 Hz	(Fluor Lock Sample)
RG 0.250	optim. by RGA	SSB 2.000	
O1P -62.800 ppm	(TFT Sample)	PH_mod 1	pk
O1P -162.2 ppm	(Fluor Lock Sample)	ME_mod 2	LPfc
SWH 396.825 Hz		NCOEF 20	
TD 1000	(TFT Sample)	ABSF1 -57.199 ppm	
TD 4096	(Fluor Lock Sample)	ABSF2 -68.200 ppm	
AQ 1.260 s		F1P -57.700 ppm	
FIDRES 0.794 Hz		F2P -67.700 ppm	
D1 4.300 s	AQ+D1=const (TFT Sample)	CY 11.000 cm	
D1 1.0 s	(Fluor Lock Sample)		
P1 14.0 us	90deg NUC1		
PLW1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22


If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.22 19F P90 pulse calibration on H-coil (NPT\_19F\_p90determinationf1\_hcoil)

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
 Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
  - 2 skip SINO check on PROCNO 11
  - 11 ignore specifications (optimize power for pulse length from prosol)
  - 12 ignore specifications and skip SINO check on PROCNO 11
  - 21 ignore specifications and optimize pulse length for power from prosol
  - 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
  - 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX  
 1000Same as xxx but skip automatic O1P determination  
 +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 19F		SI 16384	
PARMODE 0	Data Dimension	WDW 3	
PULPROG zg		LB 2.000 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 0.250	optim. by RGA	ME_mod 2	LPfc
O1P -62.800 ppm		NCOEF 20	
SWH 396.825 Hz		ABSF1 -57.199 ppm	
TD 1000		ABSF2 -68.200 ppm	
AQ 1.260 s		F1P -57.700 ppm	
FIDRES 0.794 Hz		F2P -67.700 ppm	
D 1 4.300 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.


A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.23 <sup>19</sup>F sensitivity (NPT\_19F\_sensitivity)

---

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Fluorine-19 sensitivity test.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	19F			SI	32768		
PULPROG	zg			WDW	1		
NS	1			LB	2.000	Hz	
DS	0			PC	1.000		
RG	101.000		no optim.	F1P	-58.015	ppm	
O1P	-62.766	ppm		F2P	-67.985	ppm	
SW	9.838	ppm		CY	11.000	cm	
TD	32768						
AQ	4.424	s	field dependent				
FIDRES	0.226	Hz	field dependent				
D 1	30.576	s	AQ+D1=const				
P 1	9.0	us	90deg NUC1				
PLW 1	25.1	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				


## Experiment Description

Fluorine-19 sensitivity test (no 1H decoupling). Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from -61.0 to -65.0 ppm, while the best 1 ppm noise region is determined over the range from -58.5 to -63.0 ppm.

## 5.2.24 <sup>19</sup>F sensitivity on H-coil (NPT\_19F\_sensitivity\_hcoil)

---

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Fluorine-19 sensitivity test on 1H-coil tuned to 19F.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	19F			SI	32768		
PULPROG	zg			WDW	1		
NS	1			LB	2.000	Hz	
DS	0			PC	1.000		
RG	101.000		no optim.	F1P	-58.015	ppm	
O1P	-62.766	ppm		F2P	-67.985	ppm	
SW	9.838	ppm		CY	11.000	cm	
TD	32768						
AQ	4.424	s	field dependent				
FIDRES	0.226	Hz	field dependent				
D 1	30.576	s	AQ+D1=const				
P 1	9.0	us	90deg NUC1				
PLW 1	25.1	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				


## Experiment Description

Fluorine-19 sensitivity test (no 1H decoupling). Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from -61.0 to -65.0 ppm, while the best 1 ppm noise region is determined over the range from -58.5 to -63.0 ppm.

## 5.2.25 <sup>19</sup>F sensitivity with 1H decoupling and LB=0.5 (NPT\_19F\_sensitivity\_lb05\_dec1h)

---

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Fluorine-19 sensitivity test (LB = 0.5) with 1H decoupling.

### Control Option for Acquisition (L23)

1 default



## Parameters

F1 ACQU		Parameters		F1 PROC		Parameters	
NUC1	19F			SI	32768		
NUC2	1H			WDW	1		
PULPROG	zgig			LB	0.500	Hz	
NS	1			PC	1.000		
DS	0			F1P	-58.015	ppm	
SW	19.823	ppm		F2P	-67.985	ppm	
RG	101.000		no optim.	CY	11.000	cm	
O1P	-62.000	ppm					
O2P	8.000	ppm					
SW	19.823	ppm					
TD	44776						
AQ	3.000	s	field dependent				
FIDRES	0.333	Hz	field dependent				
D 1	35.000	s					
P 1	9.0	us	90deg NUC1				
PLW 1	22.5	W	Pow@90deg(Specs) NUC1				
PLW 12	0.1	W	Pow@CPD NUC2				
CPDPRG2	waltz64		decoupl. sequence				
TE	298.000	K	default				

## Experiment Description

Fluorine-19 sensitivity test with 1H decoupling. Processing is using line broadening (LB) of 0.5 Hz and baseline correction (ABS). Evaluation is carried out by the AU `hwcal` to determine the line width at 50% of signal height and by the AU `sinocal`. The signal is searched over the range from -62.0 to -63.0 ppm, while the best 1 ppm noise region is determined over the range from -53.0 to -62.0 ppm.

## 5.2.26 13 degree pulse stability test (NPT\_1H\_13degtest)


---

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727

**Solvent:** D<sub>2</sub>O

**Lock parameter:** Lockregulation based on LGAIN=80 dB

**Sample State:** Rotation off



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the residual water signal of the sample.

### Control Option for Acquisition (L23)

1 default

## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H          PARMODE 1          PULPROG sys13deg          NS 1          DS 8          RG 0.250          O1P 4.704 ppm          SW 9.917 ppm          TD 8192          AQ 1.032 s          FIDRES 0.969 Hz          D 1 2.000 s          D 20 0.001 s          P 1 11.0 us          PLW 1 19.3 W          TE 298.000 K</p> <p><b>F2 PROC</b></p> <p>SI 16384          WDW 1          LB 1.000 Hz          SSB 0.000          PH_mod 1          F1P 5.204 ppm          F2P 4.204 ppm</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>optim. by RGA</p> <p>field dependent</p> <p>field dependent</p> <p>90deg NUC1          Pow@90deg(Specs) NUC1          default</p> <p>Parameters F2</p> <p>pk</p>
<p><b>F1 ACQU</b></p> <p>NUC1 1H          TD 32</p> <p><b>F1 PROC</b></p> <p>SI 32</p> <p><b>NMRPT</b></p> <p>CNST 45 -1.000          CNST 46 -1.000</p>	<p>Parameters F1</p> <p>Parameters F1</p> <p>Parameters          Return Mean Value          Return Std Dev</p>

## Experiment Description


Purpose of this experiment is the measurement of overall pulse stability of the spectrometer, according to the methods described in 'Assessment of Spectrometer Pulse Reproducibility' by G.A. Morris, JMR, 78, 281ff (1988).

For this hardware test (HWT) some experimental preparation is needed. Initially the exact resonance position of the residual water signal is determined after receiver gain adjustment based on a single scan acquisition followed by processing with FT, MC, PP. The processed data are stored in PROCNO=2. The setting of special LOCK settings is executed prior to the start of the main acquisition. Data are processed with LB and no baseline correction. Evaluation is based on peak picking after conversion of the processed data to 1D-mode using CONVTO1D procedure.

## 5.2.27 30 degree pulse stability test (NPT\_1H\_30degtest)

---

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** Lockregulation based on LGAIN=80 dB  
**Sample State:** Rotation off



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the residual water signal of the sample.

### Control Option for Acquisition (L23)

1 default

## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H          PARMODE 1          PULPROG npt_zg2dp0          NS 1          DS 8          RG 0.250          O1P 4.704 ppm          SW 9.917 ppm          TD 8192          AQ 1.032 s          FIDRES 0.969 Hz          D 1 2.000 s          P 0 us          P 1 11.0 us          PLW 1 19.3 W          CNST 10 30.000 deg          TE 298.000 K</p> <p><b>F2 PROC</b></p> <p>SI 16384          WDW 1          LB 1.000 Hz          SSB 0.000          PH_mod 1          F1P 5.204 ppm          F2P 4.204 ppm</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>optim. by RGA</p> <p>field dependent</p> <p>field dependent</p> <p>P 1 * CNST 10 / 90          90deg NUC1          Pow@90deg(Specs) NUC1          flip angle          default</p> <p>Parameters F2</p> <p>pk</p>
<p><b>F1 ACQU</b></p> <p>NUC1 1H          TD 32</p> <p><b>F1 PROC</b></p> <p>SI 32</p> <p><b>NMRPT</b></p> <p>CNST 45 -1.000          CNST 46 -1.000</p>	<p>Parameters F1</p> <p>Parameters F1</p> <p>Parameters          Return Mean Value          Return Std Dev</p>

## Experiment Description

Purpose of this experiment is the measurement of overall amplitude stability of the spectrometer using a 30degree excitation pulse.

For this hardware test (HWT) some experimental preparation is needed. Initially the exact resonance position of the residual water signal is determined after receiver gain adjustment based on a single scan acquisition followed by processing with FT, MC, PP. The processed data are stored in PROCNO=2. The setting of special LOCK settings is executed prior to the start of the main acquisition.

Data are processed with LB and no baseline correction. Evaluation is based on peak picking after conversion of the processed data to 1D-mode using CONVTO1D procedure.

## 5.2.28 90 degree pulse stability test (NPT\_1H\_90degtest)


---

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727

**Solvent:** D<sub>2</sub>O

**Lock parameter:** Lockregulation based on LGAIN=80 dB

**Sample State:** Rotation off



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the residual water signal of the sample.

### Control Option for Acquisition (L23)

1 default

## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H          PARAMODE 1          PULPROG npt_zg2dp0          NS 1          DS 8          RG 0.250          O1P 4.704 ppm          SW 9.917 ppm          TD 8192          AQ 1.032 s          FIDRES 0.969 Hz          D 1 2.000 s          P 0 us          P 1 11.0 us          PLW 1 19.3 W          CNST 10 90.000 deg          TE 298.000 K</p> <p><b>F2 PROC</b></p> <p>SI 16384          WDW 1          LB 1.000 Hz          SSB 0.000          PH_mod 1          F1P 5.204 ppm          F2P 4.204 ppm</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>optim. by RGA</p> <p>field dependent</p> <p>field dependent</p> <p>P 1 * CNST 10 / 90          90deg NUC1          Pow@90deg(Specs) NUC1          flip angle          default</p> <p>Parameters F2</p> <p>pk</p>
<p><b>F1 ACQU</b></p> <p>NUC1 1H          TD 32</p> <p><b>F1 PROC</b></p> <p>SI 32</p> <p><b>NMRPT</b></p> <p>CNST 45 -1.000          CNST 46 -1.000</p>	<p>Parameters F1</p> <p>Parameters F1</p> <p>Parameters          Return Mean Value          Return Std Dev</p>

## Experiment Description

Purpose of this experiment is the measurement of overall amplitude stability of the spectrometer using a 90 degree excitation pulse.


For this hardware test (HWT) some experimental preparation is needed. Initially the exact resonance position of the residual water signal is determined after receiver gain adjustment based on a single scan acquisition followed by processing with FT, MC, PP. The processed data are stored in PROCNO=2. The setting of special LOCK settings is executed prior to the start of the main acquisition.

Data are processed with LB and no baseline correction. Evaluation is based on peak picking after conversion of the processed data to 1D-mode using CONVTO1D procedure.

## 5.2.29 Amplitude stability after gradient echo with strong gradient pulses (NPT\_1H\_ampStabGradientEchoStrong)

---

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-<sup>13</sup>C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode, showing the residual water peak after a gradient echo.

### Control Option for Acquisition (L23)

1 default



## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F1 ACQU</b>			Parameters F1
NUC1	1H			NUC1	1H		
PARAMODE	1		Data Dimension	TD	32		
PULPROG	syszggegp2d			<b>F1 PROC</b>			Parameters F1
NS	1			SI	32		
DS	8						
RG	101.000		no optim.				
O1P	4.697	ppm					
SW	16.442	ppm					
TD	8192						
AQ	0.623	s	field dependent				
FIDRES	1.606	Hz	field dependent				
D 1	0.200	s					
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPNAM1	RECT.1						
GPNAM2	RECT.1						
GPZ 1	60.000	%					
GPZ 2	-60.000	%					
P 16	5000.000	us	gradient pulse				
TE	298.000	K	default				
<b>F2 PROC</b>			Parameters F2				
SI	16384						
WDW	1						
LB	1.000	Hz					
SSB	0.000						
PH_mod	0		pk				
F1P	5.250	ppm					
F2P	4.250	ppm					

## Experiment Description

The purpose of this test is the assessment of the amplitude stability after a gradient echo. After a 90 degree pulse, the generated coherence is first defocused and subsequently refocused by a pair of gradient pulses with equal but opposite strengths separated by a delay D16=500 us. The gradient echo which builds up after another delay D16 is finally recorded.


The quality criteria of this test are defined by the mean relative amplitude and the standard deviation of the relative amplitude, both in percent, of a series of identical experiments.

In this test, the pair of gradient pulse strengths are +/-60% of the maximum strength of the probe gradient.

## 5.2.30 Amplitude stability after gradient echo with weak gradient pulses (NPT\_1H\_ampStabGradientEchoWeak)

---

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-1<sup>3</sup>C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode, showing the residual water peak after a gradient echo.

### Control Option for Acquisition (L23)

1 default

## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H</p> <p>PARAMODE 1</p> <p>PULPROG syszggegp2d</p> <p>NS 1</p> <p>DS 8</p> <p>RG 101.000</p> <p>O1P 4.697 ppm</p> <p>SW 16.442 ppm</p> <p>TD 8192</p> <p>AQ 0.623 s</p> <p>FIDRES 1.606 Hz</p> <p>D 1 0.200 s</p> <p>P 1 14.0 us</p> <p>PLW 1 6.6 W</p> <p>GPNAM1 RECT.1</p> <p>GPNAM2 RECT.1</p> <p>GPZ 1 20.000 %</p> <p>GPZ 2 -20.000 %</p> <p>P 16 5000.000 us</p> <p>TE 298.000 K</p> <p><b>F2 PROC</b></p> <p>SI 16384</p> <p>WDW 1</p> <p>LB 1.000 Hz</p> <p>SSB 0.000</p> <p>PH_mod 0</p> <p>F1P 5.250 ppm</p> <p>F2P 4.250 ppm</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>no optim.</p> <p>field dependent</p> <p>field dependent</p> <p>90deg Pulse</p> <p>Pow@90deg(Specs)</p> <p>gradient pulse default</p> <p>Parameters F2</p> <p>pk</p>
<p><b>F1 ACQU</b></p> <p>NUC1 1H</p> <p>TD 32</p> <p><b>F1 PROC</b></p> <p>SI 32</p>	<p>Parameters F1</p> <p>Parameters F1</p>

## Experiment Description

The purpose of this test is the assessment of the amplitude stability after a gradient echo. After a 90 degree pulse, the generated coherence is first defocused and subsequently refocused by a pair of gradient pulses with equal but opposite strengths separated by a delay D16=500 us. The gradient echo which builds up after another delay D16 is finally recorded.

The quality criteria of this test are defined by the mean relative amplitude and the standard deviation of the relative amplitude, both in percent, of a series of identical experiments.

In this test, the pair of gradient pulse strengths are +/-20% of the maximum strength of the probe gradient.


## 5.2.31 B0 magnet drift experiment (NPT\_1H\_b0drifttest)

**Test Sample:** 0.3%, 1.0% or 3.0% Chloroform in Acetone-D6  
Z10230, Z10248, Z10701, Z100926, Z10250, Z10031, Z10030, Z10029, Z10260,  
Z10249, Z10275, Z10272, Z10717

**Solvent:** Acetone

**Lock parameter:** LOCK is off during main acquisition

**Sample State:** Rotation off



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the chloroform signal.

### Control Option for Acquisition (L23)

1 duration of drift test according to specification for minimal drift time, 6 hours if not specified  
10- duration of drift test in minutes  
6000

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	65536		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_zgp02d			LB	1.000	Hz	
NS	1			SSB	0.000		
DS	0			PH_mod	1		pk
RG	101.000		no optim.	F1P	11.004	ppm	
O1P	8.000	ppm		F2P	8.502	ppm	
SWH	8196.722	Hz		<b>F1 ACQU</b>			Parameters F1
TD	65536			NUC1	1H		
AQ	3.998	s		TD	256		
FIDRES	0.250	Hz		<b>F1 PROC</b>			Parameters F1
D 1	59.500	s		SI	256		
D 20	337.500	s	time per scan				
P 0		us	P 1 * CNST 10 / 90				
P 1	14.0	us	90deg Pulse				
PLW 1	5.6	W	Pow@90deg(Specs)				
CNST 10	45.000	deg	flip angle				
TE	298.000	K	default				


## Experiment Description

Purpose of this experiment is the measurement of the magnet drift rate in unlocked state, but the experiment starts in locked state.

For this hardware test some experimental preparation is needed. Initially the exact resonance position of the chloroform signal is determined after receiver gain adjustment based on a single scan acquisition followed by processing with FT, MC, PP. The data are stored in the derived dataset with EXPNO=1. Data are processed with LB and no baseline correction. Evaluation is based on peak picking of the first and the last acquisition. The printing includes all measurements in pseudo-2D mode.

## 5.2.32 <sup>13</sup>C B1 homogeneity integral (NPT\_1H\_b1homogeneityInt\_13c)

**Test Sample:** 100 mM Urea-15N ([<sup>15</sup>NH<sub>2</sub>]<sup>2</sup>CO) and 100 mM Methanol-<sup>13</sup>C in Dimethyl Sulfoxide-D<sub>6</sub> Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721, Z142223  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a fraction around signal maximum (-60 deg, +60 deg) is acquired and shown in the figure. The variable pulse list which is used for acquisition is shown to the right.

The B1-homogeneities based on amplitudes and integrals are computed from the peak-picking list amplitudes and the magnitude of the second complex FID data point respectively.

### Control Option for Acquisition (L23)

- 1 default PLW 2 is used for determination of B1 homogeneity.
- 2 PLW 2 will be set to CNST 10 and resultant pulses are used for determination of B1 homogeneity. Experiment will be set to irregular.
- 3 PLW 2 will be set to power corresponding to specified customer pulses and resultant pulses are used for determination of B1 homogeneity.

## Parameters

<b>F2 ACQU</b>		Parameters F2	<b>F2 PROC</b>		Parameters F2
NUC1	1H		SI	4096	
PARMODE	1	Data Dimension	WDW	1	
PULPROG	npt_p3b1hom2d		LB	1.000	Hz
NS	1		PH_mod	1	pk
DS	4		ME_mod	0	LPfc
RG	0.250	optim. by RGA	NCOEF	0	
O1P	3.080		ABSF1	3.150	ppm
SWH	230.766		ABSF2	2.850	ppm
TD	1024		F1P	3.150	ppm
AQ	2.219		F2P	2.850	ppm
FIDRES	0.451		<b>F1 ACQU</b>		Parameters F1
D 1	8.804	AQ+D1=const	NUC1	1H	
P 1	14.0	90deg NUC1	TD	43	No of incr.
PLW 1	6.6	Pow@90deg(Specs) NUC1	<b>F1 PROC</b>		Parameters F1
P 3	9.0	90deg NUC1	SI	64	
PLW 2	42.0	Pow@90deg(Specs) NUC2	<b>NMRPT</b>		Parameters
TE	298.000	default	L 4	6	integ. fraction of 90deg
			L 5	8	# of step per maxima

## Experiment Description

The B1 homogeneity measurement is normally acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range is from 0 to 720 degree. In order to speed up the measurements NMRPT is acquiring only the partial ranges around the positive signal maxima for the pulse angles 0, 360, and 720 degree.


By this approach it is possible to get a better incremental resolution and a shorter overall measuring time. Prior to the main acquisition O1P (PROCNO=2) and phase correction (PROCNO=3) will be optimized. Intermediate results of the CONVTO1D routine is stored in PROCNO=998, result of CONVTO1D in PROCNO=999.

As control of stability, reproducibility and 90 degree calibration a spectrum with 180 degree pulse is acquired before and after each maxima. If relaxation conditions and instrument stability is sufficient, the result should not differ among them.

For the prediction of the signal maxima at 360 and 720 degree the propagation delay is included in the calculation of the VP list. (PDelay is calculated as difference of 270 and 90 degree pulse lengths during the corresponding 90 degree pulse determination experiment.) In this measurement O1 is determined before the main acquisition is started. Evaluation is achieved after processing the data in F2 only (XF2) by peak picking and integral determination of the most intense signal per maxima.

## 5.2.33 15N B1 homogeneity integral (NPT\_1H\_b1homogeneityInt\_15n)

**Test Sample:** 100 mM Urea-15N ([15NH<sub>2</sub>]<sup>2</sup>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub> Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721, Z142223  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a fraction around signal maximum (-60 deg, +60 deg) is acquired and shown in the figure. The variable pulse list which is used for acquisition is shown to the right.

The B1-homogeneities based on amplitudes and integrals are computed from the peak-picking list amplitudes and the magnitude of the second complex FID data point respectively.

### Control Option for Acquisition (L23)

- 1 default PLW 2 is used for determination of B1 homogeneity.
- 2 PLW 2 will be set to CNST 10 and resultant pulses are used for determination of B1 homogeneity. Experiment will be set to irregular.
- 3 PLW 2 will be set to power corresponding to specified customer pulses and resultant pulses are used for determination of B1 homogeneity.



## Parameters

<b>F2 ACQU</b>		Parameters F2	<b>F2 PROC</b>		Parameters F2
NUC1	1H		SI	4096	
PARMODE	1	Data Dimension	WDW	1	
PULPROG	npt_p3b1hom2d		LB	1.000	Hz
NS	1		PH_mod	1	pk
DS	4		ME_mod	2	LPfc
RG	0.250	optim. by RGA	NCOEF	20	
O1P	5.500		ABSF1	1000.000	ppm
SWH	230.766		ABSF2	-1000.000	ppm
TD	1024		F1P	5.720	ppm
AQ	2.219		F2P	5.320	ppm
FIDRES	0.451		<b>F1 ACQU</b>		Parameters F1
D 1	0.381	AQ+D1=const	NUC1	1H	
P 1	14.0	90deg NUC1	TD	43	No of incr.
PLW 1	6.6	Pow@90deg(Specs) NUC1	<b>F1 PROC</b>		Parameters F1
P 3	14.0	90deg NUC1	SI	64	
PLW 2	86.0	Pow@90deg(Specs) NUC2	<b>NMRPT</b>		Parameters
TE	298.000	default	L 4	6	integ. fraction of 90deg
			L 5	8	# of step per maxima

## Experiment Description

The B1 homogeneity measurement is normally acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range is from 0 to 720 degree. In order to speed up the measurements NMRPT is acquiring only the partial ranges around the positive signal maxima for the pulse angles 0, 360, and 720 degree.


By this approach it is possible to get a better incremental resolution and a shorter overall measuring time. Prior to the main acquisition O1P (PROCNO=2) and phase correction (PROCNO=3) will be optimized. Intermediate results of the CONVTO1D routine is stored in PROCNO=998, result of CONVTO1D in PROCNO=999.

As control of stability, reproducibility and 90 degree calibration a spectrum with 180 degree pulse is acquired before and after each maxima. If relaxation conditions and instrument stability is sufficient, the result should not differ among them.

For the prediction of the signal maxima at 360 and 720 degree the propagation delay is included in the calculation of the VP list. (PDelay is calculated as difference of 270 and 90 degree pulse lengths during the corresponding 90 degree pulse determination experiment.) In this measurement O1 is determined before the main acquisition is started. Evaluation is achieved after processing the data in F2 only (XF2) by peak picking and integral determination of the most intense signal per maxima.

## 5.2.34 1H B1 homogeneity integral (NPT\_1H\_b1homogeneityInt\_1h)

**Test Sample:** 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub> Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721, Z142223  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a fraction around signal maximum (-60 deg, +60 deg) is acquired and shown in the figure. The variable pulse list which is used for acquisition is shown to the right.

The B1-homogeneities based on amplitudes and integrals are computed from the peak-picking list amplitudes and the magnitude of the second complex FID data point respectively.

### Control Option for Acquisition (L23)

- 1 default PLW 1 is used for determination of B1 homogeneity.
- 2 PLW 1 will be set to CNST 10 and resultant pulses are used for determination of B1 homogeneity. Experiment will be set to irregular.
- 3 PLW 1 will be set to power corresponding to specified customer pulses and resultant pulses are used for determination of B1 homogeneity.

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	4096		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_p1b1hom2d			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	4			ME_mod	2		LPfc
RG	0.250		optim. by RGA	NCOEF	20		
O1P	5.500	ppm		ABSF1	1000.000	ppm	
SWH	230.766	Hz		ABSF2	-1000.000	ppm	
TD	1024			F1P	5.720	ppm	
AQ	2.219	s		F2P	5.320	ppm	
FIDRES	0.451	Hz		<b>F1 ACQU</b>			Parameters F1
D 1	6.531	s	AQ+D1=const	NUC1	1H		
P 1	14.0	us	90deg NUC1	TD	43		No of incr.
PLW 1	6.6	W	Pow@90deg(Specs) NUC1	<b>F1 PROC</b>			Parameters F1
TE	298.000	K	default	SI	64		
				<b>NMRPT</b>			Parameters
				L 4	6		integ. fraction of 90deg
				L 5	8		# of step per maxima

## Experiment Description

The B1 homogeneity measurement is normally acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range is from 90 to 810 degree. In order to speed up the measurements NMRPT is acquiring only the partial ranges around the positive signal maxima for the pulse angles 90, 450, and 810 degree.


By this approach it is possible to get a better incremental resolution and a shorter overall measuring time. Prior to the main acquisition O1P (PROCNO=2) and phase correction (PROCNO=3) will be optimized. Intermediate results of the CONVTO1D routine is stored in PROCNO=998, result of CONVTO1D in PROCNO=999.

As control of stability, reproducibility and 90 degree calibration a spectrum with 180 degree pulse is acquired before and after each maxima. If relaxation conditions and instrument stability is sufficient, the result should not differ among them.

For the prediction of the signal maxima at 450 and 810 degree the propagation delay is included in the calculation of the VP list. (PDelay is calculated as difference of 360 and 180 degree pulse lengths during the corresponding 90 degree pulse determination experiment.) In this measurement O1 is determined before the main acquisition is started. Evaluation is achieved after processing the data in F2 only (XF2) by peak picking and integral determination of the most intense signal per maxima.

## 5.2.35 1H background without sample (NPT\_1H\_backgr\_nosample)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10121, Z10033, Z10270, Z10718  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

1H Background signal spectrum without sample (bottom left spectrum). No sharp signal should be present. Broad signal could arise from solid compound in the probe. Top spectrum shows expansion from 15.0 ppm to -2.0 ppm. CY is reduced by scaling factor (CNST50) compared to bottom left spectrum. 1H spectrum with sample to inspect phase correction (bottom right spectrum). Same phase correction values used for spectra with and without sample.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F1 ACQU</b>			Parameters	<b>CNST 10</b>	45.000		Flip angle for P90
NUC1	1H			<b>F1 PROC</b>			Parameters
PULPROG	npt_zg0			SI	32768		
NS	100			WDW	1		
DS	4			LB	1.000	Hz	
RG	101.000		no optim.	PC	1.000		
O1P	-2.500	ppm		F1P	66.605	ppm	
SWH	50000.000	Hz		F2P	-58.605	ppm	
TD	32768			CY	6.500	cm	
AQ	0.328	s		<b>NMRPT</b>			Parameters
FIDRES	3.052	Hz		CNST 50	0.500		Scaling factor for CY
D 1	0.872	s	AQ+D1=const				
P 0		us	P 1 * CNST 10 / 90				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
TE	298.000	K	default				

## Experiment Description


Background signal measurements are executed using a small flip angle due to two reasons. First, to get more efficient excitation bandwidth, secondly to compensate for usually long T1 relaxation time of solid signals.

Full spectrum is normally printed without any baseline correction (bottom spectrum). Using CNST 50, it is possible to scale CY for the top spectrum and adjust print out.

Experiment will be set to irregular if one or both of the options 'Skip Tuning/Matching' or 'Skip HNUC Tuning/Matching' are selected.

## 5.2.36 1H background with sample (NPT\_1H\_backgr\_withsample)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10121, Z10033, Z10270, Z10718, Z142221  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

1H Background signal spectrum with sample. Sharp signal arises from sample, broad signal could arise from solid compound in the probe.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F1 ACQU</b>			Parameters	<b>CNST 10</b>	45.000	Flip angle for P90
NUC1	1H			<b>F1 PROC</b>		Parameters
PULPROG	npt_zg0			SI	32768	
NS	10			WDW	1	
DS	4			LB	1.000	Hz
RG	101.000		no optim.	PC	1.000	
O1P	-2.500	ppm		F1P	63.434	ppm
SWH	50000.000	Hz		F2P	-55.775	ppm
TD	32768			CY	200.000	cm
AQ	0.328	s				
FIDRES	3.052	Hz				
D 1	3.837	s	AQ+D1=const			
P 0	7,9	us	P 1 * CNST 10 / 90			
P 1	14,0	us	90deg Pulse			
PLW 1	6.6	W	Pow@90deg(Specs)			
TE	298.000	K	default			

## Experiment Description


Background signal measurements are executed using a small flip angle due to two reasons. First, to get more efficient excitation bandwidth, secondly to compensate for usually long T1 relaxation time of solid signals.

Full spectrum is normally printed without any baseline correction.



## 5.2.37 2D COSY (NPT\_1H\_cosydfphpr)

**Test Sample:** 2 mM Lysozyme in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10241  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Double quantum filtered, phase sensitive COSY with presaturation during relaxation delay. 1D watersuppression experiment is shown at the left and at the top of the 2D.

### Control Option for Acquisition (L23)

- 1 default, with O1 optimization
- 2 no O1 optimization, the optimization of O1 is enforced, if O1 was not determined during a previous measurement.



## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	2048		
FnMODE	5			WDW	4		
PARMODE	1		Data Dimension	SSB	4.000		
PULPROG	npt_cosydfphpr			PH_mod	1		pk
NS	8			F1P	24.732	ppm	
DS	16			F2P	-15.316	ppm	
RG	0.250		optim. by RGA	LEV0	4.250		
TD0	1			TOPLEV	50.000	%	
SW	20.485	ppm		NLEV	20		
TD	4096			<b>F1 ACQU</b>			Parameters F1
AQ	0.250	s	field dependent	NUC1	1H		
FIDRES	4.002	Hz	field dependent	FnMODE	5		
D 1	2.000	s		O1P	4.708	ppm	
P 1	14.0	us	90deg Pulse	SW	20.485	ppm	
PLW 1	6.6	W	Pow@90deg(Specs)	TD	1024		
TE	298.000	K	default	<b>F1 PROC</b>			Parameters F1
DSPFIRM	4		rectangle	SI	2048		
DIGMOD	3		baseopt	WDW	4		
DE	40.000	us	set after getprosol	SSB	4.000		
				PH_mod	1		pk
				PHC0	90.000	deg	90deg (default)
				PHC1	-180.000	deg	180deg (default)
				F1P	0.000	ppm	
				F2P	0.000	ppm	

## Experiment Description

Double quantum filtered, phase sensitive COSY with presaturation.

Presaturation requires the exact determination of the irradiation position (O1P). The determination is executed in a derived data set using the parameter set 'NPT\_1H\_watersuppression\_recflow'.

Using the control option for acquisition L23=1, O1 is determined by an intensity profile(second point of FID) over a range of 16 Hz in 0.4 Hz steps around submitted O1.

Option L23=2 will skip the procedure just outlined above.


RG can be set in the Preparation Panel. If RG <= 1 RGA will be executed after O1 determination in the derived data set.

Processing is achieved using the phase correction values from the preparation experiment.

Contour levels for 2D spectrum are obtained by command levcalc. Levcalc sets the lowest contour level to LEV0\*S\_DEV, where S\_DEV (standard deviation) is a processing status parameter.

## 5.2.38 CPD 1H pulse calibration (NPT\_1H\_cpddeterminationf1\_1h)

**Test Sample:** 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub> Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 2048	
PARMODE 0	Data Dimension	WDW 3	
PULPROG zg		LB 0.750 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 0.250	optim. by RGA	ME_mod 2	LPfc
O1P 5.500 ppm		NCOEF 20	
SWH 230.766 Hz		ABSF1 1000.000 ppm	
TD 300		ABSF2 -1000.000 ppm	
AQ 0.650 s		F1P 5.496 ppm	
FIDRES 1.538 Hz		F2P 5.096 ppm	
D 1 1.225 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22


If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.39 Indirect CPD 13C pulse calibration (NPT\_1H\_cpddeterminationf2\_13c)

**Test Sample:** 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub> Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the decoupler pulse. On the left side the result of a CONVTO1D routine shows six experiments from 0 to 180 deg (PROCNO 100). On the right side the series goes from 180 to 360 deg (PROCNO 200).

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 2048	
PARMODE 0	Data Dimension	WDW 1	
PULPROG decp90		LB 0.500 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 0.250	optim. by RGA	ME_mod 2	LPfc
O1P 3.012 ppm		NCOEF 20	
SWH 230.766 Hz		ABSF1 1000.000 ppm	
TD 1000		ABSF2 -1000.000 ppm	
AQ 2.167 s		F1P 3.150 ppm	
FIDRES 0.462 Hz		F2P 2.850 ppm	
D 1 1.710 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
P 3 9.0 us	90deg NUC1		
PLW 2 42.0 W	Pow@90deg(Specs) NUC2		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22


If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 270 and 90 degree pulse lengths and memorised for further use in B1 homogeneity experiments.

## 5.2.40 Indirect CPD 15N pulse calibration (NPT\_1H\_cpddeterminationf2\_15n)

**Test Sample:** 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub> Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the decoupler pulse. On the left side the result of a CONVTO1D routine shows six experiments from 0 to 180 deg (PROCNO 100). On the right side the series goes from 180 to 360 deg (PROCNO 200).

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 2048	
PARMODE 0	Data Dimension	WDW 3	
PULPROG decp90		LB 0.750 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 0.250	optim. by RGA	ME_mod 2	LPfc
O1P 5.500 ppm		NCOEF 20	
SWH 230.766 Hz		ABSF1 1000.000 ppm	
TD 200		ABSF2 -1000.000 ppm	
AQ 0.433 s		F1P 5.667 ppm	
FIDRES 2.308 Hz		F2P 5.367 ppm	
D 1 0.433 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
P 3 14.0 us	90deg NUC1		
PLW 2 86.0 W	Pow@90deg(Specs) NUC2		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22


If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 270 and 90 degree pulse lengths and memorised for further use in B1 homogeneity experiments.

## 5.2.41 <sup>13</sup>C CPMG test (NPT\_1H\_cpmgtestf2\_13c)

**Test Sample:** 100 mM Urea-15N ([<sup>15</sup>NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-<sup>13</sup>C in Dimethyl Sulfoxide-D<sub>6</sub>  
Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the signal used for evaluation.

Top: Processing with WDW=QSINE and SSB=2, sharp signals which are sensitive to shim effects.

Bottom: Processing with WDW=EM and LB=10.0 Hz, broadend signals which are sensitive to tuning effects.

### Control Option for Acquisition (L23)

1 default



## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	2048			
PULPROG	npt_decp90f2cpmg2d				PH_mod	1			pk
NS	1				ME_mod	0			LPfc
DS	16				ABSF1	20.000	ppm		
RG	0.250		optim. by RGA		ABSF2	0.000	ppm		
O1P	3.200	ppm			F1P	3.096	ppm		
SW	0.753	ppm			F2P	3.049	ppm		
TD	1024				CY	13.500	cm		
AQ	1.700	s	field dependent						
FIDRES	0.588	Hz	field dependent						
D 1	0.800	s							
D 21	0.000	s	interpulse delay/2						
P 1	14.0	us	90deg Pulse						
PLW 1	6.6	W	Pow@90deg(Specs)						
P 3	14.0	us	90deg Pulse						
PLW 2	89.0	W	Pow@90deg(Specs)						
P 30	80.0	us	180deg Pulse						
PLW 20	10.9	W	Pow@90deg(P30)						
VCLIST	npt_cpmg13c								
TE	298.000	K	default						
WDW	LB		1st processing						
LB	10	Hz	1st processing						
WDW	QSINE		2nd processing						
SSB	2		2nd processing						

## Experiment Description


Pseudo 2D experiment with CPMG sequence. The interpulse delay is  $d21 \times 2$ . The number of repetitions of the CPMG sequence and hence the CPMG duration is given by VCLIST.

For evaluation the ratio of the average intensity from signals 99 to 128 and 1 to 30 is determined for two different sets of processing parameters. Using WDW=QSINE and SSB=2 will result in sharp signals, i.e. the determined ratio is sensitive for shim effects. Using WDW==EM and LB=10.0 Hz will result in broadened signals, i.e. the determined ratio is sensitive for tuning effects.

## 5.2.42 15N CPMG test (NPT\_1H\_cpmgtestf2\_15n)

---

**Test Sample:** 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub>  
Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the signal used for evaluation.

Top: Processing with WDW=QSINE and SSB=2, sharp signals which are sensitive to shim effects.

Bottom: Processing with WDW=EM and LB=10.0 Hz, broadend signals which are sensitive to tuning effects.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	1H			SI	2048		
PULPROG	npt_decp90f2cpmg2d			PH_mod	1		pk
NS	1			ME_mod	0		LPfc
DS	16			ABSF1	20.000	ppm	
RG	0.250		optim. by RGA	ABSF2	0.000	ppm	
O1P	5.400	ppm		F1P	5.517	ppm	
SW	0.753	ppm		F2P	5.424	ppm	
TD	1024			CY	13.500	cm	
AQ	1.700	s	field dependent				
FIDRES	0.588	Hz	field dependent				
D 1	0.800	s					
D 21	0.000	s	interpulse delay/2				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
P 3	14.0	us	90deg Pulse				
PLW 2	89.0	W	Pow@90deg(Specs)				
P 30	80.0	us	180deg Pulse				
PLW 20	10.9	W	Pow@90deg(P30)				
VCLIST	npt_cpmg15n						
TE	298.000	K	default				
WDW	LB		1st processing				
LB	10	Hz	1st processing				
WDW	QSINE		2nd processing				
SSB	2		2nd processing				

## Experiment Description


Pseudo 2D experiment with CPMG sequence. The interpulse delay is  $d21 \cdot 2$ . The number of repetitions of the CPMG sequence and hence the CPMG duration is given by VCLIST.

For evaluation the ratio of the average intensity from signals 99 to 128 and 1 to 30 is determined for two different sets of processing parameters. Using WDW=QSINE and SSB=2 will result in sharp signals, i.e. the determined ratio is sensitive for shim effects. Using WDW==EM and LB=10.0 Hz will result in broadened signals, i.e. the determined ratio is sensitive for tuning effects.

## 5.2.43 <sup>13</sup>C decoupler profile Chirp (NPT\_1H\_decProfile\_chirp\_13c)

---

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-<sup>13</sup>C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Decoupling profile of the methanol doublet as a function of the decoupling offset.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	1H			SI	32768		
PULPROG	sysdecpro			WDW	1		
NS	1			PH_mod	0	mc	
DS	8			F1P	3.760	ppm	
RG	101.000		no optim.	F2P	2.760	ppm	
O1P	3.260	ppm		<b>NMRPT</b>			Parameters
O2P	49.200	ppm		CNST 24	48000.0	Hz	Total Offset Range
SWH	5000.000	Hz		CNST 41	1.000		Return Value Evaluation
TD	2048						
AQ	0.205	s					
FIDRES	4.883	Hz					
D 1	2.813	s					
P 1	14.0	us	90deg NUC1				
P 3	11.0	us	90deg NUC2				
PCPD2	1500.0	us	PCPD NUC2				
PLW 1	6.7	W	Pow@90deg(Specs) NUC1				
PLW 2	26.5	W	Pow@90deg(Specs) NUC2				
PLW 12	0.56	W	Pow@CPD NUC2				
SPNAM15	crp40,1.5.hwt		shaped pulse				
CPDPRG2	p5m4sp180		decoupl. sequence				
FQ2LIST			Offset List, gen. dynamically.				
TE	298.000	K	default				


## Experiment Description

The purpose of this test is the assessment of the <sup>13</sup>C-decoupling efficiency on the methanol doublet as a function of the decoupling offset at constant decoupling power. The amplitude of the resulting singlet peak should be as high as possible over the largest offset range. The decoupling scheme used in this experiment is chirp.

## 5.2.44 <sup>13</sup>C decoupler profile Garp (NPT\_1H\_decProfile\_garp\_13c)

---

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-<sup>13</sup>C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Decoupling profile of the methanol doublet as a function of the decoupling offset.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU Parameters				F1 PROC Parameters			
NUC1	1H			SI	32768		
PULPROG	sysdecpro			WDW	1		
NS	1			PH_mod	0	mc	
DS	8			F1P	3.760	ppm	
RG	101.000		no optim.	F2P	2.760	ppm	
O1P	3.260	ppm		<b>NMRPT</b>			Parameters
O2P	49.200	ppm		CNST 24	24000.0	Hz	Total Offset Range
SWH	5000.000	Hz		CNST 41	1.000		Return Value Evaluation
TD	2048						
AQ	0.205	s					
FIDRES	4.883	Hz					
D 1	2.813	s					
P 1	14.0	us	90deg NUC1				
P 3	11.0	us	90deg NUC2				
PCPD2	70.0	us	PCPD NUC2				
PLW 1	6.7	W	Pow@90deg(Specs) NUC1				
PLW 2	26.5	W	Pow@90deg(Specs) NUC2				
PLW 12	0.56	W	Pow@CPD NUC2				
CPDPRG2	garp		decoupl. sequence				
FQ2LIST			Offset List, gen. dynamically.				
TE	298.000	K	default				

## Experiment Description

The purpose of this test is the assessment of the <sup>13</sup>C-decoupling efficiency on the methanol doublet as a function of the decoupling offset at constant decoupling power. The amplitude of the resulting singlet peak should be as high as possible over the largest offset range. The decoupling scheme used in this experiment is garp.

## 5.2.45 Low Current Diffusion Test for Z-direction (NPT\_1H\_diffusionLowCurrentZ)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top: Phase corrected first row of the pseudo 2D experiment.

Bottom left: Dosy plot.

Bottom right: Logarithmic plot of the intensities against b-values along with a linear fit.

### Control Option for Acquisition (L23)

- 1 default, execute O1P determination
- 2 skip O1P determination.



## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	diffSte			LB	1.000	Hz	
ZGOPTNS	-DLOCK			PH_mod	1		pk
	-Dspoil -			F1P	0.000	ppm	
	DSINE			F2P	0.000	ppm	
NS	8			<b>F1 ACQU</b>			Parameters F1
DS	4			NUC1	1H		
RG	0.250		optim. by RGA	TD	16		
O1P	4.700	ppm		<b>F1 PROC</b>			Parameters F1
SW	13.884	ppm		SI	16		
TD	16666		field dependent				
AQ	1.500	s					
FIDRES	0.667	Hz	field dependent				
D 1	1.450	s					
D 2	0.001	s	gradient stabilisation time				
D 5	0.014	s	big delta remainder				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPZ 5	2.41	%	41.8 gauss/cm				
GPNAM5	SINE.100						
P 19	2000.000	us	spoil gradient pulse				
GPNAM31	npt_diffusion						
CNST 3	7.40	%	128.3 gauss/cm				
D 16	0.000	s	ramp down time				
D 18	0.002	s	gradient on time				
D 60	300.000	s	temperature stabilisation				
TE	298.200	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the diffusion constant by varying the strength of the applied field gradients.

Low current diffusion experiment using the ste method. The experiment is only implemented for the gradient shape sine. The following parameters are used: gmax=128.33 gauss/cm, delta=1.0 ms, DELTA=20.0 ms, repetition time=3.0 s, ZGOPTNS=-DLOCK -Dspoil -DSINE

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.


Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the first experiment of the series using APK0 as default and applied to the series. The diffusion constant is fitted using the Steijskal-Tanner Equation. Therefore the signal intensities are determined from PROCNO 999.

The dosy plot is created with the AU programm setdiffparm and the command dosy2d.

Attention: Before running the diffusion experiment the DC offset of the gradient amplifier must be calibrated.

## 5.2.46 High Current Diffusion Test for Z-direction (NPT\_1H\_diffusionHighCurrentZ)

**Test Sample:** Dry Glycerol  
 Z10650  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Rotation off



### Example Printout

Top: Phase corrected first row of the pseudo 2D experiment.

Bottom left: Dosy plot.

Bottom right: Logarithmic plot of the intensities against b-values along with a linear fit.

### Control Option for Acquisition (L23)

- 1 default, execute O1P and pulse determination
- 2 skip O1P determination and execute pulse determination.
- 10 execute O1P determination and skip pulse determination.
- 12 skip O1P and pulse determination.

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	diffSte			LB	1.000	Hz	
	-Dspoil -			PH_mod	1		pk
ZGOPTNS	DLED -			F1P	0.000	ppm	
	DSINE			F2P	0.000	ppm	
NS	8			<b>F1 ACQU</b>			Parameters F1
DS	4			NUC1	1H		
RG	0.250		optim. by RGA	TD	16		
O1P	4.700	ppm		<b>F1 PROC</b>			Parameters F1
SW	13.884	ppm		SI	16		
TD	16666		field dependent				
AQ	1.500	s					
FIDRES	0.667	Hz	field dependent				
D 1	1.290	s					
D 2	0.001	s	gradient stabilisation time				
D 5	0.092	s	big delta remainder				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPZ 5	3.06	%	53.0 gauss/cm				
GPNAM5	SINE.100						
P 19	2000.000	us	spoil gradient pulse				
GPNAM31	npt_diffusion						
CNST 3	57.67	%	1000.0 gauss/cm				
D 16	0.000	s	ramp down time				
D 18	0.003	s	gradient on time				
D 60	300.000	s	temperature stabilisation				
TE	298.200	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the diffusion constant by varying the strength of the applied field gradients.

High current diffusion experiment using the ste method. The experiment is only implemented for the gradient shape sine. The following parameters are used: gmax=1000.00 gauss/cm, delta=2.0 ms, DELTA=100.0 ms, repetition time=3.0 s, ZGOPTNS=-Dspoil -DLED -DSINE

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

L23 dependent pulse determination using pulsecal.

Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the first experiment of the series using APK0 as default and applied to the series. The diffusion constant is fitted using the Steijskal-Tanner Equation. Therefore the signal intensities are determined from PROCNO 999.


The dosy plot is created with the AU programm setdiffparm and the command dosy2d.

Attention: Before running the diffusion experiment the DC offset of the gradient amplifier must be calibrated.

## 5.2.47 <sup>13</sup>C decoupler profile Waltz (NPT\_1H\_decProfile\_waltz\_13c)

---

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-<sup>13</sup>C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Decoupling profile of the methanol doublet as a function of the decoupling offset.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU Parameters				F1 PROC Parameters			
NUC1	1H			SI	32768		
PULPROG	sysdecpro			WDW	1		
NS	1			PH_mod	0	mc	
DS	8			F1P	3.760	ppm	
RG	101.000		no optim.	F2P	2.760	ppm	
O1P	3.260	ppm		<b>NMRPT</b>			Parameters
O2P	49.200	ppm		CNST 24	24000.0	Hz	Total Offset Range
SWH	5000.000	Hz		CNST 41	1.000		Return Value Evaluation
TD	2048						
AQ	0.205	s					
FIDRES	4.883	Hz					
D 1	2.813	s					
P 1	14.0	us	90deg NUC1				
P 3	11.0	us	90deg NUC2				
PCPD2	70.0	us	PCPD NUC2				
PLW 1	6.7	W	Pow@90deg(Specs) NUC1				
PLW 2	26.5	W	Pow@90deg(Specs) NUC2				
PLW 12	0.56	W	Pow@CPD NUC2				
CPDPRG2	waltz64		decoupl. sequence				
FQ2LIST			Offset List, gen. dynamically.				
TE	298.000	K	default				

## Experiment Description

The purpose of this test is the assessment of the <sup>13</sup>C-decoupling efficiency on the methanol doublet as a function of the decoupling offset at constant decoupling power. The amplitude of the resulting singlet peak should be as high as possible over the largest offset range. The decoupling scheme used in this experiment is waltz64.

## 5.2.48 1H detection with 13C garp decoupling (NPT\_1H\_garpdectestf2\_13c)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10033, Z10270, Z10718, Z10121  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top spectrum shows 2D projection (XFB, RHPP).  
Bottom spectrum shows 2D projection (XTRF2, RHPP).

### Control Option for Acquisition (L23)

- 70 PLW12[apparent]=PLW12[prosol], default
- 80 PLW12[apparent]=PLW12[prosol], 6 h waiting time before execution
- 71 PLW12[apparent]=PLW12[prosol]\*1.5849 (-2 dB)
- 81 PLW12[apparent]=PLW12[prosol]\*1.5849 (-2 dB), 6 h waiting time before execution
- 72 PLW12[apparent]=PLW12[prosol]\*2.0 (-3 dB)
- 82 PLW12[apparent]=PLW12[prosol]\*2.0 (-3 dB), 6 h waiting time before execution

## Parameters

<p><b>F2 ACQU</b></p> <p>Parameters F2</p> <p>NUC1 1H</p> <p>NUC2 13C</p> <p>PARMODE 1 Data Dimension</p> <p>PULPROG npt_garptest</p> <p>NS 1</p> <p>DS 0</p> <p>RG 101.000 no optim.</p> <p>TD0 1</p> <p>SW 12.132 ppm</p> <p>TD 8192</p> <p>AQ 0.844 s</p> <p>FIDRES 1.185 Hz field dependent</p> <p>D 1 2.109 s 2.5*AQ</p> <p>P 1 14.0 us 90deg NUC1</p> <p>PCPD 2 80.0 us CPD 90deg</p> <p>PLW 1 6.6 W Pow@90deg(Specs) NUC1</p> <p>PLW 12 0.5 W Pow@CPD(Specs)</p> <p>CPDPRG2 garp default CPD Seq.</p> <p>TE 298.000 K default</p>	<p><b>F2 PROC</b></p> <p>Parameters F2</p> <p>SI 4096</p> <p>WDW 0</p> <p>PH_mod 0</p> <p>BC_mod 0</p> <p>ME_mod 0</p> <p>FT_mod 0</p> <p>F1P 8.500 ppm</p> <p>F2P 0.500 ppm</p> <p><b>F1 ACQU</b></p> <p>Parameters F1</p> <p>NUC1 13C</p> <p>TD 256</p> <p><b>F1 PROC</b></p> <p>Parameters F1</p> <p>SI 256</p> <p>WDW 0</p> <p>PH_mod 0</p> <p>BC_mod 0</p> <p>ME_mod 0</p> <p>FT_mod 0</p>
--	---

## Experiment Description

The GARP decoupling test indicates the reliability of the probe with respect to decoupling. No excitation pulse is applied, only decoupling takes place during acquisition of the noise. The acquisition time is splitted up according to the scheme 25%-50%-25% thereby only during the 50% decoupling power is applied. D1 is set to 2.5\*AQ, so that the duty cycle is 15%

The different options for the experiment are self-explanatory.

The first evaluation is done by processing with XFB using the standard Fourier transform features of the TopSpin software. The horizontal positive projection (RHPP) is then evaluated by searching for the minimal and maximal intensity and forming the ratio of the two. Interpretation of the result is only possible in relation to the second evaluation.


A value close to 1.0 is optimal, higher values are worse, values < 1.0 are only meaningful, if the second evaluation results in small values, too.

The second evaluation is done by processing with XTRF2 using parameters as outlined below resulting in time domain data. The horizontal positive projection (RHPP) is then used to determine a 'noise factor' between the average noise of the decoupled and the coupled region. If the ratio is close to 1.0 the decoupling efficiency is optimal.

Options L23=70 and 80 are standard whereas all other L23 options are non-standard and will not be considered as regular test from the 'NMRPT Control Structure'.

## 5.2.49 1H detection with 13C hard pulse and 13C garp decoupling (NPT\_1H\_garp\_pulse13c\_dec13c)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10033, Z10270, Z10718, Z10121  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top spectrum shows 2D projection (XFB, RHPP).

Bottom spectrum shows 2D projection (XTRF2, RHPP).

Both projections are from the 2D experiment of the last series with the highest noise factor.

### Control Option for Acquisition (L23)

1 default



## Parameters

<p><b>F2 ACQU</b></p> <p>Parameters F2</p> <p>NUC1 1H</p> <p>NUC2 13C</p> <p>PARMODE 1 Data Dimension</p> <p>PULPROG npt_garptest</p> <p>NS 1</p> <p>DS 0</p> <p>RG 101.000 no optim.</p> <p>TD0 1</p> <p>SW 12.132 ppm</p> <p>TD 3882</p> <p>AQ 0.400 s</p> <p>FIDRES 2.501 Hz field dependent</p> <p>D 1 1.000 s 2.5*AQ</p> <p>P 4 24.0 us 180deg NUC2</p> <p>PLW 2 86.0 W Pow@90deg(Specs) NUC2</p> <p>PCPD 2 80.0 us CPD 90deg</p> <p>PLW 12 0.5 W Pow@CPD(Specs)</p> <p>CPDPRG2 garp default CPD Seq.</p> <p>TE 298.000 K default</p>	<p><b>F2 PROC</b></p> <p>Parameters F2</p> <p>SI 4096</p> <p>WDW 0</p> <p>PH_mod 0</p> <p>BC_mod 0</p> <p>ME_mod 0</p> <p>FT_mod 0</p> <p>F1P 8.500 ppm</p> <p>F2P 0.500 ppm</p> <p><b>F1 ACQU</b></p> <p>Parameters F1</p> <p>NUC1 13C</p> <p>TD 256</p> <p><b>F1 PROC</b></p> <p>Parameters F1</p> <p>SI 256</p> <p>WDW 0</p> <p>PH_mod 0</p> <p>BC_mod 0</p> <p>ME_mod 0</p> <p>FT_mod 0</p>
--	---

## Experiment Description

Garp series experiment with 13C hard pulse, 13C decoupling and 1H detection.

The garp series experiment has the following structure  $[n * (2D \text{ Garp Experiment}) - (\text{Delay } T)] * m$

The parameters m, n, and T are set via specifications. If not specified following default values will be used: m=2, n=2, and T=30 min. WRPA is used to copy each acquired experiment into a derived data set.

The results for Noise Factor and Spike Ratio are taken from the 2D experiment of the last series with the highest noise factor.

The GARP decoupling test indicates the reliability of the probe with respect to decoupling. No excitation pulse is applied, only decoupling takes place during acquisition of the noise. The acquisition time is splitted up according to the scheme 50%-25%-25% thereby only during the 50% decoupling power is applied. D1 is set to 2.5\*AQ, so that the duty cycle is 15%


The first evaluation is done by processing with XFB using the standard Fourier transform features of the TopSpin software. The horizontal positive projection (RHPP) is then evaluated by searching for the minimal and maximal intensity and forming the ratio of the two. Interpretation of the result is only possible in relation to the second evaluation.

A value close to 1.0 is optimal, higher values are worse, values < 1.0 are only meaningful, if the second evaluation results in small values, too.

The second evaluation is done by processing with XTRF2 using parameters as outlined below resulting in time domain data. The horizontal positive projection (RHPP) is then used to determine a 'noise factor' between the average noise of the decoupled and the coupled region. If the ratio is close to 1.0 the decoupling efficiency is optimal.

## 5.2.50 1H detection with 13C and 15N hard pulses and 13C garp decoupling (NPT\_1H\_garp\_simpul13c15n\_dec13c)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
 Z10120, Z100927, Z10033, Z10270, Z10718, Z10121  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top spectrum shows 2D projection (XFB, RHPP).

Bottom spectrum shows 2D projection (XTRF2, RHPP).

Both projections are from the 2D experiment of the last series with the highest noise factor.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>		Parameters F2
NUC1	1H			SI	4096	
NUC2	13C			WDW	0	
NUC3	15N			PH_mod	0	
PARMODE	1		Data Dimension	BC_mod	0	
PULPROG	npt_garptest			ME_mod	0	
NS	1			FT_mod	0	
DS	0			F1P	8.500	ppm
RG	101.000		no optim.	F2P	0.500	ppm
TD0	1			<b>F1 ACQU</b>		Parameters F1
SW	12.132	ppm		NUC1	13C	
TD	3882			TD	256	
AQ	0.400	s		<b>F1 PROC</b>		Parameters F1
FIDRES	2.501	Hz	field dependent	SI	256	
D 1	1.000	s	2.5*AQ	WDW	0	
P 4	24.0	us	180deg NUC2	PH_mod	0	
PLW 2	86.0	W	Pow@90deg(Specs) NUC2	BC_mod	0	
P 6	50.0	us	180deg NUC3	ME_mod	0	
PLW 3	155.5	W	Pow@CPD(Specs)	FT_mod	0	
PCPD 2	80.0	us	CPD 90deg			
PLW 12	0.5	W	Pow@CPD(Specs)			
CPDPRG2	garp		default CPD Seq.			
TE	298.000	K	default			

## Experiment Description

Garp series experiment with simultaneous 13C and 15N hard pulses, 13C decoupling and 1H detection. The garp series experiment has the following structure  $[n * (2D \text{ Garp Experiment}) - (\text{Delay } T)] * m$ . The parameters m, n, and T are set via specifications. If not specified following default values will be used: m=2, n=2, and T=30 min. WRPA is used to copy each acquired experiment into a derived data set. The results for Noise Factor and Spike Ratio are taken from the 2D experiment of the last series with the highest noise factor.

The GARP decoupling test indicates the reliability of the probe with respect to decoupling. No excitation pulse is applied, only decoupling takes place during acquisition of the noise. The acquisition time is splitted up according to the scheme 50%-25%-25% thereby only during the 50% decoupling power is applied. D1 is set to 2.5\*AQ, so that the duty cycle is 15%


The first evaluation is done by processing with XFB using the standard Fourier transform features of the TopSpin software. The horizontal positive projection (RHPP) is then evaluated by searching for the minimal and maximal intensity and forming the ratio of the two. Interpretation of the result is only possible in relation to the second evaluation.

A value close to 1.0 is optimal, higher values are worse, values < 1.0 are only meaningful, if the second evaluation results in small values, too.

The second evaluation is done by processing with XTRF2 using parameters as outlined below resulting in time domain data. The horizontal positive projection (RHPP) is then used to determine a 'noise factor' between the average noise of the decoupled and the coupled region. If the ratio is close to 1.0 the decoupling efficiency is optimal.

## 5.2.51 $^{13}\text{C}$ detection with $^{13}\text{C}$ and $^{15}\text{N}$ hard pulses and $^{15}\text{N}$ and $^1\text{H}$ garp decoupling (NPT\_13C\_garp\_pulses13c15n\_dec15n1h)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10033, Z10270, Z10718, Z10121  
**Solvent:**  $\text{CDCl}_3$   
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top spectrum shows 2D projection (XFB, RHPP).

Bottom spectrum shows 2D projection (XTRF2, RHPP).

Both projections are from the 2D experiment of the last series with the highest noise factor.

### Control Option for Acquisition (L23)

1 default

## Parameters

<p><b>F2 ACQU</b></p> <p>Parameters F2</p> <p>NUC1 13C</p> <p>NUC2 15N</p> <p>NUC3 1H</p> <p>PARMODE 1</p> <p>Data Dimension</p> <p>PULPROG npt_garptest</p> <p>NS 1</p> <p>DS 0</p> <p>RG 101.000</p> <p>no optim.</p> <p>TD0 1</p> <p>SW 48.246 ppm</p> <p>TD 3882</p> <p>AQ 0.400 s</p> <p>FIDRES 2.501 Hz</p> <p>field dependent</p> <p>D 1 1.000 s</p> <p>2.5*AQ</p> <p>P 2 24.0 us</p> <p>180deg NUC2</p> <p>PLW 1 86.0 W</p> <p>Pow@90deg(Specs) NUC2</p> <p>P 4 50.0 us</p> <p>180deg NUC3</p> <p>PLW 2 155.5 W</p> <p>Pow@CPD(Specs)</p> <p>PCPD 2 200.0 us</p> <p>CPD 90deg</p> <p>PLW 12 2.5 W</p> <p>Pow@CPD(Specs)</p> <p>CPDPRG2 garp</p> <p>default CPD Seq.</p> <p>PCPD 3 80.0 us</p> <p>CPD 90deg</p> <p>PLW 16 0.5 W</p> <p>Pow@CPD(Specs)</p> <p>CPDPRG3 garp</p> <p>default CPD Seq.</p> <p>TE 298.000 K</p> <p>default</p>	<p><b>F2 PROC</b></p> <p>Parameters F2</p> <p>SI 4096</p> <p>WDW 0</p> <p>PH_mod 0</p> <p>BC_mod 0</p> <p>ME_mod 0</p> <p>FT_mod 0</p> <p>F1P 8.500 ppm</p> <p>F2P 0.500 ppm</p> <p><b>F1 ACQU</b></p> <p>Parameters F1</p> <p>NUC1 13C</p> <p>TD 256</p> <p><b>F1 PROC</b></p> <p>Parameters F1</p> <p>SI 256</p> <p>WDW 0</p> <p>PH_mod 0</p> <p>BC_mod 0</p> <p>ME_mod 0</p> <p>FT_mod 0</p>
---	---

## Experiment Description

Garp series experiment with serial 13C and 15N hard pulses, 15N and 1H decoupling and 13C detection. The garp series experiment has the following structure  $[n * (2D \text{ Garp Experiment}) - (\text{Delay } T)] * m$ . The parameters m, n, and T are set via specifications. If not specified following default values will be used: m=2, n=2, and T=30 min. WRPA is used to copy each acquired experiment into a derived data set. The results for Noise Factor and Spike Ratio are taken from the 2D experiment of the last series with the highest noise factor.

The GARP decoupling test indicates the reliability of the probe with respect to decoupling. No excitation pulse is applied, only decoupling takes place during acquisition of the noise. The acquisition time is splitted up according to the scheme 50%-25%-25% thereby only during the 50% decoupling power is applied. D1 is set to 2.5\*AQ, so that the duty cycle is 15%


The first evaluation is done by processing with XFB using the standard Fourier transform features of the TopSpin software. The horizontal positive projection (RHPP) is then evaluated by searching for the minimal and maximal intensity and forming the ratio of the two. Interpretation of the result is only possible in relation to the second evaluation.

A value close to 1.0 is optimal, higher values are worse, values < 1.0 are only meaningful, if the second evaluation results in small values, too.

The second evaluation is done by processing with XTRF2 using parameters as outlined below resulting in time domain data. The horizontal positive projection (RHPP) is then used to determine a 'noise factor' between the average noise of the decoupled and the coupled region. If the ratio is close to 1.0 the decoupling efficiency is optimal.

## 5.2.52 $^{15}\text{N}$ detection with $^{15}\text{N}$ and $^{13}\text{C}$ hard pulses and $^{13}\text{C}$ and $^1\text{H}$ garp decoupling (NPT\_15N\_garp\_pulses15n13c\_dec13c1h)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10033, Z10270, Z10718, Z10121  
**Solvent:**  $\text{CDCl}_3$   
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top spectrum shows 2D projection (XFB, RHPP).

Bottom spectrum shows 2D projection (XTRF2, RHPP).

Both projections are from the 2D experiment of the last series with the highest noise factor.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	15N			SI	4096		
NUC2	13C			WDW	0		
NUC3	1H			PH_mod	0		
PARMODE	1		Data Dimension	BC_mod	0		
PULPROG	npt_garptest			ME_mod	0		
NS	1			FT_mod	0		
DS	0			F1P	8.500	ppm	
RG	101.000		no optim.	F2P	0.500	ppm	
TD0	1			<b>F1 ACQU</b>			Parameters F1
SW	119.714	ppm		NUC1	13C		
TD	3882			TD	256		
AQ	0.400	s		<b>F1 PROC</b>			Parameters F1
FIDRES	2.501	Hz	field dependent	SI	256		
D 1	1.000	s	2.5*AQ	WDW	0		
P 2	50.0	us	180deg NUC3	PH_mod	0		
PLW 1	155.5	W	Pow@CPD(Specs)	BC_mod	0		
P 4	24.0	us	180deg NUC2	ME_mod	0		
PLW 2	86.0	W	Pow@90deg(Specs) NUC2	FT_mod	0		
PCPD 2	65.0	us	CPD 90deg				
PLW 12	3.3	W	Pow@CPD(Specs)				
CPDPRG2	garp		default CPD Seq.				
PCPD 3	80.0	us	CPD 90deg				
PLW 16	0.5	W	Pow@CPD(Specs)				
CPDPRG3	garp		default CPD Seq.				
TE	298.000	K	default				

## Experiment Description

Garp series experiment with serial 15N and 13C hard pulses, 13C and 1H decoupling and 15N detection. The garp series experiment has the following structure  $[n * (2D \text{ Garp Experiment}) - (\text{Delay } T)] * m$ . The parameters m, n, and T are set via specifications. If not specified following default values will be used: m=2, n=2, and T=30 min. WRPA is used to copy each acquired experiment into a derived data set. The results for Noise Factor and Spike Ratio are taken from the 2D experiment of the last series with the highest noise factor.

The GARP decoupling test indicates the reliability of the probe with respect to decoupling. No excitation pulse is applied, only decoupling takes place during acquisition of the noise. The acquisition time is splitted up according to the scheme 50%-25%-25% thereby only during the 50% decoupling power is applied. D1 is set to 2.5\*AQ, so that the duty cycle is 15%

The first evaluation is done by processing with XFB using the standard Fourier transform features of the TopSpin software. The horizontal positive projection (RHPP) is then evaluated by searching for the minimal and maximal intensity and forming the ratio of the two. Interpretation of the result is only possible in relation to the second evaluation.

A value close to 1.0 is optimal, higher values are worse, values < 1.0 are only meaningful, if the second evaluation results in small values, too.

The second evaluation is done by processing with XTRF2 using parameters as outlined below resulting in time domain data. The horizontal positive projection (RHPP) is then used to determine a 'noise factor' between the average noise of the decoupled and the coupled region. If the ratio is close to 1.0 the decoupling efficiency is optimal.




## 5.2.53 1H Z-gradient profile [-] (NPT\_1H\_gradientprofile\_neg)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10720, Z10719,  
Z142222

**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation according to RO



### Example Printout

Proton Z-gradient profile.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip gradient functionality check during processing



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 4096	
PULPROG npt_imgegp1d		WDW 0	
NS 1		PH_mod 0	
DS 0		F1P 54.559 ppm	
RG 0.250	optim. by RGA	F2P -45.159 ppm	
O1P 4.700 ppm		CY 11.000 cm	
SWH 81967.211 Hz		<b>NMRPT</b>	Parameters
TD 1024		CNST 20 1.000 G/cm*A	gradient strength
AQ 0.006 s		CNST 37 19.400 mm	active sample size
FIDRES 160.092 Hz			
D 1 0.500 s			
D 15 0.005 s	Echo time		
D 21 0.000 s	Grad. stab.		
D 27 0.002 s	Dephas. grad.		
P 0 2.0 us	Cryo=1deg, RT=5deg		
PLW 0	adjusted power		
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
GPNAM1 gauss			
GPNAM2 gauss			
GPZ 1 -4.843 %			
GPZ 2 6.053 %			
TE 298.000 K	default		

## Experiment Description

Z-gradient profile is acquired with negative gradient in order to check the basic functionality of gradient.

GPZ 2 is calculated to obtain a profilewidth of 25000 Hz.

The profile width is determined at 50% (HR) and 15% (HRMAS) of the maximum profile intensity respectively.

Together with the length of the RF coil and gyromagnetic ratio of the observed nuclei the gradient strength (G/cm\*A) is calculated. Processing is always executed with magnitude correction (PH\_mod=MC).

## 5.2.54 1H Z-gradient profile [+] (NPT\_1H\_gradientprofile\_pos)


---

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10720, Z10719,  
Z142222

**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation according to RO



### Example Printout

Proton Z-gradient profile.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip gradient functionality check during processing

## Parameters

<p><b>F1 ACQU</b></p> <p>NUC1 1H  PULPROG npt_imgegp1d  NS 1  DS 0  RG 0.250  O1P 4.700 ppm  SWH 81967.211 Hz  TD 1024  AQ 0.006 s  FIDRES 160.092 Hz  D 1 0.500 s  D 15 0.005 s  D 21 0.000 s  D 27 0.002 s  P 0 2.0 us  PLW 0 6.6 W  P 1 14.0 us  PLW 1 6.6 W  GPNAM1 gauss  GPNAM2 gauss  GPZ 1 4.843 %  GPZ 2 -6.053 %  TE 298.000 K</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">optim. by RGA</p> <p style="text-align: center;">Echo time  Grad. stab.  Dephas. grad.  Cryo=1deg, RT=5deg  adjusted power  90deg NUC1  Pow@90deg(Specs) NUC1</p> <p style="text-align: center;">default</p>
<p><b>F1 PROC</b></p> <p>SI 4096  WDW 0  PH_mod 0  F1P 54.559 ppm  F2P -45.159 ppm  CY 11.000 cm  <b>NMRPT</b>  CNST 20 1.000 G/cm*A  CNST 37 19.400 mm</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">Parameters  gradient strength  active sample size</p>


## Experiment Description

Z-gradient profile is acquired with positive gradient in order to check the basic functionality of gradient. GPZ 2 is calculated to obtain a profilewidth of 25000 Hz. The profile width is determined at 50% (HR) and 15% (HRMAS) of the maximum profile intensity respectively. Together with the length of the RF coil and gyromagnetic ratio of the observed nuclei the gradient strength (G/cm\*A) is calculated. Processing is always executed with magnitude correction (PH\_mod=MC).

## 5.2.55 1H X-gradient profile [-] (NPT\_1H\_gradprofX\_neg)

---

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10720, Z10719  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Proton X-gradient profile.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip gradient functionality check during processing

## Parameters

<p><b>F1 ACQU</b></p> <p>NUC1 1H          PULPROG npt_imgegp1d          NS 1          DS 0          RG 0.250          O1P 4.700 ppm          SWH 81967.211 Hz          TD 1024          AQ 0.006 s          FIDRES 160.092 Hz          D 1 0.500 s          D 15 0.005 s          D 21 0.000 s          D 27 0.002 s          P 0 2.0 us          PLW 0 W          P 1 14.0 us          PLW 1 W          GPNAM1 gauss          GPNAM2 gauss          GPX 1 -0.000 %          GPX 2 0.000 %          TE 298.000 K</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">optim. by RGA</p> <p style="text-align: center;">Echo time          Grad. stab.          Dephas. grad.          Cryo=1deg, RT=5deg          adjusted power          90deg NUC1          Pow@90deg(Specs) NUC1</p> <p style="text-align: center;">default</p>
<p><b>F1 PROC</b></p> <p>SI 4096          WDW 0          PH_mod 0          F1P 54.559 ppm          F2P -45.159 ppm          CY 11.000 cm  <b>NMRPT</b>          CNST 20 1.000 G/cm*A          CNST 37 4.240 mm</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">Parameters          gradient strength          active sample size</p>

## Experiment Description

X-gradient profile is acquired with negative gradient in order to check the basic functionality of gradient.

GPX 2 is calculated to obtain a profilewidth of 25000 Hz.

The profile width is determined at 15% of the maximum profile intensity.

Together with the length of the RF coil and gyromagnetic ratio of the observed nuclei the gradient strength (G/cm\*A) is calculated. Processing is always executed with magnitude correction (PH\_mod=MC).

## 5.2.56 1H X-gradient profile [+] (NPT\_1H\_gradprofX\_pos)


---

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10720, Z10719

**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

Proton X-gradient profile.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip gradient functionality check during processing

## Parameters

<p><b>F1 ACQU</b></p> <p>NUC1 1H  PULPROG npt_imgegp1d  NS 1  DS 0  RG 0.250  O1P 4.700 ppm  SWH 81967.211 Hz  TD 1024  AQ 0.006 s  FIDRES 160.092 Hz  D 1 0.500 s  D 15 0.005 s  D 21 0.000 s  D 27 0.002 s  P 0 2.0 us  PLW 0 W  P 1 14.0 us  PLW 1 6.6 W  GPNAM1 gauss  GPNAM2 gauss  GPX 1 0.000 %  GPX 2 -0.000 %  TE 298.000 K</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">optim. by RGA</p> <p style="text-align: center;">Echo time  Grad. stab.  Dephas. grad.  Cryo=1deg, RT=5deg  adjusted power  90deg NUC1  Pow@90deg(Specs) NUC1</p> <p style="text-align: center;">default</p>
<p><b>F1 PROC</b></p> <p>SI 4096  WDW 0  PH_mod 0  F1P 54.559 ppm  F2P -45.159 ppm  CY 11.000 cm  <b>NMRPT</b>  CNST 20 1.000 G/cm*A  CNST 37 4.240 mm</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">Parameters  gradient strength  active sample size</p>

## Experiment Description

X-gradient profile is acquired with positive gradient in order to check the basic functionality of gradient.


GPX 2 is calculated to obtain a profilewidth of 25000 Hz.

The profile width is determined at 15% of the maximum profile intensity.

Together with the length of the RF coil and gyromagnetic ratio of the observed nuclei the gradient strength (G/cm\*A) is calculated. Processing is always executed with magnitude correction (PH\_mod=MC).

## 5.2.57 1H Y-gradient profile [-] (NPT\_1H\_gradprofY\_neg)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10720, Z10719  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Proton Y-gradient profile.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip gradient functionality check during processing



## Parameters

<p><b>F1 ACQU</b></p> <p>NUC1 1H  PULPROG npt_imgegp1d  NS 1  DS 0  RG 0.250  O1P 4.700 ppm  SWH 81967.211 Hz  TD 1024  AQ 0.006 s  FIDRES 160.092 Hz  D 1 0.500 s  D 15 0.005 s  D 21 0.000 s  D 27 0.002 s  P 0 2.0 us  PLW 0 W  P 1 14.0 us  PLW 1 6.6 W  GPNAM1 gauss  GPNAM2 gauss  GPY 1 -0.000 %  GPY 2 0.000 %  TE 298.000 K</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">optim. by RGA</p> <p style="text-align: center;">Echo time  Grad. stab.  Dephas. grad.  Cryo=1deg, RT=5deg  adjusted power  90deg NUC1  Pow@90deg(Specs) NUC1</p> <p style="text-align: center;">default</p>
<p><b>F1 PROC</b></p> <p>SI 4096  WDW 0  PH_mod 0  F1P 54.559 ppm  F2P -45.159 ppm  CY 11.000 cm</p> <p><b>NMRPT</b></p> <p>CNST 20 1.000 G/cm*A  CNST 37 4.240 mm</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">Parameters  gradient strength  active sample size</p>

## Experiment Description

Y-gradient profile is acquired with negative gradient in order to check the basic functionality of gradient.


GPY 2 is calculated to obtain a profilewidth of 25000 Hz.

The profile width is determined at 15% of the maximum profile intensity.

Together with the length of the RF coil and gyromagnetic ratio of the observed nuclei the gradient strength (G/cm\*A) is calculated. Processing is always executed with magnitude correction (PH\_mod=MC).

## 5.2.58 1H Y-gradient profile [+] (NPT\_1H\_gradprofY\_pos)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10720, Z10719  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Proton Y-gradient profile.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip gradient functionality check during processing

## Parameters

<p><b>F1 ACQU</b></p> <p>NUC1 1H  PULPROG npt_imgegp1d  NS 1  DS 0  RG 0.250  O1P 4.700 ppm  SWH 81967.211 Hz  TD 1024  AQ 0.006 s  FIDRES 160.092 Hz  D 1 0.500 s  D 15 0.005 s  D 21 0.000 s  D 27 0.002 s  P 0 2.0 us  PLW 0 W  P 1 14.0 us  PLW 1 6.6 W  GPNAM1 gauss  GPNAM2 gauss  GPY 1 0.000 %  GPY 2 -0.000 %  TE 298.000 K</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">optim. by RGA</p> <p>Echo time  Grad. stab.  Dephas. grad.  Cryo=1deg, RT=5deg  adjusted power  90deg NUC1  Pow@90deg(Specs) NUC1</p> <p style="text-align: center;">default</p>
<p><b>F1 PROC</b></p> <p>SI 4096  WDW 0  PH_mod 0  F1P 54.559 ppm  F2P -45.159 ppm  CY 11.000 cm</p> <p><b>NMRPT</b></p> <p>CNST 20 1.000 G/cm*A  CNST 37 4.240 mm</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">Parameters</p> <p style="text-align: center;">gradient strength active sample size</p>

## Experiment Description

Y-gradient profile is acquired with positive gradient in order to check the basic functionality of gradient. GPY 2 is calculated to obtain a profilewidth of 25000 Hz. The profile width is determined at 15% of the maximum profile intensity. Together with the length of the RF coil and gyromagnetic ratio of the observed nuclei the gradient strength (G/cm\*A) is calculated. Processing is always executed with magnitude correction (PH\_mod=MC).

## 5.2.59 1H Z-gradient profile [-] (NPT\_1H\_gradprofZ\_neg)


---

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10720, Z10719

**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation according to RO



### Example Printout

Proton Z-gradient profile.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip gradient functionality check during processing

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 4096	
PULPROG npt_imgegp1d		WDW 0	
NS 1		PH_mod 0	
DS 0		F1P 54.559 ppm	
RG 0.250	optim. by RGA	F2P -45.159 ppm	
O1P 4.700 ppm		CY 11.000 cm	
SWH 81967.211 Hz		<b>NMRPT</b>	Parameters
TD 1024		CNST 20 1.000 G/cm*A	gradient strength
AQ 0.006 s		CNST 37 19.400 mm	active sample size
FIDRES 160.092 Hz			
D 1 0.500 s			
D 15 0.005 s	Echo time		
D 21 0.000 s	Grad. stab.		
D 27 0.002 s	Dephas. grad.		
P 0 2.0 us	Cryo=1deg, RT=5deg		
PLW 0	adjusted power		
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
GPNAM1 gauss			
GPNAM2 gauss			
GPZ 1 -4.843 %			
GPZ 2 6.053 %			
TE 298.000 K	default		

## Experiment Description

Z-gradient profile is acquired with negative gradient in order to check the basic functionality of gradient.

GPZ 2 is calculated to obtain a profilewidth of 25000 Hz.

The profile width is determined at 50% (HR) and 15% (HRMAS) of the maximum profile intensity respectively.

Together with the length of the RF coil and gyromagnetic ratio of the observed nuclei the gradient strength (G/cm\*A) is calculated. Processing is always executed with magnitude correction (PH\_mod=MC).

## 5.2.60 1H Z-gradient profile [+] (NPT\_1H\_gradprofZ\_pos)


---

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10720, Z10719

**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation according to RO



### Example Printout

Proton Z-gradient profile.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip gradient functionality check during processing

## Parameters



<p><b>F1 ACQU</b></p> <p>NUC1 1H          PULPROG npt_imgegp1d          NS 1          DS 0          RG 0.250          O1P 4.700 ppm          SWH 81967.211 Hz          TD 1024          AQ 0.006 s          FIDRES 160.092 Hz          D 1 0.500 s          D 15 0.005 s          D 21 0.000 s          D 27 0.002 s          P 0 2.0 us          PLW 0 W          P 1 14.0 us          PLW 1 W          GPNAM1 gauss          GPNAM2 gauss          GPZ 1 4.843 %          GPZ 2 -6.053 %          TE 298.000 K</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">optim. by RGA</p> <p style="text-align: center;">Echo time          Grad. stab.          Dephas. grad.          Cryo=1deg, RT=5deg          adjusted power          90deg NUC1          Pow@90deg(Specs) NUC1</p> <p style="text-align: center;">default</p>
<p><b>F1 PROC</b></p> <p>SI 4096          WDW 0          PH_mod 0          F1P 54.559 ppm          F2P -45.159 ppm          CY 11.000 cm</p> <p><b>NMRPT</b></p> <p>CNST 20 1.000 G/cm*A          CNST 37 19.400 mm</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">Parameters          gradient strength          active sample size</p>

## Experiment Description

Z-gradient profile is acquired with positive gradient in order to check the basic functionality of gradient. GPZ 2 is calculated to obtain a profilewidth of 25000 Hz. The profile width is determined at 50% (HR) and 15% (HRMAS) of the maximum profile intensity respectively. Together with the length of the RF coil and gyromagnetic ratio of the observed nuclei the gradient strength (G/cm\*A) is calculated. Processing is always executed with magnitude correction (PH\_mod=MC).

## 5.2.61 Gradient recovery stability test (NPT\_1H\_gradrec\_stest\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727, Z142231  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT				
NO#	SecDel [dl6]	Scal.Int. [426]	B0-shift [Hz]	APK0 [APK #6]
1	10.00 us	99.86	4.56	3.0
2	15.85 us	99.89	4.39	0.0
3	25.12 us	99.87	4.22	0.0
4	39.81 us	99.86	4.03	0.0
5	63.10 us	99.90	3.84	3.0
6	100.0 us	99.87	3.65	1.0
7	158.5 us	99.87	3.48	0.5
8	251.2 us	99.91	3.30	1.0
9	398.1 us	99.85	3.03	1.0
10	630.9 us	99.94	2.96	1.0
11	1.000 ms	99.86	2.73	1.0
12	1.585 ms	99.91	2.58	0.0
13	2.512 ms	99.91	2.37	0.0
14	3.981 ms	99.93	2.20	0.0
15	6.310 ms	99.87	1.99	0.0
16	10.00 ms	99.88	1.82	0.5
17	15.85 ms	99.92	1.63	-0.5
18	25.12 ms	99.93	1.45	0.0
19	39.81 ms	99.87	1.27	0.0
20	63.10 ms	99.93	1.10	0.0
21	0.100 s	99.92	0.89	0.5
22	0.158 s	99.90	0.72	0.5
23	0.251 s	99.96	0.51	0.0
24	0.398 s	99.99	0.34	0.5
25	0.631 s	99.95	0.13	1.0
26	1.000 s	100.00	0.00	0.0

### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the residual water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B<sub>0</sub>), phase deviation (apk<sub>0</sub>).

### Control Option for Acquisition (L23)

1 default



## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_gradrecvd			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	2			F1P	5.720	ppm	
RG	0.250		optim. by RGA	F2P	5.320	ppm	
O1P	4.697	ppm		<b>F1 ACQU</b>			Parameters F1
SW	7.528	ppm		NUC1	1H		
TD	16506		field dependent	TD	26		
AQ	2.740	s		<b>F1 PROC</b>			Parameters F1
FIDRES	0.365	Hz		SI	32		
D 1	0.300	s					
VDLIST	npt_gradrec		default				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPNAM1	RECT.1						
GPZ 1	0.000	%					
P 16	5000.000	us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the basic instrument stability under the same conditions as used for the gradient recovery test. The experiment is normally executed just prior or after the series of gradient experiments.


Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.

Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.


The results are summarized in the text file printed together with spectrum (bottom left).

## 5.2.62 Gradient recovery test for X-direction [-] (NPT\_1H\_gradrecX\_sqn\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT					
NO#	SecDel [dl6]	Scal.Int. [#26]	B0-shift [Hz]	*APK0 [APK #6]	
1	10.00 us	99.30	4.68	4.0	
2	15.85 us	99.17	4.53	2.5	
3	25.12 us	98.28	4.30	2.0	
4	39.81 us	97.73	4.13	2.0	
5	63.10 us	97.76	3.92	2.0	
6	100.0 us	97.70	3.75	1.5	
7	158.5 us	97.62	3.54	1.5	
8	251.2 us	97.68	3.38	1.0	
9	398.1 us	97.64	3.17	0.5	
10	630.9 us	97.85	3.00	0.0	
11	1.000 ms	97.95	2.79	0.0	
12	1.585 ms	98.29	2.62	-0.5	
13	2.512 ms	98.60	2.41	1.0	
14	3.981 ms	98.78	2.24	0.0	
15	6.310 ms	99.18	2.03	0.0	
16	10.00 ms	99.58	1.86	0.5	
17	15.85 ms	99.75	1.65	-0.5	
18	25.12 ms	99.77	1.48	0.0	
19	39.81 ms	99.85	1.33	-0.5	
20	63.10 ms	99.82	1.10	-0.5	
21	0.100 s	99.95	0.93	0.5	
22	0.158 s	99.92	0.72	-0.5	
23	0.251 s	99.90	0.55	0.0	
24	0.398 s	99.97	0.38	0.0	
25	0.631 s	99.96	0.21	0.0	
26	1.000 s	100.00	0.00	0.0	



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B<sub>0</sub>), phase deviation (apk0).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_gradrecvd			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	2			F1P	5.720	ppm	
RG	0.250		optim. by RGA	F2P	5.320	ppm	
O1P	4.697	ppm		<b>F1 ACQU</b>			Parameters F1
SW	7.528	ppm		NUC1	1H		
TD	16506		field dependent	TD	26		
AQ	2.740	s		<b>F1 PROC</b>			Parameters F1
FIDRES	0.365	Hz		SI	32		
D 1	0.300	s					
VDLIST	npt_gradrec		default				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPNAME1	RECT.1						
GPX 1	-75.000	%					
P 16	5000.000	us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.


Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of range one, needs to be specified.


## 5.2.63 Gradient recovery test for X-direction [+] (NPT\_1H\_gradrecX\_sq\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT

NO#	SecDel [dl6]	Scal.Int. [426]	B0-shift [Hz]	APK0 [APK #46]
1	10.00 us	99.46	4.60	-3.5
2	15.85 us	99.62	4.43	-2.0
3	25.12 us	98.62	4.28	-1.0
4	39.82 us	98.34	4.05	-1.5
5	63.10 us	98.31	3.92	-2.0
6	100.0 us	98.24	3.73	-1.5
7	158.5 us	98.16	3.54	-1.5
8	251.2 us	98.10	3.34	-1.5
9	398.1 us	98.16	3.23	-1.0
10	630.9 us	98.04	2.96	-0.5
11	1.000 ms	98.55	2.79	0.0
12	1.585 ms	98.46	2.62	-0.5
13	2.512 ms	98.70	2.43	-0.5
14	3.981 ms	99.04	2.24	-0.5
15	6.310 ms	99.38	2.07	-0.5
16	10.00 ms	99.72	1.86	0.0
17	15.85 ms	99.63	1.69	-0.5
18	25.12 ms	99.31	1.48	-0.5
19	39.81 ms	99.33	1.35	-0.5
20	63.10 ms	99.39	1.10	-0.5
21	0.100 s	99.97	1.01	-0.5
22	0.158 s	100.00	0.76	0.0
23	0.251 s	99.92	0.59	0.0
24	0.398 s	99.95	0.42	-0.5
25	0.631 s	99.98	0.25	0.0
26	1.000 s	100.00	0.00	0.0



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B<sub>0</sub>), phase deviation (apk<sub>0</sub>).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_gradrecvd			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	2			F1P	5.720	ppm	
RG	0.250		optim. by RGA	F2P	5.320	ppm	
O1P	4.697	ppm		<b>F1 ACQU</b>			Parameters F1
SW	7.528	ppm		NUC1	1H		
TD	16506		field dependent	TD	26		
AQ	2.740	s		<b>F1 PROC</b>			Parameters F1
FIDRES	0.365	Hz		SI	32		
D 1	0.300	s					
VDLIST	npt_gradrec		default				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPX 1	75.000	%					
P 16	5000.000	us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.


Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).


Requirements: At least time of recovery delay, which defines start point of range one, needs to be specified.

## 5.2.64 Gradient recovery test for Y-direction [-] (NPT\_1H\_gradrecY\_sqn\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT					
NO#	SecDel [dl6]	Scal.Int. [#26]	B0-shift [Hz]	APK0 [APK #6]	
1	10.00 us	99.30	4.68	4.0	
2	15.85 us	99.17	4.53	2.5	
3	25.12 us	98.28	4.30	2.0	
4	39.81 us	97.73	4.13	2.0	
5	63.10 us	97.76	3.92	2.0	
6	100.0 us	97.70	3.75	1.5	
7	158.5 us	97.62	3.54	1.5	
8	251.2 us	97.68	3.38	1.0	
9	398.1 us	97.64	3.17	0.5	
10	630.9 us	97.85	3.00	0.0	
11	1.000 ms	97.95	2.79	0.0	
12	1.585 ms	98.29	2.62	-0.5	
13	2.512 ms	98.60	2.43	1.0	
14	3.981 ms	98.78	2.24	0.0	
15	6.310 ms	99.18	2.03	0.0	
16	10.00 ms	99.58	1.86	0.5	
17	15.85 ms	99.75	1.65	-0.5	
18	25.12 ms	99.77	1.48	0.0	
19	39.81 ms	99.85	1.33	-0.5	
20	63.10 ms	99.82	1.10	-0.5	
21	0.100 s	99.95	0.93	0.5	
22	0.158 s	99.92	0.72	-0.5	
23	0.251 s	99.90	0.55	0.0	
24	0.398 s	99.97	0.38	0.0	
25	0.631 s	99.96	0.21	0.0	
26	1.000 s	100.00	0.00	0.0	



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B0), phase deviation (apk0).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_gradrecvd			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	2			F1P	5.720	ppm	
RG	0.250		optim. by RGA	F2P	5.320	ppm	
O1P	4.697	ppm		<b>F1 ACQU</b>			Parameters F1
SW	7.528	ppm		NUC1	1H		
TD	16506		field dependent	TD	26		
AQ	2.740	s		<b>F1 PROC</b>			Parameters F1
FIDRES	0.365	Hz		SI	32		
D 1	0.300	s					
VDLIST	npt_gradrec		default				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPNAME1	RECT.1						
GPY 1	-75.000	%					
P 16	5000.000	us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.

Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of range one, needs to be specified.


## 5.2.65 Gradient recovery test for Y-direction [+] (NPT\_1H\_gradrecY\_sq\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727

**Solvent:** D<sub>2</sub>O


**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT

NO#	SecDel [dl6]	Scal.Int. [#26]	B0-shift [Hz]	APK0 [APK #36]
1	10.00 us	99.46	4.60	-3.5
2	15.85 us	99.62	4.43	-2.0
3	25.12 us	98.62	4.28	-1.0
4	39.82 us	98.34	4.05	-1.5
5	63.10 us	98.31	3.92	-2.0
6	100.0 us	98.24	3.73	-1.5
7	158.5 us	98.16	3.54	-1.5
8	251.2 us	98.10	3.34	-1.5
9	398.1 us	98.16	3.23	-1.0
10	630.9 us	98.04	2.96	-0.5
11	1.000 ms	98.55	2.79	0.0
12	1.585 ms	98.46	2.62	-0.5
13	2.512 ms	98.70	2.43	-0.5
14	3.981 ms	99.04	2.24	-0.5
15	6.310 ms	99.38	2.07	-0.5
16	10.00 ms	99.72	1.86	0.0
17	15.85 ms	99.63	1.69	-0.5
18	25.12 ms	99.31	1.48	-0.5
19	39.81 ms	99.33	1.35	-0.5
20	63.10 ms	99.39	1.10	-0.5
21	0.100 s	99.97	1.01	-0.5
22	0.158 s	100.00	0.76	0.0
23	0.251 s	99.92	0.59	0.0
24	0.398 s	99.95	0.42	-0.5
25	0.631 s	99.98	0.25	0.0
26	1.000 s	100.00	0.00	0.0



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B<sub>0</sub>), phase deviation (apk<sub>0</sub>).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test



## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_gradrecvd			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	2			F1P	5.720	ppm	
RG	0.250		optim. by RGA	F2P	5.320	ppm	
O1P	4.697	ppm		<b>F1 ACQU</b>			Parameters F1
SW	7.528	ppm		NUC1	1H		
TD	16506		field dependent	TD	26		
AQ	2.740	s		<b>F1 PROC</b>			Parameters F1
FIDRES	0.365	Hz		SI	32		
D 1	0.300	s					
VDLIST	npt_gradrec		default				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPNAM1	RECT.1						
GPY 1	75.000	%					
P 16	5000.000	us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.



Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of range one, needs to be specified.

## 5.2.66 Gradient recovery test for Z-direction [-] (NPT\_1H\_gradrecZ\_sqn\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727, Z142231  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT					
NO#	SecDel [dl6]	Scal.Int. [426]	B0-shift [Hz]	-APK0 [APK #6]	
1	10.00 us	99.30	4.68	4.0	
2	15.85 us	99.17	4.53	2.5	
3	25.12 us	98.28	4.30	2.0	
4	39.81 us	97.73	4.13	2.0	
5	63.10 us	97.76	3.92	2.0	
6	100.0 us	97.70	3.75	1.5	
7	158.5 us	97.62	3.54	1.5	
8	251.2 us	97.68	3.38	1.0	
9	398.1 us	97.64	3.17	0.5	
10	630.9 us	97.85	3.00	0.0	
11	1.000 ms	97.95	2.79	0.0	
12	1.585 ms	98.29	2.62	-0.5	
13	2.512 ms	98.60	2.41	1.0	
14	3.981 ms	98.78	2.24	0.0	
15	6.310 ms	99.18	2.03	0.0	
16	10.00 ms	99.58	1.86	0.5	
17	15.85 ms	99.75	1.65	-0.5	
18	25.12 ms	99.77	1.48	0.0	
19	39.81 ms	99.85	1.33	-0.5	
20	63.10 ms	99.82	1.10	-0.5	
21	0.100 s	99.95	0.93	0.5	
22	0.158 s	99.92	0.72	-0.5	
23	0.251 s	99.90	0.55	0.0	
24	0.398 s	99.97	0.38	0.0	
25	0.631 s	99.96	0.21	0.0	
26	1.000 s	100.00	0.00	0.0	

### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B<sub>0</sub>), phase deviation (apk0).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test

## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H            PARAMODE 1            PULPROG npt_gradrecvd            NS 1            DS 2            RG 0.250            O1P 4.697 ppm            SW 7.528 ppm            TD 16506            AQ 2.740 s            FIDRES 0.365 Hz            D 1 0.300 s            VDLIST npt_gradrec            P 1 14.0 us            PLW 1 6.6 W            GPNAM1 RECT.1            GPZ 1 -75.000 %            P 16 5000.000 us            TE 298.000 K</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>optim. by RGA</p> <p>field dependent</p> <p>default 90deg Pulse Pow@90deg(Specs)</p> <p>gradient pulse default</p>	<p><b>F2 PROC</b></p> <p>SI 8192            WDW 1            LB 1.000 Hz            PH_mod 1            F1P 5.720 ppm            F2P 5.320 ppm</p> <p><b>F1 ACQU</b></p> <p>NUC1 1H            TD 26</p> <p><b>F1 PROC</b></p> <p>SI 32</p>	<p>Parameters F2</p> <p>pk</p> <p>Parameters F1</p> <p>Parameters F1</p>
--	--	---	--

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.



Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of range one, needs to be specified.

## 5.2.67 Gradient recovery test for Z-direction [+] (NPT\_1H\_gradrecZ\_sq\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727, Z142231  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT					
NO#	SecDel [dl6]	Scal.Int. [#26]	B0-shift [Hz]	APK0 [APK #6]	
1	10.00 us	99.46	4.60	-3.5	
2	15.85 us	99.62	4.43	-2.0	
3	25.12 us	98.62	4.28	-1.0	
4	39.81 us	98.34	4.05	-1.5	
5	63.10 us	98.31	3.92	-2.0	
6	100.0 us	98.24	3.73	-1.5	
7	158.5 us	98.16	3.54	-1.5	
8	251.2 us	98.10	3.34	-1.5	
9	398.1 us	98.16	3.23	-1.0	
10	630.9 us	98.04	2.96	-0.5	
11	1.000 ms	98.55	2.79	0.0	
12	1.585 ms	98.46	2.62	-0.5	
13	2.512 ms	98.70	2.43	-0.5	
14	3.981 ms	99.04	2.24	-0.5	
15	6.310 ms	99.38	2.07	-0.5	
16	10.00 ms	99.72	1.86	0.0	
17	15.85 ms	99.63	1.69	-0.5	
18	25.12 ms	99.31	1.48	-0.5	
19	39.81 ms	99.33	1.35	-0.5	
20	63.10 ms	99.39	1.10	-0.5	
21	0.100 s	99.97	1.01	-0.5	
22	0.158 s	100.00	0.76	0.0	
23	0.251 s	99.92	0.59	0.0	
24	0.398 s	99.95	0.42	-0.5	
25	0.631 s	99.98	0.25	0.0	
26	1.000 s	100.00	0.00	0.0	

### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B<sub>0</sub>), phase deviation (apk<sub>0</sub>).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test

## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H            PARAMODE 1            PULPROG npt_gradrecvd            NS 1            DS 2            RG 0.250            O1P 4.697 ppm            SW 7.528 ppm            TD 16506            AQ 2.740 s            FIDRES 0.365 Hz            D 1 0.300 s            VDLIST npt_gradrec            P 1 14.0 us            PLW 1 6.6 W            GPNAM1 RECT.1            GPZ 1 75.000 %            P 16 5000.000 us            TE 298.000 K</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>optim. by RGA</p> <p>field dependent</p> <p>default 90deg Pulse Pow@90deg(Specs)</p> <p>gradient pulse default</p>
<p><b>F2 PROC</b></p> <p>SI 8192            WDW 1            LB 1.000 Hz            PH_mod 1            F1P 5.720 ppm            F2P 5.320 ppm</p> <p><b>F1 ACQU</b></p> <p>NUC1 1H            TD 26</p> <p><b>F1 PROC</b></p> <p>SI 32</p>	<p>Parameters F2</p> <p>pk</p> <p>Parameters F1</p> <p>Parameters F1</p>

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.


Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of range one, needs to be specified.


## 5.2.68 Gradient recovery test for X-direction [-] with trapezoid pulses (NPT\_1H\_gradrecX\_trapezoid\_neg\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-1<sup>3</sup>C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z10085, Z10084  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT

NO[4]	RecDel[1416]	Real.Int.[#10]	SI[SI]E[10]	PHAS[AP0] #32
1	0.10 ms			
2	0.13 ms	36.03	-1.09	16.94
3	0.16 ms	47.00	-1.14	0.70
4	0.24 ms	83.04	-1.04	0.44
5	0.33 ms	91.47	-1.20	-0.24
6	0.44 ms	95.54	-1.00	0.08
7	0.59 ms	97.31	-1.30	0.44
8	0.80 ms	97.09	-0.99	0.48
9	1.08 ms	98.09	-1.42	-0.12
10	1.45 ms	98.93	-0.77	-0.20
11	1.85 ms	98.80	-1.52	0.22
12	2.43 ms	98.78	-0.43	-0.44
13	3.54 ms	99.05	-1.47	-0.02
14	4.76 ms	99.99	-0.47	0.09
15	6.40 ms	99.29	-1.01	0.10
16	8.42 ms	99.54	-0.24	-0.12
17	11.40 ms	99.43	-1.94	-0.14
18	15.42 ms	99.74	-0.09	0.10
19	21.02 ms	99.95	-2.01	-0.14
20	28.29 ms	100.00	-0.01	-0.14
21	38.00 ms	99.94	-2.03	-0.12
22	51.25 ms	99.99	-0.00	-0.14
23	68.98 ms	100.00	-2.02	-0.10
24	92.94 ms	100.00	-0.01	-0.12
25	124.94 ms	99.94	-2.00	0.24
26	168.19 ms	99.95	0.03	-0.18
27	224.38 ms	100.01	-2.03	0.04
28	294.70 ms	99.94	0.00	0.04
29	410.11 ms	99.94	-2.00	0.14
30	552.00 ms	99.97	0.00	0.00
31	742.94 ms	99.99	-2.04	-0.08
32	1.00 w	100.00	0.00	0.00



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B0), phase deviation (apk0).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test

## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H            PARAMODE 1            PULPROG npt_gradrectrvd            NS 1            DS 2            RG 0.250            O1P 4.697 ppm            SW 7.528 ppm            TD 16506            AQ 2.740 s            FIDRES 0.365 Hz            D 1 0.300 s            VDLIST npt_gradrectTrapezoid default            P 1 14.0 us 90deg Pulse            PLW 1 6.6 W Pow@90deg(Specs)            GPNAM1 GRADREC5m            GPX 1 0.000 % 7.5 A            P 16 5000.000 us gradient pulse            TE 298.000 K default</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>optim. by RGA</p> <p>field dependent</p> <p>default</p> <p>90deg Pulse</p> <p>Pow@90deg(Specs)</p> <p>7.5 A</p> <p>gradient pulse</p> <p>default</p>
<p><b>F2 PROC</b></p> <p>SI 8192            WDW 1            LB 1.000 Hz            PH_mod 1            F1P 5.720 ppm            F2P 5.320 ppm</p> <p><b>F1 ACQU</b></p> <p>NUC1 1H            TD 32</p> <p><b>F1 PROC</b></p> <p>SI 32</p>	<p>Parameters F2</p> <p>pk</p> <p>Parameters F1</p> <p>Parameters F1</p>

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.


Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of evaluation range, needs to be specified.


## 5.2.69 Gradient recovery test for X-direction [+] with trapezoid pulses (NPT\_1H\_gradrecX\_trapezoid\_pos\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z10085, Z10084  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT

NO[4]	RecDel[14]6	Real-Int.[#10]	SI-Shift[Hz]	PHAS[AP0] #32
1	0.10 ms			
2	0.13 ms	34.33	0.86	7.36
3	0.16 ms	44.55	0.90	3.10
4	0.24 ms	82.21	0.85	1.54
5	0.33 ms	99.99	0.89	0.58
6	0.44 ms	94.84	0.88	-0.82
7	0.59 ms	94.71	0.88	-0.50
8	0.80 ms	97.70	0.90	-0.14
9	1.08 ms	97.89	0.84	0.14
10	1.45 ms	98.07	0.91	0.24
11	1.95 ms	98.14	0.83	0.24
12	2.43 ms	98.49	0.88	0.04
13	3.54 ms	98.79	0.85	-0.18
14	4.76 ms	99.04	0.84	-0.12
15	6.40 ms	99.30	0.84	0.22
16	8.42 ms	99.44	0.76	-0.08
17	11.40 ms	99.75	0.84	-0.24
18	15.42 ms	99.90	0.75	-0.20
19	21.02 ms	99.90	0.82	0.04
20	28.29 ms	100.03	0.67	0.10
21	38.00 ms	100.09	0.89	-0.08
22	51.25 ms	100.01	0.44	0.02
23	68.98 ms	100.09	0.87	0.10
24	92.84 ms	99.99	0.87	-0.12
25	124.94 ms	99.95	0.96	-0.10
26	168.19 ms	100.01	0.44	-0.12
27	224.38 ms	100.05	1.09	-0.04
28	284.70 ms	100.06	0.20	-0.02
29	410.11 ms	100.08	1.25	0.02
30	552.00 ms	100.07	0.12	0.08
31	742.94 ms	100.07	1.41	0.08
32	1.0 s	100.00	0.00	0.00



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B0), phase deviation (apk0).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test



## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_gradrectrvd			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	2			F1P	5.720	ppm	
RG	0.250		optim. by RGA	F2P	5.320	ppm	
O1P	4.697	ppm		<b>F1 ACQU</b>			Parameters F1
SW	7.528	ppm		NUC1	1H		
TD	16506		field dependent	TD	32		
AQ	2.740	s		<b>F1 PROC</b>			Parameters F1
FIDRES	0.365	Hz		SI	32		
D 1	0.300	s					
VDLIST	npt_gradrecTrapezoid		default				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPNAM1	GRADREC5m						
GPX 1	0.000	%	7.5 A				
P 16	5000.000	us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.


Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of evaluation range, needs to be specified.


## 5.2.70 Gradient recovery test for Y-direction [-] with trapezoid pulses (NPT\_1H\_gradrecY\_trapezoid\_neg\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-1<sup>3</sup>C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z10085, Z10084  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT

NO[4]	RecDel[1416]	Eval.Int.[#10]	SI[SIa]E[SIa]	PHAS[AP0] #32
1	0.10 ms			
2	0.13 ms	36.03	-1.09	16.94
3	0.16 ms	47.00	-1.14	0.70
4	0.24 ms	83.04	-1.04	0.44
5	0.33 ms	91.47	-1.20	-0.24
6	0.44 ms	95.54	-1.00	0.08
7	0.59 ms	97.31	-1.30	0.44
8	0.80 ms	97.09	-0.99	0.48
9	1.08 ms	98.09	-1.42	-0.12
10	1.45 ms	98.51	-0.77	-0.20
11	1.85 ms	98.80	-1.52	0.22
12	2.43 ms	98.78	-0.43	-0.44
13	3.54 ms	99.05	-1.47	-0.02
14	4.76 ms	99.99	-0.47	0.09
15	6.40 ms	99.29	-1.01	0.10
16	8.42 ms	99.54	-0.24	-0.12
17	11.40 ms	99.45	-1.94	-0.14
18	15.42 ms	99.74	-0.09	0.10
19	21.02 ms	99.95	-0.01	-0.14
20	28.29 ms	100.00	-0.01	-0.14
21	38.00 ms	99.94	-0.03	-0.12
22	51.25 ms	99.99	-0.00	-0.14
23	68.98 ms	100.00	-0.00	-0.10
24	92.94 ms	100.00	-0.01	-0.12
25	124.94 ms	99.94	-0.00	0.24
26	168.19 ms	99.95	0.03	-0.18
27	224.38 ms	100.01	-0.03	0.04
28	294.70 ms	99.94	0.00	0.04
29	410.11 ms	99.94	-0.00	0.14
30	552.00 ms	99.97	0.00	0.00
31	742.94 ms	99.99	-0.04	-0.08
32	1.00 w	100.00	0.00	0.00



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B0), phase deviation (ap0).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_gradrectrvd			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	2			F1P	5.720	ppm	
RG	0.250		optim. by RGA	F2P	5.320	ppm	
O1P	4.697	ppm		<b>F1 ACQU</b>			Parameters F1
SW	7.528	ppm		NUC1	1H		
TD	16506		field dependent	TD	32		
AQ	2.740	s		<b>F1 PROC</b>			Parameters F1
FIDRES	0.365	Hz		SI	32		
D 1	0.300	s					
VDLIST	npt_gradrecTrapezoid		default				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPNAM1	GRADREC5m						
GPY 1	0.000	%	7.5 A				
P 16	5000.000	us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.


Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of evaluation range, needs to be specified.


## 5.2.71 Gradient recovery test for Y-direction [+] with trapezoid pulses (NPT\_1H\_gradrecY\_trapezoid\_pos\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-1<sup>3</sup>C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z10085, Z10084  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT

NO.[#]	RecDel.[s]	Real-Int. [F10]	SI-Shift [Hz]	PHAS [AP0: #32]
1	0.10 ms	100.00	0.00	0.00
2	0.13 ms	34.33	0.86	7.36
3	0.16 ms	44.55	0.90	3.10
4	0.24 ms	82.21	0.85	1.54
5	0.33 ms	99.99	0.89	0.58
6	0.44 ms	94.84	0.88	-0.82
7	0.59 ms	94.71	0.88	-0.50
8	0.80 ms	97.70	0.90	-0.14
9	1.08 ms	97.89	0.84	0.14
10	1.45 ms	98.07	0.91	0.24
11	1.95 ms	98.14	0.83	0.24
12	2.43 ms	98.49	0.88	0.04
13	3.54 ms	98.79	0.85	-0.18
14	4.76 ms	99.04	0.84	-0.12
15	6.40 ms	99.30	0.84	0.22
16	8.42 ms	99.44	0.76	-0.08
17	11.40 ms	99.75	0.84	-0.24
18	15.42 ms	99.90	0.75	-0.20
19	21.02 ms	99.90	0.82	0.04
20	28.29 ms	100.03	0.67	0.10
21	38.00 ms	100.09	0.89	-0.08
22	51.25 ms	100.01	0.44	0.02
23	68.98 ms	100.09	0.87	0.10
24	92.84 ms	99.99	0.87	-0.12
25	124.94 ms	99.95	0.96	-0.10
26	168.19 ms	100.01	0.44	-0.12
27	224.38 ms	100.05	1.09	-0.04
28	284.70 ms	100.06	0.20	-0.02
29	410.11 ms	100.08	1.25	0.02
30	552.00 ms	100.07	0.12	0.08
31	742.94 ms	100.07	1.41	0.08
32	1.00 s	100.00	0.00	0.00



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B0), phase deviation (apk0).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test

## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H            PARAMODE 1            PULPROG npt_gradrectrvd            NS 1            DS 2            RG 0.250            O1P 4.697 ppm            SW 7.528 ppm            TD 16506            AQ 2.740 s            FIDRES 0.365 Hz            D 1 0.300 s            VDLIST npt_gradrectTrapezoid default            P 1 14.0 us 90deg Pulse            PLW 1 6.6 W Pow@90deg(Specs)            GPNAM1 GRADREC5m            GPY 1 0.000 % 7.5 A            P 16 5000.000 us gradient pulse            TE 298.000 K default</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>optim. by RGA</p> <p>field dependent</p> <p>default</p> <p>90deg Pulse</p> <p>Pow@90deg(Specs)</p> <p>7.5 A</p> <p>gradient pulse</p> <p>default</p>
<p><b>F2 PROC</b></p> <p>SI 8192            WDW 1            LB 1.000 Hz            PH_mod 1            F1P 5.720 ppm            F2P 5.320 ppm</p> <p><b>F1 ACQU</b></p> <p>NUC1 1H            TD 32</p> <p><b>F1 PROC</b></p> <p>SI 32</p>	<p>Parameters F2</p> <p>pk</p> <p>Parameters F1</p> <p>Parameters F1</p>

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.


Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of evaluation range, needs to be specified.


## 5.2.72 Gradient recovery test for Z-direction [-] with trapezoid pulses (NPT\_1H\_gradrecZ\_trapezoid\_neg\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-1<sup>3</sup>C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z10085, Z10084  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT

NO[4]	RecDel[1416]	Real.Int.[#12]	SI[SIa]E[SIa]	PHAS[AP] #322
1	0.10 ms			
2	0.13 ms	36.03	-1.09	16.94
3	0.16 ms	47.00	-1.14	0.70
4	0.24 ms	83.04	-1.04	0.44
5	0.33 ms	91.47	-1.20	-0.24
6	0.44 ms	95.54	-1.00	0.08
7	0.59 ms	97.31	-1.30	0.44
8	0.80 ms	97.09	-0.99	0.48
9	1.08 ms	98.09	-1.42	-0.12
10	1.45 ms	98.51	-0.77	-0.20
11	1.85 ms	98.80	-1.52	0.22
12	2.43 ms	98.78	-0.43	-0.44
13	3.54 ms	99.05	-1.47	-0.02
14	4.76 ms	99.99	-0.47	0.09
15	6.40 ms	99.29	-1.01	0.10
16	8.42 ms	99.54	-0.24	-0.12
17	11.40 ms	99.63	-1.94	-0.14
18	15.42 ms	99.74	-0.09	0.10
19	21.02 ms	99.95	-0.01	-0.14
20	28.29 ms	100.00	-0.01	-0.14
21	38.00 ms	99.94	-0.03	-0.12
22	51.25 ms	99.99	-0.00	-0.14
23	68.98 ms	100.00	-0.00	-0.10
24	92.94 ms	100.00	-0.01	-0.12
25	124.94 ms	99.94	-0.00	0.24
26	168.19 ms	99.95	0.03	-0.18
27	224.38 ms	100.01	-0.03	0.04
28	294.70 ms	99.94	0.00	0.04
29	410.11 ms	99.94	-0.00	0.14
30	552.00 ms	99.97	0.00	0.00
31	742.94 ms	99.99	-0.04	-0.08
32	1.00 w	100.00	0.00	0.00



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B0), phase deviation (apk0).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_gradrectrvd			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	2			F1P	5.720	ppm	
RG	0.250		optim. by RGA	F2P	5.320	ppm	
O1P	4.697	ppm		<b>F1 ACQU</b>			Parameters F1
SW	7.528	ppm		NUC1	1H		
TD	16506		field dependent	TD	32		
AQ	2.740	s		<b>F1 PROC</b>			Parameters F1
FIDRES	0.365	Hz		SI	32		
D 1	0.300	s					
VDLIST	npt_gradrecTrapezoid		default				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPNAM1	GRADREC5m						
GPZ 1	-75.000	%	7.5 A				
P 16	5000.000	us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.


Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of evaluation range, needs to be specified.


## 5.2.73 Gradient recovery test for Z-direction [+] with trapezoid pulses (NPT\_1H\_gradrecZ\_trapezoid\_pos\_1h)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-1<sup>3</sup>C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
 Z10083, Z10085, Z10084  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



EVALUATION RESULTS OF GRADIENT RECOVERY MEASUREMENT

NO[4]	RecDel[14]6	Real-Int.[#32]	SI-Shift[Hz]	PHAS[AP0] #32
1	0.10 ms	100.00	0.00	0.00
2	0.13 ms	34.33	0.86	7.36
3	0.16 ms	44.55	0.90	3.10
4	0.24 ms	82.21	0.85	1.54
5	0.33 ms	99.99	0.89	0.58
6	0.44 ms	94.84	0.88	-0.82
7	0.59 ms	94.71	0.88	-0.50
8	0.80 ms	97.70	0.90	-0.14
9	1.08 ms	97.89	0.84	0.14
10	1.45 ms	98.07	0.91	0.24
11	1.95 ms	99.14	0.83	0.24
12	2.43 ms	98.49	0.88	0.04
13	3.54 ms	98.79	0.85	-0.18
14	4.76 ms	99.04	0.84	-0.12
15	6.40 ms	99.39	0.84	0.22
16	8.42 ms	99.44	0.76	-0.08
17	11.40 ms	99.75	0.84	-0.24
18	15.42 ms	99.90	0.75	-0.20
19	21.02 ms	99.90	0.82	0.04
20	28.29 ms	100.03	0.67	0.10
21	38.00 ms	100.09	0.89	-0.08
22	51.25 ms	100.01	0.44	0.02
23	68.98 ms	100.09	0.87	0.10
24	92.84 ms	99.99	0.87	-0.12
25	124.94 ms	99.95	0.96	-0.10
26	168.19 ms	100.01	0.44	-0.12
27	224.38 ms	100.05	1.09	-0.04
28	284.70 ms	100.06	0.20	-0.02
29	410.11 ms	100.08	1.25	0.02
30	552.00 ms	100.07	0.12	0.08
31	742.94 ms	100.07	1.41	0.08
32	1.0 w	100.00	0.00	0.00



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, signal intensity, signal shift (B0), phase deviation (apk0).

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check and gradient amplifier functioning test



## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	8192		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_gradrectrvd			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	2			F1P	5.720	ppm	
RG	0.250		optim. by RGA	F2P	5.320	ppm	
O1P	4.697	ppm		<b>F1 ACQU</b>			Parameters F1
SW	7.528	ppm		NUC1	1H		
TD	16506		field dependent	TD	32		
AQ	2.740	s		<b>F1 PROC</b>			Parameters F1
FIDRES	0.365	Hz		SI	32		
D 1	0.300	s					
VDLIST	npt_gradrecTrapezoid		default				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
GPNAM1	GRADREC5m						
GPZ 1	75.000	%	7.5 A				
P 16	5000.000	us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

The next step is a gradient echo experiment on a derived data set to check the functioning of the gradient amplifier.

SINO check and gradient amplifier functioning test can be skipped with L23=10

After successful SINO and gradient checks, the pseudo 2D gradient recovery experiment will be acquired. Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.

Evaluation consists of the determination of the zero-order phase correction difference to the reference experiment (last of the series) and the determination of the intensity deviation again referenced to the last experiment.

The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of evaluation range, needs to be specified.

## 5.2.74 Inverse spin echo difference (NPT\_1H\_hmqc1df2\_13c)


---

**Test Sample:** 0.3%, 1.0% or 3.0% Chloroform in Acetone-D6  
Z10230, Z10248, Z10701, Z100926, Z10250, Z10031, Z10030, Z10029, Z10260,  
Z10249, Z10275, Z10272, Z10717

**Solvent:** Acetone

**Lock parameter:** Lockregulation based on LGAIN=80 dB and default LOCKPOWER

**Sample State:** Rotation off



### Example Printout

Inverse spin echo difference spectrum (1H) of the <sup>13</sup>C satellites of chloroform. Printing range is centered around the suppressed proton signal of chloroform attached to <sup>12</sup>C (no coupling).

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F1 ACQU</b>			Parameters	<b>F1 PROC</b>			Parameters
PARMODE	0		Data Dimension	SI	16384		
TD0	1			WDW	1		
NUC1	1H			LB	0.000	Hz	
NUC2	13C			PH_mod	1		
PULPROG	hmqcndrd1d			F1P	2475.044	ppm	
NS	8			F2P	2345.044	ppm	
DS	4			<b>NMRPT</b>			Parameters
RG	101.000		no optim.	CNST 50	2.000	min	Waittime for AQ
O1P	8.000	ppm	CHCl3 (1H)				
O2P	76.987	ppm	CDCl3 (13C)				
SWH	1000.000	Hz					
TD	16384						
AQ	8.192	s					
FIDRES	0.122	Hz					
D 1	14.000	s					
CNST 2	216.000	Hz	J[CH]				
P 1	14.0	us	90deg NUC1				
P 3	11.0	us	90deg NUC2				
PLW 1	6.6	W	Pow@P90(Specs)				
PLW 2	26.5	W	Pow@P90(Specs)				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is to show phase stability of the spectrometer. The central signal (1H attached to 12C) should ideally be suppressed. Influencing variables are lock stability, instrument environment and sample temperature besides the electronic stability and quality of the instrument itself. It is very important, that the sample is sufficiently relaxed before starting the experiment.

NMRPT is doing an O1 optimization, followed by a delay of 600 s before the main experiment.

Processing is accomplished with no line broadening. Data evaluation consists of the comparison of the intensity ratio satellite signal versus central signal under consideration of potential unlike sign.


## 5.2.75 Inverse spin echo difference experiment [2D] (NPT\_1H\_hmqc2df2\_13c)

**Test Sample:** 0.3%, 1.0% or 3.0% Chloroform in Acetone-D6  
 Z10230, Z10248, Z10701, Z100926, Z10250, Z10031, Z10030, Z10029, Z10260,  
 Z10249, Z10275, Z10272, Z10717

**Solvent:** Acetone

**Lock parameter:** Lockregulation based on LGAIN=80 dB and default LOCKPOWER

**Sample State:** Rotation off



### Example Printout

Inverse spin echo difference spectrum (1H) of the  $^{13}\text{C}$  satellites of chloroform.

The top spectrum shows a series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the left satellite and the suppressed proton signal of chloroform attached to  $^{12}\text{C}$  (no coupling). For all other spectra the printing range is centered around the suppressed proton signal.

The bottom spectrum to the left shows the pseudo-2D transformed with XF2 in magnitude mode. The middle spectrum to the right shows the row with highest suppression of the central peak. The bottom spectrum to the right shows the row with lowest suppression.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
PARMODE	1		Data Dimension	SI	65536		
TD0	1			WDW	0		no
NUC1	1H			LB	0.000	Hz	0=default
NUC2	13C			PH_mod	2		mc
PULPROG	npt_hmqcndrd2d			F1P	2475.044	ppm	
NS	2			F2P	2345.044	ppm	
DS	8			<b>F1 ACQU</b>			Parameters F1
RG	101.000		no optim.	TD	32		
O1P	8.000	ppm	CHCl3 (1H)	<b>F1 PROC</b>			Parameters F1
O2P	76.988	ppm	CDCl3 (13C)	SI	32		
SWH	1000.000	Hz		<b>NMRPT</b>			Parameters
TD	32768			CNST	50	min	Waittime for AQ
AQ	16.384	s					
FIDRES	0.061	Hz					
D 1	10.000	s					
CNST 2	214.900	Hz	J[CH]				
P 1	14.0	us	90deg NUC1				
P 3	11.0	us	90deg NUC2				
PLW 1	6.6	W	Pow@P90(Specs)				
PLW 2	26.5	W	Pow@P90(Specs)				
TE	298.000	K	default				

## Experiment Description

Purpose of this experiment is to show phase stability of the spectrometer over extended time periods. The central signal (1H attached to 12C) should ideally be suppressed. Influencing variables are lock stability, instrument environment and sample temperature besides the electronic stability and quality of the instrument itself.


NMRPT is going through the preparative steps:

- 1) LOCK optimization (regulation, lockgain, lockphase, lockgain)
- 2) Exact O1P determination
- 3) Waiting time for sample temperature stabilisation (CNST50, in minutes)
- 4) Main acquisition

Processing is accomplished with no line broadening. For each row the intensity ratio is calculated as ratio of the intensity of central peak versus the average intensity of the two satellites. For calculation of the average intensity ratio first the intensities of central peak and the average intensities of the two satellites of each row are added separately. Maximal, minimal, and average intensity ratio are stated in the experiment title.

## 5.2.76 1H of pseudo honey with noesy for solvent suppression (NPT\_1H\_honeyNoesy)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
H177072  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Bottom & Top: 1H noesy overview spectrum of pseudo honey sample with different scaling.

### Control Option for Acquisition (L23)

- 1 default, GPZ1=50, GPZ2=-10
- 2 GPZ1=-10, GPZ2=50
- 3 GPZ1=0, GPZ2=0
- 4 GPZ1 and GPZ2 can be set by user

## Parameters

F1 ACQU Parameters				F1 PROC Parameters			
NUC1	1H			SI	131072		
PULPROG	noesygppr1d			LB	0.300	Hz	
NS	32			CY	5.500	cm	
DS	4			<b>NMRPT</b>			Parameters
RG	16.000			CNST 50	1000.000		Scaling factor for CY
O1P	4.702	ppm					
SWH	8196.722	Hz					
TD	65536						
AQ	3.998	s					
FIDRES	0.250	Hz	field dependent				
D 1	4.000	s					
D 8	0.010	s					
D 16	0.000	s					
P 0	14.0	us	90deg Pulse				
P 1	14.0	us	90deg Pulse				
P 16	1000.000	us	gradient pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
PLW 9	0.015	mW	Pow@90deg(25 Hz)				
GPNAM1	SMSQ10.100						
GPZ 1	50.000	%					
GPNAM2	SMSQ10.100						
GPZ 2	-10.000	%					
TE	298.000	K	default				

## Experiment Description

1H spectrum with presaturation and noesy for solvent suppression.

Before the acquisition, the proton 90 degrees pulse is calibrated using pulscal, which result is stored in the corresponding derived dataset. The pulse determination is skipped if experiment is measured with option 'Skip Getprosol'.

After pulscal PLW 9 is calculated and set for excitation bandwidth of 25 Hz.


Processing of spectrum with window funktion EM and zero order phase correction followed by a lortentzian baseline correction.

For evaluation the RMSD ratio of baseline to noise is determined. RMSD per point of baseline is determined over the whole spectrum excluding signal regions and spectrum edges. RMSD per point of noise is determined from 14.0 to 13.0 ppm after baseline correction (ABSF) to minimize baseline offset.

## 5.2.77 1H homodecoupling (NPT\_1H\_homodec)

---

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10121, Z10033, Z10270, Z10718  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Proton spectrum of 0.1% Ethylbenzene with homodecoupled methyl group at 1.26 ppm.

### Control Option for Acquisition (L23)

1 default



## Parameters


F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	1H			SI	32768		
NUC2	1H			WDW	1		
PULPROG	npt_zghdp0.2			LB	1.000	Hz	
NS	8			PC	1.000		
DS	4			F1P	8.000	ppm	
RG	101.000		no optim.	F2P	-2.000	ppm	
O1P	3.000	ppm		SIGF1	1.500	ppm	
O1P	3.000	ppm		SIGF2	0.250	ppm	
SW	9.917	ppm		NOISF1	0.200	ppm	
TD	32768			NOISF2	-1.800	ppm	
AQ	4.129	s	field dependent	CY	11.000	cm	
FIDRES	0.242	Hz	field dependent				
D 1	10.871	s	AQ+D1=constant				
DIGMOD	1		default				
HDDUTY	20.000	%	hd duty cycle				
HDRATE	20.000		oversampl. hd				
P 0		us	P 1 * CNST 10 / 90				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
PLW 24	0.004	W	Pow@hd(Specs) NUC2				
CNST 10	30.000	deg	flip angle				
TE	298.000	K	default				

## Experiment Description

Homodecoupling test is performed to check apparent spectrometer performance with respect to switching times. The irradiation power is proportional to ~2 ms pulse length. The experimental setup consists of the exact determination of the irradiation position (O2). The processed data are stored in PROCNO=2. Processing is accomplished with exponential multiplication using line broadening. Evaluation consists of signal-to-noise determination using fixed noise region.

## 5.2.78 2D 1H-13C HSQC (NPT\_1H\_hsqc\_10EB\_13c\_2d)

**Test Sample:** 10% Ethylbenzene (EB) in Chloroform-D  
Z10153, Z10154, Z10034, Z10253, Z10292, Z10723  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

2D 1H-13C HSQC of 10% ethyl benzene in CDCl<sub>3</sub>. The F2 projection is shown at the top, the F1 projection at the left.

### Control Option for Acquisition (L23)

- 1 default
- 2 use gradients with inverted sign (GPZ 1 = -80.0%, GPZ 2 = -20.1%)

## Parameters

<p><b>F2 ACQU</b></p> <p>PARAMODE 1          TD0 1          NUC1 1H          NUC2 13C          PULPROG hsqcedetgppsp.3          NS 1          DS 32          RG 0.250          O1P 4.700 ppm          O2P 70.000 ppm          SW 12.132 ppm          TD 2912          AQ 0.300 s          FIDRES 3.334 Hz          D 1 2.000 s          CNST 2 145.000 Hz          D 1 2.000 s          D 16 0.000 s          D 21 0.004 s          P 1 14.0 us          P 3 11.0 us          P 14 500.0 us          P 24 2000.0 us          P 28 0.1 us          PLW 1 16.2 W          PLW 2 54.35 W          PLW 12 0.8 W          PCPD 2 80 us          CPDPRG2 p5m4sp180.2          CPDPRG2 garp4          SPNAM3 Crp60,0.5,20.1          SPOAL 3 0.500          SPW 3 8.3 W          SPNAM18 Crp60_xfilt.2          SPOAL 18 0.500          SPW 18 1.6 W          SPNAM31 Crp32,1.5,20.2          SPOAL 31 0.500          SPW 31 0.9 W          GPNAM1 SMSQ10.100          GPNAM2 SMSQ10.100          GPZ 1 80.000 %          GPZ 2 20.100 %          P 16 1000 us          TE 298.000 K</p>	<p>Parameters F2          Data Dimension</p> <p>optim. by RGA          1H          13C)</p> <p>field dependent</p> <p>J[CH]</p> <p>gradrec del.          1/2 J[CH]</p> <p>90deg NUC1          90deg NUC2          sh.pul.invers.          sh.pul.refoc.          trimpul.          Pow@P90(Specs)          Pow@P90(Specs)          Pow@CPD(Specs)          90deg CPD NUC2          &gt; 600 MHz          &lt;= 600 MHz</p> <p>0.5          default</p> <p>default</p> <p>default</p> <p>default</p> <p>grad,pulse          default</p>
<p><b>F1 ACQU</b></p> <p>NUC1 13C          FnMODE 6          SW 165.000 ppm          TD 256          O1P 70.000 ppm</p>	<p>Parameters F1</p> <p>Parameters F1</p> <p>13C          Parameters F1</p>
<p><b>F2 PROC</b></p> <p>SI 8192          WDW 4          SSB 2.000          PH_mod 1          F1P 10.707 ppm          F2P -1.307 ppm</p>	<p>Parameters F2</p> <p>QSINE</p> <p>pk</p> <p>Parameters F1</p> <p>QSINE</p>
<p><b>F1 PROC</b></p> <p>SI 1024          WDW 4          SSB 2.000          PH_mod 0          F1P 152.405 ppm          F2P -12.405 ppm</p>	<p>Parameters F1</p> <p>QSINE</p>

## Experiment Description

2D 1H-13C HSQC is acquired phase-sensitive using shaped pulses for all 180 degree 13C pulses.

Before the acquisition, the proton 90 degrees pulse is calibrated using pulscal, which result is stored in the corresponding derived dataset. The pulse determination is skipped if experiment is measured with option 'Skip Getprosol'.

Contour levels for 2D spectrum are obtained by command levcalc. Levcalc sets the lowest contour level to LEV0\*S\_DEV, where S\_DEV (standard deviation) is a processing status parameter.


## 5.2.79 2D 1H-13C HSQC (NPT\_1H\_hsqc\_etsisp\_13c\_2d)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10902, Z10246, Z100930, Z10268, Z10036, Z10247, Z10267, Z10720, Z10719

**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

2D 1H-13C HSQC of 2 mM sucrose in 90%/10% H<sub>2</sub>O/D<sub>2</sub>O. The F2 projection from 95 to 50 ppm is shown at the top.

### Control Option for Acquisition (L23)

- 1 default
- 20 no decoupling during acquisition (PLW12=0)

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
PARMODE	1		Data Dimension	SI	2048		
TD0	1			WDW	4		QSINE
NUC1	1H			SSB	2.000		
NUC2	13C			PH_mod	1		pk
PULPROG	hsqcetgpsisp.2			SIGF1	5.500	ppm	
NS	4			SIGF2	5.100	ppm	
DS	64			NOISF1	9.580	ppm	
RG	101.000		no optim.	NOISF2	6.560	ppm	
TD	2048			F1P	0.000	ppm	
O1P	4.700	ppm	1H	F2P	0.000	ppm	
O2P	74.987	ppm	13C)	<b>F1 ACQU</b>			Parameters F1
SW	12.132	ppm		NUC1	13C		
TD	2048			FnMODE	6		
AQ	0.211	s	field dependent	SW	60.008	ppm	
FIDRES	4.741	Hz	field dependent	TD	256		
D 1	1.189	s		O1P	74.987	ppm	13C
CNST 2	145.000	Hz	J[CH]	<b>F1 PROC</b>			Parameters F1
CNST 17	-0.500		-0.5, Crp60comp.4	SI	512		
D 1	1.189	s		WDW	4		QSINE
D 16	0.000	s	gradrec del.	SSB	2.000		
D 24	0.001	s	1/8*J[CH]	PH_mod	0		no=default
P 1	14.0	us	90deg NUC1	SIGF1	93.000	ppm	
P 3	11.0	us	90deg NUC2	SIGF2	91.000	ppm	
P 14	500.0	us	sh.pul.invers.	NOISF1	81.600	ppm	
P 24	2000.0	us	sh.pul.refoc.	NOISF2	66.700	ppm	
P 28	0.1	us	trimpul.	F1P	0.000	ppm	
PLW 1	6.6	W	Pow@P90(Specs)	F2P	0.000	ppm	
PLW 2	26.8	W	Pow@P90(Specs)				
PLW 12	0.3	W	Pow@CPD(Specs)				
PCPD 2	102.5	us	90deg CPD NUC2				
CPDPRG2	garp		cpd seq.				
SPNAM3	Crp60,0.5,20.1						
SPOAL 3	0.500		0.5				
SPW 3	0.000	W	default				
SPNAM7	Crp60comp.4						
SPOAL 7	0.500		0.5				
SPW 7	0.000	W	default				
GPNAM1	SINE.100						
GPNAM2	SINE.100						
GPZ 1	80.000	%					
GPZ 2	20.100	%					
P 16	0.000	us	grad.pulse				
TE	298.000	K	default				

## Experiment Description


2D 1H-13C HSQC is acquired phase-sensitive using shaped pulses for all 180 degree 13C pulses. Evaluation is achieved by determining the 2-dimensional signal-to-noise ratio according to the parameters section. The evaluation routine is part of the NMRPT package.

Option L23=1 is standard whereas option L23=20 is non-standard and will not be considered as regular test from the 'NMRPT Control Structure'.

Contour levels for 2D spectrum are obtained by command levcalc. Levcalc sets the lowest contour level to LEV0\*S\_DEV, where S\_DEV (standard deviation) is a processing status parameter.

## 5.2.80 2D 1H-13C HSQC with adiabatic 13C decoupling (NPT\_1H\_hsqc\_etsisp\_adia13c\_2d)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10902, Z10246, Z100930, Z10268, Z10036, Z10247, Z10267, Z10719, Z10720  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

2D 1H-13C HSQC of 2 mM sucrose in 90%/10% H<sub>2</sub>O/D<sub>2</sub>O. The F2 projection from 95 to 50 ppm is shown at the top.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
PARMODE	1		Data Dimension	SI	2048		
TD0	1			WDW	4		QSINE
NUC1	1H			SSB	2.000		
NUC2	13C			PH_mod	1		pk
PULPROG	hsqcetgpsisp.2			SIGF1	5.500	ppm	
NS	4			SIGF2	5.100	ppm	
DS	64			NOISF1	9.580	ppm	
RG	101.000		no optim.	NOISF2	6.560	ppm	
TD	2426			F1P	0.000	ppm	
O1P	4.700	ppm	1H	F2P	0.000	ppm	
O2P	74.987	ppm	13C)	<b>F1 ACQU</b>			Parameters F1
SW	12.132	ppm		NUC1	13C		
TD	2426			FnMODE	6		
AQ	0.250	s	field dependent	SW	60.008	ppm	
FIDRES	4.002	Hz	field dependent	TD	256		
D 1	1.150	s		O1P	74.987	ppm	13C
CNST 2	145.000	Hz	J[CH]	<b>F1 PROC</b>			Parameters F1
CNST 17	-0.500		-0.5, Crp60comp.4	SI	512		
D 1	1.150	s		WDW	4		QSINE
D 16	0.000	s	gradrec del.	SSB	2.000		
D 24	0.001	s	1/8*J[CH]	PH_mod	0		no=default
P 1	14.0	us	90deg NUC1	SIGF1	93.000	ppm	
P 3	11.0	us	90deg NUC2	SIGF2	91.000	ppm	
P 14	500.0	us	sh.pul.invers.	NOISF1	81.600	ppm	
P 24	2000.0	us	sh.pul.refoc.	NOISF2	66.700	ppm	
P 28	0.1	us	trimpul.	F1P	0.000	ppm	
PLW 1	6.6	W	Pow@P90(Specs)	F2P	0.000	ppm	
PLW 2	26.8	W	Pow@P90(Specs)				
PLW 12	0.3	W	Pow@CPD(Specs)				
PCPD 2	102.5	us	90deg CPD NUC2				
CPDPRG2	bi_p5m4sp_4sp.2		cpd seq.				
SPNAM3	Crp60,0.5,20.1						
SPOAL 3	0.500		0.5				
SPW 3	0.000	W	default				
SPNAM7	Crp60comp.4						
SPOAL 7	0.500		0.5				
SPW 7	0.000	W	default				
GPNAM1	SINE.100						
GPNAM2	SINE.100						
GPZ 1	80.000	%					
GPZ 2	20.100	%					
P 16	0.000	us	grad.pulse				
TE	298.000	K	default				

## Experiment Description

2D 1H-13C HSQC is acquired phase-sensitive using shaped pulses for all 180 degree 13C pulses.


For decoupling adiabatic sequence bi\_p5m4sp\_4sp.2 is used.

Evaluation is achieved by determining the 2-dimensional signal-to-noise ratio according to the parameters section. The evaluation routine is part of the NMRPT package.

Contour levels for 2D spectrum are obtained by command levcalc. Levcalc sets the lowest contour level to LEV0\*S\_DEV, where S\_DEV (standard deviation) is a processing status parameter.

## 5.2.81 1H integral sensitivity (NPT\_1H\_inno)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10121, Z10033, Z10270, Z10718, Z142221  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

1H spectrum of ethylbenzene processed without line broadening. Top right shows the methylene (CH<sub>2</sub>) group used in evaluation for signal-to-noise and integral-to-noise ratio.

### Control Option for Acquisition (L23)

1 default



## Parameters


F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	131072			
PULPROG	zg				WDW	1			
NS	1				LB	0.000	Hz		
DS	0				PC	1.000			
RG	101.000		no optim.		F1P	8.520	ppm		
O1P	4.000	ppm			F2P	0.480	ppm		
SW	20.485	ppm			CY	100.000	cm		
TD	264292								
AQ	16.122	s	field dependent						
FIDRES	0.062	Hz	field dependent						
D 1	113.574	s							
P 1	14.0	us	90deg Pulse						
PLW 1	6.6	W	Pow@90deg(Specs)						
TE	298.000	K	default						

## Experiment Description

The experiment determines SINO and INNO (Integral-to-noise ratio, I/N). The determination of I/N allows to obtain a measure for the sensitivity of the probe independent of the actual shim.

## 5.2.82 Linearity test with constant flip angle (NPT\_1H\_linearityConstFlipAngle)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



Exp #	Pulse Length [us]	Power	Intensity [%]
1	14.0	7.24 W	99.97
2	28.0	1.81 W	99.85
3	56.0	452 mW	99.70
4	112.0	113 mW	99.53
5	224.0	28.3 mW	99.88
6	448.0	7.07 mW	99.94
7	896.0	1.77 mW	100.00
8	1792.0	442 uW	99.46

### Example Printout

Plot of the water peak intensities corresponding to increasing pulse lengths while keeping a constant flip angle of 90 degrees.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
PULPROG	syszg4				WDW	1			
NS	1				LB	1.000	Hz		
DS	0				PC	1.000			
RG	101.000		no optim.		F1P	5.250	ppm		
O1P	4.688	ppm			F2P	4.250	ppm		
SW	16.442	ppm			CY	15.000	cm		
TD	8192								
AQ	0.623	s	field dependent						
FIDRES	1.606	Hz	field dependent						
D 1	0.200	s							
P 1	14.0	us	90deg NUC1						
PLW 1	6.6	W	Pow@90deg(Specs) NUC1						
TE	298.000	K	default						

## Experiment Description

The purpose of this test is the assessment of the spectrometer response to a series of pulses of increasing length and constant flip angle of 90 degrees.


The experiment starts with the pulse length and power level from prosol. The offset O1 is first optimized (PROCNOs 10 and 11).

Afterwards, a series of experiments where the pulse length is successively doubled and the power level divided by four is acquired, processed and stored in PROCNO 999.

The quality criterion of the test is the minimum peak intensity relative to the maximum peak intensity (normalized to 100 percent) and should be as high as possible.

## 5.2.83 Linearity test with decreasing power (NPT\_1H\_linearityDecreasingPower)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top: Plot of the water peak intensities as a function of the logarithmically decreasing power level (dBW) for a constant pulse length.

Top detail: 50% width zoom plot of the lower intensities to the right.

Bottom: Logarithmic plot of the upper plot along with a linear fit.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU Parameters				F1 PROC Parameters			
NUC1	1H			SI	32768		
PULPROG	syszg4			WDW	1		
NS	1			LB	1.000	Hz	
DS	0			PC	1.000		
RG	101.000		no optim.	F1P	5.250	ppm	
O1P	4.688	ppm		F2P	4.250	ppm	
SW	16.442	ppm		CY	15.000	cm	
TD	8192						
AQ	0.623	s	field dependent				
FIDRES	1.606	Hz	field dependent				
D 1	0.200	s					
P 1	3.1	us	20deg Pulse				
PLW 1		W	adjusted power				
TE	298.000	K	default				

## Experiment Description

Purpose of this test is the assessment of the spectrometer response to a series of pulses of constant duration and logarithmically decreasing power.

The experiment starts with the power level from prosol and a pulse length corresponding to a 20 degrees pulse. The offset O1 is first optimized (PROCNOs 10 and 11). Second, a series of experiments with decreasing power level (1dB-steps) is acquired, processed and stored in PROCNO 999.

Simfit is then used to obtain a linear fit of  $\log_{10}(\text{peak amplitude})$  as a function of the power level in dBW.

The quality criterion of the test is the standard deviation of the linear fit which should be as small as possible.


## 5.2.84 1H lineshape without sample rotation (NPT\_1H\_lineshape\_nrot)

**Test Sample:** 0.3%, 1.0% or 3.0% Chloroform in Acetone-D6  
Z10230, Z10248, Z10701, Z100926, Z10250, Z10702, Z10031, Z10030, Z10029,  
Z10249, Z10275, Z10272, Z10717

**Solvent:** Acetone

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

Proton line shape spectrum with offset calibration to O1P as overview spectrum. Printing is from +125.0 Hz to -125.0 Hz. The expansion plot to the left shows chloroform signal with higher resolution (+7.0 Hz, -7.0 Hz).

### Control Option for Acquisition (L23)

- 1 default
- 2 write default shimfile, in case of successful evaluation

## Parameters

<p><b>F1 ACQU</b></p> <p>NUC1 1H          PULPROG zg30          NS 1          DS 0          RG 101.000          O1P 7.700 ppm          SWH 1000.000 Hz          TD 32768          AQ 16.384 s          FIDRES 0.061 Hz          D 1 9.116 s          P 1 14.0 us          PLW 1 6.6 W          TE 298.000 K</p> <p style="text-align: right;">Parameters</p> <p style="text-align: right;">no optim.</p> <p style="text-align: right;">AQ+D1=const          90deg Pulse          Pow@90deg(Specs)          default</p>	<p><b>F1 PROC</b></p> <p>SI 16384          WDW 0          LB 0.000 Hz          PC 1.000          F1P 8.640 ppm          F2P 7.440 ppm          CY 1000.000 cm</p> <p><b>NMRPT</b></p> <p>CNST 50 0.200</p> <p style="text-align: right;">Parameters          Scaling factor for CY</p>
--	--

## Experiment Description

Before acquiring the final spectrum, the exact signal position is determined with NS=1. The carrier position is afterwards set to the position optimized O1 = peak frequency [Hz] - (SWH/4)

This procedure allows the application of a strip transformation with the parameters STSR=0, STSI=SI/2 resulting in the signal position on-resonance and in most cases resulting in an optimal automatic phase correction.

Setting L23=2, it is possible to store the standard shimfile provided the evaluation of the experiment is successful. This event takes place during acquisition only. During regular processing of the data no shimfile is stored.


## 5.2.85 1H lineshape with sample rotation and NS = 4 (NPT\_1H\_lineshape\_wrot)

**Test Sample:** 0.3%, 1.0% or 3.0% Chloroform in Acetone-D6  
Z10230, Z10248, Z10701, Z100926, Z10250, Z10702, Z10031, Z10030, Z10029,  
Z10249, Z10275, Z10272, Z10717

**Solvent:** Acetone

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation according to RO



### Example Printout

Proton line shape spectrum with offset calibration to O1P as overview spectrum. Printing is from +125.0 Hz to -125.0 Hz. The expansion plot to the left shows chloroform signal with higher resolution (+7.0 Hz, -7.0 Hz).

### Control Option for Acquisition (L23)

1 default



## Parameters

<p><b>F1 ACQU</b></p> <p>NUC1 1H          PULPROG zg30          NS 4          DS 0          RG 101.000          O1P 7.700 ppm          SWH 1000.000 Hz          TD 32768          AQ 16.384 s          FIDRES 0.061 Hz          D 1 9.116 s          P 1 14.0 us          PLW 1 6.6 W          TE 298.000 K</p> <p style="text-align: right;">Parameters</p> <p style="text-align: right;">no optim.</p> <p style="text-align: right;">AQ+D1=const          90deg Pulse          Pow@90deg(Specs)          default</p>	<p><b>F1 PROC</b></p> <p>SI 16384          WDW 0          LB 0.000 Hz          PC 1.000          F1P 8.640 ppm          F2P 7.440 ppm          CY 1000.000 cm</p> <p><b>NMRPT</b></p> <p>CNST 50 0.200</p> <p style="text-align: right;">Parameters</p> <p style="text-align: right;">Scaling factor for CY</p>
--	---

## Experiment Description

Before acquiring the final spectrum, the exact signal position is determined with NS=1. The carrier position is afterwards set to the position optimized O1 = peak frequency [Hz] - (SWH/4)

This procedure allows the application of a strip transformation with the parameters STSR=0, STSI=SI/2 resulting in the signal position on-resonance and in most cases resulting in an optimal automatic phase correction.


## 5.2.86 1H lineshape with sample rotation and NS = 1 (NPT\_1H\_lineshape\_wrot\_ns1)

**Test Sample:** 0.3%, 1.0% or 3.0% Chloroform in Acetone-D6  
Z10230, Z10248, Z10701, Z100926, Z10031, Z10030, Z10029, Z10249, Z10275,  
Z10717

**Solvent:** Acetone

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation according to RO



### Example Printout

Proton line shape spectrum with offset calibration to O1P as overview spectrum. Printing is from +125.0 Hz to -125.0 Hz. The expansion plot to the left shows chloroform signal with higher resolution (+7.0 Hz, -7.0 Hz).

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	16384			
PULPROG	zg30				WDW	0			
NS	1				LB	0.000	Hz		
DS	0				PC	1.000			
RG	101.000		no optim.		F1P	8.640	ppm		
O1P	7.700	ppm			F2P	7.440	ppm		
SWH	1000.000	Hz			CY	1000.000	cm		
TD	32768				<b>NMRPT</b>				Parameters
AQ	16.384	s			CNST 50	0.200			Scaling factor for CY
FIDRES	0.061	Hz							
D 1	9.116	s	AQ+D1=const						
P 1	14.0	us	90deg Pulse						
PLW 1	6.6	W	Pow@90deg(Specs)						
TE	298.000	K	default						

## Experiment Description

Before acquiring the final spectrum, the exact signal position is determined with NS=1. The carrier position is afterwards set to the position optimized O1 = peak frequency [Hz] - (SWH/4)

This procedure allows the application of a strip transformation with the parameters STSR=0, STSI=SI/2 resulting in the signal position on-resonance and in most cases resulting in an optimal automatic phase correction.

In order to get better statistics for spinning sideband values, the number of scans is reduced to 1. Evaluation is the same as for the standard experiment.


## 5.2.87 1H lineshape stability test (NPT\_1H\_lineshapeStability)

**Test Sample:** 0.3%, 1.0% or 3.0% Chloroform in Acetone-D6  
Z10230, Z10248, Z10701, Z100926, Z10250, Z10702, Z10031, Z10030, Z10029,  
Z10249, Z10275, Z10272, Z10717

**Solvent:** Acetone

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

The 2D plot to the left shows all measured line shape experiments as pseudo 2D and their F1 projection. The plot to the right shows the experiment with the highest (black) and lowest (blue) resolution.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	1H			SI	262144		
PULPROG	zg2d			LB	0.000	Hz	
NS	1			F1P	5.201	ppm	
DS	0			F2P	4.201	ppm	
RG	101.000		no optim.	CY	11.000	cm	
O1P	8.000	ppm					
SWH	1315.789	Hz					
TD	65536						
AQ	24.904	s					
FIDRES	0.040	Hz					
D 20	600.000	s	repetition rate				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
TE	298.000	K	default				

## Experiment Description

In the line shape stability test a certain number of 1D spectra will be acquired on a CHCl<sub>3</sub> sample. The repetition rate, given by delay D20, must be high enough (10 min) to guarantee complete relaxation.

## 5.2.88 2D NOESY (NPT\_1H\_noesyphpr)

**Test Sample:** 2 mM Lysozyme in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10241  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Phase sensitive NOESY with presaturation during mixing and relaxation delay. 1D watersuppression experiment is shown at the left and at the top of the 2D.

### Control Option for Acquisition (L23)

- 1 default, with O1 optimization
- 2 no O1 optimization, the optimization of O1 is enforced, if O1 was not determined during a previous measurement.

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	1H			SI	2048		
FnMODE	5			WDW	4		
PARMODE	1		Data Dimension	SSB	4.000		
PULPROG	noesyphpr			PH_mod	1		pk
NS	8			F1P	24.732	ppm	
DS	16			F2P	-15.316	ppm	
RG	0.250		optim. by RGA	LEV0	4.250		
TD0	1			TOPLEV	50.000	%	
SW	20.485	ppm		NLEV	20		
TD	4096			<b>F1 ACQU</b>			Parameters F1
AQ	0.250	s	field dependent	NUC1	1H		
FIDRES	4.002	Hz	field dependent	FnMODE	5		
D 1	2.000	s		O1P	4.708	ppm	
D 8	0.150	s	mixing time	SW	20.485	ppm	
P 1	14.0	us	90deg Pulse	TD	512		
PLW 1	6.6	W	Pow@90deg(Specs)	<b>F1 PROC</b>			Parameters F1
TE	298.000	K	default	SI	2048		
DSPFIRM	4		rectangle	WDW	4		
DIGMOD	3		baseopt	SSB	4.000		
DE	40.000	us	set after getprosol	PH_mod	1		pk
				PHC0	90.000	deg	90deg (default)
				PHC1	-180.000	deg	180deg (default)
				F1P	0.000	ppm	
				F2P	0.000	ppm	

## Experiment Description

Phase sensitive NOESY with presaturation.

Presaturation requires the exact determination of the irradiation position (O1P). The determination is executed in a derived data set using the parameter set 'NPT\_1H\_watersuppression\_recflow'.

Using the control option for acquisition L23=1, O1 is determined by an intensity profile(second point of FID) over a range of 16 Hz in 0.4 Hz steps around submitted O1.

Option L23=2 will skip the procedure just outlined above.


RG can be set in the Preparation Panel. If RG <= 1 RGA will be executed after O1 determination in the derived data set.

Processing is achieved using the phase correction values from the preparation experiment.

Contour levels for 2D spectrum are obtained by command levcalc. Levcalc sets the lowest contour level to LEV0\*S\_DEV, where S\_DEV (standard deviation) is a processing status parameter.

## 5.2.89 P90 1H pulse calibration 0.5M NaCl (NPT\_1H\_p90\_05M\_NaCl\_1h)

**Test Sample:** 0.5 M Sodium Chloride (NaCl) in D2O  
 Z10288, Z101716, Z101712, Z100376, Z100372, Z10730  
**Solvent:** D2O\_salt  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



```

Report of All program poptau
F12=4.772ppm F22=4.622ppm
Linear optimization of P1 in 6 steps,
PROCNO=939
starting at 14.00000000 endval= 42.00000000
OPTIMON= ZEBO
VARMOD= L10
Experiment  P1      Maximum point  Minimum point  Integral
1  14.000000000000  372690885     -506292        587881.807988
2  19.000000000000  225120794     -148768        3926851.889504
3  25.000000000000  84499496      -75113         186328.728111
4  30.000000000000  8108898      -48188         43048.215054
5  36.000000000000  4859075      -56832381     -379861.417819
6  42.000000000000  3553925     -115605191    -1377969.499232
poptau for P1 finished.
ZEBO at experiment 4.124862; P1 = 31.499229
  
```

### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg.

### Control Option for Acquisition (L23)

1 default  
 100 Same as xx but skip automatic phase correction and apply manually set values.  
 +XX  
 1000 Same as xxx but skip automatic O1P determination  
 +XXX



## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	1H		Data Dimension	SI	2048		
PARMODE	0			WDW	3		
PULPROG	zg			LB	0.750	Hz	
NS	1			SSB	2.000		
DS	0		optim. by RGA	PH_mod	1		pk
RG	101.000			ME_mod	2		LPfc
O1P	4.700	ppm		NCOEF	20		
SWH	230.766	Hz		ABSF1	1000.000	ppm	
TD	300			ABSF2	-1000.000	ppm	
AQ	0.650	s		F1P	5.496	ppm	
FIDRES	1.538	Hz		F2P	5.096	ppm	
D 1	5.000	s	90deg Pulse	CY	11.000	cm	
P 1	14.0	us	Pow@90deg(Specs)				
PLW 1	6.6	W	default				
TE	298.000	K					

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL generic table. Prior to the execution of P90 determination for a given solvent, the standard p90 experiment on the urea sample must be executed.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. The information about O1, RG, PHC0, PHC1 is stored to a separate file to be used during the a possible repetition of the determination.

Before the acquisition, the proton 90 degrees pulse is calibrated using POPT, which result is stored in the corresponding derived dataset. The calibrated pulse, which is not set in prosol, is written in the acquisition title along with the starting pulse length and power from prosol.

The determination of the pulse is executed once in six steps. For the determined pulse at the used power the PROSOL table for the given solvent is updated. Results are stored under PROCNO 999.


## 5.2.90 P90 1H pulse calibration acetone (NPT\_1H\_p90\_acetone\_1h)

**Test Sample:** 0.3%, 1.0% or 3.0% Chloroform in Acetone-D6  
 Z10230, Z100926, Z10031, Z10030, Z10029, Z10249, Z10275, Z10717

**Solvent:** Acetone

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation according to RO



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg.

### Control Option for Acquisition (L23)

- 1 default
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 2048	
PARMODE 0	Data Dimension	WDW 3	
PULPROG zg		LB 0.750 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 101.000	optim. by RGA	ME_mod 2	LPfc
O1P 8.020 ppm		NCOEF 20	
SWH 230.766 Hz		ABSF1 1000.000 ppm	
TD 300		ABSF2 -1000.000 ppm	
AQ 0.650 s		F1P 5.496 ppm	
FIDRES 1.538 Hz		F2P 5.096 ppm	
D 1 255.000 s		CY 11.000 cm	
P 1 14.0 us	90deg Pulse		
PLW 1 6.6 W	Pow@90deg(Specs)		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL generic table. Prior to the execution of P90 determination for a given solvent, the standard p90 experiment on the urea sample must be executed.


As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. The information about O1, RG, PHC0, PHC1 is stored to a separate file to be used during the a possible repetition of the determination.

The determination of the pulse is executed once in six steps. For the determined pulse at the used power the PROSOL table for the given solvent is updated. Results are stored under PROCNO 999.

## 5.2.91 P90 1H pulse calibration (NPT\_1H\_p90det\_astm\_1h)

**Test Sample:** 40% Dioxane in Benzene-D6 (ASTM)  
 Z10163, Z100929, Z10164, Z10035, Z10255, Z10242, Z10724  
**Solvent:** C6D6  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments around 360 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 1024	
PARMODE 0	Data Dimension	LB 2.000	Hz
PULPROG zg		F1P 3.625	ppm
NS 1		F2P 3.125	ppm
DS 0		CY 5.500	cm
RG 0.250	optim. by RGA		
SWH 1250.000	Hz		
TD 1048			
AQ 0.419	s		
FIDRES 2.385	Hz		
O1P 3.375	ppm		
P 1 14.0	us		
PLW 1 6.6	W		
DIGMOD 3	90deg Pulse		
DSPFIRM 4	Pow@90deg(Specs)		
TE 298.000	baseopt		
	rectangle		
	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skiped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).


## 5.2.92 P90 1H pulse calibration (NPT\_1H\_p90determinationf1\_1h)

**Test Sample:** (a) 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D6  
 (b) 40% Dioxane in Benzene-D6 (ASTM)  
 Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721, Z142223

**Solvent:** (a) DMSO  
 (b) C6D6

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 2048	
PARMODE 0	Data Dimension	WDW 3	(Urea Sample)
PULPROG zg		WDW 1	(Dioxane Sample)
NS 1		LB 0.750 Hz	(Urea Sample)
DS 0		LB 10.0 Hz	(Dioxane Sample)
RG 0.250	optim. by RGA	SSB 2.000	
O1P 5.500 ppm	(Urea Sample)	PH_mod 1	pk
O1P 3.38 ppm	(Dioxane Sample)	ME_mod 2	LPfc
SWH 230.766 Hz	(Urea Sample)	NCOEF 20	
SWH 1500.0 Hz	(Dioxane Sample)	ABSF1 1000.000 ppm	
TD 300	(Urea Sample)	ABSF2 -1000.000 ppm	
TD 1024	(Dioxane Sample)	F1P 5.496 ppm	
AQ 0.650 s	(Urea Sample)	F2P 5.096 ppm	
AQ 0.341 s	(Dioxane Sample)	CY 11.000 cm	
FIDRES 1.538 Hz	(Urea Sample)		
FIDRES 2.93 Hz	(Dioxane Sample)		
D1 1.225 s	AQ+D1=const (Urea)		
D1 30.0 s	(Dioxane Sample)		
P1 14.0 us	90deg NUC1		
PLW1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22


If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.93 Indirect P90 13C pulse calibration (NPT\_1H\_p90determinationf2\_13c)

**Test Sample:** 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub> Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721, Z142223  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the decoupler pulse. On the left side the result of a CONVTO1D routine shows six experiments from 0 to 180 deg (PROCNO 100). On the right side the series goes from 180 to 360 deg (PROCNO 200).

### Control Option for Acquisition (L23)

- 1 default
  - 2 skip SINO check on PROCNO 11
  - 11 ignore specifications (optimize power for pulse length from prosol)
  - 12 ignore specifications and skip SINO check on PROCNO 11
  - 21 ignore specifications and optimize pulse length for power from prosol
  - 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
  - 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX  
 1000Same as xxx but skip automatic O1P determination  
 +XXX



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 2048	
PARMODE 0	Data Dimension	WDW 1	
PULPROG decp90		LB 0.500 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 0.250	optim. by RGA	ME_mod 2	LPfc
O1P 3.012 ppm		NCOEF 20	
SWH 230.766 Hz		ABSF1 1000.000 ppm	
TD 1000		ABSF2 -1000.000 ppm	
AQ 2.167 s		F1P 3.150 ppm	
FIDRES 0.462 Hz		F2P 2.850 ppm	
D 1 1.710 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
P 3 9.0 us	90deg NUC1		
PLW 2 42.0 W	Pow@90deg(Specs) NUC2		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22


If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 270 and 90 degree pulse lengths and memorised for further use in B1 homogeneity experiments.

## 5.2.94 Indirect P90 15N pulse calibration (NPT\_1H\_p90determinationf2\_15n)

**Test Sample:** 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub> Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721, Z142223  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the decoupler pulse. On the left side the result of a CONVTO1D routine shows six experiments from 0 to 180 deg (PROCNO 100). On the right side the series goes from 180 to 360 deg (PROCNO 200).

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 2048	
PARMODE 0	Data Dimension	WDW 3	
PULPROG decp90		LB 0.750 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 0.250	optim. by RGA	ME_mod 2	LPfc
O1P 5.500 ppm		NCOEF 20	
SWH 230.766 Hz		ABSF1 1000.000 ppm	
TD 200		ABSF2 -1000.000 ppm	
AQ 0.433 s		F1P 5.667 ppm	
FIDRES 2.308 Hz		F2P 5.367 ppm	
D 1 0.433 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
P 3 14.0 us	90deg NUC1		
PLW 2 86.0 W	Pow@90deg(Specs) NUC2		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 270 and 90 degree pulse lengths and memorised for further use in B1 homogeneity experiments.

## 5.2.95 Phase propagation test (NPT\_1H\_phase\_propagation)


---

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727

**Solvent:** D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

Series of spectra showing the water peak where, while maintaining the phase of the first pulse of the receiver to 0 degree, the phase of the second pulse is varied from 0 degree to 360 degrees in 10 degree steps.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU Parameters				F1 PROC Parameters			
NUC1	1H			SI	32768		
PULPROG	sysphas2f1			WDW	1		
NS	1			LB	1.000	Hz	
DS	4			PC	1.000		
RG	101.000		optim. by RGA	F1P	5.250	ppm	
O1P	4.698	ppm		F2P	4.250	ppm	
SW	16.442	ppm		CY	15.000	cm	
TD	8192						
AQ	0.623	s	field dependent				
FIDRES	1.606	Hz	field dependent				
D 1	0.200	s					
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
TE	298.000	K	default				


## Experiment Description

This test consists in a series of pairs of 90 degree pulses of constant power and duration recorded in a two-dimensional fashion. The phase of the second pulse is varied from 0 to 360 degrees in 10 degree steps, while maintaining the phases of the first pulse and of the receiver set to zero degree. The 2D-spectrum is stored in procno 1 and the result as a series of 1D spectra in procno 999.

## 5.2.96 Phase shifting test (NPT\_1H\_phase\_shifting)

---

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Series of spectra showing the water peak where the phase of the pulse is varied from 0 degree to 360 degrees in 10 degree steps, while maintaining the phase of the receiver to 0 degree.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
PULPROG	sysphasf1				WDW	1			
NS	1				LB	1.000	Hz		
DS	4				PC	1.000			
RG	101.000		optim. by RGA		F1P	5.250	ppm		
O1P	4.698	ppm			F2P	4.250	ppm		
SW	16.442	ppm			CY	15.000	cm		
TD	8192								
AQ	0.623	s	field dependent						
FIDRES	1.606	Hz	field dependent						
D 1	0.200	s							
P 1	14.0	us	90deg Pulse						
PLW 1	6.6	W	Pow@90deg(Specs)						
TE	298.000	K	default						

## Experiment Description

This test consists in a series of 90 degree pulses of constant power and duration recorded in a two-dimensional fashion. The phase of the pulse is varied from 0 to 360 degrees in 10 degree steps, while maintaining the phase of the receiver set to zero degree. The 2D-spectrum is stored in procno 1 and the result as a series of 1D spectra in procno 999.

## 5.2.97 Phase cycle cancelation (NPT\_1H\_phaseCycleCancelation)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top: Series of twelve spectra showing the residual water peak in phase-sensitive mode with NS=1,1,1,1,2,2,2,2,4,4,4 and 4.

Bottom: Same series as the top but in magnitude mode.

### Control Option for Acquisition (L23)

1 default



## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
PULPROG	syscancel				WDW	1			
NS				varied parameter	LB	1.000	Hz		
DS	0				PC	1.000			
RG	101.000			no optim.	F1P	5.250	ppm		
O1P	4.697	ppm			F2P	4.250	ppm		
SW	16.442	ppm			CY	15.000	cm		
TD	8192								
AQ	0.623	s		field dependent					
FIDRES	1.606	Hz		field dependent					
D 1	0.200	s							
P 1	14.0	us		90deg Pulse					
PLW 1	6.6	W		Pow@90deg(Specs)					
TE	298.000	K		default					


## Experiment Description

The aim of this test is the assessment of the signal cancelation by a two-steps and a four-steps phase cycle. The test consists of a 90 degree pulse which is phase-cycled, while maintaining the receiver phase set to zero.

The test comprises twelve experiments. The first four are recorded with NS=1, i.e. without phase cycling suppression. The average signal intensity of these four experiments is set to 100 percent. The next four experiments are acquired with NS=2, meaning a two-steps phase cycle (0 and 180 degrees). Finally, the last four experiments are recorded with NS=4 which corresponds to a four-steps phase-cycling (0, 180, 90 and 270 degrees).

## 5.2.98 Phase cycle cancelation after gradient pulse (NPT\_1H\_phaseCycleCancelationGrad)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top: Series of twelve spectra showing the residual water peak in phase-sensitive mode with NS=1,1,1,1,2,2,2,2,4,4,4 and 4.

Bottom: Same series as the top but in magnitude mode.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
PULPROG	sysgrcan				WDW	1			
NS				varied parameter	LB	1.000	Hz		
DS	0				PC	1.000			
RG	101.000			no optim.	F1P	5.250	ppm		
O1P	4.697	ppm			F2P	4.250	ppm		
SW	16.442	ppm			CY	15.000	cm		
TD	8192								
AQ	0.623	s		field dependent					
FIDRES	1.606	Hz		field dependent					
D 1	0.200	s							
GPNAM1	RECT.1								
GPZ 1	14.286	%							
P 1	14.0	us		90deg Pulse					
PLW 1	6.6	W		Pow@90deg(Specs)					
TE	298.000	K		default					

## Experiment Description

The aim of this test is the assessment of the signal cancellation by a two-steps and a four-steps phase cycle following a gradient pulse. The test consists of a Z-gradient pulse followed by a delay D16=100 us and finally a 90 degree pulse which is phase-cycled, while maintaining the receiver phase set to zero. The test comprises twelve experiments. The first four are recorded with NS=1, i.e. without phase cycling suppression. The average signal intensity of these four experiments is set to 100 percent. The next four experiments are acquired with NS=2, meaning a two-steps phase cycle (0 and 180 degrees). Finally, the last four experiments are recorded with NS=4 which corresponds to a four-steps phase-cycling (0, 180, 90 and 270 degrees).

## 5.2.99 1H quantification reference (NPT\_1H\_quant\_ref)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10121, Z10033, Z10270, Z10718, Z142221  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

1H overview spectrum of ethylbenzene.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	65536			
PULPROG	zg30				WDW	1			
NS	16				LB	0.300	Hz		
DS	2				PC	1.000			
RG	32.000		no optim.		F1P	0.000	ppm		
O1P	6.175	ppm			F2P	0.000	ppm		
SW	20.485	ppm			CY	11.000	cm		
TD	65536								
AQ	3.998	s	field dependent						
FIDRES	0.250	Hz	field dependent						
D 1	30.000	s							
P 1	14.0	us	90deg Pulse						
PLW 1	6.6	W	Pow@90deg(Specs)						
TE	298.000	K	default						


## Experiment Description

Quantification experiment using the ethylbenzen sample. This experiment is used as default reference by the eretic procedure. The expansion plot shows the quantified signal with higher resolution  
 Processing is done by the AU program cmc\_proc1h\_usup. For this reason the NMRPT processing options are ignored.

## 5.2.100 1H low flipangle single scan experiment (NPT\_1H\_rd)

---

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10719  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

1H spectrum of H<sub>2</sub>O signal showing radiation damping effect (line broadening). Top left shows H<sub>2</sub>O signal expanded.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	1H			SI	32768		
PULPROG	zg0			WDW	1		
NS	1			LB	0.000	Hz	
DS	0			F1P	10.806	ppm	
RG	2.000		no optim.	F2P	-1.227	ppm	
O1P	4.700	ppm		CY	11.000	cm	
SW	20.485	ppm					
TD	32768						
AQ	1.999	s	field dependent				
FIDRES	0.500	Hz	field dependent				
D 1	5.000	s					
P 0	1.0	us	1 us=1 deg				
PLW 1	0.005	W	Pow@1 deg=1 us NUC1				
TE	298.000	K	default				

## Experiment Description

Radiation damping experiment is executed with small flip angle as single scan experiment. NMRPT is calculating PLW1 according to the equation  $1\text{deg}=1\text{us}$ . Processing is achieved with Fourier transformation without line broadening (LB=0). Evaluation consists of line width determination at 50% of signal height using the TopSpin AU program 'hwcal'.

## 5.2.101 1H selective excitation (NPT\_1H\_selex)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10121, Z10033, Z10270, Z10718  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Bottom: 1H overview spectrum of ethylbenzene using hard pulse. Center: 1H signal of methylene signal using selective excitation. Top: Selectively excited 1H signal expanded including the noise region for evaluation.

### Control Option for Acquisition (L23)

1 default



## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	1H			SI	65536		
PULPROG	selzg			LB	0.000	Hz	
NS	1			SIGF1	3.500	ppm	
DS	0			SIGF2	2.000	ppm	
RG	101.000		no optim.	NOISF1	0.000	ppm	
O1P	2.673	ppm		NOISF2	-9.000	ppm	
SW	24.992	ppm		F1P	8.000	ppm	
TD	65536			F2P	-9.200	ppm	
AQ	3.277	s	field dependent	CY	11.000	cm	
FIDRES	0.305	Hz	field dependent				
D 1	119.000	s					
P 11	20000.000	us	P90 selective				
SPW 1	0.000	W	Pow@P90 sel. NUC1				
SPNAM1	Gaus1.1000		shape				
SPOAL 1	0.500		shape				
SPOFFS 1	0.000		shape				
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
TE	298.000	K	default				

## Experiment Description


Selective excitation experiment is carried out to show the quality with respect to selectivity of the frequency generation and the noise excitation of the amplifier. The experiment consists of two parts:

- 1) Reference acquisition using hard pulse, stored in derived dataset (EXPNO=1).
- 2) Selectively excited methylene signal using the same spectral parameters as for 1), stored in the main dataset.

Processing is for both spectra the same. Evaluation consists of signal-to-noise determinations for both experiments and the formation of the ratio of the two. The noise range (9 ppm) is fixed thereby the comparison becomes possible.

## 5.2.102 1H sensitivity (NPT\_1H\_sensitivity)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10121, Z10033, Z10270, Z10718, Z142221  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Bottom: 1H overview spectrum of ethylbenzene.

Top right: Expanded region showing the methylene signal used for evaluation.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	16384			
PULPROG	zg				LB	1.000	Hz	1.0	
NS	1				SIGF1	3.000	ppm		
DS	0				SIGF2	2.000	ppm		
RG	101.000		no optim.		NOISF1	7.000	ppm		
O1P	4.000	ppm			NOISF2	2.800	ppm		
SW	10.159	ppm			F1P	8.520	ppm		
TD	32768				F2P	0.480	ppm		
AQ	4.030	s	field dependent		CY	11.000	cm		
FIDRES	0.248	Hz	field dependent		<b>NMRPT</b>				Parameters
D 1	113.574	s			CNST 57	1.000			Return Value Intensity
P 1	14.0	us	90deg Pulse		CNST 58	1.000			InnoSinoShim
PLW 1	6.6	W	Pow@90deg(Specs)						Return Value Integral
TE	298.000	K	default						InnoSinoShim

## Experiment Description

Proton sensitivity is measured using the ethylbenzene sample. Processing is using LB. The signal-to-noise is determined using the methylene signal (CH<sub>2</sub>) of the molecule. The signal is searched over the range from 3.0 to 2.0 ppm, while the best 2 ppm or 200 Hz noise region is determined over the range from 7.0 to 2.8 ppm.

## 5.2.103 Triple resonance (NPT\_1H\_sensitivity\_dec13c15n)

**Test Sample:** 100 mM Urea-15N ([<sup>15</sup>NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub>  
 Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

1H spectra of 100 mM 15N labelled urea and 100 mM 13C labelled methanol. Spectra at the top without decoupling, spectra at the bottom with simultaneous 13C and 15N decoupling. The extensions on the left show the urea signals, the extensions on the right th methanol signals.

### Control Option for Acquisition (L23)

1 default

## Parameters


<b>F1 ACQU</b>			Parameters	<b>CNST 10</b>	60.000		flip angle
NUC1	1H			<b>F1 PROC</b>			Parameters
NUC2	13C			SI	32768		
NUC3	15N			LB	0.750	Hz	
PULPROG	npt_zg0fbig			SIGF1	0.000	ppm	
NS	8			SIGF2	0.000	ppm	
DS	2			NOISF1	0.000	ppm	
RG	1.000		no optim.	NOISF2	-9.000	ppm	
O1P	2.670	ppm		F1P	6.500	ppm	
O2P	49.487	ppm		F2P	1.500	ppm	
O3P	75.986	ppm		CY	11.000	cm	
SW	24.992	ppm		<b>NMRPT</b>			Parameters
TD	32768						
AQ	1.638	s	field dependent				
FIDRES	0.610	Hz	field dependent				
D 1	30.000	s	>10*T1				
P 0	6.0	us	P 1 * CNST 10 / 90				
P 1	10.0	us	90deg NUC1				
PLW 1	15.0	W	Pow@P90(Specs)				
PLW 12	9.0	W	Pow@CPD(Specs) NUC2				
PLW 16	8.2	W	Pow@CPD(Specs) NUC3				
PCPD 2	50.0	us	90deg CPD NUC2				
CPDPRG2	garp		cpd seq.				
PCPD 3	200.0	us	90deg CPD NUC3				
CPDPRG3	garp		cpd seq.				
TE	298.000	K	default				

## Experiment Description

Simultaneous decoupling on two nuclei (13C, 15N) is executed with the 15N-urea/13C-methanol sample. Two spectra are acquired (coupled, decoupled) using derived dataset with EXNPO=1 for the coupled spectrum and the current dataset for the decoupled spectrum. Processing is executed with LB. Evaluation is comparing the signal-to-noise ratios (coupled/decoupled) for methylene group (13C, 3.6 to 2.8 ppm) and amino group (15N, 6.8 to 5.0 ppm) and the noise region from 0.0 to -9.0 ppm.

## 5.2.104 1H sensitivity with 19F GARP decoupling (NPT\_1H\_sensitivity\_dec19f)

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
Z10234, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Bottom: <sup>1</sup>H overview spectrum of trifluorotoluene with <sup>19</sup>F decoupling.  
Top left: Expanded region showing the signal region of the aromatic part of the molecule used for evaluation.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	1H			SI	32768		
NUC2	19F			SIGF1	8.000	ppm	
PULPROG	zgig			SIGF2	7.400	ppm	
NS	1			NOISF1	18.200	ppm	
DS	0			NOISF2	9.200	ppm	
RG	101.000		no optim.	F1P	19.000	ppm	
O1P	7.500	ppm		F2P	0.500	ppm	
O2P	-62.498	ppm		CY	11.000	cm	
SW	24.992	ppm					
TD	32768		32768 <500MHz, 65536 >=500 MHz				
AQ	1.638	s	field dependent				
FIDRES	0.610	Hz	field dependent				
D 1	52.500	s					
P 1	14.0	us	90deg NUC1				
PLW 1	6.5	W	Pow@P90(Specs)				
PLW 12	0.064	W	Pow@CPD(Specs)				
CPDPRG2	garp4		cpd seq.				
TE	298.000	K	default				

## Experiment Description

Proton sensitivity with <sup>19</sup>F decoupling is measured using the trifluorotoluene sample. Processing is using no LB. The signal-to-noise is determined using the highest signal of the aromatic region (~7.64 ppm). The noise region is fixed from 18.2 to 9.2 ppm.

## 5.2.105 $^1\text{H}$ sensitivity with HSQC selection and $^{13}\text{C}$ garp decoupling (NPT\_1H\_sensitivity\_hsqc13c)

**Test Sample:** 10% Ethylbenzene (EB) in Chloroform-D  
Z10153, Z10154, Z10034, Z10253, Z10292, Z10723  
**Solvent:**  $\text{CDCl}_3$   
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Bottom: 1D  $^1\text{H}$ - $^{13}\text{C}$  HSQC overview spectrum of ethylbenzene with  $^{13}\text{C}$  decoupling.  
Top left: Expanded region showing the signal region of the methylene group ( $\text{CH}_2$ ) used for evaluation.

### Control Option for Acquisition (L23)

1 default



## Parameters

F1 ACQU			Parameters	F1 PROC		
NUC1	1H			GPNAM1	SMSQ10.32	
NUC2	13C			GPNAM2	SMSQ10.32	
PULPROG	hsqcetgpsisp2.2			GPNAM3	SMSQ10.32	
NS	8			GPNAM4	SMSQ10.32	
DS	4			GPZ 1	80.000	%
RG	101.000		no optim.	GPZ 2	20.100	%
O1P	4.000	ppm		GPZ 3	11.000	%
O2P	79.987	ppm		GPZ 4	-5.000	%
SW	10.159	ppm		P 16	0.000	us
TD	4096			P 19	0.000	us
AQ	0.504	s	field dependent	TE	298.000	K
FIDRES	1.985	Hz	field dependent	<b>F1 PROC</b>		
D 1	35.196	s	AQ+D1=const	SI	16384	
D 16	0.001	s	gradrec del.	LB	1.000	Hz
D 24	0.001	s	1/8*J[CH]	SIGF1	3.000	ppm
P 1	12.9	us	90deg NUC1	SIGF2	2.000	ppm
P 3	11.0	us	90deg NUC2	NOISF1	7.000	ppm
P 14	500.0	us	180deg NUC2 Inversion	NOISF2	2.800	ppm
P 24	2000.0	us	180deg NUC2 Refocussing	F1P	8.520	ppm
P 28	0.1	us	trimpul.	F2P	0.480	ppm
PLW 1	7.0	W	Pow@P90(Specs)	CY	11.000	cm
PLW 2	27.3	W	Pow@P90(Specs)			
PLW 12	0.24	W	Pow@CPD(Specs)			
PCPD 2	102.5	us	90deg CPD NUC2			
CPDPRG2	garp4		cpd seq.			
CNST 2	135.000	Hz	J[CH]			
CNST 17	-0.500		-0.5 for SPNAM7			
SPNAM3	Gaus1.1000		Crp60,0.5,20.1			
SPOAL 3	0.500		phase align.			
SPOFFS 3	0.000		offset freq.			
SPW 3	0.000	W	default			
SPNAM7	Gaus1.1000		Crp60comp.4			
SPOAL 7	0.500		phase align.			
SPOFFS 7	0.000		offset freq.			
SPW 7	0.000	W	default			

## Experiment Description


Proton sensitivity with 13C decoupling is measured using the ethylbenzene sample. Method used: Phase sensitive HSQC with echo/antiecho TPPI gradient selection.

Before the acquisition, the proton 90 degrees pulse is calibrated using pulscal, which result is stored in the corresponding derived dataset. The pulse determination is skipped if experiment is measured with option 'Skip Getprosol'.

Processing is using LB. The signal-to-noise is determined using the methylene signal (CH2) of the molecule. The signal is searched over the range from 3.0 to 2.0 ppm, whereas the best 2 ppm region over the range from 7.0 to 2.8 ppm.

## 5.2.106 Simultaneous hard pulses on $^{13}\text{C}$ and $^{15}\text{N}$ (NPT\_1H\_simpul\_13c15n)

**Test Sample:** 100 mM Urea- $^{15}\text{N}$  ( $[\text{15NH}_2\text{]}_2\text{CO}$ ) and 100 mM Methanol- $^{13}\text{C}$  in Dimethyl Sulfoxide- $\text{D}_6$   
 Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Left: Pseudo-2D showing expanded region of the methyl group of methanol. Experiment consists of a modified decp90 method using 360 deg ( $^{13}\text{C}$ ) and 180 deg ( $^{15}\text{N}$ ) giving in-phase signal for the methyl group.

Right: 1D representation of the pseudo-2D, generated with CONVTO1D procedure.

### Control Option for Acquisition (L23)

1 default

## Parameters


<p><b>F2 ACQU</b></p> <p>NUC1 1H          NUC2 13C          NUC3 15N          PULPROG npt_simpul13c15n          NS 1          DS 8          RG 0.250          O1P 3.080 ppm          O2P 49.430 ppm          O3P 75.921 ppm          SWH 230.766 Hz          TD 1000          AQ 2.167 s          FIDRES 0.462 Hz          D 1 1.710 s          P 1 9.5 us          P 3 14.0 us          P 21 30.0 us          PLW 1 0.7 W          PLW 2 72.2 W          PLW 3 158.4 W          CNST 2 139.000 Hz          TE 298.000 K</p>	<p style="text-align: center;">Parameters F2</p> <p style="text-align: center;">optim. by RGA</p> <p>AQ+D1=const          90deg NUC1          90deg NUC2          90deg NUC3          Pow@P90(Specs)          Pow@P90(Specs)          Pow@P90(Specs)          J[XH]          default</p>
<p><b>F2 PROC</b></p> <p>SI 4096          LB 0.500 Hz          SIGF1 0.000 ppm          SIGF2 0.000 ppm          NOISF1 0.000 ppm          NOISF2 0.000 ppm          F1P 3.200 ppm          F2P 2.950 ppm          CY 11.000 cm</p>	<p style="text-align: center;">Parameters F2</p> <p style="text-align: center;">Parameters F1</p> <p style="text-align: center;">Parameters F1</p> <p><b>F1 ACQU</b>          NUC1 1H          TD 64</p> <p><b>F1 PROC</b>          SI 64</p>

## Experiment Description

Simultaneous application of pulses are common especially in multi-dimensional and multi-nuclear applications. The simultaneous pulse test checks the stability of the probe under these conditions by repetitively applying pulses concurrently on two decoupling channels. Processing is using LB. Evaluation consist of determination of intensity deviation based on peak picking intensities.

## 5.2.107 Simultaneous hard pulses on $^{15}\text{N}$ and $^{13}\text{C}$ (NPT\_1H\_simpul\_15n13c)

**Test Sample:** 100 mM Urea- $^{15}\text{N}$  ( $[\text{15NH}_2]\text{2CO}$ ) and 100 mM Methanol- $^{13}\text{C}$  in Dimethyl Sulfoxide- $\text{D}_6$   
 Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Left: Pseudo-2D showing expanded region of the amino group of urea. Experiment consists of a modified decp90 method using 360 deg ( $^{15}\text{N}$ ) and 180 deg ( $^{13}\text{C}$ ) giving in-phase signal for the amino group.  
 Right: 1D representation of the pseudo-2D, generated with CONVTO1D procedure.

### Control Option for Acquisition (L23)

1 default

## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H          NUC2 13C          NUC3 15N          PULPROG npt_simpul15n13c          NS 1          DS 8          RG 0.250          O1P 5.500 ppm          O2P 49.430 ppm          O3P 75.921 ppm          SWH 230.766 Hz          TD 200          AQ 0.433 s          FIDRES 2.308 Hz          D 1 0.433 s          P 1 9.5 us          P 3 14.0 us          P 21 30.0 us          PLW 1 0.7 W          PLW 2 72.2 W          PLW 3 158.4 W          CNST 2 88.500 Hz          TE 298.000 K</p>	<p style="text-align: center;">Parameters F2</p> <p style="text-align: center;">optim. by RGA</p> <p>AQ+D1=const          90deg NUC1          90deg NUC2          90deg NUC3          Pow@P90(Specs)          Pow@P90(Specs)          Pow@P90(Specs)          J[XH]          default</p>
<p><b>F2 PROC</b></p> <p>SI 4096          LB 0.500 Hz          SIGF1 0.000 ppm          SIGF2 0.000 ppm          NOISF1 0.000 ppm          NOISF2 0.000 ppm          F1P 5.670 ppm          F2P 5.370 ppm          CY 11.000 cm</p>	<p style="text-align: center;">Parameters F2</p> <p style="text-align: center;">Parameters F1</p> <p style="text-align: center;">Parameters F1</p>
<p><b>F1 ACQU</b></p> <p>NUC1 1H          TD 64</p>	<p><b>F1 PROC</b></p> <p>SI 64</p>


## Experiment Description

Simultaneous application of pulses are common especially in multi-dimensional and multi-nuclear applications. The simultaneous pulse test checks the stability of the probe under these conditions by repetitively applying pulses concurrently on two decoupling channels. Processing is using LB. Evaluation consist of determination of intensity deviation based on peak picking intensities.

## 5.2.108 1H temperature calibration with 99.8% MeOD (NPT\_1H\_tempcalib\_998meod)

---

**Test Sample:** Temperature Calibration 99.8% Methanol-D4  
Z10627, Z10628, Z10053, Z10734  
**Solvent:** MeOD  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Spectrum of 0.2% MeOH at standard probe temperature.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU Parameters				F1 PROC Parameters			
NUC1	1H			SI	16384		
PULPROG	zg30			LB	0.500	Hz	0.5
NS	1			SIGF1	0.000	ppm	
DS	0			SIGF2	0.000	ppm	
RG	0.250		optim. by RGA	NOISF1	0.000	ppm	
O1P	4.600	ppm		NOISF2	0.000	ppm	
SW	4.165	ppm		F1P	5.501	ppm	
TD	8192			F2P	2.499	ppm	
AQ	2.458	s	field dependent	CY	11.000	cm	
FIDRES	0.407	Hz	field dependent	<b>NMRPT</b>			
D 1	2.000	s		CNST 50	1.000	K	Parameters
P 1	14.0	us	90deg Pulse	CNST 51	1.000	K	min TE for calibr.
PLW 1	6.6	W	Pow@90deg(Specs)	CNST 52	1.000		max TE for calibr.
TE	298.000	K	default				no. points TE calibr.

## Experiment Description

Temperature calibration experiment expects a temperature unit. The procedure is calibrating at least two TE settings according to the entries in the parameter set (CNST 50 to 53).

The result of the experiment is the determination of the offset and slope of the correction curve ( $y=ax + b$ ). For a BSVT the correction will be set in the temperature unit if the experiment was not measured with option 'Skip FLOW'.


CNST 50: lowest TE used for calibration, default value of 1 means that the standard temperature will be used.

CNST 51: highest TE used for calibration, default value of 1 means that lowest TE + 10 K will be used.

CNST 52: number of points (TE settings) used for calibration. The default is 2.

## 5.2.109 Vibration Test using Doped Water Sample (NPT\_1H\_vibration\_doped\_water)

**Test Sample:** 0.1 mg/ml Gadolinium Chloride (GdCl<sub>3</sub>), 0.1% Methanol-13C, and 1% H<sub>2</sub>O in D<sub>2</sub>O  
Z10083, Z100933, Z10085, Z10046, Z10084, Z10082, Z10727, Z142231  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top: Expansion plot of the arithmetic Mean Projection of pseudo-2D data.

Other Plots: Calculated symmetrized standard deviation for the different frequency ranges.

### Control Option for Acquisition (L23)

1 default



## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H</p> <p>PARMODE 1</p> <p>PULPROG zg2d_random</p> <p>NS 1</p> <p>DS 0</p> <p>RG 0.250</p> <p>TD0 1</p> <p>SWH 10000.000 Hz</p> <p>TD 32768</p> <p>AQ 1.638 s</p> <p>FIDRES 0.610 Hz</p> <p>D 1 1.000 s</p> <p>P 1 14.0 us</p> <p>PLW 1 6.6 W</p> <p>TE 298.000 K</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>optim. by RGA</p> <p>90deg Pulse Pow@90deg(Specs) default</p>	<p><b>F2 PROC</b></p> <p>SI 32768</p> <p>WDW 1</p> <p>SSB 0.000</p> <p>PH_mod 1</p> <p>PHC0 0.000 deg</p> <p>PHC1 0.000 deg</p> <p>F1P 0.000 ppm</p> <p>F2P 0.000 ppm</p> <p><b>F1 ACQU</b></p> <p>NUC1 1H</p> <p>TD 256</p> <p><b>F1 PROC</b></p> <p>SI 256</p>	<p>Parameters F2</p> <p>pk</p> <p>Parameters F1</p> <p>Parameters F1</p>
---	---	--	--

## Experiment Description

The standard deviation is computed from the rows of the pseudo 2D spectrum. The first 8 largest peaks are listed for each frequency range in the tables on the right.


## 5.2.110 Vibration Test using Lineshape Sample (NPT\_1H\_vibration\_lineshape)

**Test Sample:** 0.3%, 1.0% or 3.0% Chloroform in Acetone-D6  
 Z10230, Z10248, Z10701, Z100926, Z10031, Z10030, Z10029, Z10249, Z10275,  
 Z10717, Z142220

**Solvent:** Acetone

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

Top Right: Expansion plot of the arithmetic Mean Projection of pseudo-2D data.  
 Other Plots: Calculated symmetrized standard deviation for the different frequency ranges.

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F2 ACQU</b>		Parameters F2	<b>F2 PROC</b>		Parameters F2
NUC1	1H		SI	65536	
PARMODE	1	Data Dimension	WDW	1	
PULPROG	zg2d_random		SSB	0.000	
NS	1		PH_mod	1	pk
DS	6		PHC0	0.000	deg
RG	0.250	optim. by RGA	PHC1	0.000	deg
TD0	1		F1P	0.000	ppm
SWH	1612.903	Hz	F2P	0.000	ppm
TD	65536		<b>F1 ACQU</b>		Parameters F1
AQ	20.316	s	NUC1	1H	
FIDRES	0.049	Hz	TD	128	
D 1	1.000	s	<b>F1 PROC</b>		Parameters F1
P 1	8.6	us	SI	128	
PLW 1	6.6	W			
TE	298.000	K			
		55deg Pulse Pow@90deg(Specs) default			

## Experiment Description

The standard deviation is computed from the rows of the pseudo 2D spectrum. The first 8 largest peaks are listed for each frequency range in the tables on the right.


## 5.2.111 Watersuppression NaCl with recommended gas flow (NPT\_1H\_watersupp\_NaCl\_recflow)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O with salt  
Z100384, Z100385, Z100386, Z100387, Z100388, Z100389, Z107150, Z107151,  
Z107152

**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O\_salt

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

Water suppression spectrum printed from 10.5 ppm to -1.0 ppm as overview spectrum. The expansion plot to the right shows the DSS (sodium 2,2-dimethyl-2-silapentane-5-sulphonate) signal, indicative for the quality of the spectral resolution. The quality of the water suppression is determined as linewidth of the residual water signal using the 10% and 50% intensity of the DSS signal.

The expansion plot to the left shows the signal of the anomeric proton (alpha-C1) of the sucrose (2-beta-D-fructofuranosyl-1-alpha-D-glucopyranosid). The anomeric proton intensity is used for the signal-to-noise calculation and the splitting of the signal is analysed to give a measure for spectral resolution.

### Control Option for Acquisition (L23)

- 1 with O1 optimization, with PULPROG=zgpr
- 20 O1 from parameter set, no optimization, with PULPROG=zgpr
- 23 O1 from parameter set, no optimization, with PULPROG=npt\_zggppr
- 25 O1 from parameter set, no optimization, with PULPROG=zgcppr
- 27 O1 from parameter set, no optimization, with PULPROG=zgcpppr

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
PULPROG	zgpr				WDW	1			
NS	8				LB	0.000	Hz		
DS	4				PC	0.100			
RG	0.250		optim. by NMRPT		F1P	10.806	ppm		
O1P	4.699	ppm			F2P	-1.227	ppm		
SW	12.132	ppm			CY	111.000	cm		
TD	10194		field dependent						
AQ	1.050	s							
FIDRES	0.952	Hz	field dependent						
D 1	5.000	s							
P 1	14.0	us	90deg						
PLW 1	7.5	W	Pow@90deg(Specs)						
PLW 9	0.00006	W	Pow@90deg(5000u)						
TE	298.000	K	default						

## Experiment Description

For the watersuppression experiment the carrier offset O1 needs to be very carefully adjusted. Using the control option for acquisition L23=1, O1 is determined by an intensity profile(second point of FID) over a range of 16 Hz in 0.4 Hz steps around submitted O1.

RG adjustment is executed by running experiments with DS=0, NS=4 and a reduced PLW1. The highest RG without signal overflow in these experiments will be used.

Before the acquisition, the proton 90 degrees pulse is calibrated using POPT, which result is stored in the corresponding derived dataset. The calibrated pulse, which is not set in prosol, is written in the acquisition title along with the starting pulse length and power from prosol.


The pulse determination is skiped if experiment is measured with option 'Skip Getprosol'.

Options L23=1, 20 and 30 are standard whereas Options L23=23, 25, 27, 33, 35, and 37 are non-standard and will not be considered as regular test from the 'NMRPT Control Structure'. Their existence is purely diagnostic.

Experiment will be set to irregular if option 'Skip Temperature' is selected.

## 5.2.112 Watersuppression with 270 l/h gas flow (NPT\_1H\_watersuppression\_270l)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
 Z10902, Z10246, Z100930, Z10268, Z10036, Z10267, Z10719, Z10720  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Water suppression spectrum printed from 10.5 ppm to -1.0 ppm as overview spectrum. The expansion plot to the right shows the DSS (sodium 2,2-dimethyl-2-silapentane-5-sulphonate) signal, indicative for the quality of the spectral resolution. The quality of the water suppression is determined as linewidth of the residual water signal using the 10% and 50% intensity of the DSS signal. The expansion plot to the left shows the signal of the anomeric proton (alpha-C1) of the sucrose (2-beta-D-fructofuranosyl-1-alpha-D-glucopyranosid). The anomeric proton intensity is used for the signal-to-noise calculation and the splitting of the signal is analysed to give a measure for spectral resolution.

### Control Option for Acquisition (L23)

- 1 with O1 optimization, with PULPROG=zgpr
- 20 O1 from parameter set, no optimization, with PULPROG=zgpr
- 23 O1 from parameter set, no optimization, with PULPROG=npt\_zggppr
- 25 O1 from parameter set, no optimization, with PULPROG=zgcppr
- 27 O1 from parameter set, no optimization, with PULPROG=zgcppppr
- 30 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgpr
- 33 O1 from previous watersuppression, no optimization[\*], with PULPROG=npt\_zggppr
- 35 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppr
- 37 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppppr
- [\*] The optimization of O1 is enforced, if O1 was not determined during a previous measurement.

## Parameters

F1 ACQU				F1 PROC			
NUC1	1H		Parameters	SI	32768		Parameters
PULPROG	zgpr			WDW	1		
NS	8			LB	0.000	Hz	
DS	4			PC	0.100		
RG	0.250		optim. by NMRPT	F1P	10.806	ppm	
O1P	4.699	ppm		F2P	-1.227	ppm	
SW	12.132	ppm		CY	111.000	cm	
TD	10194		field dependent				
AQ	1.050	s					
FIDRES	0.952	Hz	field dependent				
D 1	5.000	s					
P 1	14.0	us	90deg				
PLW 1	7.5	W	Pow@90deg(Specs)				
PLW 9	0.00006	W	Pow@90deg(5000u)				
TE	298.000	K	default				

## Experiment Description


For the watersuppression experiment the carrier offset O1 needs to be very carefully adjusted. Using the control option for acquisition L23=1, O1 is determined by an intensity profile(second point of FID) over a range of 16 Hz in 0.4 Hz steps around submitted O1.

RG adjustment is executed by running experiments with DS=0, NS=4 and a reduced PLW1. The highest RG without signal overflow in these experiments will be used.

Options L23=1 and 20 are standard whereas Options L23=23, 25, and 27 are non-standard and will not be considered as regular test from the 'NMRPT Control Structure'. Their existence is purely diagnostic. Experiment will be set to irregular if one or both of the options 'Skip Flow' or 'Skip Temperature' are selected.

## 5.2.113 Watersuppression with 400 l/h gas flow (NPT\_1H\_watersuppression\_400l)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
 Z10902, Z10246, Z100930, Z10036, Z10247, Z10267, Z10720, Z10719  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Water suppression spectrum printed from 10.5 ppm to -1.0 ppm as overview spectrum. The expansion plot to the right shows the DSS (sodium 2,2-dimethyl-2-silapentane-5-sulphonate) signal, indicative for the quality of the spectral resolution. The quality of the water suppression is determined as linewidth of the residual water signal using the 10% and 50% intensity of the DSS signal.

The expansion plot to the left shows the signal of the anomeric proton (alpha-C1) of the sucrose (2-beta-D-fructofuranosyl-1-alpha-D-glucopyranosid). The anomeric proton intensity is used for the signal-to-noise calculation and the splitting of the signal is analysed to give a measure for spectral resolution.

### Control Option for Acquisition (L23)

- 1 with O1 optimization, with PULPROG=zgpr
- 20 O1 from parameter set, no optimization, with PULPROG=zgpr
- 23 O1 from parameter set, no optimization, with PULPROG=npt\_zggppr
- 25 O1 from parameter set, no optimization, with PULPROG=zgcppr
- 27 O1 from parameter set, no optimization, with PULPROG=zgcppppr
- 30 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgpr
- 33 O1 from previous watersuppression, no optimization[\*], with PULPROG=npt\_zggppr
- 35 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppr
- 37 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppppr
- [\*] The optimization of O1 is enforced, if O1 was not determined during a previous measurement.



## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
PULPROG	zgpr				WDW	1			
NS	8				LB	0.000	Hz		
DS	4				PC	0.100			
RG	0.250		optim. by NMRPT		F1P	10.806	ppm		
O1P	4.699	ppm			F2P	-1.227	ppm		
SW	12.132	ppm			CY	111.000	cm		
TD	10194		field dependent						
AQ	1.050	s							
FIDRES	0.952	Hz	field dependent						
D 1	5.000	s							
P 1	14.0	us	90deg						
PLW 1	7.5	W	Pow@90deg(Specs)						
PLW 9	0.00006	W	Pow@90deg(5000u)						
TE	298.000	K	default						

## Experiment Description


For the watersuppression experiment the carrier offset O1 needs to be very carefully adjusted. Using the control option for acquisition L23=1, O1 is determined by an intensity profile(second point of FID) over a range of 16 Hz in 0.4 Hz steps around submitted O1.

RG adjustment is executed by running experiments with DS=0, NS=4 and a reduced PLW1. The highest RG without signal overflow in these experiments will be used.

Options L23=1 and 20 are standard whereas Options L23=23, 25, and 27 are non-standard and will not be considered as regular test from the 'NMRPT Control Structure'. Their existence is purely diagnostic. Experiment will be set to irregular if one or both of the options 'Skip Flow' or 'Skip Temperature' are selected.

## 5.2.114 Watersuppression with 535 l/h gas flow (NPT\_1H\_watersuppression\_535l)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
 Z10902, Z10246, Z100930, Z10268, Z10036, Z10267, Z10719, Z10720  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Water suppression spectrum printed from 10.5 ppm to -1.0 ppm as overview spectrum. The expansion plot to the right shows the DSS (sodium 2,2-dimethyl-2-silapentane-5-sulphonate) signal, indicative for the quality of the spectral resolution. The quality of the water suppression is determined as linewidth of the residual water signal using the 10% and 50% intensity of the DSS signal.

The expansion plot to the left shows the signal of the anomeric proton (alpha-C1) of the sucrose (2-beta-D-fructofuranosyl-1-alpha-D-glucopyranosid). The anomeric proton intensity is used for the signal-to-noise calculation and the splitting of the signal is analysed to give a measure for spectral resolution.

### Control Option for Acquisition (L23)

- 1 with O1 optimization, with PULPROG=zgpr
- 20 O1 from parameter set, no optimization, with PULPROG=zgpr
- 23 O1 from parameter set, no optimization, with PULPROG=npt\_zggppr
- 25 O1 from parameter set, no optimization, with PULPROG=zgcppr
- 27 O1 from parameter set, no optimization, with PULPROG=zgcppppr
- 30 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgpr
- 33 O1 from previous watersuppression, no optimization[\*], with PULPROG=npt\_zggppr
- 35 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppr
- 37 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppppr
- [\*] The optimization of O1 is enforced, if O1 was not determined during a previous measurement.

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
PULPROG	zgpr				WDW	1			
NS	8				LB	0.000	Hz		
DS	4				PC	0.100			
RG	0.250		optim. by NMRPT		F1P	10.806	ppm		
O1P	4.699	ppm			F2P	-1.227	ppm		
SW	12.132	ppm			CY	111.000	cm		
TD	10194		field dependent						
AQ	1.050	s							
FIDRES	0.952	Hz	field dependent						
D 1	5.000	s							
P 1	14.0	us	90deg						
PLW 1	7.5	W	Pow@90deg(Specs)						
PLW 9	0.00006	W	Pow@90deg(5000u)						
TE	298.000	K	default						

## Experiment Description


For the watersuppression experiment the carrier offset O1 needs to be very carefully adjusted. Using the control option for acquisition L23=1, O1 is determined by an intensity profile(second point of FID) over a range of 16 Hz in 0.4 Hz steps around submitted O1.

RG adjustment is executed by running experiments with DS=0, NS=4 and a reduced PLW1. The highest RG without signal overflow in these experiments will be used.

Options L23=1 and 20 are standard whereas Options L23=23, 25, and 27 are non-standard and will not be considered as regular test from the 'NMRPT Control Structure'. Their existence is purely diagnostic. Experiment will be set to irregular if one or both of the options 'Skip Flow' or 'Skip Temperature' are selected.

## 5.2.115 Watersuppression with 670 l/h gas flow (NPT\_1H\_watersuppression\_670l)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
 Z10902, Z10246, Z100930, Z10268, Z10036, Z10247, Z10267, Z10719, Z10720  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Water suppression spectrum printed from 10.5 ppm to -1.0 ppm as overview spectrum. The expansion plot to the right shows the DSS (sodium 2,2-dimethyl-2-silapentane-5-sulphonate) signal, indicative for the quality of the spectral resolution. The quality of the water suppression is determined as linewidth of the residual water signal using the 10% and 50% intensity of the DSS signal.

The expansion plot to the left shows the signal of the anomeric proton (alpha-C1) of the sucrose (2-beta-D-fructofuranosyl-1-alpha-D-glucopyranosid). The anomeric proton intensity is used for the signal-to-noise calculation and the splitting of the signal is analysed to give a measure for spectral resolution.

### Control Option for Acquisition (L23)

- 1 with O1 optimization, with PULPROG=zgpr
- 20 O1 from parameter set, no optimization, with PULPROG=zgpr
- 23 O1 from parameter set, no optimization, with PULPROG=npt\_zggppr
- 25 O1 from parameter set, no optimization, with PULPROG=zgcppr
- 27 O1 from parameter set, no optimization, with PULPROG=zgcppppr
- 30 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgpr
- 33 O1 from previous watersuppression, no optimization[\*], with PULPROG=npt\_zggppr
- 35 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppr
- 37 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppppr
- [\*] The optimization of O1 is enforced, if O1 was not determined during a previous measurement.

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
PULPROG	zgpr				WDW	1			
NS	8				LB	0.000	Hz		
DS	4				PC	0.100			
RG	0.250		optim. by NMRPT		F1P	10.806	ppm		
O1P	4.699	ppm			F2P	-1.227	ppm		
SW	12.132	ppm			CY	111.000	cm		
TD	10194		field dependent						
AQ	1.050	s							
FIDRES	0.952	Hz	field dependent						
D 1	5.000	s							
P 1	14.0	us	90deg						
PLW 1	7.5	W	Pow@90deg(Specs)						
PLW 9	0.00006	W	Pow@90deg(5000u)						
TE	298.000	K	default						

## Experiment Description

For the watersuppression experiment the carrier offset O1 needs to be very carefully adjusted. Using the control option for acquisition L23=1, O1 is determined by an intensity profile(second point of FID) over a range of 16 Hz in 0.4 Hz steps around submitted O1.

RG adjustment is executed by running experiments with DS=0, NS=4 and a reduced PLW1. The highest RG without signal overflow in these experiments will be used.

Options L23=1 and 20 are standard whereas Options L23=23, 25, and 27 are non-standard and will not be considered as regular test from the 'NMRPT Control Structure'. Their existence is purely diagnostic. Experiment will be set to irregular if one or both of the options 'Skip Flow' or 'Skip Temperature' are selected.


## 5.2.116 Watersuppression with recommended gas flow (NPT\_1H\_watersuppression\_recflow)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10902, Z10246, Z180181, Z100930, Z10268, Z10036, Z10247, Z10267, Z10719,  
Z10720, Z142222

**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

Water suppression spectrum printed from 10.5 ppm to -1.0 ppm as overview spectrum. The expansion plot to the right shows the DSS (sodium 2,2-dimethyl-2-silapentane-5-sulphonate) signal, indicative for the quality of the spectral resolution. The quality of the water suppression is determined as linewidth of the residual water signal using the 10% and 50% intensity of the DSS signal.

The expansion plot to the left shows the signal of the anomeric proton (alpha-C1) of the sucrose (2-beta-D-fructofuranosyl-1-alpha-D-glucopyranosid). The anomeric proton intensity is used for the signal-to-noise calculation and the splitting of the signal is analysed to give a measure for spectral resolution.

### Control Option for Acquisition (L23)

- 1 with O1 optimization, with PULPROG=zgpr
- 20 O1 from parameter set, no optimization, with PULPROG=zgpr
- 23 O1 from parameter set, no optimization, with PULPROG=npt\_zggppr
- 25 O1 from parameter set, no optimization, with PULPROG=zgcppr
- 27 O1 from parameter set, no optimization, with PULPROG=zgcpgppr
- 30 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgpr
- 33 O1 from previous watersuppression, no optimization[\*], with PULPROG=npt\_zggppr
- 35 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppr
- 37 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcpgppr
- [\*] The optimization of O1 is enforced, if O1 was not determined during a previous measurement.

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
PULPROG	zgpr				WDW	1			
NS	8				LB	0.000	Hz		
DS	4				PC	0.100			
RG	0.250		optim. by NMRPT		F1P	10.806	ppm		
O1P	4.699	ppm			F2P	-1.227	ppm		
SW	12.132	ppm			CY	111.000	cm		
TD	10194		field dependent						
AQ	1.050	s							
FIDRES	0.952	Hz	field dependent						
D 1	5.000	s							
P 1	14.0	us	90deg						
PLW 1	7.5	W	Pow@90deg(Specs)						
PLW 9	0.00006	W	Pow@90deg(5000u)						
TE	298.000	K	default						

## Experiment Description


For the watersuppression experiment the carrier offset O1 needs to be very carefully adjusted. Using the control option for acquisition L23=1, O1 is determined by an intensity profile(second point of FID) over a range of 16 Hz in 0.4 Hz steps around submitted O1.

RG adjustment is executed by running experiments with DS=0, NS=4 and a reduced PLW1. The highest RG without signal overflow in these experiments will be used.

Options L23=1 and 20 are standard whereas Options L23=23, 25, and 27 are non-standard and will not be considered as regular test from the 'NMRPT Control Structure'. Their existence is purely diagnostic. Experiment will be set to irregular if option 'Skip Temperature' is selected.

## 5.2.117 <sup>29</sup>Si background without sample (NPT\_29Si\_backgr\_nosample)

**Test Sample:** 85% Hexamethyldisiloxane (HMDSO, [[CH<sub>3</sub>]<sub>3</sub>Si]<sub>2</sub>O) in Benzene-D<sub>6</sub>  
Z10209, Z10210  
**Solvent:** C<sub>6</sub>D<sub>6</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Top: <sup>29</sup>Si spectrum with sample to estimate spectrum quality.

Bottom: <sup>29</sup>Si Background signal spectrum without sample. No sharp signal should be present. Broad signal could arise from solid compound in the probe.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on experiment with sample



## Parameters

<b>F1 ACQU</b>			Parameters	<b>PLW 2</b>	6.6	W	Pow@90deg(Specs) NUC2
NUC1	29Si			<b>F1 PROC</b>			Parameters
PULPROG	zgig30			SI	32768		
NS	500			WDW	1		
DS	0			LB	5.000	Hz	
RG	101.000		no optim.	PC	1.400		
O1P	6.518	ppm		F1P	10.000	ppm	
SW	499.182	ppm		F2P	0.000	ppm	
TD	65536			CY	8.000	cm	
D 1	4.874	s	AQ+D1=const				
P 1	20.0	us	90deg NUC1				
PLW 1	14.3	W	Pow@90deg(Specs) NUC1				
PLW 12	0.1	W	Pow@90deg(CPD)				
CPDPRG2	waltz64		decoupl. sequence				
TE	298.000	K	default				

## Experiment Description


Background signal measurements are executed using a small flip angle due to two reasons. First, to get more efficient excitation bandwidth, secondly to compensate for usually long T1 relaxation time of solid signals.

Experiment will be set to irregular if one or both of the options 'Skip Tuning/Matching' or 'Skip XNUC Tuning/Matching' are selected.

A spectrum with sample and NS=1 will be acquired on a derived data set. This experiment can be used to estimate the spectrum quality. After phase correction SINO will be executed on the spectrum with sample. The experiment will be aborted, if a minimal SINO of 20 is not achieved. SINO check can be skipped with L23=2.

## 5.2.118 P90 29Si pulse calibration (NPT\_29Si\_p90determination\_29si)

**Test Sample:** 85% Hexamethyldisiloxane (HMDSO,  $[(CH_3)_3Si]_2O$ ) in Benzene-D6  
 Z10209, Z142229, Z10210  
**Solvent:** C6D6  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 29Si		SI 4096	
PARMODE 0	Data Dimension	WDW 1	
PULPROG zg		LB 5.000 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 101.000	optim. by RGA	ME_mod 0	LPfc
O1P 6.512 ppm		NCOEF 20	
SWH 396.825 Hz		ABSF1 1000.000 ppm	
TD 1000		ABSF2 -1000.000 ppm	
AQ 1.260 s		F1P 10.000 ppm	
FIDRES 0.794 Hz		F2P -38.000 ppm	
D 1 62.415 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.


A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.119 <sup>29</sup>Si sensitivity (NPT\_29Si\_sensitivity)

---

**Test Sample:** 85% Hexamethyldisiloxane (HMDSO, [[CH<sub>3</sub>]<sub>3</sub>Si]<sub>2</sub>O) in Benzene-D<sub>6</sub>  
Z10209, Z142229, Z10210  
**Solvent:** C<sub>6</sub>D<sub>6</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Silicon-29 sensitivity test.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	29Si			SI	65536		
PULPROG	zg			WDW	1		
NS	1			LB	5.000	Hz	
DS	0			PC	1.400		
RG	101.000		no optim.	F1P	10.000	ppm	
O1P	6.512	ppm		F2P	0.000	ppm	
SW	61.065	ppm		CY	11.000	cm	
TD	32768						
AQ	3.375	s	field dependent				
FIDRES	0.296	Hz	field dependent				
D 1	415.525	s	AQ+D1=const				
P 1	20.0	us	90deg NUC1				
PLW 1	14.3	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				

## Experiment Description

Silicon-29 sensitivity test (no 1H decoupling). Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from 10.0 to 0.0 ppm, while the best 10 ppm noise region is determined over the range from 37.0 to 9.0 ppm.

## 5.2.120 31P B1 homogeneity integral (NPT\_31P\_b1homogeneityInt\_31p)

**Test Sample:** 0.0485 M Triphenyl Phosphate (TPP, [C<sub>6</sub>H<sub>5</sub>]<sub>3</sub>PO<sub>4</sub>) in Acetone-D<sub>6</sub>  
 Z10201, Z100934, Z10202, Z10038, Z10257, Z10276, Z10722, Z142226  
**Solvent:** Acetone  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a fraction around signal maximum (-60 deg, +60 deg) is acquired and shown in the figure. The variable pulse list which is used for acquisition is shown to the right.

The B1-homogeneities based on amplitudes and integrals are computed from the peak-picking list amplitudes and the magnitude of the second complex FID data point respectively.

### Control Option for Acquisition (L23)

- 1 default PLW 1 is used for determination of B1 homogeneity.
- 2 PLW 1 will be set to CNST 10 and resultant pulses are used for determination of B1 homogeneity. Experiment will be set to irregular.
- 3 PLW 1 will be set to power corresponding to specified customer pulses and resultant pulses are used for determination of B1 homogeneity.

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	31P			SI	4096		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_p1b1hom2d			LB	5.000	Hz	
NS	1			PH_mod	1		pk
DS	0			ME_mod	2		LPfc
RG	101.000		optim. by RGA	NCOEF	20		
O1P	-17.609	ppm		ABSF1	1000.000	ppm	
SWH	396.825	Hz		ABSF2	-1000.000	ppm	
TD	1024			F1P	5.720	ppm	
AQ	1.290	s		F2P	5.320	ppm	
FIDRES	0.775	Hz		<b>F1 ACQU</b>			Parameters F1
D 1	83.727	s	AQ+D1=const	NUC1	31P		
P 1	14.0	us	90deg NUC1	TD	43		No of incr.
PLW 1	6.6	W	Pow@90deg(Specs) NUC1	<b>F1 PROC</b>			Parameters F1
TE	298.000	K	default	SI	64		
				<b>NMRPT</b>			Parameters
				L 4	15		integ. fraction of 90deg
				L 5	4		# of step per maxima

## Experiment Description

The B1 homogeneity measurement is normally acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range is from 90 to 810 degree. In order to speed up the measurements NMRPT is acquiring only the partial ranges around the positive signal maxima for the pulse angles 90, 450, and 810 degree.

By this approach it is possible to get a better incremental resolution and a shorter overall measuring time. Prior to the main acquisition O1P (PROCNO=2) and phase correction (PROCNO=3) will be optimized. Intermediate results of the CONVTO1D routine is stored in PROCNO=998, result of CONVTO1D in PROCNO=999.


As control of stability, reproducibility and 90 degree calibration a spectrum with 180 degree pulse is acquired before and after each maxima. If relaxation conditions and instrument stability is sufficient, the result should not differ among them.

For the prediction of the signal maxima at 450 and 810 degree the propagation delay is included in the calculation of the VP list. (PDelay is calculated as difference of 360 and 180 degree pulse lengths during the corresponding 90 degree pulse determination experiment.) In this measurement O1 is determined before the main acquisition is started. Evaluation is achieved after processing the data in F2 only (XF2) by peak picking and integral determination of the most intense signal per maxima.

## 5.2.121 31P background with sample (NPT\_31P\_backgr\_withsample)

---

**Test Sample:** 0.0485 M Triphenyl Phosphate (TPP,  $[C_6H_5]_3PO_4$ ) in Acetone-D6  
Z10201, Z100934, Z10202, Z10038, Z10257, Z10276, Z10722  
**Solvent:** Acetone  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

31P Background signal spectrum with sample. Sharp signal arises from sample, broad signal could arise from solid compound in the probe.

### Control Option for Acquisition (L23)

1 default



## Parameters

<b>F1 ACQU</b>			Parameters	<b>CNST 10</b>	20.000		Flip angle for P90
NUC1	31P			<b>F1 PROC</b>			Parameters
NUC2	1H			SI	32768		
PULPROG	npt_zg0ig			WDW	1		
NS	1000			LB	5.000	Hz	
DS	4			PC	1.400		
RG	101.000		no optim.	F1P	24.747	ppm	
O1P	-17.609	ppm		F2P	-150.759	ppm	
O2P	5.000	ppm					
CPDPRG2	waltz64		decoupl. sequence				
SW	308.694	ppm					
TD	32768						
AQ	0.328	s	field dependent				
FIDRES	3.052	Hz	field dependent				
D 1	1.487	s	AQ+D1=const				
P 0		us	P 1 * CNST 10 / 90				
P 1	11.0	us	90deg NUC1				
PLW 1	15.5	W	Pow@90deg(Specs) NUC1				
PLW 12	0.1	W	Pow@CPD NUC2				
TE	298.000	K	default				


## Experiment Description

Background signal measurements are executed using a small flip angle due to two reasons. First, to get more efficient excitation bandwidth, secondly to compensate for usually long T1 relaxation time of solid signals.

Full spectrum is normally printed without any baseline correction.

## 5.2.122 CPD 31P pulse calibration (NPT\_31P\_cpddeterminationf1\_31p)

**Test Sample:** 0.0485 M Triphenyl Phosphate (TPP, [C<sub>6</sub>H<sub>5</sub>]<sub>3</sub>PO<sub>4</sub>) in Acetone-D<sub>6</sub>  
 Z10201, Z100934, Z10202, Z10038, Z10257, Z10276, Z10722  
**Solvent:** Acetone  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 31P		SI 4096	
PARMODE 0	Data Dimension	WDW 1	
PULPROG zg		LB 2.000 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 101.000	optim. by RGA	ME_mod 0	LPfc
O1P -17.609 ppm		NCOEF 20	
SWH 396.825 Hz		ABSF1 1000.000 ppm	
TD 1000		ABSF2 -1000.000 ppm	
AQ 1.260 s		F1P 10.000 ppm	
FIDRES 0.794 Hz		F2P -38.000 ppm	
D 1 17.200 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.


A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.123 31P test for artifacts (NPT\_31P\_fullsw\_dec1h)

---

**Test Sample:** 0.0485 M Triphenyl Phosphate (TPP, [C<sub>6</sub>H<sub>5</sub>]<sub>3</sub>PO<sub>4</sub>) in Acetone-D<sub>6</sub>  
Z10201, Z100934, Z10202, Z10038, Z10257, Z10276, Z10722  
**Solvent:** Acetone  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Full range 31P spectrum with 1H decoupling.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	31P				SI	131072			
NUC2	1H				WDW	1			
PULPROG	npt_zg0ig				LB	5.000	Hz		
NS	64				PC	1.400			
DS	4				F1P	24.747	ppm		
RG	101.000		no optim.		F2P	-150.759	ppm		
O1P	-17.609	ppm			CY	11.000	cm		
O2P	5.000	ppm							
SW	308.694	ppm							
TD	65536								
AQ	0.655	s	field dependent						
FIDRES	1.526	Hz	field dependent						
D 1	3.545	s	AQ+D1=const						
P 1	11.0	us	90deg NUC1						
PLW 1	18.0	W	Pow@90deg(Specs) NUC1						
PLW 12	0.2	W	Pow@CPD NUC2						
CPDPRG2	waltz64		decoupl. sequence						
TE	298.000	K	default						

## Experiment Description

Full range spectra (spectroscopically useful range) are acquired as test for RF artifacts (spikes). Currently no evaluation is provided and visual inspection of the data is required. Data are treated with baseline correction.

## 5.2.124 P90 31P pulse calibration (NPT\_31P\_p90determinationf1\_31p)

**Test Sample:** 0.0485 M Triphenyl Phosphate (TPP, [C<sub>6</sub>H<sub>5</sub>]<sub>3</sub>PO<sub>4</sub>) in Acetone-D<sub>6</sub>  
 Z10201, Z100934, Z10202, Z10038, Z10257, Z10276, Z10722, Z142226  
**Solvent:** Acetone  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 31P		SI 4096	
PARMODE 0	Data Dimension	WDW 1	
PULPROG zg		LB 2.000 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 101.000	optim. by RGA	ME_mod 0	LPfc
O1P -17.609 ppm		NCOEF 20	
SWH 396.825 Hz		ABSF1 1000.000 ppm	
TD 1000		ABSF2 -1000.000 ppm	
AQ 1.260 s		F1P 10.000 ppm	
FIDRES 0.794 Hz		F2P -38.000 ppm	
D 1 17.200 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.


A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.125 <sup>31</sup>P sensitivity (NPT\_31P\_sensitivity)

---

**Test Sample:** 0.0485 M Triphenyl Phosphate (TPP, [C<sub>6</sub>H<sub>5</sub>]<sub>3</sub>PO<sub>4</sub>) in Acetone-D<sub>6</sub>  
Z10201, Z100934, Z10202, Z10038, Z10257, Z10276, Z10722, Z142226  
**Solvent:** Acetone  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Phosphorous-31 sensitivity test.

### Control Option for Acquisition (L23)

1 default



## Parameters

F1 ACQU		Parameters		F1 PROC		Parameters	
NUC1	31P			SI	16384		
PULPROG	zg			WDW	1		
NS	1			LB	5.000	Hz	
DS	0			PC	1.400		
RG	101.000		no optim.	F1P	6.168	ppm	
O1P	-17.609	ppm		F2P	-43.919	ppm	
SW	50.606	ppm		CY	11.000	cm	
TD	32768						
AQ	1.999	s	field dependent				
FIDRES	0.500	Hz	field dependent				
D 1	119.001	s	AQ+D1=const				
P 1	9.0	us	90deg NUC1				
PLW 1	27.4	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				


## Experiment Description

Phosphorous-31 sensitivity test (no 1H decoupling). Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from -15.0 to -18.0 ppm, while the best 5 ppm noise region is determined over the range from 7.0 to -15.0 ppm.

## 5.2.126 <sup>31</sup>P sensitivity with 1H decoupling (NPT\_31P\_sensitivity\_dec1h)

---

**Test Sample:** 0.0485 M Triphenyl Phosphate (TPP, [C<sub>6</sub>H<sub>5</sub>]<sub>3</sub>PO<sub>4</sub>) in Acetone-D<sub>6</sub>  
Z10201, Z100934, Z10202, Z10038, Z10257, Z10276, Z10722, Z142226  
**Solvent:** Acetone  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Phosphorous-31 sensitivity test with 1H decoupling.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	31P			SI	16384		
NUC2	1H			WDW	1		
PULPROG	zgig			LB	5.000	Hz	
NS	1			PC	1.400		
DS	0			F1P	6.168	ppm	
RG	101.000		no optim.	F2P	-43.919	ppm	
O1P	-17.609	ppm		CY	11.000	cm	
O2P	5.000	ppm					
SW	50.606	ppm					
TD	32768						
AQ	1.999	s	field dependent				
FIDRES	0.500	Hz	field dependent				
D 1	119.001	s	AQ+D1=const				
P 1	9.0	us	90deg NUC1				
PLW 1	27.4	W	Pow@90deg(Specs) NUC1				
PLW 12	0.35	W	Pow@CPD NUC2				
CPDPRG2	waltz64		decoupl. sequence				
TE	298.000	K	default				

## Experiment Description

Phosphorous-31 sensitivity test with 1H decoupling. Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from -15.0 to -18.0 ppm, while the best 5 ppm noise region is determined over the range from 7.0 to -15.0 ppm.

## 5.2.127 P90 39K pulse calibration (NPT\_39K\_p90determination\_39k)

**Test Sample:** 1 M Potassium Chloride (KCl) in D2O  
 Z10075, Z10076  
**Solvent:** D2O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
  - 2 skip SINO check on PROCNO 11
  - 11 ignore specifications (optimize power for pulse length from prosol)
  - 12 ignore specifications and skip SINO check on PROCNO 11
  - 21 ignore specifications and optimize pulse length for power from prosol
  - 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
  - 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX  
 1000Same as xxx but skip automatic O1P determination  
 +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 39K		SI 4096	
PARMODE 0	Data Dimension	WDW 3	
PULPROG zg		LB 6.000 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 101.000	optim. by RGA	ME_mod 2	LPfc
O1P 1.341 ppm		NCOEF 20	
SWH 588.235 Hz		ABSF1 2.652 ppm	
TD 1000		ABSF2 -1.964 ppm	
AQ 0.850 s		F1P 2.907 ppm	
FIDRES 1.176 Hz		F2P -2.221 ppm	
D 1 0.250 s	AQ+D1=const	CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.


A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.128 39K sensitivity (NPT\_39K\_sensitivity)

---

**Test Sample:** 1 M Potassium Chloride (KCl) in D2O  
Z10075, Z10076  
**Solvent:** D2O\_salt  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Potassium-39 sensitivity test.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	39K				SI	8192			
PULPROG	zg				WDW	1			
NS	1				LB	6.000	Hz		
DS	0				PC	1.000			
RG	101.000		no optim.		F1P	50.100	ppm		
O1P	1.341	ppm			F2P	-50.100	ppm		
SW	111.578	ppm			CY	11.000	cm		
TD	4096								
AQ	0.983	s	field dependent						
FIDRES	1.017	Hz	field dependent						
D 1	5.000	s	AQ+D1=const						
P 1	35.0	us	90deg NUC1						
PLW 1	70.2	W	Pow@90deg(Specs) NUC1						
TE	298.000	K	default						

## Experiment Description

Potassium-39 sensitivity test. Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from 10.0 to -10.0 ppm, while the best 40 ppm noise region is determined over the range from 50.0 to 10.0 ppm.

## 5.2.129 Atma test (NPT\_prep\_atma\_test)

---

**Test Sample:** Water Sample with 0.0, 0.25, or 1.0 M Sodium Chloride (NaCl).  
Z142222, Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10719,  
Z10720, Z10191, Z101717, Z101714, Z100377, Z10731, Z10284, Z101715, Z10285,  
Z100375, Z101710, Z10192, Z10729, Z10288, Z101716, Z101712, Z100376,  
Z100372, Z10730

**Solvent:** H2O+D2O or D2O\_salt

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off

```
----- RESULTS of Tuning Matching with sample -----
ATMA_ERR= 0 : ATMcu= 130 ATMChannel= f2 -->ATMProfileName= NPT_1H_p90determinationf2_130 -->TuningFreq= 100.411000 MHz
ATMA_ERR= 0 : ATMcu= 18 ATMChannel= f1 -->ATMProfileName= NPT_1H_p90determinationf1_18 -->TuningFreq= 400.130000 MHz
ATMA_ERR= 0 : ATMcu= 158 ATMChannel= f2 -->ATMProfileName= NPT_1H_p90determinationf2_158 -->TuningFreq= 40.545000 MHz
ATMA_ERR= 0 : ATMcu= 18 ATMChannel= f1 -->ATMProfileName= NPT_1H_p90determinationf1_18 -->TuningFreq= 400.130000 MHz
ATMA_ERR= 0 : ATMcu= 119 ATMChannel= f1 -->ATMProfileName= NPT_1H_p90determinationf1_119 -->TuningFreq= 376.498000 MHz
ATMA_ERR= 0 : ATMcu= 18 ATMChannel= f1 -->ATMProfileName= NPT_1H_p90determinationf1_18 -->TuningFreq= 400.130000 MHz
ATMA_ERR= 0 : ATMcu= 218 ATMChannel= f1 -->ATMProfileName= NPT_1H_p90determinationf1_218 -->TuningFreq= 161.974000 MHz
ATMA_ERR= 0 : ATMcu= 18 ATMChannel= f1 -->ATMProfileName= NPT_1H_p90determinationf1_18 -->TuningFreq= 400.130000 MHz
ATMA_ERR= 0 : ATMcu= 198 ATMChannel= f1 -->ATMProfileName= NPT_1H_p90determinationf1_198 -->TuningFreq= 16.670000 MHz
ATMA_ERR= 0 : ATMcu= 18 ATMChannel= f1 -->ATMProfileName= NPT_1H_p90determinationf1_18 -->TuningFreq= 400.130000 MHz
ATMA_ERR= 0 : ATMcu= 2981 ATMChannel= f1 -->ATMProfileName= NPT_1H_p90determinationf1_2981 -->TuningFreq= 73.491000 MHz
ATMA_ERR= 0 : ATMcu= 18 ATMChannel= f1 -->ATMProfileName= NPT_1H_p90determinationf1_18 -->TuningFreq= 400.130000 MHz
```

### Example Printout

List of ATM events.

### Control Option for Acquisition (L23)

1 default



## Parameters

<b>F1 ACQU</b> NUC1 1H PULPROG npt_zgnopul TE 298.000 K	Parameters  default	<b>F1 PROC</b> CY 11.000 cm	Parameters
--	---------------------------	--------------------------------	------------

## Experiment Description

ATM test is executed for nuclei 1H, 2H, 13C, 15N, 19F, 29Si, 31P, 39K and 79Br, if specified for the given probe.

The tuning/matching order is defined by the ATMA\_SEQUENCE specified.

The tuning/matching frequency is always referenced to the standard basic proton frequency of BF.13 MHz.


The probe will be tuned and matched to (standard frequency - 0.5 \* specified value).

The wobble curves will be stored with procno 700 + number of atma attempt.

Requirements: Ratios for nuclei to be tested need to be specified.

## 5.2.130 2H B1 homogeneity integral (NPT\_prep\_b1homogeneityInt\_d)

**Test Sample:** 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub>  
 Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721, Z142223  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a fraction around signal maximum (-60 deg, +60 deg) is acquired and shown in the figure. The variable pulse list which is used for acquisition is shown to the right.

The B1-homogeneities based on amplitudes and integrals are computed from the peak-picking list amplitudes and the magnitude of the second complex FID data point respectively.

### Control Option for Acquisition (L23)

- 1 default PLW 1 is used for determination of B1 homogeneity.
- 2 PLW 1 will be set to CNST 10 and resultant pulses are used for determination of B1 homogeneity. Experiment will be set to irregular.
- 3 PLW 1 will be set to power corresponding to specified customer pulses and resultant pulses are used for determination of B1 homogeneity.

## Parameters

<b>F2 ACQU</b>			Parameters F2	<b>F2 PROC</b>			Parameters F2
NUC1	2H			SI	4096		
PARMODE	1		Data Dimension	WDW	1		
PULPROG	npt_p1b1hom2h2d			LB	1.000	Hz	
NS	1			PH_mod	1		pk
DS	4			ME_mod	2		LPfc
RG	0.250		optim. by RGA	NCOEF	20		
O1P	2.056	ppm		ABSF1	1000.000	ppm	
SWH	230.766	Hz		ABSF2	-1000.000	ppm	
TD	1024			F1P	5.720	ppm	
AQ	2.219	s		F2P	5.320	ppm	
FIDRES	0.451	Hz		<b>F1 ACQU</b>			Parameters F1
D 1	3.683	s	AQ+D1=const	NUC1	2H		
P 1	14.0	us	90deg NUC1	TD	43		No of incr.
PLW 1	6.6	W	Pow@90deg(Specs) NUC1	<b>F1 PROC</b>			Parameters F1
TE	298.000	K	default	SI	64		
				<b>NMRPT</b>			Parameters
				L 4	6		integ. fraction of 90deg
				L 5	8		# of step per maxima

## Experiment Description

The B1 homogeneity measurement is normally acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range is from 90 to 810 degree. In order to speed up the measurements NMRPT is acquiring only the partial ranges around the positive signal maxima for the pulse angles 90, 450, and 810 degree.


By this approach it is possible to get a better incremental resolution and a shorter overall measuring time. Prior to the main acquisition O1P (PROCNO=2) and phase correction (PROCNO=3) will be optimized. Intermediate results of the CONVTO1D routine is stored in PROCNO=998, result of CONVTO1D in PROCNO=999.

As control of stability, reproducibility and 90 degree calibration a spectrum with 180 degree pulse is acquired before and after each maxima. If relaxation conditions and instrument stability is sufficient, the result should not differ among them.

For the prediction of the signal maxima at 450 and 810 degree the propagation delay is included in the calculation of the VP list. (PDelay is calculated as difference of 360 and 180 degree pulse lengths during the corresponding 90 degree pulse determination experiment.) In this measurement O1 is determined before the main acquisition is started. Evaluation is achieved after processing the data in F2 only (XF2) by peak picking and integral determination of the most intense signal per maxima.

## 5.2.131 CPD 2H pulse calibration (NPT\_prep\_cpddeterminationf1\_d)

**Test Sample:** 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub> Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721  
**Solvent:** DMSO  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 2H		SI 2048	
PARMODE 0	Data Dimension	WDW 1	
PULPROG npt_zg2h		LB 2.000 Hz	
LOCNUC off		SSB 2.000	
NS 1		PH_mod 1	pk
DS 0		ME_mod 2	LPfc
RG 1.000	optim. by RGA	NCOEF 20	
O1P 2.509 ppm		ABSF1 1000.000 ppm	
SWH 326.797 Hz		ABSF2 -1000.000 ppm	
TD 1024		F1P 5.000 ppm	
AQ 1.567 s		F2P 0.000 ppm	
FIDRES 0.638 Hz		CY 11.000 cm	
D 1 0.350 s	AQ+D1=const		
P 1 180.0 us	90deg NUC1		
PLW 1 6.0 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skiped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.132 Optimization of 2H frequency (NPT\_prep\_fieldsetting\_d)


---

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10719, Z10720,  
Z142222

**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation according to RO



### Example Printout

2H spectrum after field optimization

### Control Option for Acquisition (L23)

1 default

## Parameters

<p><b>F1 ACQU</b></p> <p>NUC1 2H          PARMODE 0          PULPROG npt_zg2h          LOCNUC off          NS 8          DS 0          RG 0.250          O1P 0.000 ppm          SWH 74626.867 Hz          TD 131072          AQ 0.878 s          FIDRES 1.139 Hz          D 1 0.500 s          P 1 180.0 us          PLW 1 6.0 W          TE 298.000 K          DE 100.000 us</p>	<p>Parameters</p> <p>Data Dimension</p> <p>optim. by RGA</p> <p>90deg NUC1          Pow@90deg(Specs) NUC1          default          set after getprosol</p>
<p><b>F1 PROC</b></p> <p>SI 131072          WDW 1          PH_mod 2          F1P 813.947 ppm          F2P -813.947 ppm          CY 11.000 cm</p> <p><b>NMRPT</b></p> <p>CNST 50 0.450          CNST 51 2.500          CNST 52 -75.000 dB          CNST 53 -59.900 dB          CNST 54 -500.000          CNST 55 105.000 dB</p>	<p>Parameters</p> <p>MC</p> <p>Parameters</p> <p>LOCK(sweep rate)          LOCK(sweep amplitude)          LOCKDC (default)          min LOCKPOWER          delta FIELD for determ.          LOCKGAIN reference</p>

## Experiment Description


The experiment is used for determination of the current FIELD value based on the SOLVENT which is defined.

The experimental procedure includes two 2H acquisitions with constant 01 and LOCKSHIFT at two known FIELD positions. Using 4.7 ppm as chemical shift of 2H in H2O+D2O the correct FIELD value can be calculated from these measurements.

NMRPT stores the resulting FIELD value in the BSMS.

## 5.2.133 2H lineshape with sample rotation (NPT\_prep\_lineshape\_wrot)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10902, Z10246, Z10268, Z10247  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation according to RO



### Example Printout

Deuterium line shape spectrum with offset calibration to O1P as overview spectrum. Printing is from +125.0 Hz to -125.0 Hz. The expansion plot to the left shows deuterium signal with higher resolution (+7.0 Hz, -7.0 Hz).

### Control Option for Acquisition (L23)

1 default



## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	2H			SI	16384		
PULPROG	zg2h			WDW	0		
LOCNUC	off			LB	0.000	Hz	
NS	1			PC	0.400		
DS	0			F1P	14.693	ppm	
RG	0.250		optim. by RGA	F2P	-5.305	ppm	
O1P	0.642	ppm		CY	1000.000	cm	
SWH	1000.000	Hz					
TD	16384						
AQ	8.192	s					
FIDRES	0.122	Hz					
D 1	10.000	s					
P 1	20.0	us	90deg NUC1				
PLW 1	64.0	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				
DE	100.000	us	set after getprosol				

## Experiment Description

Before acquiring the final spectrum, the exact signal position is determined. The carrier position is afterwards set to the position

optimized  $O1 = \text{peak frequency [Hz]} - (\text{SWH}/4)$

This procedure allows the application of a strip transformation with the parameters  $\text{STSR}=0$ ,  $\text{STSI}=\text{SI}/2$  resulting in the signal position on-resonance and in most cases resulting in an optimal automatic phase correction.

## 5.2.134 Optimization of <sup>19</sup>F locksetting (NPT\_prep\_locksettings\_19f)

---

**Test Sample:** 45% Chloroform-D (CDCl<sub>3</sub>) and 45% Chloroform (CHCl<sub>3</sub>) in 10% Hexafluorobenzene (C<sub>6</sub>F<sub>6</sub>),  
Z10078, Z10079  
**Solvent:** C<sub>6</sub>F<sub>6</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off

```
=====
Current LOCK Settings (according to 19Flock)
Solvent      FIELD    LOCKPOWER  LOOPGAIN  LOOPTIME  LOOPFILTER  LOCKPHASE
-----
C6F6         5763.0   -50.0      -3.4      0.3122    96.3        169.3
=====
```

### Example Printout

Listing of entries from edlock (TopSpin configuration file, accessible by edlock).

### Control Option for Acquisition (L23)

- 1 Update LOCK control parameters for current SOLVENT (locked after experiment, default)
- 11 same as L23=1 but enforces a more complex phase optimization algorithm

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	2H				WDW	1			
PULPROG	zg2h				PH_mod	1			
LOCNUC	19F				F1P	14.693	ppm		
NS	1				F2P	-5.305	ppm		
DS	0				CY	11.000	cm		
RG	101.000			optim. by RGA					
O1P	7.000	ppm							
SWH	7462.687	Hz							
TD	74626								
AQ	5.000	s							
FIDRES	0.200	Hz							
D 1	1.000	s							
P 1	14	us		90deg NUC1					
PLW 1	6.6	W		Pow@90deg(Specs) NUC1					
TE	298.000	K		default					

## Experiment Description

This experiment is used for the initial setup of the LOCK environment and the update of the lock table of individual SOLVENT.

In case of successful LOCK and after AUTOGAIN the LOCKGAIN is memorized. The best LOCKGAIN value must be nearest to the optimal LOCKPHASE. Using these settings the AUTOPHASE routine determines the exact LOCKPHASE.

The final task of the experiment after storage of all information is the execution of a regular LOCK procedure.

## 5.2.135 Optimization of 2H locksetting (NPT\_prep\_locksettings\_d)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
 Z10246, Z10902, Z100930, Z10268, Z10036, Z10247, Z10267, Z10719, Z10720,  
 Z142222  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off

```
-----
Current LOCK Settings (according to 2Hlock)
-----
```

Solvent	FIELD	LOCKPOWER	LOOPGAIN	LOOPTIME	LOOPFILTER	LOCKPHASE
H2O+D2O	7855.0	-18.0	-9.4	0.4640	50.0	33.2
D2O_S0	7856.0	-18.0	5.9	0.1256	334.3	32.5
D2O	7859.0	-18.0	5.4	0.1329	312.4	32.7
CDCl3	7849.0	-30.0	-9.4	0.4640	50.0	32.4
C6D6	7861.0	-26.0	-1.5	0.2670	122.9	32.5
DMSO	7880.0	-20.0	-9.4	0.4640	50.0	32.2
D2O_S1000	7851.0	-18.0	5.7	0.1285	325.4	32.2
Acetone	7891.0	-38.0	2.1	0.1925	195.4	27.0
MeOD	7886.0	-35.0	2.0	0.1943	193.0	32.7

```
-----
```

### Example Printout

Listing of entries from edlock (TopSpin configuration file, accessible by edlock).

### Control Option for Acquisition (L23)

- 1 Update LOCK control parameters for current SOLVENT (locked after experiment, default)
- 11 same as L23=1 but enforces a more complex phase optimization algorithm

## Parameters

<b>F1 ACQU</b>		Parameters	<b>F1 PROC</b>		Parameters
NUC1	1H		CY	11.000	cm
PULPROG	npt_zgnopul				
LOCNUC	2H				
RG	101.000	no optim.			

## Experiment Description

This experiment is used for the initial setup of the LOCK environment and the update of the lock table of individual SOLVENT.


In case of successful LOCK and after AUTOGAIN the LOCKGAIN is memorized. The best LOCKGAIN value must be nearest to the optimal LOCKPHASE. Using these settings the AUTOPHASE routine determines the exact LOCKPHASE.

NMRPT stores the resulting LOCKPHASE value in the edlock of TopSpin.

The final task of the experiment after storage of all information is the execution of a regular LOCK procedure.

## 5.2.136 P90 2H pulse calibration (NPT\_prep\_p90det\_astm\_d)

**Test Sample:** 40% Dioxane in Benzene-D6 (ASTM)  
 Z10163, Z100929, Z10164, Z10035, Z10255, Z10242, Z10724  
**Solvent:** C6D6  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments around 360 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	2H				SI	1024			
PARMODE	0			Data Dimension	LB	0.500	Hz		
PULPROG	npt_zg2h				F1P	8.095	ppm		
LOCNUC	off				F2P	6.305	ppm		
NS	1				CY	5.500	cm		
DS	0								
RG	0.250			optim. by RGA					
SWH	357.143	Hz							
TD	1048								
AQ	1.467	s							
FIDRES	0.682	Hz							
O1P	7.200	ppm							
P 1	14.0	us		90deg Pulse					
PLW 1	6.6	W		Pow@90deg(Specs)					
DIGMOD	3			baseopt					
DSPFIRM	4			rectangle					
TE	298.000	K		default					

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).


## 5.2.137 P90 2H pulse calibration (NPT\_prep\_p90det\_d)

**Test Sample:** (a) 100 mM Urea-15N ([15NH<sub>2</sub>]<sub>2</sub>CO) and 100 mM Methanol-13C in Dimethyl Sulfoxide-D<sub>6</sub>  
 (b) 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
 Z10263, Z100932, Z10265, Z10037, Z10264, Z10274, Z10721, Z142223

**Solvent:** (a) DMSO  
 (b) H<sub>2</sub>O+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. On the left side the result of a CONVTO1D routine shows six experiments from 90 to 270 deg. On the right side the series goes from 270 deg to 450 deg.

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX



## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	2H				SI	2048			
PARMODE	0			Data Dimension	WDW	1			
PULPROG	npt_zg2h				LB	2.000	Hz		
LOCNUC	off				SSB	2.000			
NS	1				PH_mod	1		pk	
DS	0				ME_mod	2		LPfc	
RG	0.250			optim. by RGA	NCOEF	20			
O1P	2.509	ppm		(Urea Sample)	ABSF1	1000.000	ppm		
O1P	4.7	ppm		(Sucrose Sample)	ABSF2	-1000.000	ppm		
SWH	326.797	Hz			F1P	5.000	ppm		
TD	1024				F2P	0.000	ppm		
AQ	1.567	s			CY	11.000	cm		
FIDRES	0.638	Hz							
D 1	0.350	s		AQ+D1=const					
P 1	180.0	us		90deg NUC1					
PLW 1	6.0	W		Pow@90deg(Specs) NUC1					
TE	298.000	K		default					

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.


A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

PDelay (propagation delay) is calculated as difference of 360 and 180 degree pulse lengths- and memorised for further use in B1 homogeneity experiments.

## 5.2.138 2H sensitivity, 1% D2O (NPT\_prep\_sensitivity\_1\_d)

---

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 99% H<sub>2</sub>O + 1% D<sub>2</sub>O  
Z10908, Z10610, Z10056  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Deuterium sensitivity test.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	2H			SI	262144		
PULPROG	zg2h			WDW	1		
LOCNUC	off			LB	0.000	Hz	
NS	1			PC	0.400		
DS	0			F1P	14.693	ppm	
RG	101.000		no optim.	F2P	-5.305	ppm	
O1P	4.716	ppm		SIGF1	5.000	ppm	
SW	121.497	ppm		SIGF2	4.000	ppm	
AQ	5.000	s		NOISF1	60.000	ppm	
FIDRES	0.200	Hz		NOISF2	10.000	ppm	
D 1	10.000	s		CY	11.000	cm	
P 1	180.0	us	90deg NUC1				
PLW 1	6.0	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				


## Experiment Description

Deuterium sensitivity test. Processing is using no line broadening (LB=0) and baseline correction (ABS). Evaluation is carried out by the command SINO. The signal is searched over the range from 5.0 to 4.0 ppm. For HR probes the noise region is determined over the range from 60.0 to 10.0 ppm, while for CMP and HRMAS probes a noise region of 2000 Hz will be used to exclude spinning side bands .

## 5.2.139 2H sensitivity, 10% D2O (NPT\_prep\_sensitivity\_10\_d)

---

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
Z10902, Z10246, Z142222, Z100930, Z10268, Z10036, Z10247, Z10719, Z10720  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Deuterium sensitivity test.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	2H			SI	262144		
PULPROG	zg2h			WDW	1		
LOCNUC	off			LB	0.000	Hz	
NS	1			PC	0.400		
DS	0			F1P	14.693	ppm	
RG	101.000		no optim.	F2P	-5.305	ppm	
O1P	4.716	ppm		SIGF1	5.000	ppm	
SW	121.497	ppm		SIGF2	4.000	ppm	
AQ	5.000	s		NOISF1	60.000	ppm	
FIDRES	0.200	Hz		NOISF2	10.000	ppm	
D 1	10.000	s		CY	11.000	cm	
P 1	180.0	us	90deg NUC1				
PLW 1	6.0	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				

## Experiment Description

Deuterium sensitivity test. Processing is using no line broadening (LB=0) and baseline correction (ABS). Evaluation is carried out by the command SINO. The signal is searched over the range from 5.0 to 4.0 ppm. For HR probes the noise region is determined over the range from 60.0 to 10.0 ppm, while for CMP and HRMAS probes a noise region of 2000 Hz will be used to exclude spinning side bands .


## 5.2.140 Automated shim optimization (NPT\_prep\_tsopt)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
 Z10246, Z10902, Z180181, Z100930, Z10268, Z10036, Z10247, Z10267, Z10720, Z10719

**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation according to RO



### Example Printout

Top: Result of last water suppression experiment  
 Bottom: Protocol of shim sequences used

### Control Option for Acquisition (L23)

1 default

## Parameters

<b>F1 ACQU</b>			Parameters	<b>NMRPT</b>			Parameters
NUC1	1H			CNST 55	1.000		L23 watersuppr.
PULPROG	npt_zgnopul						
TE	298.000	K	default				
<b>F1 PROC</b>			Parameters				
CY	11.000	cm					

## Experiment Description


This experiment is used to initialize some settings at the beginning of nmrpt. As pre-condition the 1H pulse must be set with edprosol.

The composite experiment consists of the following steps:

- 1) For BB probes the BB channel will be tuned and matched.
- 2) The correct FIELD value will be determined using the experiment NPT\_prep\_fieldsetting\_d and stored into the BSMS.
- 3) Shimming by TopShim. As default 'topshim initial cal' will be used.
- 4) LOCK parameters are optimized using NPT\_prep\_locksettings.
- 5) Watersuppression experiment is executed (including O1 and RG optimization) and evaluated. Evaluation is based on the interpretation of the splitting which is obtained in the water suppression experiment.

## 5.2.141 1H inno (NPT\_1H\_sensitivity\_inno)

**Test Sample:** 0.1% Ethylbenzene (EB) in Chloroform-D  
Z10120, Z100927, Z10121, Z10033, Z10270, Z10718, Z142221  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Bottom: 1H overview spectrum of ethylbenzene processed without line broadening.  
Top left: Expanded region showing the methylene group used for evaluation.

### Control Option for Acquisition (L23)

1- number of cycles of phase and base line correction  
10



## Parameters


F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	1H			SI	131072		
PULPROG	zg			WDW	0		
NS	1			LB	0.000	Hz	
DS	0			PC	1.000		
RG	101.000		no optim.	F1P	0.000	ppm	
O1P	4.000	ppm		F2P	0.000	ppm	
SWH	11904.762	Hz		CY	11.000	cm	
TD	262144						
AQ	11.010	s					
FIDRES	0.091	Hz					
D 1	113.574	s					
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
TE	298.000	K	default				

## Experiment Description

The experiment determines INNO (Integral-to-noise ratio, I/N). Before the INNO determination baseline and phase correction is executed in an iterative manner. The noise is determined over the whole spectrum, excluding signals and spectrum edges. The determination of I/N allows to obtain a measure for the sensitivity of the probe independent of the actual shim. The noise (and integral) determination is executed after scaling the data stored as intergers in 1r with pow(2.0, NC\_proc).

## 5.2.142 <sup>13</sup>C inno (NPT\_13C\_sensitivity\_inno)

**Test Sample:** 40% Dioxane in Benzene-D6 (ASTM)  
Z10163, Z100929, Z10164, Z10035, Z10255, Z10242, Z10724, Z142224  
**Solvent:** C6D6  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Bottom: <sup>13</sup>C overview spectrum of benzene-d6 (no <sup>1</sup>H decoupling) processed without line broadening.  
Top left: Expanded region showing the CH group used for evaluation.

### Control Option for Acquisition (L23)

1- number of cycles of phase and base line correction  
10

## Parameters


F1 ACQU		Parameters		F1 PROC		Parameters	
NUC1	13C			SI	131072		
PULPROG	zg			WDW	0		
NS	1			LB	0.000	Hz	
DS	0			PC	1.400		
RG	101.000		no optim.	F1P	0.000	ppm	
O1P	127.620	ppm		F2P	0.000	ppm	
SWH	11904.762	Hz		CY	11.000	cm	
TD	262144						
AQ	11.010	s					
FIDRES	0.091	Hz					
D 1	828.899	s					
P 1	9.0	us	90deg NUC1				
PLW 1	39.6	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				

## Experiment Description

The experiment determines INNO (Integral-to-noise ratio, I/N). Before the INNO determination baseline and phase correction is executed in an iterative manner. The noise is determined over the whole spectrum, excluding signals and spectrum edges. The determination of I/N allows to obtain a measure for the sensitivity of the probe independent of the actual shim. The noise (and integral) determination is executed after scaling the data stored as intergers in 1r with pow(2.0, NC\_proc).

## 5.2.143 19F inno (NPT\_19F\_sensitivity\_inno)

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Bottom: <sup>19</sup>F overview spectrum of trifluorotoluene (no <sup>1</sup>H decoupling) processed without line broadening.  
Top left: Expanded region showing the CF<sub>3</sub> group used for evaluation.

### Control Option for Acquisition (L23)

1- number of cycles of phase and base line correction  
10

## Parameters


F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	19F			SI	131072		
PULPROG	zg			WDW	0		
NS	1			LB	0.000	Hz	
DS	0			PC	1.000		
RG	101.000		no optim.	F1P	0.000	ppm	
O1P	-62.766	ppm		F2P	0.000	ppm	
SWH	11904.762	Hz		CY	11.000	cm	
TD	262144						
AQ	11.010	s					
FIDRES	0.091	Hz					
D 1	32.090	s					
P 1	9.0	us	90deg NUC1				
PLW 1	25.1	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				

## Experiment Description

The experiment determines INNO (Integral-to-noise ratio, I/N). Before the INNO determination baseline and phase correction is executed in an iterative manner. The noise is determined over the whole spectrum, excluding signals and spectrum edges. The determination of I/N allows to obtain a measure for the sensitivity of the probe independent of the actual shim. The noise (and integral) determination is executed after scaling the data stored as intergers in 1r with pow(2.0, NC\_proc).

## 5.2.144 19F inno on 1H/19F-coil (NPT\_19F\_sensitivity\_inno\_hcoil)

**Test Sample:** 0.05% Trifluorotoluene (TFT, a,a,a-CF<sub>3</sub>C<sub>6</sub>H<sub>5</sub>) in Chloroform-D  
Z10234, Z100937, Z10235, Z10040, Z10728, Z142228  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Bottom: <sup>19</sup>F overview spectrum of trifluorotoluene (no <sup>1</sup>H decoupling) processed without line broadening.  
Top left: Expanded region showing the CF<sub>3</sub> group used for evaluation.

### Control Option for Acquisition (L23)

1- number of cycles of phase and base line correction  
10

## Parameters


F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	19F			SI	131072		
PULPROG	zg			WDW	0		
NS	1			LB	0.000	Hz	
DS	0			PC	1.000		
RG	101.000		no optim.	F1P	0.000	ppm	
O1P	-62.766	ppm		F2P	0.000	ppm	
SWH	11904.762	Hz		CY	11.000	cm	
TD	262144						
AQ	11.010	s					
FIDRES	0.091	Hz					
D 1	32.090	s					
P 1	9.0	us	90deg NUC1				
PLW 1	25.1	W	Pow@90deg(Specs) NUC1				
TE	298.000	K	default				

## Experiment Description

The experiment determines INNO (Integral-to-noise ratio, I/N). Before the INNO determination baseline and phase correction is executed in an iterative manner. The noise is determined over the whole spectrum, excluding signals and spectrum edges. The determination of I/N allows to obtain a measure for the sensitivity of the probe independent of the actual shim. The noise (and integral) determination is executed after scaling the data stored as intergers in 1r with pow(2.0, NC\_proc).

## 5.2.145 31P inno (NPT\_31P\_sensitivity\_inno)

**Test Sample:** 0.0485 M Triphenyl Phosphate (TPP, [C<sub>6</sub>H<sub>5</sub>]<sub>3</sub>PO<sub>4</sub>) in Acetone-D<sub>6</sub>  
Z10201, Z100934, Z10202, Z10038, Z10257, Z10276, Z10722, Z142226  
**Solvent:** Acetone  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Bottom: 31P overview spectrum of triphenyl phosphate (no 1H decoupling) processed without line broadening.

Top left: Expanded region showing the PO<sub>4</sub> group used for evaluation.

### Control Option for Acquisition (L23)

1- number of cycles of phase and base line correction  
10



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1	31P	SI	131072
PULPROG	zg	WDW	0
NS	1	LB	0.000 Hz
DS	0	PC	1.400
RG	101.000	F1P	0.000 ppm
O1P	-17.609 ppm	F2P	0.000 ppm
SWH	11904.762 Hz	CY	11.000 cm
TD	262144		
AQ	11.010 s		
FIDRES	0.091 Hz		
D 1	119.654 s		
P 1	9.0 us		
PLW 1	27.4 W		
TE	298.000 K		
	no optim.		
	90deg NUC1		
	Pow@90deg(Specs) NUC1		
	default		

## Experiment Description

The experiment determines INNO (Integral-to-noise ratio, I/N). Before the INNO determination baseline and phase correction is executed in an iterative manner. The noise is determined over the whole spectrum, excluding signals and spectrum edges. The determination of I/N allows to obtain a measure for the sensitivity of the probe independent of the actual shim. The noise (and integral) determination is executed after scaling the data stored as intergers in 1r with pow(2.0, NC\_proc).

## 5.2.146 1H Z-gradient profile [+] (NPT\_1H\_CMV\_gradientprofile\_pos)


---

**Test Sample:** 5% H<sub>2</sub>O, 0.6mM CuSO<sub>4</sub> in D<sub>2</sub>O  
Z10688

**Solvent:** D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

Proton Z-gradient profile.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 2048	
PULPROG npt_imgegp1d		WDW 0	
NS 16		PH_mod 0	
DS 0		F1P 16.243 ppm	
RG 1.000	no optim.	F2P -3.893 ppm	
O1P 4.700 ppm		CY 11.000 cm	
SWH 2500000.000 Hz		<b>NMRPT</b>	Parameters
TD 1024		CNST 37 19.400 mm	active sample size
AQ 0.000 s			
FIDRES 4882.812 Hz			
D 1 10.000 s			
D 15 0.015 s	Echo time		
D 21 0.000 s	Grad. stab.		
D 27 0.002 s	Dephas. grad.		
P 0 11.0 us	90 degree		
PLW 0 1.1 W	Pow@90deg(Specs) NUC1		
P 1 11.0 us	90deg NUC1		
PLW 1 1.1 W	Pow@90deg(Specs) NUC1		
GPNAM1			
GPNAM2			
GPZ 1 25.000 %			
GPZ 2 -100.000 %			
TE 298.000 K	default		

## Experiment Description

Z-gradient profile is acquired with positive gradient in order to check the basic functionality of gradient. The profile width is determined at 50% of the maximum profile intensity respectively. Together with the length of the RF coil and gyromagnetic ratio of the observed nuclei the gradient strength (G/cm\*A) is calculated. Processing is always executed with magnitude correction (PH\_mod=MC).

## 5.2.147 1H Z-gradient profile [-] (NPT\_1H\_CMV\_gradientprofile\_neg)


---

**Test Sample:** 5% H<sub>2</sub>O, 0.6mM CuSO<sub>4</sub> in D<sub>2</sub>O  
Z10688

**Solvent:** D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Rotation off



### Example Printout

Proton Z-gradient profile.

### Control Option for Acquisition (L23)

1 default

## Parameters


<p><b>F1 ACQU</b></p> <p>NUC1 1H  PULPROG npt_imgegp1d  NS 16  DS 0  RG 1.000  O1P 4.700 ppm  SWH 2500000.000 Hz  TD 1024  AQ 0.000 s  FIDRES 4882.812 Hz  D 1 10.000 s  D 15 0.015 s  D 21 0.000 s  D 27 0.002 s  P 0 11.0 us  PLW 0 1.1 W  P 1 11.0 us  PLW 1 1.1 W  GPNAM1  GPNAM2  GPZ 1 -25.000 %  GPZ 2 100.000 %  TE 298.000 K</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">no optim.</p> <p>Echo time  Grad. stab.  Dephas. grad.  90 degree  Pow@90deg(Specs) NUC1  90deg NUC1  Pow@90deg(Specs) NUC1</p> <p style="text-align: center;">default</p>
<p><b>F1 PROC</b></p> <p>SI 2048  WDW 0  PH_mod 0  F1P 16.243 ppm  F2P -3.893 ppm  CY 11.000 cm  <b>NMRPT</b>  CNST 37 19.400 mm</p>	<p style="text-align: center;">Parameters</p> <p style="text-align: center;">Parameters active sample size</p>

## Experiment Description

Z-gradient profile is acquired with negative gradient in order to check the basic functionality of gradient. The profile width is determined at 50% of the maximum profile intensity respectively. Together with the length of the RF coil and gyromagnetic ratio of the observed nuclei the gradient strength (G/cm\*A) is calculated. Processing is always executed with magnitude correction (PH\_mod=MC).

## 5.2.148 Gradient recovery test for Z-direction [+] (NPT\_1H\_CMV\_gradientrecovery\_pos)

**Test Sample:** 5% H<sub>2</sub>O, 0.6mM CuSO<sub>4</sub> in D<sub>2</sub>O  
 Z10688  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, and signal intensity.

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check

## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H          PARAMODE 1          PULPROG npt_gradrecvd          NS 1          DS 0          RG 1.000          O1P 3.000 ppm          SW 20.485 ppm          TD 16384          AQ 0.999 s          FIDRES 1.001 Hz          D 1 10.000 s          VDLIST npt_gradrecCMR          P 1 11.0 us          PLW 1 1.1 W          GPNAM1 SINE.100          GPZ 1 100.000 %          P 16 1000.000 us          TE 298.000 K</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>no optim.</p> <p>field dependent</p> <p>default 90deg Pulse Pow@90deg(Specs)</p> <p>7.5 A gradient pulse default</p>	<p><b>F2 PROC</b></p> <p>SI 8192          WDW 1          LB 1.000 Hz          PH_mod 1          F1P 5.720 ppm          F2P 5.320 ppm</p> <p><b>F1 ACQU</b></p> <p>NUC1 1H          TD 12</p> <p><b>F1 PROC</b></p> <p>SI 12</p>	<p>Parameters F2</p> <p>pk</p> <p>Parameters F1</p> <p>Parameters F1</p>
---	--	---	--

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

SINO check can be skipped with L23=10

After successful SINO check, the pseudo 2D gradient recovery experiment will be acquired.

Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.


Evaluation consists of the determination of the intensity deviation to the reference experiment (last of the series)

The results are summarized in the text file printed together with spectrum (bottom left).


Requirements: At least time of recovery delay, which defines start point of range where specified intensity must be reached, needs to be specified.

## 5.2.149 Gradient recovery test for Z-direction [-] (NPT\_1H\_CMV\_gradientrecovery\_neg)

**Test Sample:** 5% H<sub>2</sub>O, 0.6mM CuSO<sub>4</sub> in D<sub>2</sub>O  
 Z10688  
**Solvent:** D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Rotation off



Experiment Number	Recovery Delay	Scaled Intensity
1	100.0 us	86.01%
2	200.0 us	85.64%
3	300.0 us	89.47%
4	400.0 us	92.16%
5	500.0 us	94.12%
6	600.0 us	95.64%
7	700.0 us	96.47%
8	800.0 us	97.72%
9	900.0 us	98.47%
10	1.0 ms	99.19%
11	1.1 ms	99.65%
12	1.2 ms	100.00%



### Example Printout

Series of 1D spectra acquired in pseudo-2D mode. The printed segment shows the water signal of the sample. Top left side shows the same spectra in 2D mode. Bottom left side shows table output of the evaluation, containing experiment number, recovery delay, and signal intensity.

### Control Option for Acquisition (L23)

- 1 default
- 10 skip sino check



## Parameters

<p><b>F2 ACQU</b></p> <p>NUC1 1H          PARAMODE 1          PULPROG npt_gradrecvd          NS 1          DS 0          RG 1.000          O1P 3.000 ppm          SW 20.485 ppm          TD 16384          AQ 0.999 s          FIDRES 1.001 Hz          D 1 10.000 s          VDLIST npt_gradrecCMR          P 1 11.0 us          PLW 1 1.1 W          GPNAM1 SINE.100          GPZ 1 -100.000 %          P 16 1000.000 us          TE 298.000 K</p>	<p>Parameters F2</p> <p>Data Dimension</p> <p>no optim.</p> <p>field dependent</p> <p>default 90deg Pulse Pow@90deg(Specs)</p> <p>7.5 A gradient pulse default</p>	<p><b>F2 PROC</b></p> <p>SI 8192          WDW 1          LB 1.000 Hz          PH_mod 1          F1P 5.720 ppm          F2P 5.320 ppm</p> <p><b>F1 ACQU</b></p> <p>NUC1 1H          TD 12</p> <p><b>F1 PROC</b></p> <p>SI 12</p>	<p>Parameters F2</p> <p>pk</p> <p>Parameters F1</p> <p>Parameters F1</p>
--	--	---	--

## Experiment Description

Purpose of this experiment is the measurement of the recovery delay after applying a field gradient. The exact signal offset (O1P) is first determined with a recovery delay of 1 s. The resulting spectrum is stored in PROCNO 2, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved.

SINO check can be skipped with L23=10

After successful SINO check, the pseudo 2D gradient recovery experiment will be acquired.

Processing is accomplished by Fourier transformation in observation domain only (XF2). Phase correction is determined on the last experiment of the series and applied to the series.

Evaluation consists of the determination of the intensity deviation to the reference experiment (last of the series)


The results are summarized in the text file printed together with spectrum (bottom left).

Requirements: At least time of recovery delay, which defines start point of range where specified intensity must be reached, needs to be specified.

## 5.3 Experiments for HR-Probes with Flow Inserts (LC)

## 5.3.1 1H lineshape (NPT\_1H\_LC\_lineshape)

**Test Sample:** 0.5%, 1.0% or 3.0% Chloroform in Acetone-D6  
H7284, H7284-01, H7284-02  
**Solvent:** Acetone  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Flow cell



### Example Printout

Proton line shape spectrum with offset calibration to O1P as overview spectrum. Printing is from +125.0 Hz to -125.0 Hz. The expansion plot to the left shows chloroform signal with higher resolution (+7.0 Hz, -7.0 Hz).

### Control Option for Acquisition (L23)

- 1 default
- 2 write default shimfile, in case of successful evaluation

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	16384			
PULPROG	zg30				WDW	0			
NS	1				LB	0.000	Hz		
DS	0				PC	1.000			
RG	0.250		optim. by RGA		F1P	8.640	ppm		
O1P	7.700	ppm			F2P	7.440	ppm		
SWH	1000.000	Hz			CY	1000.000	cm		
TD	32768				<b>NMRPT</b>			Parameters	
AQ	16.384	s			CNST 50	0.200		Scaling factor for CY	
FIDRES	0.061	Hz							
D 1	9.116	s	AQ+D1=const						
P 1	14.0	us	90deg Pulse						
PLW 1	6.6	W	Pow@90deg(Specs)						
TE	298.000	K	default						

## Experiment Description

Before acquiring the final spectrum, the exact signal position is determined with NS=1. The carrier position is afterwards set to the position optimized O1 = peak frequency [Hz] - (SWH/4)


This procedure allows the application of a strip transformation with the parameters STSR=0, STSI=SI/2 resulting in the signal position on-resonance and in most cases resulting in an optimal automatic phase correction.

Setting L23=2, it is possible to store the standard shimfile provided the evaluation of the experiment is successful. This event takes place during acquisition only. During regular processing of the data no shimfile is stored.

## 5.3.2 loop transfer time determination starting with empty cell (NPT\_1H\_LC\_loopTransferTimeEmptyCell)

---

**Test Sample:** Mixture of solvents Acetonitrile and D2O  
CH3CN\_D2O  
**Solvent:** CH3CN+D2O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Flow cell



### Example Printout

Bottom: Time curve of the integral of the proton Z gradient profile. At time=0 s transfer was started. Top: Proton Z gradient profile, row where the maximum integral/broadest gradient profile was observed.

### Control Option for Acquisition (L23)

- 1 Execution of preparation steps (shimming, lock, tuning/matching)
- 2 Skip preparation steps

## Parameters

<p><b>F1 ACQU</b></p> <p>NUC1 1H          NUC2 off          PULPROG imgegp2d          NS 1          DS 0          RG 0.250          O1P 2.000 ppm          O2P 2.000 ppm          SW 204.851 ppm          TD 1024          AQ 0.006 s          FIDRES 160.092 Hz          D 1 0.250 s          D 20 1.000 s          P 1 14.0 us          PLW 1 6.6 W          TE 298.000 K</p> <p style="text-align: right;">Parameters</p> <p style="text-align: right;">optim. by RGA</p> <p style="text-align: right;">field dependent field dependent</p> <p style="text-align: right;">time per row 90deg NUC1 Pow@90deg(Specs) NUC1 default</p>	<p><b>F1 PROC</b></p> <p>SI 1024          LB 1.000 Hz          CY 11.000 cm</p> <p><b>NMRPT</b></p> <p>CNST 20 0.000          L 27 0          L 28 0</p> <p style="text-align: right;">Parameters</p> <p style="text-align: right;">used flow rate row of UV detection row with maximum integral</p>
--	--

## Experiment Description

Experiment to determine the transfer time from the BPSU Loop starting with an empty flow cell by continuous acquisition of proton gradient profiles into a 2D file. The preparation of the experiment (temperature equilibration, tuning/matching, locking, shimming, and RGA) is executed after BPSU transfer is finished and the flow cell is filled with solvent. After the preparation the user will be guided to empty the flow cell, so that the main NMR experiment starts with an empty flow cell and to execute the transfer. The row at which 'Valve Position/Transfer to Probehead' was executed will be saved in L 27. The flow rate will be saved in CNST 20. After the user reports that BPSU transfer finished or the flow cell is completely filled NMR experiment will be stopped.


The time resolution of the experiment (time per row) depends on parameter D20.

The row at which the flow cell is filled completely will be saved in L 28. If L 28 equals 0 (e.g. first execution of processing) the user will be asked to manually determine the row where flow cell is filled completely. If necessary the user may change parameters L 27, and L 28 in procno 1 before reprocessing the experiment.

## 5.3.3 loop transfer time determination starting with filled cell (NPT\_1H\_LC\_loopTransferTimeFilledCell)

---

**Test Sample:** 5 ug/ul of 1,3,5-Trimethoxybenzene in Acetonitrile/D2O 70/30  
H5798  
**Solvent:** CH3CN+D2O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Flow cell



### Example Printout

Bottom: Time curve of the integral of the Tri-methoxy benzene signal. At time=0 s the BPSU switched to transfer. Top: 1H signal of Tri-methoxy benzene, row where the maximum integral was observed.

### Control Option for Acquisition (L23)

- 1 Optimization of O1 and O2 for solvent suppression
- 2 No optimization of O1 and O2 for solvent suppression

## Parameters

F1 ACQU Parameters				F1 PROC Parameters			
NUC1	1H			SI	8192		
NUC2	1H			LB	2.000	Hz	
PULPROG	lc2prf2			SIGF1	7.200	ppm	
NS	1			SIGF2	5.800	ppm	
DS	0			NOISF1	-2.000	ppm	
RG	0.250		optim. by RGA	NOISF2	-3.000	ppm	
O1P	2.000	ppm		CY	11.000	cm	
O2P	4.700	ppm		<b>NMRPT</b>			Parameters
SW	20.485	ppm		CNST 20	0.000		used flow rate
TD	8192			CNST 21	1.000		time per row
AQ	0.500	s	field dependent	L 27	0		row where BPSU switched to transfer
FIDRES	2.001	Hz	field dependent	L 28	0		row with maximum integral
D 1	0.468	s					
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
PLW 9	0.000020	W	Pow@90deg(10000u)				
PLW 21	0.000005	W	Pow@90deg(20000u)				
TE	298.000	K	default				

## Experiment Description

Experiment to determine the loop transfer time starting with a filled flow cell. The preparation of the experiment is executed after the user started the LC-pump and the flow cell is filled with solvent. The flow rate will be saved in CNST 20. After the preparation the user will be prompted to report the time at which the BPSU switches to transfer. The preflow time must be large enough to encounter preparation of the NMR (temperature equilibration, tuning/matching, locking, shimming, O1- and O2-determination, and RGA). The main NMR experiment starts automatically after these preparation steps. The row at which BPSU switched to transfer will be saved in L 27. After the user reports that the substance has passed the flow cell NMR experiment will be stopped.

The time resolution of the loop transfer time determination (time per row) depends mainly on the parameters NS, TD and D1. The default time per row is around one second.

The time resolution of the experiment (time per row) depends mainly on the parameters NS, TD and D1. The default time per row is around one second.

The row at which the flow cell is filled completely will be saved in L 28. If L 28 equals 0 (e.g. first execution of processing) the user will be asked to manually determine the row where flow cell is filled completely. If necessary the user may change parameters L 27, and L 28 in procno 1 before reprocessing the experiment.


## 5.3.4 1H integrated system performance peak multi-trapping/transfe (NPT\_1H\_LC\_multiTrappingTransfer\_spe)

**Test Sample:** Mixture of 4 PHBA Esters in CH<sub>3</sub>CN/D<sub>2</sub>O separated by HPLC (i.e. only one of these esters is present in the spectrum)  
H5799-SPE, H5799-SPE-5, H5799-SPE-3, H5799-SPE-1.7

**Solvent:** CD<sub>3</sub>CN\_SPE

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Flow cell



### Example Printout

Proton spectrum with double solvent suppression as overview spectrum.

The expansion plot to the left shows the aromatic signals of the PHBA ester of selected chromatographic fraction.

### Control Option for Acquisition (L23)

- 1 Optimization of O1 and O2 for solvent suppression
- 2 No optimization of O1 and O2 for solvent suppression



## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
NUC2	1H				LB	1.000	Hz		
PULPROG	lc1pngpf2				SIGF1	8.700	ppm		
NS	24				SIGF2	6.000	ppm		
DS	4				NOISF1	-1.000	ppm		
RG	0.250			NOISF2	-7.000	ppm			
O1P	2.000	ppm		CY	11.000	cm			
O2P	4.700	ppm							
SW	20.485	ppm							
TD	32768								
AQ	1.999	s	field dependent						
FIDRES	0.500	Hz	field dependent						
D 1	10.000	s							
P 1	14.0	us	90deg Pulse						
PLW 1	6.6	W	Pow@90deg(Specs)						
PLW 9	0.000005	W	Pow@90deg(20000u)						
PLW 21	0.000005	W	Pow@90deg(20000u)						
GPNAM1	SMSQ10.100								
GPNAM2	SMSQ10.100								
GPZ 1	50.000	%							
GPZ 2	-10.000	%							
P 16		us	gradient pulse						
TE	298.000	K	default						

## Experiment Description

Proton Sensitivity test with cw presaturation on channels F1 and F2. A preparation experiment is acquired for referencing and the search of the two biggest solvent signals (O1 and O2 optimization). For baseline correction AU program lcabsf is used.


Integrated system performance is proofed with this test. In advance to this test the mixture of four PHBA esters is separated twice by LC and trapped on the same spe cartrdiges (multi trapping).

The resulting sensitivity should be around twice the sensitivity observed in

NPT\_1H\_LC\_performanceTransfer\_spe\_peak\_A or NPT\_1H\_LC\_performanceTransfer\_spe\_peak\_B experiment.

## 5.3.5 Indirect P90 13C pulse calibration, LC (NPT\_1H\_LC\_p90det\_13c)

**Test Sample:** 0.3%, 1.0% or 3.0% Chloroform in Acetone-D6  
 H7284, H7284-01, H7284-02  
**Solvent:** CDCl3  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Flow cell



### Example Printout

P90 pulse determination using 1D method by varying the decoupler pulse. The result of a CONVTO1D routine shows five experiments from 45 to 135 deg (PROCNO 999).

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 32768	
PARMODE 0	Data Dimension	WDW 1	
PULPROG decp90		LB 0.500 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 0.250	optim. by RGA	ME_mod 2	LPfc
O1P 8.000 ppm		NCOEF 20	
SWH 5000.000 Hz		ABSF1 1000.000 ppm	
TD 16384		ABSF2 -1000.000 ppm	
AQ 1.638 s		F1P 8.340 ppm	
FIDRES 0.610 Hz		F2P 8.190 ppm	
D 1 60.000 s		CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
P 3 9.0 us	90deg NUC1		
PLW 2 42.0 W	Pow@90deg(Specs) NUC2		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase with absf using the corrected O1P. F1P, F2P, ABSF1 and ABSF2 are set symmetrically around the phase corrected Signal. The result is stored in PROCNO 11, the FID is discarded.


The pulse will be determined on the signal(s) between F1P and F2P, i.e the left <sup>13</sup>C satellite. The pulse determination may be erroneous if the slope of the main signal is included in the plot region. If so, please select a smaller plot region by setting F1P and F2P while the experiment is in the preparation queue.

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates based on the specified pulse length a new power value, updates the PROSOL table and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n.

The PROSOL table is then updated with the determined pulse and the used power. Results are stored under PROCNO 999.

## 5.3.6 P90 1H pulse calibration, LC (NPT\_1H\_LC\_p90det\_1h)

**Test Sample:** 2 mM Sucrose in D2O  
 H7285  
**Solvent:** D2O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Flow cell



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 131072	
PARMODE 0	Data Dimension	WDW 1	
PULPROG zg		LB 1.000 Hz	
NS 1		SSB 0.000	
DS 0		PH_mod 1	pk
RG 0.250	optim. by RGA	ME_mod 0	LPfc
O1P 4.700 ppm		NCOEF 0	
SWH 11904.762 Hz		ABSF1 5.600 ppm	
TD 65536		ABSF2 5.100 ppm	
AQ 2.753 s		F1P 5.500 ppm	
FIDRES 0.363 Hz		F2P 5.200 ppm	
D 1 30.000 s		CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase with absf using the corrected O1P. F1P, F2P, ABSF1 and ABSF2 are set symmetrically around the phase corrected Signal. The result is stored in PROCNO 11, the FID is discarded.

The pulse will be determined on the signal(s) between F1P and F2P, i.e the anomeric proton.


If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates based on the specified pulse length a new power value, updates the PROSOL table and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n.

The PROSOL table is then updated with the determined pulse and the used power. Results are stored under PROCNO 999.

## 5.3.7 <sup>1</sup>H system performance, LC (NPT\_1H\_LC\_performance)

---

**Test Sample:** 800 ng of 1,3,5-Trimethoxybenzene in Acetonitrile/D<sub>2</sub>O 50/50 (concentration depends on flowcell size)  
H9630, H9630-01, H9630-02, H9630-06  
**Solvent:** CH<sub>3</sub>CN+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Flow cell



### Example Printout

Proton spectrum with double solvent suppression as overview spectrum.

### Control Option for Acquisition (L23)

- 1 Optimization of O1 and O2 for solvent suppression
- 2 No optimization of O1 and O2 for solvent suppression

## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	1H			SI	32768		
NUC2	1H			LB	1.000	Hz	
PULPROG	lc1pngpf2			SIGF1	3.900	ppm	
NS	16			SIGF2	3.400	ppm	
DS	4			NOISF1	-1.000	ppm	
RG	0.250		optim. by RGA	NOISF2	-7.000	ppm	
O1P	2.000	ppm		CY	11.000	cm	
O2P	4.700	ppm					
SW	20.485	ppm					
TD	32768						
AQ	1.999	s	field dependent				
FIDRES	0.500	Hz	field dependent				
D 1	10.000	s					
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
PLW 9	0.000020	W	Pow@90deg(10000u)				
PLW 21	0.000005	W	Pow@90deg(20000u)				
GPNAM1	SMSQ10.100						
GPNAM2	SMSQ10.100						
GPZ 1	50.000	%					
GPZ 2	-10.000	%					
P 16		us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Proton Sensitivity test with cw presaturation on channels F1 and F2. A preparation experiment is acquired for referencing and the search of the two biggest solvent signals (O1 and O2 optimization). For baseline correction AU program lcabsf is used. Performance of FlowProbe/CryoFit is proofed with this test. A defined sample is injected directly into the flow cell (without LC purification).


## 5.3.8 <sup>1</sup>H integrated system performance stop-flow peak A (NPT\_1H\_LC\_performanceStopFlow\_d2o\_peak\_A)

**Test Sample:** Mixture of 4 PHBA Esters in CH<sub>3</sub>CN/D<sub>2</sub>O separated by HPLC (i.e. only one of these esters is present in the spectrum)  
H5799-D<sub>2</sub>O

**Solvent:** CH<sub>3</sub>CN+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Flow cell



### Example Printout

Proton spectrum with double solvent suppression as overview spectrum.

The expansion plot to the left shows the aromatic signals of the PHBA ester of selected chromatographic fraction.

### Control Option for Acquisition (L23)

- 1 Optimization of O1 and O2 for solvent suppression
- 2 No optimization of O1 and O2 for solvent suppression



## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
NUC2	1H				LB	1.000	Hz		
PULPROG	lc1pngpf2				SIGF1	8.700	ppm		
NS	24				SIGF2	6.000	ppm		
DS	4				NOISF1	-1.000	ppm		
RG	0.250			optim. by RGA	NOISF2	-7.000	ppm		
O1P	2.000	ppm			CY	11.000	cm		
O2P	4.700	ppm							
SW	20.485	ppm							
TD	32768								
AQ	1.999	s	field dependent						
FIDRES	0.500	Hz	field dependent						
D 1	10.000	s							
P 1	14.0	us	90deg Pulse						
PLW 1	6.6	W	Pow@90deg(Specs)						
PLW 9	0.000020	W	Pow@90deg(10000u)						
PLW 21	0.000005	W	Pow@90deg(20000u)						
GPNAM1	SMSQ10.100								
GPNAM2	SMSQ10.100								
GPZ 1	50.000	%							
GPZ 2	-10.000	%							
P 16		us	gradient pulse						
TE	298.000	K	default						

## Experiment Description

Proton Sensitivity test with cw presaturation on channels F1 and F2. A preparation experiment is acquired for referencing and the search of the two biggest solvent signals (O1 and O2 optimization). For baseline correction AU program lcabsf is used.

Integrated system performance is proofed with this test. A mixture of four PHBA esters is separated by LC in stop-flow mode.

The resulting sensitivities observed in NPT\_1H\_LC\_performanceStopFlow\_d2o\_peak\_A and NPT\_1H\_LC\_performanceStopFlow\_d2o\_peak\_B should be equal.


## 5.3.9 <sup>1</sup>H integrated system performance stop-flow peak B (NPT\_1H\_LC\_performanceStopFlow\_d2o\_peak\_B)

**Test Sample:** Mixture of 4 PHBA Esters in CH<sub>3</sub>CN/D<sub>2</sub>O separated by HPLC (i.e. only one of these esters is present in the spectrum)  
H5799-D<sub>2</sub>O

**Solvent:** CH<sub>3</sub>CN+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Flow cell



### Example Printout

Proton spectrum with double solvent suppression as overview spectrum.

The expansion plot to the left shows the aromatic signals of the PHBA ester of selected chromatographic fraction.

### Control Option for Acquisition (L23)

- 1 Optimization of O1 and O2 for solvent suppression
- 2 No optimization of O1 and O2 for solvent suppression

## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	1H			SI	32768		
NUC2	1H			LB	1.000	Hz	
PULPROG	lc1pngpf2			SIGF1	8.700	ppm	
NS	24			SIGF2	6.000	ppm	
DS	4			NOISF1	-1.000	ppm	
RG	0.250		optim. by RGA	NOISF2	-7.000	ppm	
O1P	2.000	ppm		CY	11.000	cm	
O2P	4.700	ppm					
SW	20.485	ppm					
TD	32768						
AQ	1.999	s	field dependent				
FIDRES	0.500	Hz	field dependent				
D 1	10.000	s					
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
PLW 9	0.000020	W	Pow@90deg(10000u)				
PLW 21	0.000005	W	Pow@90deg(20000u)				
GPNAM1	SMSQ10.100						
GPNAM2	SMSQ10.100						
GPZ 1	50.000	%					
GPZ 2	-10.000	%					
P 16		us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Proton Sensitivity test with cw presaturation on channels F1 and F2. A preparation experiment is acquired for referencing and the search of the two biggest solvent signals (O1 and O2 optimization). For baseline correction AU program lcabsf is used.

Integrated system performance is proofed with this test. A mixture of four PHBA esters is separated by LC in stop-flow mode.

The resulting sensitivities observed in NPT\_1H\_LC\_performanceStopFlow\_d2o\_peak\_A and NPT\_1H\_LC\_performanceStopFlow\_d2o\_peak\_B should be equal.


## 5.3.10 <sup>1</sup>H integrated system performance loop-sampling/transfer peak A (NPT\_1H\_LC\_performanceTransfer\_d2o\_peak\_A)

**Test Sample:** Mixture of 4 PHBA Esters in CH<sub>3</sub>CN/D<sub>2</sub>O separated by HPLC (i.e. only one of these esters is present in the spectrum)  
H5799-D<sub>2</sub>O

**Solvent:** CH<sub>3</sub>CN+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Flow cell



### Example Printout

Proton spectrum with double solvent suppression as overview spectrum.

The expansion plot to the left shows the aromatic signals of the PHBA ester of selected chromatographic fraction.

### Control Option for Acquisition (L23)

- 1 Optimization of O1 and O2 for solvent suppression
- 2 No optimization of O1 and O2 for solvent suppression

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
NUC2	1H				LB	1.000	Hz		
PULPROG	lc1pngpf2				SIGF1	8.700	ppm		
NS	24				SIGF2	6.000	ppm		
DS	4				NOISF1	-1.000	ppm		
RG	0.250			optim. by RGA	NOISF2	-7.000	ppm		
O1P	2.000	ppm			CY	11.000	cm		
O2P	4.700	ppm							
SW	20.485	ppm							
TD	32768								
AQ	1.999	s	field dependent						
FIDRES	0.500	Hz	field dependent						
D 1	10.000	s							
P 1	14.0	us	90deg Pulse						
PLW 1	6.6	W	Pow@90deg(Specs)						
PLW 9	0.000020	W	Pow@90deg(10000u)						
PLW 21	0.000005	W	Pow@90deg(20000u)						
GPNAM1	SMSQ10.100								
GPNAM2	SMSQ10.100								
GPZ 1	50.000	%							
GPZ 2	-10.000	%							
P 16		us	gradient pulse						
TE	298.000	K	default						

## Experiment Description

Proton Sensitivity test with cw presaturation on channels F1 and F2. A preparation experiment is acquired for referencing and the search of the two biggest solvent signals (O1 and O2 optimization). For baseline correction AU program lcabsf is used.

Integrated system performance is proofed with this test. In advance to this test the mixture of four PHBA esters is separated by LC and stored in loops.

The resulting sensitivities observed in NPT\_1H\_LC\_performanceTransfer\_d2o\_peak\_A and NPT\_1H\_LC\_performanceTransfer\_d2o\_peak\_B should be equal.


## 5.3.11 <sup>1</sup>H integrated system performance loop-sampling/transfer peak B (NPT\_1H\_LC\_performanceTransfer\_d2o\_peak\_B)

**Test Sample:** Mixture of 4 PHBA Esters in CH<sub>3</sub>CN/D<sub>2</sub>O separated by HPLC (i.e. only one of these esters is present in the spectrum)  
H5799-D2O

**Solvent:** CH<sub>3</sub>CN+D<sub>2</sub>O

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Flow cell



### Example Printout

Proton spectrum with double solvent suppression as overview spectrum.

The expansion plot to the left shows the aromatic signals of the PHBA ester of selected chromatographic fraction.

### Control Option for Acquisition (L23)

- 1 Optimization of O1 and O2 for solvent suppression
- 2 No optimization of O1 and O2 for solvent suppression

## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	1H			SI	32768		
NUC2	1H			LB	1.000	Hz	
PULPROG	lc1pngpf2			SIGF1	8.700	ppm	
NS	24			SIGF2	6.000	ppm	
DS	4			NOISF1	-1.000	ppm	
RG	0.250		optim. by RGA	NOISF2	-7.000	ppm	
O1P	2.000	ppm		CY	11.000	cm	
O2P	4.700	ppm					
SW	20.485	ppm					
TD	32768						
AQ	1.999	s	field dependent				
FIDRES	0.500	Hz	field dependent				
D 1	10.000	s					
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
PLW 9	0.000020	W	Pow@90deg(10000u)				
PLW 21	0.000005	W	Pow@90deg(20000u)				
GPNAM1	SMSQ10.100						
GPNAM2	SMSQ10.100						
GPZ 1	50.000	%					
GPZ 2	-10.000	%					
P 16		us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Proton Sensitivity test with cw presaturation on channels F1 and F2. A preparation experiment is acquired for referencing and the search of the two biggest solvent signals (O1 and O2 optimization). For baseline correction AU program lcabsf is used.

Integrated system performance is proofed with this test. In advance to this test the mixture of four PHBA esters is separated by LC and stored in loops.

The resulting sensitivities observed in NPT\_1H\_LC\_performanceTransfer\_d2o\_peak\_A and NPT\_1H\_LC\_performanceTransfer\_d2o\_peak\_B should be equal.


## 5.3.12 <sup>1</sup>H integrated system performance peak-trapping/transfer peak A (NPT\_1H\_LC\_performanceTransfer\_spe\_peak\_A)

**Test Sample:** Mixture of 4 PHBA Esters in CH<sub>3</sub>CN/D<sub>2</sub>O separated by HPLC (i.e. only one of these esters is present in the spectrum)  
H5799-SPE, H5799-SPE-5, H5799-SPE-3, H5799-SPE-1.7

**Solvent:** CD<sub>3</sub>CN\_SPE

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Flow cell



### Example Printout

Proton spectrum with double solvent suppression as overview spectrum.

The expansion plot to the left shows the aromatic signals of the PHBA ester of selected chromatographic fraction.

### Control Option for Acquisition (L23)

- 1 Optimization of O1 and O2 for solvent suppression
- 2 No optimization of O1 and O2 for solvent suppression



## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	1H			SI	32768		
NUC2	1H			LB	1.000	Hz	
PULPROG	lc1pngpf2			SIGF1	8.700	ppm	
NS	24			SIGF2	6.000	ppm	
DS	4			NOISF1	-1.000	ppm	
RG	0.250		optim. by RGA	NOISF2	-7.000	ppm	
O1P	2.000	ppm		CY	11.000	cm	
O2P	4.700	ppm					
SW	20.485	ppm					
TD	32768						
AQ	1.999	s	field dependent				
FIDRES	0.500	Hz	field dependent				
D 1	10.000	s					
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
PLW 9	0.000005	W	Pow@90deg(20000u)				
PLW 21	0.000005	W	Pow@90deg(20000u)				
GPNAM1	SMSQ10.100						
GPNAM2	SMSQ10.100						
GPZ 1	50.000	%					
GPZ 2	-10.000	%					
P 16		us	gradient pulse				
TE	298.000	K	default				

## Experiment Description

Proton Sensitivity test with cw presaturation on channels F1 and F2. A preparation experiment is acquired for referencing and the search of the two biggest solvent signals (O1 and O2 optimization). For baseline correction AU program lcabsf is used.

Integrated system performance is proofed with this test. In advance to this test the mixture of four PHBA esters is separated by LC and trapped on spe cartrdiges.

The resulting sensitivities observed in NPT\_1H\_LC\_performanceTransfer\_spe\_peak\_A and NPT\_1H\_LC\_performanceTransfer\_spe\_peak\_B should be equal.


## 5.3.13 <sup>1</sup>H integrated system performance peak-trapping/transfer peak B (NPT\_1H\_LC\_performanceTransfer\_spe\_peak\_B)

**Test Sample:** Mixture of 4 PHBA Esters in CH<sub>3</sub>CN/D<sub>2</sub>O separated by HPLC (i.e. only one of these esters is present in the spectrum)  
H5799-SPE, H5799-SPE-5, H5799-SPE-3, H5799-SPE-1.7

**Solvent:** CD<sub>3</sub>CN\_SPE

**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table

**Sample State:** Flow cell



### Example Printout

Proton spectrum with double solvent suppression as overview spectrum.

The expansion plot to the left shows the aromatic signals of the PHBA ester of selected chromatographic fraction.

### Control Option for Acquisition (L23)

- 1 Optimization of O1 and O2 for solvent suppression
- 2 No optimization of O1 and O2 for solvent suppression

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
NUC2	1H				LB	1.000	Hz		
PULPROG	lc1pngpf2				SIGF1	8.700	ppm		
NS	24				SIGF2	6.000	ppm		
DS	4				NOISF1	-1.000	ppm		
RG	0.250			optim. by RGA	NOISF2	-7.000	ppm		
O1P	2.000	ppm			CY	11.000	cm		
O2P	4.700	ppm							
SW	20.485	ppm							
TD	32768								
AQ	1.999	s		field dependent					
FIDRES	0.500	Hz		field dependent					
D 1	10.000	s							
P 1	14.0	us		90deg Pulse					
PLW 1	6.6	W		Pow@90deg(Specs)					
PLW 9	0.000005	W		Pow@90deg(20000u)					
PLW 21	0.000005	W		Pow@90deg(20000u)					
GPNAM1	SMSQ10.100								
GPNAM2	SMSQ10.100								
GPZ 1	50.000	%							
GPZ 2	-10.000	%							
P 16		us		gradient pulse					
TE	298.000	K		default					

## Experiment Description

Proton Sensitivity test with cw presaturation on channels F1 and F2. A preparation experiment is acquired for referencing and the search of the two biggest solvent signals (O1 and O2 optimization). For baseline correction AU program lcabsf is used.


Integrated system performance is proofed with this test. In advance to this test the mixture of four PHBA esters is separated by LC and trapped on spe cartrdiges.

The resulting sensitivities observed in NPT\_1H\_LC\_performanceTransfer\_spe\_peak\_A and NPT\_1H\_LC\_performanceTransfer\_spe\_peak\_B should be equal.

## 5.3.14 1H sensitivity, LC (NPT\_1H\_LC\_sensitivity)

---

**Test Sample:** 2 mM Sucrose in D2O  
H7285  
**Solvent:** D2O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Flow cell



### Example Printout

Bottom: 1H overview spectrum of sucrose.

Top left: Expanded region showing the anomeric proton used for evaluation.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	65536			
PULPROG	zg				LB	1.000	Hz	1.0	
NS	1				SIGF1	3.000	ppm		
DS	0				SIGF2	2.000	ppm		
RG	0.250		optim. by RGA		NOISF1	6.000	ppm		
O1P	4.700	ppm			NOISF2	4.000	ppm		
SW	29.752	ppm			F1P	8.520	ppm		
TD	65536				F2P	0.480	ppm		
AQ	2.753	s	field dependent		CY	100.000	cm		
FIDRES	0.363	Hz	field dependent						
D 1	30.000	s							
P 1	14.0	us	90deg Pulse						
PLW 1	6.6	W	Pow@90deg(Specs)						
TE	298.000	K	default						

## Experiment Description

Proton sensitivity is measured using the sucrose in D2O sample. Processing is using LB. The signal-to-noise is determined using the signal of the anomeric proton. The signal is searched over the range from 5.6 to 5.3 ppm, while the best 200 Hz noise region is determined over the complete spectrum.

## 5.3.15 stop-flow time determination (NPT\_1H\_LC\_stopFlowTime)

**Test Sample:** 5 ug/ul of 1,3,5-Trimethoxybenzene in Acetonitrile/D2O 70/30  
H5798  
**Solvent:** CH3CN+D2O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Flow cell



### Example Printout

Bottom: Time curve of the integral of the Tri-methoxy benzene signal. At time=0 s the UV maximum was detected. Top: 1H signal of Tri-methoxy benzene, row where the maximum integral was observed.

### Control Option for Acquisition (L23)

- 1 Optimization of O1 and O2 for solvent suppression
- 2 No optimization of O1 and O2 for solvent suppression

## Parameters

<b>F1 ACQU</b>			Parameters	<b>F1 PROC</b>			Parameters
NUC1	1H			SI	8192		
NUC2	1H			LB	2.000	Hz	
PULPROG	lc2prf2			SIGF1	7.200	ppm	
NS	1			SIGF2	5.800	ppm	
DS	0			NOISF1	-2.000	ppm	
RG	0.250		optim. by RGA	NOISF2	-3.000	ppm	
O1P	2.000	ppm		CY	11.000	cm	
O2P	4.700	ppm		<b>NMRPT</b>			Parameters
SW	20.485	ppm		CNST 20	0.000		used flow rate
TD	8192			CNST 21	1.000		time per row
AQ	0.500	s	field dependent	L 27	0		row of UV detection
FIDRES	2.001	Hz	field dependent	L 28	0		row with maximum integral
D 1	0.468	s					
P 1	14.0	us	90deg Pulse				
PLW 1	6.6	W	Pow@90deg(Specs)				
PLW 9	0.000020	W	Pow@90deg(10000u)				
PLW 21	0.000005	W	Pow@90deg(20000u)				
TE	298.000	K	default				

## Experiment Description


Experiment to determine the stop-flow time. The preparation of the experiment (temperature equilibration, tuning/matching, locking, shimming, O1- and O2-determination, and RGA) is executed after the user started the LC-pump and the flow cell is filled with solvent. The flow rate will be saved in CNST 20. After the preparation the user will be prompted to inject the sample and to report the time at which the peak appears at the UV detector. The main NMR experiment starts at time of sample injection. The row at which UV maximum was observed will be saved in L 27. After the user reports that the substance has passed the flow cell NMR experiment will be stopped.

The time resolution of the experiment (time per row) depends mainly on the parameters NS, TD and D1. The default time per row is around one second.

The row at which the flow cell is filled completely will be saved in L 28. If L 28 equals 0 (e.g. first execution of processing) the user will be asked to manually determine the row where flow cell is filled completely. If necessary the user may change parameters L 27, and L 28 in procno 1 before reprocessing the experiment.

## 5.3.16 P90 2H pulse calibration, LC (NPT\_prep\_LC\_p90det\_d)

**Test Sample:** 2 mM Sucrose in D2O  
 H7285  
**Solvent:** D2O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Flow cell



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

1 default



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 2H		SI 8192	
PARMODE 0	Data Dimension	WDW 1	
PULPROG zg2h		LB 0.000 Hz	
NS 1		SSB 0.000	
DS 0		PH_mod 1	pk
RG 0.250	optim. by RGA	ME_mod 0	LPfc
O1P 4.700 ppm		NCOEF 0	
SWH 1000.000 Hz		ABSF1 10.000 ppm	
TD 4096		ABSF2 0.000 ppm	
AQ 2.048 s		F1P 5.200 ppm	
FIDRES 0.488 Hz		F2P 4.200 ppm	
D 1 10.000 s		CY 11.000 cm	
P 1 14.0 us	90deg NUC1		
PLW 1 6.6 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

The next step is the determination of the correct phase with absf using the corrected O1P. F1P, F2P, ABSF1 and ABSF2 are set symmetrically around the phase corrected Signal. The result is stored in PROCNO 11, the FID is discarded.

The pulse will be determined on the signal(s) between F1P and F2P.


If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates based on the specified pulse length a new power value, updates the PROSOL table and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n.

The PROSOL table is then updated with the determined pulse and the used power. Results are stored under PROCNO 999.

## 5.3.17 peak transfer volume determination starting with empty cell (NPT\_prep\_LC\_peakTransferVolume)

---

**Test Sample:** Solvent Acetonitrile-D3 or Methanol-D4  
MeOD, CD3CN  
**Solvent:** CD3OD\_SPE or CD3CN\_SPE  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Flow cell



### Example Printout

Bottom: Time curve of the integral of the deuterium Z gradient profile. At volume=0 ul transfer was started. Top: Deuterium Z gradient profile, row where the maximum integral/broadest gradient profile was observed.

### Control Option for Acquisition (L23)

- 1 Execution of preparation steps (shimming, lock, tuning/matching)
- 2 Skip preparation steps

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	2H				SI	1024			
NUC2	off				LB	1.000	Hz		
PULPROG	imgegp2d2h				CY	11.000	cm		
NS	1				<b>NMRPT</b>				Parameters
DS	0				CNST 20	0.000			used flow rate
RG	0.250		optim. by RGA		L 27	0			row where BPSU switched to
O1P	2.000	ppm			L 28	0			transfer
O2P	2.000	ppm							row with maximum integral
SW	203.508	ppm							
TD	1024								
AQ	0.041	s	field dependent						
FIDRES	24.414	Hz	field dependent						
D 1	0.750	s							
D 20	1.000	s	time per row						
P 1	180.0	us	90deg NUC1						
PLW 1	6.0	W	Pow@90deg(Specs) NUC1						
TE	298.000	K	default						

## Experiment Description

Experiment to determine the peak transfer volume from SPE to flow cell by continuous acquisition of deuterium gradient profiles into a 2D file. The preparation of the experiment (temperature equilibration, tuning/matching, shimming, and RGA) is executed after HyStar transfer is finished and the flow cell is filled with solvent. After the preparation the user will be prompted to start another transfer. The transfer step includes probe drying, so that the main NMR experiment starts with an empty flow cell. The row at which HyStar shows TRANSFER will be saved in L 27. The flow rate will be saved in CNST 20. After the user reports that HyStar transfer finished or the flow cell is completely filled NMR experiment will be stopped. The time resolution of the experiment (time per row) depends on parameter D20. The row at which the flow cell is filled completely will be saved in L 28. If L 28 equals 0 (e.g. first execution of processing) the user will be asked to manually determine the row where flow cell is filled completely. If necessary the user may change parameters L 27, and L 28 in procno 1 before reprocessing the experiment.

## 5.4 Experiments for Magic Angle Spinning Probes (Solids)

In *NMRPT* the default rotational spinning frequency of the rotors depends on sample, experiments and base frequency of the magnet. The default values for each experiment are provided in the experiment description in this chapter.

It is recommended to adjust the FIELD value of the spectrometer using KBr or Adamantane sample. If using KBr for field adjustment, set the 79Br signal to 59.70 ppm. If using Adamantane set the 13C methylene (low field) signal to 38.48 ppm. With this FIELD value calibration done, carrier frequency optimisation (O1P and O2P) during experiment setup should not be necessary.

## 5.4.1 <sup>13</sup>C B1 homogeneity, MAS (NPT\_13C\_MAS\_b1homogeneity\_13c)


---

**Test Sample:** Adamantane  
Z151221, Z151271, Z151261, Z151251, Z151231, Z151241, Z151211, Z163274,  
Z183104

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a region around each main signal is shown in the figure.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 8192	
NUC2 1H		LB 0.000	Hz
PARMODE 1	Data Dimension	PH_mod 1	pk
PULPROG npt_p1nuthpdec2d		CY 11.000	cm
NS 1			
DS 0			
RG 101.000	no optim.		
O1P 34.000 ppm			
O2P 2.460 ppm			
SWH 10000.000 Hz			
TD 19998			
AQ 1.000 s			
FIDRES 1.000 Hz			
D 1 15.000 s			
CPDPRG2 cw	decoupl. sequence		
P 1 4.0 us	90deg NUC1		
PCPD2 17.0 us	PCPD NUC2		
PLW 1 125 W	Pow@90deg NUC1		
PLW 12 0.2 W	B1(NUC2) = MASR/4		
TE 298.000 K	default		

## Experiment Description

The B1 homogeneity measurement is acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range reaches from 15 to approx. 810 degrees. The B1-homogeneities based on integrals are computed from the magnitude of the second complex FID data point in PROCNO=2.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized. O2P is determined for decoupling by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, and 3 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

The result of CONVTO1D is stored in PROCNO=999. This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Requirements: The corresponding 90° pulse determination experiment (NPT\_X\_MAS\_p90det\_YZ) is mandatory for B1 homogeneity measurement.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	111000	67000	42000	35000	24000	15000	7000
400	111000	67000	42000	35000	24000	15000	7000
500	111000	67000	42000	35000	24000	15000	7000
600	111000	67000	42000	35000	24000	15000	7000
700	111000	67000	42000	35000	24000	15000	7000
750	111000	67000	42000	35000	24000	15000	7000
800	111000	67000	42000	35000	24000	15000	7000
850	111000	67000	42000	35000	24000	15000	7000
900	111000	67000	42000	35000	24000	15000	7000
950	111000	67000	42000	35000	24000	15000	7000
1000	111000	67000	42000	35000	24000	15000	7000


## 5.4.2 Double CP 1H-15N-13C, MAS (NPT\_13C\_MAS\_double\_cp1h15n\_13c)

**Test Sample:** Alpha-crystalline 2-13C, 15N Glycine  
Z151223, Z151273, Z151263, Z151253, Z151233, Z151243, Z151213, Z163276,  
Z183106

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Top: 13C spectrum with 1H 15N double cross polarization. Bottom: 13C spectrum with 1H cross polarization.

### Control Option for Acquisition (L23)

- 1 default, automatic parameter optimization
- 11 Several manual user interaction steps for parameter optimization

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 32768	
NUC2 1H		LB 0.000	Hz
NUC3 15N		PH_mod 1	pk
PARMODE 0	Data Dimension	ABSF1 96.013	ppm
PULPROG doubcp		ABSF2 -8.769	ppm
NS 16		F1P 91.249	ppm
DS 0		F2P -4.011	ppm
RG 101.000	no optim.	CY 11.000	cm
O1P 110.000	ppm		
O2P 6.200	ppm		
O3P 35.000	ppm		
SW 299.337	ppm		
TD 3012			
AQ 0.050	s		
FIDRES 20.000	Hz		
D 1 5.000	s		
CNST 9 110.000	ppm		13C carrier pos.
CNST 10 20.000	ppm		13C carrier pos. (optimized)
CNST 53 5.50	us		90 deg. 15N pulse
CNST 54 42	W		max. Pow@90deg 1H
P 1 8.33	us		90deg 13C
P 3 2.27	us		1H max. dec. field
P 15 4000.0	us		1H 15N contact time (optimized)
P 16 10000.0	us		15N 13C contact time (optimized)
PCPD 2 4.2	us		PCPD2 1H
PLW 1 20.5	W		Pow@90deg 13C
PLW 3 162.0	W		Pow@90deg 15N
PLW 5 140.9	W		Pow@90deg 15N
PLW 11 20.5	W		Pow@90degCP(Specs) 13C
PLW 12 230.0	W		Pow@90deg 1H
PLW 13 230.0	W		Pow@90deg 1H (optimized)
SPW 0 100.0	W		Pow 1H contact (optimized)
SPW 1 27.5	W		Pow 13C contact (optimized)
TE 298.000	K		default
SPNAM 0 ramp.100			
SPNAM 1 tacn80			
CPDPRG 2 spinal64			

## Experiment Description

13C sensitivity experiment with 1H 15n double cross polarization including automatic parameter optimization.

First step: parameter optimization for 1H 15N cross polarization in a derived 15N 1H data set (expno=1, NS=4).

Second step: 1H 13C cross polarization are optimized in a derived 13C 1H data set (expno=2, NS=4).

Third step: acquisition of the 1H 13C cross polarization experiment as reference for evaluating the efficiency of the double cross polarization (NS=16).

Fourth step: parameter optimization of the 15N 13C cross polarization using the double cross polarization experiment (NS=4, optimized parameters from first step).

Fifth step: acquisition of the 1H 15N 13C double cross polarization experiment (NS=16).

The transfer efficiency in percent is determined as ratio of sino of 1H 15N 13C double cross polarization and 1H 13C cross polarization (derived data set, expno=2).

For MASR < 20000 the start values of SPW1 and PLW5 are calculated for B1(NUC1) = (5/2 \* MASR) and B1(NUC3) = (7/2 \* MASR) respectively.

For MASR >= 20000 the start values of SPW1 and PLW5 are calculated for B1(NUC1) = (1/3 \* MASR) and B1(NUC3) = (2/3 \* MASR) respectively.

Start value of SPW0 is calculated for absolute value of B1(NUC2) = (B1(NUC1) + sideBandCondition \* MASR) \* rampFactor.

For MASR < 60000 sideBandCondition = 1, for MASR >=60000 sideBandCondition = -1 is used.

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	111000	12000	12000	12000	12000	12000	7000
400	111000	12000	12000	12000	12000	12000	7000
500	111000	12000	12000	12000	12000	12000	7000
600	111000	11000	11000	11000	11000	11000	7000
700	111000	12000	12000	12000	12000	12000	7000
750	111000	12000	12000	12000	12000	12000	7000
800	111000	12000	12000	12000	12000	12000	7000
850	111000	12000	12000	12000	12000	12000	7000
900	111000	12000	12000	12000	12000	12000	7000
950	111000	12000	12000	12000	12000	12000	7000
1000	111000	12000	12000	12000	12000	12000	7000






## 5.4.3 Optimization of $^{13}\text{C}$ frequency (NPT\_13C\_MAS\_fieldsetting\_dec1h)

---

**Test Sample:** Adamantane  
Z151221, Z151271, Z151261, Z151251, Z151231, Z151241, Z151211, Z163274,  
Z183104  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

$^{13}\text{C}$  spectrum with  $^1\text{H}$  CW decoupling after field optimization

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO and line width check on PROCNO 2

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 8192	
NUC2 1H		WDW 0	
PARMODE 0	Data Dimension	PH_mod 0	
PULPROG hpdec		F1P 0.000	ppm
NS 4		F2P 0.000	ppm
DS 0		CY 100.000	cm
RG 101.000	no optim.		
O1P 34.000	ppm		
O2P 2.460	ppm		
SWH 10000.000	Hz		
TD 4000			
AQ 0.200	s		
FIDRES 5.000	Hz		
D 1 15.000	s		
P 1 4.0	us		
CPDPRG2 cw	90deg NUC1 decoupl. sequence		
PLW 1 52	W		
PLW 12 0.05	W		
TE 298.000	K		
	B1(NUC2) = MASR/4 default		

## Experiment Description

The experimental procedure includes two <sup>13</sup>C acquisitions with constant O1 at two known FIELD positions. Using 38.46 ppm as <sup>13</sup>C chemical shift of the CH<sub>2</sub> groups Adamantane the correct FIELD value can be calculated from these measurements.

O2P is determined for each FIELD position by acquisition and processing of a spectrum of NUC 2 in a derived data set.

After acquisition of the first <sup>13</sup>C spectrum SINO and line width determination will be executed. The experiment will be aborted, if a minimal SINO of 10 and a minimal line width of 15 Hz are not achieved. SINO and line width check can be skipped with L23=2.

Before acquisition of the final nmr spectrum NMRPT stores the resulting FIELD value in the BSMS.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	111000	67000	42000	35000	24000	15000	7000
400	111000	67000	42000	35000	24000	15000	7000
500	111000	67000	42000	35000	24000	15000	7000
600	111000	67000	42000	35000	24000	15000	7000
700	111000	67000	42000	35000	24000	15000	7000
750	111000	67000	42000	35000	24000	15000	7000
800	111000	67000	42000	35000	24000	15000	7000
850	111000	67000	42000	35000	24000	15000	7000
900	111000	67000	42000	35000	24000	15000	7000
950	111000	67000	42000	35000	24000	15000	7000
1000	111000	67000	42000	35000	24000	15000	7000

## 5.4.4 P90 13C pulse calibration, MAS (NPT\_13C\_MAS\_p90det\_13c)

**Test Sample:** Adamantane  
 Z151221, Z151271, Z151261, Z151251, Z151231, Z151241, Z151211, Z163274,  
 Z183104  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination.
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
 +xxx

## Parameters

<p><b>F1 ACQU</b></p> <p>Parameters</p> <p>NUC1 13C</p> <p>NUC2 1H</p> <p>PARMODE 0 Data Dimension</p> <p>PULPROG hpdec</p> <p>NS 4</p> <p>DS 0</p> <p>RG 101.000 no optim.</p> <p>O1P 34.000 ppm</p> <p>O2P 2.460 ppm</p> <p>SWH 10000.000 Hz</p> <p>TD 4000</p> <p>AQ 0.200 s</p> <p>FIDRES 5.000 Hz</p> <p>D 1 15.000 s</p> <p>P 1 4.0 us 90deg NUC1</p> <p>CPDPRG2 cw decoupl. sequence</p> <p>PLW 1 125 W Pow@90deg(Specs) NUC1</p> <p>PLW 12 0.2 W B1(NUC2) = MASR/4</p> <p>TE 298.000 K default</p>	<p><b>F1 PROC</b></p> <p>Parameters</p> <p>SI 8192</p> <p>LB 0.000 Hz</p> <p>PH_mod 1 pk</p>
--	--

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

O2P is determined for decoupling by acquisition and processing of a spectrum of NUC 2 in a derived data set.

Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated with the determined pulse and the power used. Results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	111000	67000	42000	35000	24000	15000	7000
400	111000	67000	42000	35000	24000	15000	7000
500	111000	67000	42000	35000	24000	15000	7000
600	111000	67000	42000	35000	24000	15000	7000
700	111000	67000	42000	35000	24000	15000	7000
750	111000	67000	42000	35000	24000	15000	7000
800	111000	67000	42000	35000	24000	15000	7000
850	111000	67000	42000	35000	24000	15000	7000
900	111000	67000	42000	35000	24000	15000	7000
950	111000	67000	42000	35000	24000	15000	7000
1000	111000	67000	42000	35000	24000	15000	7000


## 5.4.5 P90 13C 19F-13C CP pulse calibration, MAS (NPT\_13C\_MAS\_p90det\_cp19f\_13c)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

P90 CP pulse determination using 1D method by varying the excitation pulse P1. The result of a CONVTO1D routine shows six experiments P1\*0.5 to P1\*1.5 (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
+XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 4096	
NUC2 19F		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp90			
NS 4			
DS 0			
RG 101.000	no optim.		
O1P 140.000	ppm		
O2P -74.000	ppm		
SWH 30120.482	Hz		
TD 3012			
AQ 0.050	s		
FIDRES 20.000	Hz		
D 1 2.000	s		
P 1 4.0	us	90deg NUC1	
P 3 3.5	us	max. dec. field	
P 15 5000.0	us	HH NUC2-NUC1	
PCPD 2 6.8	us	PCPD2 NUC2	
PLW 1 125.0	W	Pow@90deg NUC1	
PLW 11 125.0	W	Pow@90degCP(Specs)	
PLW 12 49.0	W	NUC1	
SPW 0 49.0	W	Pow@90deg NUC2	
TE 298.000	K	Pow@HHshaped NUC2	
SPNAM 0 ramp50100.100		default	
CPDPRG 2 spinal64			

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal (maximum intensity) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set. Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + 2 * MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition 1, -1, and -2 until power is in limits.

The selection of a certain side band condition can be forced by setting CNST 50 to 2, 1, -1, or -2.

The results of the parameter optimization for PLW11 are stored under PROCNO 100.

Experiment results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

The PROSOL table will not be updated during this experiment.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000






## 5.4.6 P90 13C 19F-13C CP pulse calibration using H-coil, MAS (NPT\_13C\_MAS\_p90det\_cp19f\_13c\_hcoil)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

P90 CP pulse determination using 1D method by varying the excitation pulse P1. The result of a CONVTO1D routine shows six experiments P1\*0.5 to P1\*1.5 (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
+XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 4096	
NUC2 19F		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp90			
NS 4			
DS 0			
RG 101.000	no optim.		
O1P 140.000 ppm			
O2P -74.000 ppm			
SWH 30120.482 Hz			
TD 3012			
AQ 0.050 s			
FIDRES 20.000 Hz			
D 1 2.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 5000.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 125.0 W	Pow@90deg NUC1		
PLW 11 125.0 W	Pow@90degCP(Specs)		
PLW 12 49.0 W	NUC1		
SPW 0 49.0 W	Pow@90deg NUC2		
TE 298.000 K	Pow@HHshaped NUC2		
SPNAM 0 ramp50100.100	default		
CPDPRG 2 spinal64			

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal (maximum intensity) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set. Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + 2 * MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition 1, -1, and -2 until power is in limits.

The selection of a certain side band condition can be forced by setting CNST 50 to 2, 1, -1, or -2.

The results of the parameter optimization for PLW11 are stored under PROCNO 100.

Experiment results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

The PROSOL table will not be updated during this experiment.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000




## 5.4.7 P90 13C 1H-13C CP pulse calibration, MAS (NPT\_13C\_MAS\_p90det\_cp1h\_13c)

**Test Sample:** Alpha-crystalline Glycine (not isotope-labeled or 2-13C, 15N Glycine)  
Z151222, Z151272, Z151262, Z151252, Z151232, Z151242, Z151212, Z183105,  
Z151223, Z151273, Z151263, Z151253, Z151233, Z151243, Z151213, Z163276,  
Z183106

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



```
Report of AD program poptau
FIP=50.037ppm F2P=40.037ppm
PROCNO=999
Linear optimization of P1M1 in 6 steps,
starting at 37.04349451 endval= 81.539703349
OPTIMIZE ZERO
VARIABLE= L23
Experiment P1M1 Maximum point Minimum point Integral
1 37.0435 27849745 -22544608 2327377.536
2 45.3507 20495046 -18423053 1958866.831
3 54.8540 130794500 -35847005 9061262.938
4 63.7492 81714707 -24061449 5451099.338
5 72.4445 45938775 -20712092 3208236.944
6 81.5397 51806392 -78741408 -2447711.451
poptau For P1M1 finished.
ZERO at experiment 5.374859; P1M1 = 75.975266
```

### Example Printout

P90 CP pulse determination using 1D method by varying the excitation pulse P1. The result of a CONVTO1D routine shows six experiments P1\*0.5 to P1\*1.5 (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
+XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 4096	
NUC2 1H		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp90			
NS 4			
DS 0			
RG 101.000	no optim.		
O1P 43.000 ppm			
O2P 6.200 ppm			
SWH 30120.482 Hz			
TD 3012			
AQ 0.050 s			
FIDRES 20.000 Hz			
D 1 5.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 2000.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 125.0 W	Pow@90deg NUC1		
PLW 11 125.0 W	Pow@90degCP(Specs)		
	NUC1		
PLW 12 54.0 W	Pow@90deg NUC2		
SPW 0 54.0 W	Pow@HHshaped NUC2		
TE 298.000 K	default		
SPNAM 0 ramp50100.100			
CPDPRG 2 spinal64			

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal (maximum intensity) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set. Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + 2 * MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition 1, -1, and -2 until power is in limits.

The selection of a certain side band condition can be forced by setting CNST 50 to 2, 1, -1, or -2.

The results of the parameter optimization for PLW11 are stored under PROCNO 100.

Experiment results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

The PROSOL table will be updated with the determined pulse and the power used.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000





## 5.4.8 CP 1H-13C parameter optimization, MAS (NPT\_13C\_MAS\_paropt\_cp1h\_13c)


---

**Test Sample:** Alpha-crystalline 2-13C, 15N Glycine  
Z151223, Z151273, Z151263, Z151253, Z151233, Z151243, Z151213, Z163276,  
Z183106

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 32768	
NUC2 1H		TDeff 2048	
PARMODE 0	Data Dimension	LB 0.000 Hz	
PULPROG cp	set according specs	PH_mod 1	pk
NS 4		ABSF1 1000.000 ppm	
DS 0		ABSF2 -1000.000 ppm	
RG 101.000	no optim.	F1P 0.000 ppm	
O1P 110.000 ppm		F2P 0.000 ppm	
O2P 6.200 ppm		CY 11.000 cm	
SW 299.337 ppm			
TD 3012			
AQ 0.050 s			
FIDRES 20.000 Hz			
D 1 5.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 2000.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 125.0 W	Pow@90deg NUC1		
PLW 11 125.0 W	Pow@90degCP(Specs)		
PLW 12 54.0 W	NUC1		
SPW 0 54.0 W	Pow@90deg NUC2		
TE 298.000 K	Pow@HHshaped NUC2		
SPNAM 0 ramp50100.100	default		
CPDPRG 2 spinal64			

## Experiment Description

Experiment for parameter optimization (PLW12 and SPW0) for mas experiments with cp. Signal to noise value is calculated using noise range of 20 ppm. All ranges are provided in plot title.

L23 dependent offset optimization: Prior to the main acquisition O1P (PROCNO=10) will be optimized. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, and 3 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

Start value of SPW0 is calculated for B1(NUC2) = (B1(NUC1) + MASR) \* rampFactor.

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000


## 5.4.9 P90 13C 19F-13C CP shortest pulse calibration, MAS (NPT\_13C\_MAS\_shortestPulse\_cp19f\_13c)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
 Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Shortest pulse determination with CP using 1D method by varying the excitation pulse P1. The result of a CONVTO1D routine shows 21 experiments P1\*0.5 to P1\*1.5(PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
 +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 4096	
NUC2 19F		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp90			
NS 4			
DS 0			
RG 101.000	no optim.		
O1P 140.000 ppm			
O2P -74.000 ppm			
SWH 30120.482 Hz			
TD 3012			
AQ 0.050 s			
FIDRES 20.000 Hz			
D 1 2.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 5000.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 125.0 W	Pow@90deg NUC1		
PLW 11 125.0 W	Pow@90degCP(Specs)		
PLW 12 49.0 W	NUC1		
SPW 0 49.0 W	Pow@90deg NUC2		
TE 298.000 K	Pow@HHshaped NUC2		
SPNAM 0 ramp50100.100	default		
CPDPRG 2 spinal64			

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal (maximum intensity) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set. Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + 2 * MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition 1, -1, and -2 until power is in limits.

The selection of a certain side band condition can be forced by setting CNST 50 to 2, 1, -1, or -2.

The results of the parameter optimization for PLW11 are stored under PROCNO 100.

Experiment results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

The PROSOL table will not be updated during this experiment.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000




## 5.4.10 P90 13C 19F-13C CP shortest pulse calibration using H-coil, MAS (NPT\_13C\_MAS\_shortestPulse\_cp19f13c\_hcoil)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
 Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Shortest pulse determination with CP using 1D method by varying the excitation pulse P1. The result of a CONVTO1D routine shows 21 experiments P1\*0.5 to P1\*1.5(PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
 +XXX



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 4096	
NUC2 19F		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp90			
NS 4			
DS 0			
RG 101.000	no optim.		
O1P 140.000 ppm			
O2P -74.000 ppm			
SWH 30120.482 Hz			
TD 3012			
AQ 0.050 s			
FIDRES 20.000 Hz			
D 1 2.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 5000.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 125.0 W	Pow@90deg NUC1		
PLW 11 125.0 W	Pow@90degCP(Specs)		
PLW 12 49.0 W	NUC1		
SPW 0 49.0 W	Pow@90deg NUC2		
TE 298.000 K	Pow@HHshaped NUC2		
SPNAM 0 ramp50100.100	default		
CPDPRG 2 spinal64			

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal (maximum intensity) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set. Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

$SPW0$  is calculated for  $B1(NUC2) = (B1(NUC1) + 2 * MASR) * rampFactor$ .

If the resulting power  $SPW0$  exceeds the power limits of the probe,  $SPW0$  will be recalculated for side band condition 1, -1, and -2 until power is in limits.

The selection of a certain side band condition can be forced by setting CNST 50 to 2, 1, -1, or -2.

The results of the parameter optimization for PLW11 are stored under PROCNO 100.

Experiment results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

The PROSOL table will not be updated during this experiment.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000




## 5.4.11 P90 13C 1H-13C CP shortest pulse calibration, MAS (NPT\_13C\_MAS\_shortestPulse\_cp1h\_13c)

**Test Sample:** Alpha-crystalline Glycine (not isotope-labeled or 2-13C, 15N Glycine)  
 Z151222, Z151272, Z151262, Z151252, Z151232, Z151242, Z151212, Z183105,  
 Z151223, Z151273, Z151263, Z151253, Z151233, Z151243, Z151213, Z163276,  
 Z183106

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Shortest pulse determination with CP using 1D method by varying the excitation pulse P1. The result of a CONVTO1D routine shows 21 experiments P1\*0.5 to P1\*1.5(PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
 +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 4096	
NUC2 1H		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp90			
NS 4			
DS 0			
RG 101.000	no optim.		
O1P 43.000 ppm			
O2P 6.200 ppm			
SWH 30120.482 Hz			
TD 3012			
AQ 0.050 s			
FIDRES 20.000 Hz			
D 1 5.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 2000.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 125.0 W	Pow@90deg NUC1		
PLW 11 125.0 W	Pow@90degCP(Specs)		
	NUC1		
PLW 12 54.0 W	Pow@90deg NUC2		
SPW 0 54.0 W	Pow@HHshaped NUC2		
TE 298.000 K	default		
SPNAM 0 ramp50100.100			
CPDPRG 2 spinal64			

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal (maximum intensity) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set. Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + 2 * MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition 1, -1, and -2 until power is in limits.

The selection of a certain side band condition can be forced by setting CNST 50 to 2, 1, -1, or -2.

The results of the parameter optimization for PLW11 are stored under PROCNO 100.

Experiment results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

The PROSOL table will not be updated during this experiment.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency


	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000



## 5.4.12 <sup>13</sup>C sensitivity, MAS (NPT\_13C\_MAS\_sino\_13c)

---

**Test Sample:** Adamantane  
Z151221, Z151271, Z151261, Z151251, Z151231, Z151241, Z151211, Z163274,  
Z183104  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity and line width determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 32768	
NUC2 1H		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG hpdec		ABSF1 1000.000	ppm
NS 1		ABSF2 -1000.000	ppm
DS 0		F1P 0.000	ppm
RG 101.000	no optim.	F2P 0.000	ppm
O1P 34.000		CY 11.000	cm
O2P 2.460			
SWH 10000.000			
TD 19998			
AQ 1.000			
FIDRES 1.000			
D 1 15.000			
P 1 4.0	us		
CPDPRG2 cw	90deg NUC1 decoupl. sequence		
PLW 1 125	W	Pow@90deg(Specs) NUC1	
PLW 12 0.2	W	B1(NUC2) = MASR/4	
TE 298.000	K	default	

## Experiment Description

Sensitivity determination on solid probes including line width determination. Signal to noise value is calculated using noise range of 20 ppm set apart possible MAS spinning side bands. All ranges are provided in plot title.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized. O2P is determined for decoupling by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, and 3 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	111000	67000	42000	35000	24000	15000	7000
400	111000	67000	42000	35000	24000	15000	7000
500	111000	67000	42000	35000	24000	15000	7000
600	111000	67000	42000	35000	24000	15000	7000
700	111000	67000	42000	35000	24000	15000	7000
750	111000	67000	42000	35000	24000	15000	7000
800	111000	67000	42000	35000	24000	15000	7000
850	111000	67000	42000	35000	24000	15000	7000
900	111000	67000	42000	35000	24000	15000	7000
950	111000	67000	42000	35000	24000	15000	7000
1000	111000	67000	42000	35000	24000	15000	7000

## 5.4.13 CP 19F-13C sensitivity, MAS (NPT\_13C\_MAS\_sino\_cp19f\_13c)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 Same as 1 but with manual user interaction to set SPW0 and PLW12
- 12 Same as 2 but with manual user interaction to set SPW0 and PLW12
- 13 Same as 3 but with manual user interaction to set SPW0 and PLW12
- 14 Same as 4 but with manual user interaction to set SPW0 and PLW12
- 15 Same as 5 but with manual user interaction to set SPW0 and PLW12
- 21 Same as 1 but with forced automatic optimization of SPW0 and PLW12
- 22 Same as 2 but with forced automatic optimization of SPW0 and PLW12
- 23 Same as 3 but with forced automatic optimization of SPW0 and PLW12
- 24 Same as 4 but with forced automatic optimization of SPW0 and PLW12
- 25 Same as 5 but with forced automatic optimization of SPW0 and PLW12

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 32768	
NUC2 19F		LB 0.000 Hz	
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp		ABSF1 1000.000 ppm	
NS 64	set according specs	ABSF2 -1000.000 ppm	
DS 0		F1P 0.000 ppm	
RG 101.000	no optim.	F2P 0.000 ppm	
O1P 140.000 ppm		CY 11.000 cm	
O2P -74.000 ppm			
SW 299.328 ppm			
TD 3012			
AQ 0.050 s			
FIDRES 20.000 Hz			
D 1 5.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 5000.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 125.0 W	Pow@90deg NUC1		
PLW 11 125.0 W	Pow@90degCP(Specs)		
	NUC1		
PLW 12 49.0 W	Pow@90deg NUC2		
SPW 0 49.0 W	Pow@HHshaped NUC2		
TE 298.000 K	default		
SPNAM 0 ramp50100.100			
CPDPRG 2 spinal64			

## Experiment Description

Sensitivity determination on solid probes including steps for optimization of PLW12 and SPW0. Signal to noise value is calculated using noise range of 20 ppm. All ranges are provided in plot title.

If number of scans is specified, NS will be set accordingly during acquisition.

L23 dependent offset optimization: Prior to the main acquisition O1P (PROCNO=10) will be optimized.

O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, and 3 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

For L23 = 1, 2, 3, 4, and 5 the values of SPW0 and PLW12 will be taken from an already acquired sine experiment.

For L23 = 11, 12, 13, 14 and 15 or if the sine experiment has not been measured a manual user interaction part to set or optimize SPW0, and PLW12 is included.

For L23 = 21, 22, 23, 24 and 25 the optimization of SPW0 and PLW12 will be enforced.

SPW0 is calculated for B1(NUC2) = (B1(NUC1) + 2\*MASR) \* rampFactor.

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition 1, -1, and -2 until power is in limits.

The selection of a certain side band condition can be forced by setting CNST 50 to 2, 1, -1, or -2.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

## Spinrate


MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000



## 5.4.14 CP 19F-13C sensitivity using H-coil, MAS (NPT\_13C\_MAS\_sino\_cp19f\_13c\_hcoil)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 Same as 1 but with manual user interaction to set SPW0 and PLW12
- 12 Same as 2 but with manual user interaction to set SPW0 and PLW12
- 13 Same as 3 but with manual user interaction to set SPW0 and PLW12
- 14 Same as 4 but with manual user interaction to set SPW0 and PLW12
- 15 Same as 5 but with manual user interaction to set SPW0 and PLW12
- 21 Same as 1 but with forced automatic optimization of SPW0 and PLW12
- 22 Same as 2 but with forced automatic optimization of SPW0 and PLW12
- 23 Same as 3 but with forced automatic optimization of SPW0 and PLW12
- 24 Same as 4 but with forced automatic optimization of SPW0 and PLW12
- 25 Same as 5 but with forced automatic optimization of SPW0 and PLW12

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 32768	
NUC2 19F		LB 0.000 Hz	
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp		ABSF1 1000.000 ppm	
NS 64	set according specs	ABSF2 -1000.000 ppm	
DS 0		F1P 0.000 ppm	
RG 101.000	no optim.	F2P 0.000 ppm	
O1P 140.000 ppm		CY 11.000 cm	
O2P -74.000 ppm			
SW 299.328 ppm			
TD 3012			
AQ 0.050 s			
FIDRES 20.000 Hz			
D 1 5.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 5000.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 125.0 W	Pow@90deg NUC1		
PLW 11 125.0 W	Pow@90degCP(Specs)		
PLW 12 49.0 W	NUC1		
SPW 0 49.0 W	Pow@90deg NUC2		
TE 298.000 K	Pow@HHshaped NUC2		
SPNAM 0 ramp50100.100	default		
CPDPRG 2 spinal64			

## Experiment Description

Sensitivity determination on solid probes including steps for optimization of PLW12 and SPW0. Signal to noise value is calculated using noise range of 20 ppm. All ranges are provided in plot title.

If number of scans is specified, NS will be set accordingly during acquisition.

L23 dependent offset optimization: Prior to the main acquisition O1P (PROCNO=10) will be optimized.

O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, and 3 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

For L23 = 1, 2, 3, 4, and 5 the values of SPW0 and PLW12 will be taken from an already acquired sine experiment.

For L23 = 11, 12, 13, 14 and 15 or if the sine experiment has not been measured a manual user interaction part to set or optimize SPW0, and PLW12 is included.

For L23 = 21, 22, 23, 24 and 25 the optimization of SPW0 and PLW12 will be enforced.

SPW0 is calculated for B1(NUC2) = (B1(NUC1) + 2\*MASR) \* rampFactor.

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition 1, -1, and -2 until power is in limits.

The selection of a certain side band condition can be forced by setting CNST 50 to 2, 1, -1, or -2.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000






## 5.4.15 CP 1H-13C sensitivity, MAS (NPT\_13C\_MAS\_sino\_cp1h\_13c)

**Test Sample:** Alpha-crystalline Glycine (weighted sample depending on rotor diameter)  
Z151222, Z151272, Z151262, Z151252, Z151232, Z151242, Z151212, Z163275,  
Z183105

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 Same as 1 but with manual user interaction to set SPW0 and PLW12
- 12 Same as 2 but with manual user interaction to set SPW0 and PLW12
- 13 Same as 3 but with manual user interaction to set SPW0 and PLW12
- 14 Same as 4 but with manual user interaction to set SPW0 and PLW12
- 15 Same as 5 but with manual user interaction to set SPW0 and PLW12
- 21 Same as 1 but with forced automatic optimization of SPW0 and PLW12
- 22 Same as 2 but with forced automatic optimization of SPW0 and PLW12
- 23 Same as 3 but with forced automatic optimization of SPW0 and PLW12
- 24 Same as 4 but with forced automatic optimization of SPW0 and PLW12
- 25 Same as 5 but with forced automatic optimization of SPW0 and PLW12

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 32768	
NUC2 1H		TDeff 2048	
PARMODE 0	Data Dimension	LB 0.000 Hz	
PULPROG cp		PH_mod 1	pk
NS 64	set according specs	ABSF1 1000.000 ppm	
DS 0		ABSF2 -1000.000 ppm	
RG 101.000	no optim.	F1P 0.000 ppm	
O1P 110.000 ppm		F2P 0.000 ppm	
O2P 6.200 ppm		CY 11.000 cm	
SW 299.337 ppm			
TD 3012			
AQ 0.050 s			
FIDRES 20.000 Hz			
D 1 5.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 2000.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 125.0 W	Pow@90deg NUC1		
PLW 11 125.0 W	Pow@90degCP(Specs)		
	NUC1		
PLW 12 54.0 W	Pow@90deg NUC2		
SPW 0 54.0 W	Pow@HHshaped NUC2		
TE 298.000 K	default		
SPNAM 0 ramp50100.100			
CPDPRG 2 spinal64			

## Experiment Description

Sensitivity determination on solid probes including steps for optimization of PLW12 and SPW0. Signal to noise value is calculated using noise range of 20 ppm. All ranges are provided in plot title.

If number of scans is specified, NS will be set accordingly during acquisition.

L23 dependent offset optimization: Prior to the main acquisition O1P (PROCNO=10) will be optimized.

O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, and 3 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

For L23 = 1, 2, 3, 4, and 5 the values of SPW0 and PLW12 will be taken from an already acquired paropt\_cp1h experiment.

For L23 = 11, 12, 13, 14 and 15 or if the paropt\_cp1h experiment has not been measured a manual user interaction part to set or optimize SPW0, and PLW12 is included.

For L23 = 21, 22, 23, 24 and 25 the optimization of SPW0 and PLW12 will be enforced.

SPW0 is calculated for B1(NUC2) = (B1(NUC1) + MASR) \* rampFactor.

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

## Spinrate


MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000



## 5.4.16 CP 19F-13C power stability MAS (NPT\_13C\_MAS\_stab\_cp19f\_13c)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Pseudo 2D CP spectrum to observe power stability.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 Same as 1 but with manual user interaction to set SPW0 and PLW12
- 12 Same as 2 but with manual user interaction to set SPW0 and PLW12
- 13 Same as 3 but with manual user interaction to set SPW0 and PLW12
- 14 Same as 4 but with manual user interaction to set SPW0 and PLW12
- 15 Same as 5 but with manual user interaction to set SPW0 and PLW12

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 16384	
NUC2 19F		LB 0.000	Hz
PARMODE 1	Data Dimension	PH_mod 1	pk
PULPROG npt_cp2d		ABSF1 1000.000	ppm
NS 16		ABSF2 -1000.000	ppm
DS 32		F1P 0.000	ppm
RG 101.000	no optim.	F2P 0.000	ppm
O1P 140.000			
O2P -74.000			
SWH 30120.482			
TD 3012			
AQ 0.050			
FIDRES 20.000			
D 1 5.000			
P 1 4.0	us	90deg NUC1	
P 3 3.5	us	max. dec. field	
P 15 8500.0	us	HH NUC2-NUC1	
PCPD 2 6.8	us	PCPD2 NUC2	
PLW 1 125.0	W	Pow@90deg NUC1	
PLW 11 125.0	W	Pow@90degCP(Specs)	
PLW 12 49.0	W	NUC1	
SPW 0 49.0	W	Pow@90deg NUC2	
TE 298.000	K	Pow@HHshaped NUC2	
SPNAM 0 ramp50100.100		default	
CPDPRG 2 spinal64			

## Experiment Description

Determination of power stability for CP experiments on solid probes using maximum decoupling field and maximum acquisition time provided as mandatory specifications. The average signal-to-noise ratio (20 ppm noise range)

is evaluated considering all rows of the pseudo-2D measurement. Also the maximum deviation of the results is determined, along with the rows of maximum and minimum result.

For power stability measurement the specs for acquisition time and decoupling pulse length must be specified.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized.

O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

For L23 < 10 the values of SPW0 and PLW12 will be taken from the corresponding sino experiment.

PLW12 will be optimized if the specified PCPD2 pulse differs from PCPD2 used in the corresponding sino experiment.

For L23 > 10 or if the corresponding sino experiment has not been measured a manual user interaction part to set or optimize SPW0 and PLW12 is included.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + 2 * MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition 1, -1, and -2 until power is in limits.

The selection of a certain side band condition can be forced by setting CNST 50 to 2, 1, -1, or -2.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency


	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000





## 5.4.17 CP 19F-13C power stability using H-coil, MAS (NPT\_13C\_MAS\_stab\_cp19f\_13c\_hcoil)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Pseudo 2D CP spectrum to observe power stability.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 Same as 1 but with manual user interaction to set SPW0 and PLW12
- 12 Same as 2 but with manual user interaction to set SPW0 and PLW12
- 13 Same as 3 but with manual user interaction to set SPW0 and PLW12
- 14 Same as 4 but with manual user interaction to set SPW0 and PLW12
- 15 Same as 5 but with manual user interaction to set SPW0 and PLW12

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 16384	
NUC2 19F		LB 0.000	Hz
PARMODE 1	Data Dimension	PH_mod 1	pk
PULPROG npt_cp2d		ABSF1 1000.000	ppm
NS 16		ABSF2 -1000.000	ppm
DS 32		F1P 0.000	ppm
RG 101.000	no optim.	F2P 0.000	ppm
O1P 140.000			
O2P -74.000			
SWH 30120.482			
TD 3012			
AQ 0.050			
FIDRES 20.000			
D 1 5.000			
P 1 4.0	us	90deg NUC1	
P 3 3.5	us	max. dec. field	
P 15 8500.0	us	HH NUC2-NUC1	
PCPD 2 6.8	us	PCPD2 NUC2	
PLW 1 125.0	W	Pow@90deg NUC1	
PLW 11 125.0	W	Pow@90degCP(Specs)	
PLW 12 49.0	W	NUC1	
SPW 0 49.0	W	Pow@90deg NUC2	
TE 298.000	K	Pow@HHshaped NUC2	
SPNAM 0 ramp50100.100		default	
CPDPRG 2 spinal64			

## Experiment Description

Determination of power stability for CP experiments on solid probes using maximum decoupling field and maximum acquisition time provided as mandatory specifications. The average signal-to-noise ratio (20 ppm noise range)

is evaluated considering all rows of the pseudo-2D measurement. Also the maximum deviation of the results is determined, along with the rows of maximum and minimum result.

For power stability measurement the specs for acquisition time and decoupling pulse length must be specified.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized.

O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

For L23 < 10 the values of SPW0 and PLW12 will be taken from the corresponding sino experiment.

PLW12 will be optimized if the specified PCPD2 pulse differs from PCPD2 used in the corresponding sino experiment.

For L23 > 10 or if the corresponding sino experiment has not been measured a manual user interaction part to set or optimize SPW0 and PLW12 is included.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + 2 * MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition 1, -1, and -2 until power is in limits.

The selection of a certain side band condition can be forced by setting CNST 50 to 2, 1, -1, or -2.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000




## 5.4.18 CP 1H-13C power stability MAS (NPT\_13C\_MAS\_stab\_cp1h\_13c)

**Test Sample:** Alpha-crystalline Glycine (not isotope-labeled or 2-13C, 15N Glycine)  
Z151222, Z151272, Z151262, Z151252, Z151232, Z151242, Z151212, Z183105,  
Z151223, Z151273, Z151263, Z151253, Z151233, Z151243, Z151213, Z163276,  
Z183106

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Pseudo 2D CP spectrum to observe power stability.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 Same as 1 but with manual user interaction to set SPW0 and PLW12
- 12 Same as 2 but with manual user interaction to set SPW0 and PLW12
- 13 Same as 3 but with manual user interaction to set SPW0 and PLW12
- 14 Same as 4 but with manual user interaction to set SPW0 and PLW12
- 15 Same as 5 but with manual user interaction to set SPW0 and PLW12

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 16384	
NUC2 1H		LB 0.000	Hz
PARMODE 1	Data Dimension	PH_mod 1	pk
PULPROG npt_cp2d		ABSF1 1000.000	ppm
NS 1		ABSF2 -1000.000	ppm
DS 32		F1P 0.000	ppm
RG 101.000	no optim.	F2P 0.000	ppm
O1P 110.000			
O2P 6.200			
SWH 30120.482			
TD 3012			
AQ 0.050			
FIDRES 20.000			
D 1 5.000			
P 1 4.0	us	90deg NUC1	
P 3 3.5	us	max. dec. field	
P 15 10000.0	us	HH NUC2-NUC1	
PCPD 2 6.8	us	PCPD2 NUC2	
PLW 1 125.0	W	Pow@90deg NUC1	
PLW 11 125.0	W	Pow@90degCP(Specs)	
PLW 12 54.0	W	NUC1	
SPW 0 54.0	W	Pow@90deg NUC2	
TE 298.000	K	Pow@HHshaped NUC2	
SPNAM 0 ramp50100.100		default	
CPDPRG 2 spinal64			

## Experiment Description

Determination of power stability for CP experiments on solid probes using maximum decoupling field and maximum acquisition time provided as mandatory specifications. The average signal-to-noise ratio (20 ppm noise range)

is evaluated considering all rows of the pseudo-2D measurement. Also the maximum deviation of the results is determined, along with the rows of maximum and minimum result.

For power stability measurement the specs for acquisition time and decoupling pulse length must be specified.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

For L23 < 10 the values of SPW0 and PLW12 will be taken from the corresponding paropt\_cp1h experiment.

PLW12 will be optimized if the specified PCPD2 pulse differs from PCPD2 used in the corresponding paropt\_cp1h experiment.

For L23 > 10 or if the corresponding sino experiment has not been measured a manual user interaction part to set or optimize SPW0 and PLW12 is included.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000






## 5.4.19 P90 15N pulse calibration, MAS (NPT\_15N\_MAS\_p90det\_15n)

**Test Sample:** Alpha-crystalline Glycine (not isotope-labeled or 2-13C, 15N Glycine)  
 Z151222, Z151272, Z151262, Z151252, Z151232, Z151242, Z151212, Z183105,  
 Z151223, Z151273, Z151263, Z151253, Z151233, Z151243, Z151213, Z163276,  
 Z183106

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
- 12 skip automatic phase correction and apply manually set values, execute O1P determination.
- 13 skip automatic phase correction and apply manually set values, execute O2P determination.
- 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
- 15 Same as 14 but with RGA during O2P determination.
- 100 same as xx but skip of SINO check on PROCNO 11
- +XX
- 1000 same as xxx but ignore specifications (optimize power for pulse length from prosol)
- +XXX

## Parameters

<p><b>F1 ACQU</b></p> <p>Parameters</p> <p>NUC1 15N</p> <p>NUC2 1H</p> <p>PARMODE 0 Data Dimension</p> <p>PULPROG hpdec</p> <p>NS 128</p> <p>DS 0</p> <p>RG 101.000 no optim.</p> <p>O1P 35.000 ppm</p> <p>O2P 6.200 ppm</p> <p>SWH 38461.539 Hz</p> <p>TD 1510</p> <p>AQ 0.020 s</p> <p>FIDRES 50.942 Hz</p> <p>D 1 15.000 s</p> <p>P 1 4.5 us 90deg NUC1</p> <p>CPDPRG2 spinal64 decoupl. sequence</p> <p>PCPD2 8.3 us PCPD NUC2</p> <p>PLW 1 236.0 W Pow@90deg(Specs) NUC1</p> <p>PLW 12 54.0 W</p> <p>TE 298.000 K default</p>	<p><b>F1 PROC</b></p> <p>Parameters</p> <p>SI 4096</p> <p>LB 0.300 Hz</p> <p>PH_mod 1 pk</p>
--	--

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table. L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined for decoupling by acquisition and processing of a spectrum of NUC 2 in a derived data set.

Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated with the determined pulse and the power used. Results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000


## 5.4.20 P90 15N 1H-15N CP pulse calibration, MAS (NPT\_15N\_MAS\_p90det\_cp1h\_15n)

**Test Sample:** Alpha-crystalline Glycine (not isotope-labeled or 2-13C, 15N Glycine)  
Z151222, Z151272, Z151262, Z151252, Z151232, Z151242, Z151212, Z183105,  
Z151223, Z151273, Z151263, Z151253, Z151233, Z151243, Z151213, Z163276,  
Z183106

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



```
Report of AD program poptau
FIP=30.254ppm F2P=20.254ppm
PROCNO=999
Linear optimisation of P1M1 in 6 steps,
starting at 64.500000000 andval= 199.500000000
OPTIMIZE ZERO
VARIABLE= P1M1
Experiment P1M1 Maximum point Minimum point Integral
1 64.5000 355395388 -4517408 2449326.477
2 93.1000 234320411 -3947284 1466934.913
3 119.7000 142196019 -3985879 1002226.904
4 148.3000 94802193 -3554389 397263.441
5 172.9000 41134332 -97224507 -2038971.974
6 199.5000 24309180 -149929017 -724266.331
poptau For P1M1 finished.
SSCO at experiment 4.498957, P1M1 = 159.570915
```

### Example Printout

P90 CP pulse determination using 1D method by varying the excitation pulse P1. The result of a CONVTO1D routine shows six experiments P1\*0.5 to P1\*1.5 (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
+XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 15N		SI 4096	
NUC2 1H		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp90			
NS 4			
DS 0			
RG 101.000	no optim.		
O1P 35.000 ppm			
O2P 6.200 ppm			
SWH 16129.032 Hz			
TD 1612			
AQ 0.050 s			
FIDRES 20.011 Hz			
D 1 5.000 s			
P 1 4.5 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 3500.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 236.0 W	Pow@90deg NUC1		
PLW 11 236.0 W	Pow@90degCP(Specs)		
	NUC1		
PLW 12 54.0 W	Pow@90deg NUC2		
SPW 0 26.0 W	Pow@HHshaped NUC2		
TE 298.000 K	default		
SPNAM 0 ramp50100.100			
CPDPRG 2 spinal64			

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal (maximum intensity) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set. Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

The results of the parameter optimization for PLW11 are stored under PROCNO 100.

Experiment results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

The PROSOL table will be updated with the determined pulse and the power used.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000



## 5.4.21 CP 1H-15N parameter optimization, MAS (NPT\_15N\_MAS\_paropt\_cp1h\_15n)


---

**Test Sample:** Alpha-crystalline 2-13C, 15N Glycine  
Z151223, Z151273, Z151263, Z151253, Z151233, Z151243, Z151213, Z163276,  
Z183106

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 15N		SI 32768	
NUC2 1H		LB 0.000 Hz	
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp		ABSF1 1000.000 ppm	
NS 4	set according specs	ABSF2 -1000.000 ppm	
DS 0		F1P 0.000 ppm	
RG 101.000	no optim.	F2P 0.000 ppm	
O1P 35.000 ppm		CY 11.000 cm	
O2P 6.200 ppm			
SW 742.867 ppm			
TD 3012	field dependent		
AQ 0.050 s			
FIDRES 20.000 Hz			
D 1 5.000 s			
P 1 4.5 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 3500.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 236.0 W	Pow@90deg NUC1		
PLW 11 236.0 W	Pow@90degCP(Specs)		
PLW 12 54.0 W	NUC1		
SPW 0 26.0 W	Pow@90deg NUC2		
TE 298.000 K	Pow@HHshaped NUC2		
SPNAM 0 ramp50100.100	default		
CPDPRG 2 spinal64			

## Experiment Description

Experiment for parameter optimization (PLW12 and SPW0) for mas experiments with cp. Signal to noise value is calculated using noise range of 20 ppm. All ranges are provided in plot title.

L23 dependent offset optimization: Prior to the main acquisition O1P (PROCNO=10) will be optimized. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, and 3 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

Start value of SPW0 is calculated for B1(NUC2) = (B1(NUC1) + MASR) \* rampFactor.

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000


## 5.4.22 P90 15N 1H-15N CP shortest pulse calibration, MAS (NPT\_15N\_MAS\_shortestPulse\_cp1h\_15n)

**Test Sample:** Alpha-crystalline Glycine (not isotope-labeled or 2-13C, 15N Glycine)  
 Z151222, Z151272, Z151262, Z151252, Z151232, Z151242, Z151212, Z183105,  
 Z151223, Z151273, Z151263, Z151253, Z151233, Z151243, Z151213, Z163276,  
 Z183106

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Shortest pulse determination with CP using 1D method by varying the excitation pulse P1. The result of a CONVTO1D routine shows 21 experiments P1\*0.5 to P1\*1.5(PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
 +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 15N		SI 4096	
NUC2 1H		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp90			
NS 4			
DS 0			
RG 101.000	no optim.		
O1P 35.000 ppm			
O2P 6.200 ppm			
SWH 16129.032 Hz			
TD 1612			
AQ 0.050 s			
FIDRES 20.011 Hz			
D 1 5.000 s			
P 1 4.5 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 3500.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 236.0 W	Pow@90deg NUC1		
PLW 11 236.0 W	Pow@90degCP(Specs)		
PLW 12 54.0 W	NUC1		
SPW 0 26.0 W	Pow@90deg NUC2		
TE 298.000 K	Pow@HHshaped NUC2		
SPNAM 0 ramp50100.100	default		
CPDPRG 2 spinal64			

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal (maximum intensity) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set. Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

The results of the parameter optimization for PLW11 are stored under PROCNO 100.

Experiment results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

The PROSOL table will not be updated during this experiment.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate


MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000



## 5.4.23 CP 1H-15N sensitivity, MAS (NPT\_15N\_MAS\_sino\_cp1h\_15n)

**Test Sample:** Alpha-crystalline Glycine (weighted sample depending on rotor diameter)  
Z151222, Z151272, Z151262, Z151252, Z151232, Z151242, Z151212, Z163275,  
Z183105  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 Same as 1 but with manual user interaction to set SPW0 and PLW12
- 12 Same as 2 but with manual user interaction to set SPW0 and PLW12
- 13 Same as 3 but with manual user interaction to set SPW0 and PLW12
- 14 Same as 4 but with manual user interaction to set SPW0 and PLW12
- 15 Same as 5 but with manual user interaction to set SPW0 and PLW12
- 21 Same as 1 but with forced automatic optimization of SPW0 and PLW12
- 22 Same as 2 but with forced automatic optimization of SPW0 and PLW12
- 23 Same as 3 but with forced automatic optimization of SPW0 and PLW12
- 24 Same as 4 but with forced automatic optimization of SPW0 and PLW12
- 25 Same as 5 but with forced automatic optimization of SPW0 and PLW12

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 15N		SI 32768	
NUC2 1H		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp		ABSF1 1000.000	ppm
NS 64	set according specs	ABSF2 -1000.000	ppm
DS 0		F1P 0.000	ppm
RG 101.000	no optim.	F2P 0.000	ppm
O1P 35.000		CY 11.000	cm
O2P 6.200			
SW 397.794			
TD 1612	field dependent		
AQ 0.050			
FIDRES 20.011			
D 1 5.000			
P 1 4.5	us 90deg NUC1		
P 3 3.5	us max. dec. field		
P 15 3500.0	us HH NUC2-NUC1		
PCPD 2 6.8	us PCPD2 NUC2		
PLW 1 236.0	W Pow@90deg NUC1		
PLW 11 236.0	W Pow@90degCP(Specs)		
PLW 12 54.0	W NUC1		
SPW 0 26.0	W Pow@90deg NUC2		
TE 298.000	K Pow@HHshaped NUC2		
SPNAM 0 ramp50100.100	default		
CPDPRG 2 spinal64			

## Experiment Description

Sensitivity determination on solid probes including steps for optimization of PLW12 and SPW0. Signal to noise value is calculated using noise range of 20 ppm. All ranges are provided in plot title.

If number of scans is specified, NS will be set accordingly during acquisition.

L23 dependent offset optimization: Prior to the main acquisition O1P (PROCNO=10) will be optimized.

O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, and 3 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

For L23 = 1, 2, 3, 4, and 5 the values of SPW0 and PLW12 will be taken from an already acquired paropt\_cp1h experiment.

For L23 = 11, 12, 13, 14 and 15 or if the paropt\_cp1h experiment has not been measured a manual user interaction part to set or optimize SPW0, and PLW12 is included.

For L23 = 21, 22, 23, 24 and 25 the optimization of SPW0 and PLW12 will be enforced.

SPW0 is calculated for B1(NUC2) = (B1(NUC1) + MASR) \* rampFactor.

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000






## 5.4.24 CP 1H-15N power stability MAS (NPT\_15N\_MAS\_stab\_cp1h\_15n)

**Test Sample:** Alpha-crystalline Glycine (not isotope-labeled or 2-13C, 15N Glycine)  
Z151222, Z151272, Z151262, Z151252, Z151232, Z151242, Z151212, Z183105,  
Z151223, Z151273, Z151263, Z151253, Z151233, Z151243, Z151213, Z163276,  
Z183106

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Pseudo 2D CP spectrum to observe power stability.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 Same as 1 but with manual user interaction to set SPW0 and PLW12
- 12 Same as 2 but with manual user interaction to set SPW0 and PLW12
- 13 Same as 3 but with manual user interaction to set SPW0 and PLW12
- 14 Same as 4 but with manual user interaction to set SPW0 and PLW12
- 15 Same as 5 but with manual user interaction to set SPW0 and PLW12

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 15N		SI 16384	
NUC2 1H		LB 0.000	Hz
PARMODE 1	Data Dimension	PH_mod 1	pk
PULPROG npt_cp2d		ABSF1 1000.000	ppm
NS 1		ABSF2 -1000.000	ppm
DS 32		F1P 0.000	ppm
RG 101.000	no optim.	F2P 0.000	ppm
O1P 35.000			
O2P 6.200			
SWH 16129.032			
TD 1612			
AQ 0.050			
FIDRES 20.011			
D 1 5.000			
P 1 4.5	us	90deg NUC1	
P 3 3.5	us	max. dec. field	
P 15 10000.0	us	HH NUC2-NUC1	
PCPD 2 6.8	us	PCPD2 NUC2	
PLW 1 236.0	W	Pow@90deg NUC1	
PLW 11 236.0	W	Pow@90degCP(Specs)	
PLW 12 54.0	W	NUC1	
SPW 0 26.0	W	Pow@90deg NUC2	
TE 298.000	K	Pow@HHshaped NUC2	
SPNAM 0 ramp50100.100		default	
CPDPRG 2 spinal64			

## Experiment Description

Determination of power stability for CP experiments on solid probes using maximum decoupling field and maximum acquisition time provided as mandatory specifications. The average signal-to-noise ratio (20 ppm noise range)

is evaluated considering all rows of the pseudo-2D measurement. Also the maximum deviation of the results is determined, along with the rows of maximum and minimum result.

For power stability measurement the specs for acquisition time and decoupling pulse length must be specified.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

For L23 < 10 the values of SPW0 and PLW12 will be taken from the corresponding paropt\_cp1h experiment.

PLW12 will be optimized if the specified PCPD2 pulse differs from PCPD2 used in the corresponding paropt\_cp1h experiment.

For L23 > 10 or if the corresponding sino experiment has not been measured a manual user interaction part to set or optimize SPW0 and PLW12 is included.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

## Spinrate


MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	4800	4000	4000	3000
400	40000	8000	5000	4800	4000	4000	4000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	6000	6000	6000	6000	6000
700	40000	8000	7000	7000	7000	7000	7000
750	40000	8000	7500	7500	7500	7500	7000
800	40000	8000	8000	8000	8000	8000	7000
850	40000	8500	8500	8500	8500	8500	7000
900	40000	9000	9000	9000	9000	9000	7000
950	40000	9500	9500	9500	9500	9500	7000
1000	40000	10000	10000	10000	10000	10000	7000



## 5.4.25 <sup>19</sup>F B1 homogeneity, MAS (NPT\_19F\_MAS\_b1homogeneity\_19f)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a region around each main signal is shown in the figure.

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.

## Parameters

<b>F1 ACQU</b> NUC1 19F PARMODE 1 PULPROG npt_p1b1hom2d_sol NS 1 DS 0 RG 16.000 O1P -74.000 ppm SWH 100000.000 Hz TD 2048 AQ 0.010 s FIDRES 97.656 Hz D 1 5.000 s P 1 3.5 us PLW 1 68.0 W TE 298.000 K	Parameters Data Dimension no optim. 90deg NUC1 Pow@90deg NUC1 default
<b>F1 PROC</b> SI 2048 LB 0.000 Hz PH_mod 1 CY 11.000 cm	Parameters pk

## Experiment Description

The B1 homogeneity measurement is acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range reaches from 15 to approx. 810 degrees. The B1-homogeneities based on integrals are computed from the magnitude of the second complex FID data point in PROCNO=2.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized. With the option L23 = 1 O1P determination step can be skipped to apply manually set value, i.e. value will be taken from data set.

The result of CONVTO1D is stored in PROCNO=999. This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Requirements: The corresponding 90° pulse determination experiment (NPT\_X\_MAS\_p90det\_YZ) is mandatory for B1 homogeneity measurement.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000

## 5.4.26 <sup>19</sup>F B1 homogeneity on H-coil, MAS (NPT\_19F\_MAS\_b1homogeneity\_19f\_hcoil)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a region around each main signal is shown in the figure.

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.



## Parameters

<b>F1 ACQU</b> NUC1 19F PARMODE 1 PULPROG npt_p1b1hom2d_sol NS 1 DS 0 RG 16.000 O1P -74.000 ppm SWH 100000.000 Hz TD 2048 AQ 0.010 s FIDRES 97.656 Hz D 1 5.000 s P 1 3.5 us PLW 1 68.0 W TE 298.000 K	Parameters  Data Dimension  no optim.  90deg NUC1 Pow@90deg NUC1 default
<b>F1 PROC</b> SI 2048 LB 0.000 Hz PH_mod 1 CY 11.000 cm	Parameters  pk

## Experiment Description

The B1 homogeneity measurement is acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range reaches from 15 to approx. 810 degrees. The B1-homogeneities based on integrals are computed from the magnitude of the second complex FID data point in PROCNO=2.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized. With the option L23 = 1 O1P determination step can be skipped to apply manually set value, i.e. value will be taken from data set.

The result of CONVTO1D is stored in PROCNO=999. This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Requirements: The corresponding 90° pulse determination experiment (NPT\_X\_MAS\_p90det\_YZ) is mandatory for B1 homogeneity measurement.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000

## 5.4.27 P90 19F pulse calibration, MAS (NPT\_19F\_MAS\_p90det\_19f)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
 Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.
- 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
- 12 skip automatic phase correction and apply manually set values, execute O1P determination.
- 100 same as xx but skip of SINO check on PROCNO 11
- +xx
- 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)
- +xxx

## Parameters

<b>F1 ACQU</b> NUC1 19F PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 16.000 O1P -74.000 ppm SWH 100000.000 Hz TD 2988 AQ 0.015 s FIDRES 66.934 Hz D 1 5.000 s P 1 3.5 us PLW 1 68.0 W TE 298.000 K	Parameters Data Dimension no optim. 90deg NUC1 Pow@90deg(Specs) NUC1 default
<b>F1 PROC</b> SI 4096 LB 0.000 Hz PH_mod 1	Parameters pk

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table. L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. Phase correction is done applying APK as default method. With the option L23 = 2 and 12 O1P determination step can be skipped to apply manually set values. The options L23 = 11 and 12 skip automatic phase correction for manual settings to be applied. After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options. If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated updated with the determined pulse and the power used. Results are stored under PROCNO 999. This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required. Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000


## 5.4.28 P90 19F pulse calibration on h-coil, MAS (NPT\_19F\_MAS\_p90det\_19f\_hcoil)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.
- 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
- 12 skip automatic phase correction and apply manually set values, execute O1P determination.
- 100 same as xx but skip of SINO check on PROCNO 11
- +XX
- 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)
- +XXX

## Parameters

<b>F1 ACQU</b> NUC1 19F PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 16.000 O1P -74.000 ppm SWH 100000.000 Hz TD 2988 AQ 0.015 s FIDRES 66.934 Hz D 1 5.000 s P 1 3.5 us PLW 1 68.0 W TE 298.000 K	Parameters Data Dimension no optim. 90deg NUC1 Pow@90deg(Specs) NUC1 default
<b>F1 PROC</b> SI 4096 LB 0.000 Hz PH_mod 1	Parameters pk

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.  
 L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.  
 Phase correction is done applying APK as default method. With the option L23 = 2 and 12 O1P determination step can be skipped to apply manually set values. The options L23 = 11 and 12 skip automatic phase correction for manual settings to be applied.  
 After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.  
 If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated with the determined pulse and the power used. Results are stored under PROCNO 999.  
 This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.  
 Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000

## 5.4.29 P90 19F shortest pulse calibration, MAS (NPT\_19F\_MAS\_shortestPulse\_19f)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
 Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows 21 experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
  - 2 execute O1P determination.
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 100 same as xx but skip of SINO check on PROCNO 11
- +xx

## Parameters

<b>F1 ACQU</b> NUC1 19F PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 16.000 O1P -74.000 ppm SWH 100000.000 Hz TD 2988 AQ 0.015 s FIDRES 66.934 Hz D 1 5.000 s TE 298.000 K	Parameters Data Dimension no optim. default
<b>F1 PROC</b> SI 4096 LB 0.000 Hz PH_mod 1	Parameters pk

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

Phase correction is done applying APK as default method. With the option L23 = 2 and 12 O1P determination step can be skipped to apply manually set values. The options L23 = 11 and 12 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated with the determined pulse and the power used. Results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000

## 5.4.30 P90 19F shortest pulse calibration on h-coil, MAS (NPT\_19F\_MAS\_shortestPulse\_19f\_hcoil)

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
 Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows 21 experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
  - 2 execute O1P determination.
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 100 same as xx but skip of SINO check on PROCNO 11
- +xx



## Parameters

<b>F1 ACQU</b> NUC1 19F PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 16.000 O1P -74.000 ppm SWH 100000.000 Hz TD 2988 AQ 0.015 s FIDRES 66.934 Hz D 1 5.000 s TE 298.000 K	<b>Parameters</b>  Data Dimension  no optim.  default	<b>F1 PROC</b> SI 4096 LB 0.000 Hz PH_mod 1	<b>Parameters</b>  pk
--	---	--	-----------------------------

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

Phase correction is done applying APK as default method. With the option L23 = 2 and 12 O1P determination step can be skipped to apply manually set values. The options L23 = 11 and 12 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated with the determined pulse and the power used. Results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate


MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000

## 5.4.31 <sup>19</sup>F sensitivity, MAS (NPT\_19F\_MAS\_sino\_19f)

---

**Test Sample:** Ammonium Trifluoroacetate (CF<sub>3</sub>CO<sub>2</sub>NH<sub>4</sub>)  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity and line width determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.

## Parameters

<b>F1 ACQU</b> NUC1 19F PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 16.000 O1P -74.000 ppm SWH 100000.000 Hz TD 2048 AQ 0.010 s FIDRES 97.656 Hz D 1 5.000 s P 1 3.5 us PLW 1 68.0 W TE 298.000 K	Parameters  Data Dimension  no optim.  90deg NUC1 Pow@90deg(Specs) NUC1 default
<b>F1 PROC</b> SI 16384 LB 0.300 Hz PH_mod 1 ABSF1 1000.000 ppm ABSF2 -1000.000 ppm F1P 156.761 ppm F2P -140.761 ppm CY 11.000 cm	Parameters  pk

## Experiment Description

Sensitivity determination on solid probes including line width determination. Signal to noise value is calculated using noise range of 20 ppm set apart possible MAS spinning side bands. All ranges are provided in plot title.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized.

With the option L23 = 1 O1P determination step can be skipped to apply manually set value, i.e. value will be taken from data set.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000

## 5.4.32 $^{19}\text{F}$ sensitivity on h-coil, MAS (NPT\_19F\_MAS\_sino\_19f\_hcoil)

**Test Sample:** Ammonium Trifluoroacetate ( $\text{CF}_3\text{CO}_2\text{NH}_4$ )  
Z151226, Z151276, Z151266, Z151256, Z151236, Z151246, Z151216, Z163278  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity and line width determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.

## Parameters

<b>F1 ACQU</b> NUC1 19F PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 16.000 O1P -74.000 ppm SWH 100000.000 Hz TD 2048 AQ 0.010 s FIDRES 97.656 Hz D 1 5.000 s P 1 3.5 us PLW 1 68.0 W TE 298.000 K	Parameters Data Dimension no optim. 90deg NUC1 Pow@90deg(Specs) NUC1 default
<b>F1 PROC</b> SI 16384 LB 0.300 Hz PH_mod 1 ABSF1 1000.000 ppm ABSF2 -1000.000 ppm F1P 156.761 ppm F2P -140.761 ppm CY 11.000 cm	Parameters pk

## Experiment Description

Sensitivity determination on solid probes including line width determination. Signal to noise value is calculated using noise range of 20 ppm set apart possible MAS spinning side bands. All ranges are provided in plot title.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized.

With the option L23 = 1 O1P determination step can be skipped to apply manually set value, i.e. value will be taken from data set.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	24000	24000	24000	24000	15000	7000
400	40000	24000	24000	24000	24000	15000	7000
500	40000	24000	24000	24000	24000	15000	7000
600	40000	24000	24000	24000	24000	15000	7000
700	40000	24000	24000	24000	24000	15000	7000
750	40000	24000	24000	24000	24000	15000	7000
800	40000	24000	24000	24000	24000	15000	7000
850	40000	24000	24000	24000	24000	15000	7000
900	40000	24000	24000	24000	24000	15000	7000
950	40000	24000	24000	24000	24000	15000	7000
1000	40000	24000	24000	24000	24000	15000	7000

## 5.4.33 1H B1 homogeneity, MAS (NPT\_1H\_MAS\_b1homogeneity\_1h)


---

**Test Sample:** Adamantane  
Z151221, Z151271, Z151261, Z151251, Z151231, Z151241, Z151211, Z163274,  
Z183104

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

B1 homogeneity determination using a pseudo 2D method. Only a region around each main signal is shown in the figure.

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.

## Parameters

<b>F1 ACQU</b> NUC1 1H PARMODE 1 PULPROG npt_p1b1hom2d_sol NS 1 DS 4 RG 8.000 O1P 2.460 ppm SWH 100000.000 Hz TD 2048 AQ 0.010 s FIDRES 97.656 Hz D 1 5.000 s P 1 3.5 us PLW 1 74.0 W TE 298.000 K	<b>Parameters</b>  Data Dimension  no optim.  90deg NUC1 Pow@90deg NUC1 default
<b>F1 PROC</b> SI 2048 LB 0.000 Hz PH_mod 1 CY 11.000 cm	<b>Parameters</b>  pk

## Experiment Description

The B1 homogeneity measurement is acquired by a series of 1D spectra whereby the excitation pulse is step-by-step incremented. The acquired range reaches from 15 to approx. 810 degrees. The B1-homogeneities based on integrals are computed from the magnitude of the second complex FID data point in PROCNO=2.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized. With the option L23 = 1 O1P determination step can be skipped to apply manually set value, i.e. value will be taken from data set.

The result of CONVTO1D is stored in PROCNO=999. This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Requirements: The corresponding 90° pulse determination experiment (NPT\_X\_MAS\_p90det\_YZ) is mandatory for B1 homogeneity measurement.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	111000	67000	42000	35000	24000	15000	7000
400	111000	67000	42000	35000	24000	15000	7000
500	111000	67000	42000	35000	24000	15000	7000
600	111000	67000	42000	35000	24000	15000	7000
700	111000	67000	42000	35000	24000	15000	7000
750	111000	67000	42000	35000	24000	15000	7000
800	111000	67000	42000	35000	24000	15000	7000
850	111000	67000	42000	35000	24000	15000	7000
900	111000	67000	42000	35000	24000	15000	7000
950	111000	67000	42000	35000	24000	15000	7000
1000	111000	67000	42000	35000	24000	15000	7000

## 5.4.34 P90 1H pulse calibration, MAS (NPT\_1H\_MAS\_p90det\_1h)

**Test Sample:** Adamantane  
 Z151221, Z151271, Z151261, Z151251, Z151231, Z151241, Z151211, Z163274,  
 Z183104  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.
- 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
- 12 skip automatic phase correction and apply manually set values, execute O1P determination.
- 100 same as xx but skip of SINO check on PROCNO 11
- +XX
- 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)
- +XXX



## Parameters

<b>F1 ACQU</b> NUC1 1H PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 8.000 O1P 2.460 ppm SWH 100000.000 Hz TD 2988 AQ 0.015 s FIDRES 66.934 Hz D 1 5.000 s P 1 3.5 us PLW 1 74.0 W TE 298.000 K	Parameters Data Dimension no optim. 90deg NUC1 Pow@90deg(Specs) NUC1 default
<b>F1 PROC</b> SI 4096 LB 0.000 Hz PH_mod 1	Parameters pk

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table. L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. Phase correction is done applying APK as default method. With the option L23 = 2 and 12 O1P determination step can be skipped to apply manually set values. The options L23 = 11 and 12 skip automatic phase correction for manual settings to be applied. After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options. If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated updated with the determined pulse and the power used. Results are stored under PROCNO 999. This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required. Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	111000	67000	42000	35000	24000	15000	7000
400	111000	67000	42000	35000	24000	15000	7000
500	111000	67000	42000	35000	24000	15000	7000
600	111000	67000	42000	35000	24000	15000	7000
700	111000	67000	42000	35000	24000	15000	7000
750	111000	67000	42000	35000	24000	15000	7000
800	111000	67000	42000	35000	24000	15000	7000
850	111000	67000	42000	35000	24000	15000	7000
900	111000	67000	42000	35000	24000	15000	7000
950	111000	67000	42000	35000	24000	15000	7000
1000	111000	67000	42000	35000	24000	15000	7000


## 5.4.35 P90 1H shortest pulse calibration, MAS (NPT\_1H\_MAS\_shortestPulse\_1h)

**Test Sample:** Adamantane  
 Z151221, Z151271, Z151261, Z151251, Z151231, Z151241, Z151211, Z163274,  
 Z183104

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows 21 experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
  - 2 execute O1P determination.
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX

## Parameters

<b>F1 ACQU</b> NUC1 1H PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 8.000 O1P 2.460 ppm SWH 100000.000 Hz TD 2988 AQ 0.015 s FIDRES 66.934 Hz D 1 5.000 s TE 298.000 K	Parameters  Data Dimension  no optim.   default
<b>F1 PROC</b> SI 4096 LB 0.000 Hz PH_mod 1	Parameters  pk

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

Phase correction is done applying APK as default method. With the option L23 = 2 and 12 O1P determination step can be skipped to apply manually set values. The options L23 = 11 and 12 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated with the determined pulse and the power used. Results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	111000	67000	42000	35000	24000	15000	7000
400	111000	67000	42000	35000	24000	15000	7000
500	111000	67000	42000	35000	24000	15000	7000
600	111000	67000	42000	35000	24000	15000	7000
700	111000	67000	42000	35000	24000	15000	7000
750	111000	67000	42000	35000	24000	15000	7000
800	111000	67000	42000	35000	24000	15000	7000
850	111000	67000	42000	35000	24000	15000	7000
900	111000	67000	42000	35000	24000	15000	7000
950	111000	67000	42000	35000	24000	15000	7000
1000	111000	67000	42000	35000	24000	15000	7000

## 5.4.36 1H sensitivity, MAS (NPT\_1H\_MAS\_sino\_1h)


---

**Test Sample:** Adamantane  
Z151221, Z151271, Z151261, Z151251, Z151231, Z151241, Z151211, Z163274,  
Z183104

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity and line width determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.

## Parameters

<b>F1 ACQU</b> NUC1 1H PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 8.000 O1P 2.460 ppm SWH 100000.000 Hz TD 2048 AQ 0.010 s FIDRES 97.656 Hz D 1 5.000 s P 1 3.5 us PLW 1 74.0 W TE 298.000 K	Parameters  Data Dimension   no optim.   90deg NUC1 Pow@90deg(Specs) NUC1 default
<b>F1 PROC</b> SI 16384 LB 0.000 Hz PH_mod 1 ABSF1 1000.000 ppm ABSF2 -1000.000 ppm F1P 156.761 ppm F2P -140.761 ppm CY 11.000 cm	Parameters   pk

## Experiment Description

Sensitivity determination on solid probes including line width determination. Signal to noise value is calculated using noise range of 20 ppm set apart possible MAS spinning side bands. All ranges are provided in plot title.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized.

With the option L23 = 1 O1P determination step can be skipped to apply manually set value, i.e. value will be taken from data set.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	111000	67000	42000	35000	24000	15000	7000
400	111000	67000	42000	35000	24000	15000	7000
500	111000	67000	42000	35000	24000	15000	7000
600	111000	67000	42000	35000	24000	15000	7000
700	111000	67000	42000	35000	24000	15000	7000
750	111000	67000	42000	35000	24000	15000	7000
800	111000	67000	42000	35000	24000	15000	7000
850	111000	67000	42000	35000	24000	15000	7000
900	111000	67000	42000	35000	24000	15000	7000
950	111000	67000	42000	35000	24000	15000	7000
1000	111000	67000	42000	35000	24000	15000	7000

## 5.4.37 P90 31P pulse calibration, MAS (NPT\_31P\_MAS\_p90det\_31p)

**Test Sample:** Ammonium Dihydrogenphosphate (NH<sub>4</sub>H<sub>2</sub>PO<sub>4</sub>)  
 Z151224, Z151274, Z151264, Z151254, Z151234, Z151244, Z151214, Z163277  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination.
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
 +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 31P		SI 8192	
NUC2 1H		LB 0.300	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG hpdec			
NS 1			
DS 0			
RG 101.000	no optim.		
O1P 2.000	ppm		
O2P 7.200	ppm		
SWH 40650.406	Hz		
TD 3256			
AQ 0.040	s		
FIDRES 24.970	Hz		
D 1 10.000	s		
P 1 4.0	us	90deg NUC1	
CPDPRG2 spinal64		decoupl. sequence	
PCPD2 8.3	us	PCPD NUC2	
PLW 1 135	W	Pow@90deg(Specs) NUC1	
PLW 12 54.0	W		
TE 298.000	K	default	

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

O2P is determined for decoupling by acquisition and processing of a spectrum of NUC 2 in a derived data set.

Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated with the determined pulse and the power used. Results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	15000	15000	10000	10000	5000	4000
400	40000	15000	15000	10000	10000	5000	4000
500	40000	15000	15000	10000	10000	5000	4000
600	40000	15000	15000	10000	10000	5000	4000
700	40000	15000	15000	10000	10000	5000	4000
750	40000	15000	15000	10000	10000	5000	4000
800	40000	15000	15000	10000	10000	5000	4000
850	40000	15000	15000	10000	10000	5000	4000
900	40000	15000	15000	10000	10000	5000	4000
950	40000	15000	15000	10000	10000	5000	4000
1000	40000	15000	15000	10000	10000	5000	4000


## 5.4.38 P90 31P 1H-31P CP pulse calibration, MAS (NPT\_31P\_MAS\_p90det\_cp1h\_31p)

**Test Sample:** Ammonium Dihydrogenphosphate (NH<sub>4</sub>H<sub>2</sub>PO<sub>4</sub>)  
Z151224, Z151274, Z151264, Z151254, Z151234, Z151244, Z151214, Z163277

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

P90 CP pulse determination using 1D method by varying the excitation pulse P1. The result of a CONVTO1D routine shows six experiments P1\*0.5 to P1\*1.5 (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
+XXX



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 31P		SI 8192	
NUC2 1H		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp90			
NS 4			
DS 0			
RG 101.000	no optim.		
O1P 2.000 ppm			
O2P 7.200 ppm			
SWH 48543.688 Hz			
TD 4854			
AQ 0.050 s			
FIDRES 20.002 Hz			
D 1 5.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 3500.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 135.0 W	Pow@90deg NUC1		
PLW 11 135.0 W	Pow@90degCP(Specs)		
	NUC1		
PLW 12 54.0 W	Pow@90deg NUC2		
SPW 0 54.0 W	Pow@HHshaped NUC2		
TE 298.000 K	default		
SPNAM 0 ramp50100.100			
CPDPRG 2 spinal64			

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal (maximum intensity) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set. Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

The results of the parameter optimization for PLW11 are stored under PROCNO 100.

Experiment results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

The PROSOL table will be updated with the determined pulse and the power used.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	15000	15000	10000	10000	5000	4000
400	40000	15000	15000	10000	10000	5000	4000
500	40000	15000	15000	10000	10000	5000	4000
600	40000	15000	15000	10000	10000	5000	4000
700	40000	15000	15000	10000	10000	5000	4000
750	40000	15000	15000	10000	10000	5000	4000
800	40000	15000	15000	10000	10000	5000	4000
850	40000	15000	15000	10000	10000	5000	4000
900	40000	15000	15000	10000	10000	5000	4000
950	40000	15000	15000	10000	10000	5000	4000
1000	40000	15000	15000	10000	10000	5000	4000




## 5.4.39 P90 31P 1H-31P CP shortest pulse calibration, MAS (NPT\_31P\_MAS\_shortestPulse\_cp1h\_31p)

**Test Sample:** Ammonium Dihydrogenphosphate (NH<sub>4</sub>H<sub>2</sub>PO<sub>4</sub>)  
Z151224, Z151274, Z151264, Z151254, Z151234, Z151244, Z151214, Z163277

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

Shortest pulse determination with CP using 1D method by varying the excitation pulse P1. The result of a CONVTO1D routine shows 21 experiments P1\*0.5 to P1\*1.5 (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
  - 2 execute O1P determination.
  - 3 execute O2P determination.
  - 4 execute O1P and O2P determination.
  - 5 Same as 4 but with RGA during O2P determination
  - 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
  - 12 skip automatic phase correction and apply manually set values, execute O1P determination.
  - 13 skip automatic phase correction and apply manually set values, execute O2P determination.
  - 14 skip automatic phase correction and apply manually set values, execute O1P and O2P determination.
  - 15 Same as 14 but with RGA during O2P determination
  - 100 same as xx but skip of SINO check on PROCNO 11
- +XX  
1000same as xxx but ignore specifications (optimize power for pulse length from prosol)  
+XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 31P		SI 8192	
NUC2 1H		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp90			
NS 4			
DS 0			
RG 101.000	no optim.		
O1P 2.000 ppm			
O2P 7.200 ppm			
SWH 48543.688 Hz			
TD 4854			
AQ 0.050 s			
FIDRES 20.002 Hz			
D 1 5.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 3500.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 135.0 W	Pow@90deg NUC1		
PLW 11 135.0 W	Pow@90degCP(Specs)		
	NUC1		
PLW 12 54.0 W	Pow@90deg NUC2		
SPW 0 54.0 W	Pow@HHshaped NUC2		
TE 298.000 K	default		
SPNAM 0 ramp50100.100			
CPDPRG 2 spinal64			

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal (maximum intensity) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set. Phase correction is done applying APK as default method. With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set. The options L23 = 11, 12, 13, 14 and 15 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

The results of the parameter optimization for PLW11 are stored under PROCNO 100.

Experiment results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

The PROSOL table will not be updated during this experiment.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.

## Spinrate


MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	15000	15000	10000	10000	5000	4000
400	40000	15000	15000	10000	10000	5000	4000
500	40000	15000	15000	10000	10000	5000	4000
600	40000	15000	15000	10000	10000	5000	4000
700	40000	15000	15000	10000	10000	5000	4000
750	40000	15000	15000	10000	10000	5000	4000
800	40000	15000	15000	10000	10000	5000	4000
850	40000	15000	15000	10000	10000	5000	4000
900	40000	15000	15000	10000	10000	5000	4000
950	40000	15000	15000	10000	10000	5000	4000
1000	40000	15000	15000	10000	10000	5000	4000



## 5.4.40 CP 1H-31P sensitivity, MAS (NPT\_31P\_MAS\_sino\_cp1h\_31p)

**Test Sample:** Ammonium Dihydrogenphosphate (NH<sub>4</sub>H<sub>2</sub>PO<sub>4</sub>)  
Z151224, Z151274, Z151264, Z151254, Z151234, Z151244, Z151214, Z163277  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Spectrum of sensitivity determination.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 Same as 1 but with manual user interaction to set SPW0 and PLW12
- 12 Same as 2 but with manual user interaction to set SPW0 and PLW12
- 13 Same as 3 but with manual user interaction to set SPW0 and PLW12
- 14 Same as 4 but with manual user interaction to set SPW0 and PLW12
- 15 Same as 5 but with manual user interaction to set SPW0 and PLW12
- 21 Same as 1 but with forced automatic optimization of SPW0 and PLW12
- 22 Same as 2 but with forced automatic optimization of SPW0 and PLW12
- 23 Same as 3 but with forced automatic optimization of SPW0 and PLW12
- 24 Same as 4 but with forced automatic optimization of SPW0 and PLW12
- 25 Same as 5 but with forced automatic optimization of SPW0 and PLW12



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 31P		SI 32768	
NUC2 1H		LB 0.000	Hz
PARMODE 0	Data Dimension	PH_mod 1	pk
PULPROG cp		ABSF1 1000.000	ppm
NS 4	set according specs	ABSF2 -1000.000	ppm
DS 0		F1P 0.000	ppm
RG 101.000	no optim.	F2P 0.000	ppm
O1P 2.000		CY 11.000	cm
O2P 7.200			
SW 299.697			
TD 4854			
AQ 0.050			
FIDRES 20.002			
D 1 5.000			
P 1 4.0	us		
P 3 3.5	us		
P 15 3500.0	us		
PCPD 2 6.8	us		
PLW 1 135.0	W		
PLW 11 135.0	W		
PLW 12 54.0	W		
SPW 0 54.0	W		
TE 298.000	K		
SPNAM 0 ramp50100.100			
CPDPRG 2 spinal64			

## Experiment Description

Sensitivity determination on solid probes including steps for optimization of PLW12 and SPW0. Signal to noise value is calculated using noise range of 20 ppm. All ranges are provided in plot title.

If number of scans is specified, NS will be set accordingly during acquisition.

L23 dependent offset optimization: Prior to the main acquisition O1P (PROCNO=10) will be optimized.

O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, and 3 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

For L23 = 1, 2, 3, 4, and 5 the values of SPW0 and PLW12 will be taken from an already acquired sine experiment.

For L23 = 11, 12, 13, 14 and 15 or if the sine experiment has not been measured a manual user interaction part to set or optimize SPW0, and PLW12 is included.

For L23 = 21, 22, 23, 24 and 25 the optimization of SPW0 and PLW12 will be enforced.

SPW0 is calculated for B1(NUC2) = (B1(NUC1) + MASR) \* rampFactor.

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

## Spinrate


MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	15000	15000	10000	10000	5000	4000
400	40000	15000	15000	10000	10000	5000	4000
500	40000	15000	15000	10000	10000	5000	4000
600	40000	15000	15000	10000	10000	5000	4000
700	40000	15000	15000	10000	10000	5000	4000
750	40000	15000	15000	10000	10000	5000	4000
800	40000	15000	15000	10000	10000	5000	4000
850	40000	15000	15000	10000	10000	5000	4000
900	40000	15000	15000	10000	10000	5000	4000
950	40000	15000	15000	10000	10000	5000	4000
1000	40000	15000	15000	10000	10000	5000	4000



## 5.4.41 CP 1H-31P power stability MAS (NPT\_31P\_MAS\_stab\_cp1h\_31p)

**Test Sample:** Ammonium Dihydrogenphosphate (NH<sub>4</sub>H<sub>2</sub>PO<sub>4</sub>)  
Z151224, Z151274, Z151264, Z151254, Z151234, Z151244, Z151214, Z163277  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Pseudo 2D CP spectrum to observe power stability.

### Control Option for Acquisition (L23)

- 1 default, skip O1P and O2P determination.
- 2 execute O1P determination.
- 3 execute O2P determination.
- 4 execute O1P and O2P determination.
- 5 Same as 4 but with RGA during O2P determination
- 11 Same as 1 but with manual user interaction to set SPW0 and PLW12
- 12 Same as 2 but with manual user interaction to set SPW0 and PLW12
- 13 Same as 3 but with manual user interaction to set SPW0 and PLW12
- 14 Same as 4 but with manual user interaction to set SPW0 and PLW12
- 15 Same as 5 but with manual user interaction to set SPW0 and PLW12

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 31P		SI 32768	
NUC2 1H		LB 0.000 Hz	
PARMODE 1	Data Dimension	PH_mod 1	pk
PULPROG npt_cp2d		ABSF1 1000.000 ppm	
NS 1		ABSF2 -1000.000 ppm	
DS 32		F1P 0.000 ppm	
RG 101.000	no optim.	F2P 0.000 ppm	
O1P 2.000 ppm			
O2P 7.200 ppm			
SWH 48543.688 Hz			
TD 4854			
AQ 0.050 s			
FIDRES 20.002 Hz			
D 1 5.000 s			
P 1 4.0 us	90deg NUC1		
P 3 3.5 us	max. dec. field		
P 15 3500.0 us	HH NUC2-NUC1		
PCPD 2 6.8 us	PCPD2 NUC2		
PLW 1 135.0 W	Pow@90deg NUC1		
PLW 11 135.0 W	Pow@90degCP(Specs)		
	NUC1		
PLW 12 54.0 W	Pow@90deg NUC2		
SPW 0 54.0 W	Pow@HHshaped NUC2		
TE 298.000 K	default		
SPNAM 0 ramp50100.100			
CPDPRG 2 spinal64			

## Experiment Description

Determination of power stability for CP experiments on solid probes using maximum decoupling field and maximum acquisition time provided as mandatory specifications. The average signal-to-noise ratio (10 ppm noise range)

is evaluated considering all rows of the pseudo-2D measurement. Also the maximum deviation of the results is determined, along with the rows of maximum and minimum result.

For power stability measurement the specs for acquisition time and decoupling pulse length must be specified.

L23 dependent offset optimization: prior to the main acquisition O1P (PROCNO=10) will be optimized.

O2P is determined by acquisition and processing of a spectrum of NUC 2 in a derived data set.

With the option L23 = 1, 2, 3, 11, 12, and 13 O1P and O2P determination steps can be skipped to apply manually set values, i.e. values will be taken from data set.

For L23 < 10 the values of SPW0 and PLW12 will be taken from the corresponding sino experiment.

PLW12 will be optimized if the specified PCPD2 pulse differs from PCPD2 used in the corresponding sino experiment.

For L23 > 10 or if the corresponding sino experiment has not been measured a manual user interaction part to set or optimize SPW0 and PLW12 is included.

SPW0 is calculated for  $B1(NUC2) = (B1(NUC1) + MASR) * rampFactor$ .

If the resulting power SPW0 exceeds the power limits of the probe, SPW0 will be recalculated for side band condition -1.

The selection of a certain side band condition can be forced by setting CNST 50 to 1 or -1.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	15000	15000	10000	10000	5000	4000
400	40000	15000	15000	10000	10000	5000	4000
500	40000	15000	15000	10000	10000	5000	4000
600	40000	15000	15000	10000	10000	5000	4000
700	40000	15000	15000	10000	10000	5000	4000
750	40000	15000	15000	10000	10000	5000	4000
800	40000	15000	15000	10000	10000	5000	4000
850	40000	15000	15000	10000	10000	5000	4000
900	40000	15000	15000	10000	10000	5000	4000
950	40000	15000	15000	10000	10000	5000	4000
1000	40000	15000	15000	10000	10000	5000	4000



## 5.4.42 Optimization of $^{79}\text{Br}$ frequency (NPT\_79Br\_MAS\_fieldsetting)


---

**Test Sample:** Potassium Bromide (KBr)  
Z151220, Z151270, Z151260, Z151250, Z151230, Z151240, Z151210, Z163271,  
Z183103

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

$^{79}\text{Br}$  spectrum after field optimization

### Control Option for Acquisition (L23)

1 default



## Parameters

<b>F1 ACQU</b> NUC1 79Br PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 101.000 O1P 59.700 ppm SWH 44247.789 Hz TD 4096 AQ 0.046 s FIDRES 21.605 Hz D 1 0.500 s P 1 4.0 us PLW 1 72.0 W TE 298.000 K	<b>Parameters</b>  Data Dimension   no optim.   90deg NUC1 Pow@90deg(Specs) NUC1 default
<b>F1 PROC</b> SI 8192 WDW 0 PH_mod 2 F1P 0.000 ppm F2P 0.000 ppm CY 11.000 cm	<b>Parameters</b>

## Experiment Description

The experimental procedure includes two 79Br acquisitions with constant O1 at two known FIELD positions. Using 59.7 ppm as chemical shift of 79Br in KBr the correct FIELD value can be calculated from these measurements. The determined field is only an approximation due to the temperature dependence of the chemical shift of 79Br in KBr.

Before acquisition of the final nmr spectrum NMRPT stores the resulting FIELD value in the BSMS.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	5000	5000	5000	5000
400	40000	8000	5000	5000	5000	5000	5000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	5000	5000	5000	5000	5000
700	40000	8000	5000	5000	5000	5000	5000
750	40000	8000	5000	5000	5000	5000	5000
800	40000	8000	5000	5000	5000	5000	5000
850	40000	8000	5000	5000	5000	5000	5000
900	40000	8000	5000	5000	5000	5000	5000
950	40000	8000	5000	5000	5000	5000	5000
1000	40000	8000	5000	5000	5000	5000	5000

## 5.4.43 Magic Angle setting, MAS (NPT\_79Br\_MAS\_magicAngle)


---

**Test Sample:** Potassium Bromide (KBr)  
Z151220, Z151270, Z151260, Z151250, Z151230, Z151240, Z151210, Z163271,  
Z183103

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

79Br spectrum of KBr with spinning side bands - documentation of magic angle adjustment.

### Control Option for Acquisition (L23)

- 1 default
- 2 manual adjustment of magic angle for probes with stator controlled by ATM

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1	79Br	SI	131072
PULPROG	onepulse	LB	0.000 Hz
NS	16	PH_mod	1 pk
DS	0	ABSF1	1000.000 ppm
RG	101.000	ABSF2	-1000.000 ppm
O1P	59.700 ppm	F1P	0.000 ppm
SWH	100000.000 Hz	F2P	0.000 ppm
TD	8192	CY	10.000 cm
AQ	0.041 s		
FIDRES	24.414 Hz		
D 1	0.250 s		
P 1	4.0 us		
PLW 1	126.0 W		
TE	298.000 K		
	no optim.		
	90deg Pulse		
	Pow@90deg(Specs)		
	default		

## Experiment Description

The experiment proves the accuracy of the magic angle which is to be optimized within first step. The line width of the main signal is calculated. Further spinning side bands to be analyzed may be specified by their side band number. A spectrum is plotted for documentation.

## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	5000	5000	5000	5000
400	40000	8000	5000	5000	5000	5000	5000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	5000	5000	5000	5000	5000
700	40000	8000	5000	5000	5000	5000	5000
750	40000	8000	5000	5000	5000	5000	5000
800	40000	8000	5000	5000	5000	5000	5000
850	40000	8000	5000	5000	5000	5000	5000
900	40000	8000	5000	5000	5000	5000	5000
950	40000	8000	5000	5000	5000	5000	5000
1000	40000	8000	5000	5000	5000	5000	5000

## 5.4.44 Maximum spin rate testing, MAS (NPT\_79Br\_MAS\_maxSpinRate)


---

**Test Sample:** Potassium Bromide (KBr)  
Z151220, Z151270, Z151260, Z151250, Z151230, Z151240, Z151210, Z163271,  
Z183103

**Solvent:** None

**Lock parameter:** None

**Sample State:** Magic Angle Spinning



### Example Printout

79Br spectrum of KBr with spinning side bands for documentation of maximum rotation frequency

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1	79Br	SI	32768
PULPROG	onepulse	LB	0.000 Hz
NS	16	PH_mod	1 pk
DS	0	ABSF1	1000.000 ppm
RG	101.000	ABSF2	-1000.000 ppm
O1P	59.700 ppm	F1P	0.000 ppm
SWH	250000.000 Hz	F2P	0.000 ppm
TD	16384	CY	10.000 cm
AQ	0.033 s		
FIDRES	30.518 Hz		
D 1	0.250 s		
P 1	4.0 us		
PLW 1	126.0 W		
TE	298.000 K		
	no optim.		
	90deg Pulse		
	Pow@90deg(Specs)		
	default		

## Experiment Description

The experiment proves that maximum rotation frequency (default or as specified) can be reached. If the maximum rotation frequency is stable, maximum deviation during a period of about 3 minutes is evaluated and a spectrum is acquired. A spectrum showing the rotation side bands is plotted for documentation.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	111000	67000	42000	35000	24000	15000	7000
400	111000	67000	42000	35000	24000	15000	7000
500	111000	67000	42000	35000	24000	15000	7000
600	111000	67000	42000	35000	24000	15000	7000
700	111000	67000	42000	35000	24000	15000	7000
750	111000	67000	42000	35000	24000	15000	7000
800	111000	67000	42000	35000	24000	15000	7000
850	111000	67000	42000	35000	24000	15000	7000
900	111000	67000	42000	35000	24000	15000	7000
950	111000	67000	42000	35000	24000	15000	7000
1000	111000	67000	42000	35000	24000	15000	7000

## 5.4.45 P90 79Br pulse calibration, MAS (NPT\_79Br\_MAS\_p90det\_79br)

**Test Sample:** Potassium Bromide (KBr)  
 Z151220, Z151270, Z151260, Z151250, Z151230, Z151240, Z151210, Z163271,  
 Z183103  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.
- 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
- 12 skip automatic phase correction and apply manually set values, execute O1P determination.
- 100 same as xx but skip of SINO check on PROCNO 11
- +XX
- 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)
- +XXX

## Parameters

<b>F1 ACQU</b> NUC1 79Br PARMODE 0 PULPROG onepulse NS 1 DS 0 RG 101.000 O1P 59.700 ppm SWH 100000.000 Hz TD 2048 AQ 0.010 s FIDRES 97.656 Hz D 1 0.250 s P 1 4.0 us PLW 1 106 W TE 298.000 K	Parameters Data Dimension no optim. 90deg NUC1 Pow@90deg(Specs) NUC1 default
<b>F1 PROC</b> SI 4096 LB 0.000 Hz PH_mod 1	Parameters pk

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table. L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded. Phase correction is done applying APK as default method. With the option L23 = 2 and 12 O1P determination step can be skipped to apply manually set values. The options L23 = 11 and 12 skip automatic phase correction for manual settings to be applied. After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options. If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated with the determined pulse and the power used. Results are stored under PROCNO 999. This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required. Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed.


## Spinrate

MASR [Hz] as function of sample diameter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	5000	5000	5000	5000
400	40000	8000	5000	5000	5000	5000	5000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	5000	5000	5000	5000	5000
700	40000	8000	5000	5000	5000	5000	5000
750	40000	8000	5000	5000	5000	5000	5000
800	40000	8000	5000	5000	5000	5000	5000
850	40000	8000	5000	5000	5000	5000	5000
900	40000	8000	5000	5000	5000	5000	5000
950	40000	8000	5000	5000	5000	5000	5000
1000	40000	8000	5000	5000	5000	5000	5000

## 5.4.46 Temperature controll test on KBr, MAS (NPT\_79Br\_MAS\_temperatureTestKBr)

**Test Sample:** Potassium Bromide (KBr)  
Z151220, Z151270, Z151260, Z151250, Z151230, Z151240, Z151210, Z163271,  
Z183103  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Top: <sup>79</sup>Br spectrum of KBr with spinning side bands at lowest measured temperature.  
Bottom: <sup>79</sup>Br spectrum of KBr with spinning side bands at highest measured temperature.

### Control Option for Acquisition (L23)

- 1 default
- 2 ignore specifications and take parameters from D 60, CNST 50, 51, and 52 instead
- 11 activation of newly determined temperature correction in the BSVT
- 12 ignore specification and activation of newly determined temperature correction in the BSVT



## Parameters

<b>F1 ACQU</b> NUC1 79Br PULPROG onepulse NS 8 DS 0 RG 101.000 O1P 59.700 ppm SWH 5000.000 Hz TD 4096 AQ 0.410 s FIDRES 2.441 Hz D 1 0.250 s D 60 180.000 s P 1 4.0 us PLW 1 126.0 W TE 298.000 K	Parameters  no optim.  teready waiting time 90deg Pulse Pow@90deg(Specs) default
<b>F1 PROC</b> SI 4096 PH_mod 1 F1P 65.000 ppm F2P 55.000 ppm CY 10.000 cm <b>NMRPT</b> CNST 50 1.000 K CNST 51 1.000 K CNST 52 1.000	Parameters  pk  Parameters min TE for calibr. max TE for calibr. no. points TE calibr.

## Experiment Description

Temperature test expects a temperature unit. The procedure is acquiring at least two experiments with different TE settings according to the specifications or for L23==2 entries in the parameter set (CNST 50 to 53).

The result of the experiment is the determination of the offset and slope of the correction curve ( $y=ax+b$ ). The correction will not be set in the temperature unit.

Due to the fact that only the chemical shift is used for temperature determination, the true temperature can not be determined with this experiment. Hence, this is no temperature calibration experiment.

Temperature correction will not be disabled during the experiments. If L23==1 or L23==2 the original values will be retained in the BSVT at the end of the test. If L23==11 odr L23==12 the newly determined correction values will be activated in the BSVT.

If L23==2 or L23==12 the specifications for highest and lowest TE and number of points are ignored and will be taken from the following constants:

CNST 50: lowest TE used for calibration.

CNST 51: highest TE used for calibration.

CNST 52: number of points (TE settings) used for calibration.

If no specifications are given and L23==1 the following default values will be used:

Three experiments with 298, 308, and 318 K are acquired.

## Spinrate

MASR [Hz] as function of sample diamter and spectrometer frequency

	0.7	1.3	1.9	2.5	3.2	4.0	7.0
300	40000	8000	5000	5000	5000	5000	5000
400	40000	8000	5000	5000	5000	5000	5000
500	40000	8000	5000	5000	5000	5000	5000
600	40000	8000	5000	5000	5000	5000	5000
700	40000	8000	5000	5000	5000	5000	5000
750	40000	8000	5000	5000	5000	5000	5000
800	40000	8000	5000	5000	5000	5000	5000
850	40000	8000	5000	5000	5000	5000	5000
900	40000	8000	5000	5000	5000	5000	5000
950	40000	8000	5000	5000	5000	5000	5000
1000	40000	8000	5000	5000	5000	5000	5000

## 5.5 Experiments for High Resolution Magic Angle Spinning Probes (HRMAS)

### Sample Rotation Frequency

The parameter MASR is set to 4 kHz by *NMRPT*. The parameter RO is set to zero, commands ROTON and ROTOFF are ignored.


The experiments described in this section as well as the following HR experiments are provided for HRMAS probes:

NPT\_13C\_sensitivity  
NPT\_13C\_sensitivity\_inno  
NPT\_15N\_fullsw\_inept  
NPT\_15N\_sensitivity\_dec1h  
NPT\_15N\_sensitivity\_inept  
NPT\_1H\_b1homogeneityInt\_13c  
NPT\_1H\_b1homogeneityInt\_15n  
NPT\_1H\_b1homogeneityInt\_1h  
NPT\_1H\_backgr\_withsample  
NPT\_1H\_cpmgtestf2\_13c  
NPT\_1H\_cpmgtestf2\_15n  
NPT\_1H\_gradientprofile\_neg  
NPT\_1H\_gradientprofile\_pos  
NPT\_1H\_gradrec\_stest\_1h  
NPT\_1H\_gradrecZ\_sqn\_1h  
NPT\_1H\_gradrecZ\_sqp\_1h  
NPT\_1H\_inno  
NPT\_1H\_inno  
NPT\_1H\_p90determinationf1\_1h  
NPT\_1H\_p90determinationf2\_13c  
NPT\_1H\_p90determinationf2\_15n  
NPT\_1H\_quant\_ref  
NPT\_1H\_sensitivity  
NPT\_1H\_sensitivity\_dec19f  
NPT\_1H\_sensitivity\_inno  
NPT\_1H\_vibration\_doped\_water  
NPT\_1H\_vibration\_lineshape  
NPT\_19F\_b1homogeneityInt\_19f  
NPT\_19F\_backgr\_withsample  
NPT\_19F\_fullsw\_dec1h  
NPT\_19F\_p90determinationf1\_19f  
NPT\_19F\_sensitivity  
NPT\_19F\_sensitivity\_inno  
NPT\_19F\_sensitivity\_lb05\_dec1h  
NPT\_31P\_b1homogeneityInt\_31p  
NPT\_31P\_p90determinationf1\_31p  
NPT\_31P\_sensitivity  
NPT\_31P\_sensitivity\_dec1h  
NPT\_31P\_sensitivity\_inno  
NPT\_prep\_atma\_test  
NPT\_prep\_fieldsetting\_d  
NPT\_prep\_locksettings\_d  
NPT\_prep\_b1homogeneityInt\_d  
NPT\_prep\_p90det\_d  
NPT\_prep\_sensitivity\_10\_d  
NPT\_29Si\_p90determination\_29si  
NPT\_29Si\_sensitivity



## 5.5.1 <sup>1</sup>H lineshape with magic angle spinning (NPT\_1H\_HRMAS\_lineshape)

**Test Sample:** 1.0% Chloroform in Acetone-D6  
Z142220  
**Solvent:** Acetone  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Magic Angle Spinning



### Example Printout

Proton line shape spectrum with offset calibration to O1P as overview spectrum. Printing is from +125.0 Hz to -125.0 Hz. The expansion plot to the left shows chloroform signal with higher resolution (+7.0 Hz, -7.0 Hz).

### Control Option for Acquisition (L23)

- 1 default
- 2 write default shimfile, in case of successful evaluation

## Parameters

<p><b>F1 ACQU</b></p> <p>NUC1 1H          PULPROG zg30          NS 4          DS 0          RG 101.000          O1P 7.700 ppm          SWH 1000.000 Hz          TD 32768          AQ 16.384 s          FIDRES 0.061 Hz          D 1 9.116 s          P 1 4.5 us          PLW 1 16.6 W          TE 298.000 K</p> <p style="text-align: right;">Parameters</p> <p style="text-align: right;">no optim.</p> <p style="text-align: right;">AQ+D1=const          90deg Pulse          Pow@90deg(Specs)          default</p>	<p><b>F1 PROC</b></p> <p>SI 16384          WDW 0          LB 0.000 Hz          PC 1.000          F1P 8.640 ppm          F2P 7.440 ppm          CY 1000.000 cm</p> <p><b>NMRPT</b></p> <p>CNST 50 0.200</p> <p style="text-align: right;">Parameters</p> <p style="text-align: right;">Scaling factor for CY</p>
--	---

## Experiment Description


Before acquiring the final spectrum, the exact signal position is determined with NS=1. The carrier position is afterwards set to the position optimized O1 = peak frequency [Hz] - (SWH/4)

This procedure allows the application of a strip transformation with the parameters STSR=0, STSI=SI/2 resulting in the signal position on-resonance and in most cases resulting in an optimal automatic phase correction.

Setting L23=2, it is possible to store the standard shimfile provided the evaluation of the experiment is successful. This event takes place during acquisition only. During regular processing of the data no shimfile is stored.

## 5.5.2 Watersuppression (NPT\_1H\_HRMAS\_watersuppression)

**Test Sample:** 2 mM Sucrose, 0.5 mM DSS, and 2 mM NaN<sub>3</sub> in 90% H<sub>2</sub>O + 10% D<sub>2</sub>O  
 Z142222  
**Solvent:** H<sub>2</sub>O+D<sub>2</sub>O  
**Lock parameter:** AUTOGAIN, lock regulation according to actual Edlock Table  
**Sample State:** Magic Angle Spinning



### Example Printout

Water suppression spectrum printed from 10.5 ppm to -1.0 ppm as overview spectrum. The expansion plot to the right shows the DSS (sodium 2,2-dimethyl-2-silapentane-5-sulphonate) signal, indicative for the quality of the spectral resolution. The quality of the water suppression is determined as linewidth of the residual water signal using the 10% and 50% intensity of the DSS signal.

The expansion plot to the left shows the signal of the anomeric proton (alpha-C1) of the sucrose (2-beta-D-fructofuranosyl-1-alpha-D-glucopyranosid). The anomeric proton intensity is used for the signal-to-noise calculation and the splitting of the signal is analysed to give a measure for spectral resolution.

### Control Option for Acquisition (L23)

- 1 with O1 optimization, with PULPROG=zgpr
- 20 O1 from parameter set, no optimization, with PULPROG=zgpr
- 23 O1 from parameter set, no optimization, with PULPROG=npt\_zggppr
- 25 O1 from parameter set, no optimization, with PULPROG=zgcppr
- 27 O1 from parameter set, no optimization, with PULPROG=zgcppppr
- 30 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgpr
- 33 O1 from previous watersuppression, no optimization[\*], with PULPROG=npt\_zggppr
- 35 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppr
- 37 O1 from previous watersuppression, no optimization[\*], with PULPROG=zgcppppr
- [\*] The optimization of O1 is enforced, if O1 was not determined during a previous measurement.

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	32768			
PULPROG	zgpr				WDW	1			
NS	8				LB	0.000	Hz		
DS	4				PC	0.100			
RG	0.250		optim. by NMRPT		F1P	10.806	ppm		
O1P	4.699	ppm			F2P	-1.227	ppm		
SW	12.132	ppm			CY	111.000	cm		
TD	10194		field dependent						
AQ	1.050	s							
FIDRES	0.952	Hz	field dependent						
D 1	5.000	s							
P 1	14.0	us	90deg						
PLW 1	7.5	W	Pow@90deg(Specs)						
PLW 9	0.00006	W	Pow@90deg(5000u)						
TE	298.000	K	default						

## Experiment Description

For the watersuppression experiment the carrier offset O1 needs to be very carefully adjusted. Using the control option for acquisition L23=1, O1 is determined by an intensity profile(second point of FID) over a range of 16 Hz in 0.4 Hz steps around submitted O1.


RG adjustment is executed by running experiments with DS=0, NS=4 and a reduced PLW1. The highest RG without signal overflow in these experiments will be used.

Options L23=1 and 20 are standard whereas Options L23=23, 25, and 27 are non-standard and will not be considered as regular test from the 'NMRPT Control Structure'. Their existence is purely diagnostic. Experiment will be set to irregular if option 'Skip Temperature' is selected.

## 5.5.3 Magic Angle setting, HRMAS (NPT\_79Br\_HRMAS\_magicAngle)

---

**Test Sample:** Potassium Bromide (KBr)  
Z151220  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

79Br spectrum of KBr with spinning side bands - documentation of magic angle adjustment.

### Control Option for Acquisition (L23)

- 1 default
- 2 manual adjustment of magic angle for probes with stator controlled by ATM



## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	79Br				SI	131072			
PULPROG	zg				LB	0.000	Hz		
NS	16				PH_mod	1		pk	
DS	0				ABSF1	1000.000	ppm		
RG	101.000		no optim.		ABSF2	-1000.000	ppm		
O1P	59.700	ppm			F1P	0.000	ppm		
SWH	100000.000	Hz			F2P	0.000	ppm		
TD	8192				CY	10.000	cm		
AQ	0.041	s							
FIDRES	24.414	Hz							
D 1	0.250	s							
P 1	4.0	us	90deg Pulse						
PLW 1	126.0	W	Pow@90deg(Specs)						
TE	298.000	K	default						


## Experiment Description

The experiment proves the accuracy of the magic angle which is to be optimized within first step. The line width of the main signal is calculated. Further spinning side bands to be analyzed may be specified by their side band number. A spectrum is plotted for documentation. LOCNUC is set to off as a default.

## 5.5.4 Maximum spin rate testing, HRMAS (NPT\_79Br\_HRMAS\_maxSpinRate)

---

**Test Sample:** Potassium Bromide (KBr)  
Z151220  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

79Br spectrum of KBr with spinning side bands for documentation of maximum rotation frequency

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1	79Br	SI	16384
PULPROG	zg	LB	0.000 Hz
NS	16	PH_mod	1 pk
DS	0	ABSF1	1000.000 ppm
RG	101.000	ABSF2	-1000.000 ppm
O1P	59.700 ppm	F1P	0.000 ppm
SWH	200000.000 Hz	F2P	0.000 ppm
TD	8192	CY	10.000 cm
AQ	0.020 s		
FIDRES	48.828 Hz		
D 1	0.250 s		
P 1	4.0 us		
PLW 1	126.0 W		
TE	298.000 K		
	no optim.		
	90deg Pulse		
	Pow@90deg(Specs)		
	default		

## Experiment Description

The experiment proves that maximum rotation frequency (default or as specified) can be reached. If the maximum rotation frequency is stable, maximum deviation during a period of about 3 minutes is evaluated and a spectrum is acquired. A spectrum showing the rotation side bands is plotted for documentation. LOCNUC is set to off as a default.

## 5.5.5 P90 79Br pulse calibration, HRMAS (NPT\_79Br\_HRMAS\_p90det\_79br)

**Test Sample:** Potassium Bromide (KBr)  
 Z151220  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



```

Report of AD program poptau
F1F=71.556ppm F2F=51.556ppm
Linear optimisation of P1 in 6 steps.
PROCNO=999
Starting at 4.000000000 endval= 12.000000000
OPTIMIZE ZERO
VARIABLE LIST
Experiment P1 Maximum point Minimum point Integral
1 4.0000 312019178 -2311476 21709686.281
2 5.4000 322893506 -2205094 32981440.482
3 7.2000 83444711 -1127739 5634521.880
4 8.9000 2745491 -7536243 -4460044.504
5 10.4000 1817774 -140946870 -10243848.193
6 12.0000 2439368 -155546443 -9851386.747
poptau for P1 finished.
ZERO at experiment 3.514300: P1 = 8.024093
  
```



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 999).

### Control Option for Acquisition (L23)

- 1 default, skip O1P determination.
- 2 execute O1P determination.
- 11 skip automatic phase correction and apply manually set values, skip O1P and O2P determination.
- 12 skip automatic phase correction and apply manually set values, execute O1P determination.
- 100 same as xx but skip of SINO check on PROCNO 11
- +XX
- 1000same as xxx but ignore specifications (optimize power for pulse length from prosol)
- +XXX

## Parameters

<b>F1 ACQU</b> NUC1 79Br PARMODE 0 PULPROG zg NS 1 DS 0 RG 101.000 O1P 59.700 ppm SWH 100000.000 Hz TD 2048 AQ 0.010 s FIDRES 97.656 Hz D 1 0.250 s P 1 4.0 us PLW 1 106 W TE 298.000 K	<b>Parameters</b>  Data Dimension  no optim.  90deg NUC1 Pow@90deg(Specs) NUC1 default
<b>F1 PROC</b> SI 4096 LB 0.000 Hz PH_mod 1	<b>Parameters</b>  pk

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

L23 dependent offset optimization: the exact signal offset (O1P) of the main signal is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.

Phase correction is done applying APK as default method. With the option L23 = 2 and 12 O1P determination step can be skipped to apply manually set values. The options L23 = 11 and 12 skip automatic phase correction for manual settings to be applied.

After phase correction the resulting spectrum is stored in PROCNO 12, the FID is discarded. SINO will be executed on PROCNO 11. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped using the corresponding L23 options.


If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [90%, 100%], it calculates a new power value based on the specified pulse length and repeats the measurement with the new power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n. The PROSOL table will be updated with the determined pulse and the power used. Results are stored under PROCNO 999.

This experiment includes a manual user interaction part, e.g. to manual match/tune or prepare acquisition as required.

Phase correction is applied during acquisition and will not take any effect if experiment is reprocessed. LOCNUC is set to off as a default.

## 5.5.6 Temperature controll test on KBr, HRMAS (NPT\_79Br\_HRMAS\_temperatureTestKBr)

**Test Sample:** Potassium Bromide (KBr)  
Z151220  
**Solvent:** None  
**Lock parameter:** None  
**Sample State:** Magic Angle Spinning



### Example Printout

Top: <sup>79</sup>Br spectrum of KBr with spinning side bands at lowest measured temperature.  
Bottom: <sup>79</sup>Br spectrum of KBr with spinning side bands at highest measured temperature.

### Control Option for Acquisition (L23)

- 1 default
- 2 ignore specifications and take parameters from D 60, CNST 50, 51, and 52 instead
- 11 activation of newly determined temperature correction in the BSVT
- 12 ignore specification and activation of newly determined temperature correction in the BSVT

## Parameters

<b>F1 ACQU</b>				Parameters			
NUC1	79Br						
PULPROG	zg						
NS	8						
DS	0						
RG	101.000		no optim.				
O1P	59.700	ppm					
SWH	5000.000	Hz					
TD	4096						
AQ	0.410	s					
FIDRES	2.441	Hz					
D 1	0.250	s					
D 60	180.000	s	teready waiting time				
P 1	4.0	us	90deg Pulse				
PLW 1	126.0	W	Pow@90deg(Specs)				
TE	298.000	K	default				

<b>F1 PROC</b>				Parameters			
SI	4096						
PH_mod	1					pk	
F1P	65.000	ppm					
F2P	55.000	ppm					
CY	10.000	cm					
<b>NMRPT</b>				Parameters			
CNST 50	1.000	K	min TE for calibr.				
CNST 51	1.000	K	max TE for calibr.				
CNST 52	1.000		no. points TE calibr.				

## Experiment Description

Temperature test expects a temperature unit. The procedure is acquiring at least two experiments with different TE settings according to the specifications or for L23==2 entries in the parameter set (CNST 50 to 53).

The result of the experiment is the determination of the offset and slope of the correction curve ( $y=ax+b$ ). The correction will not be set in the temperature unit.

Due to the fact that only the chemical shift is used for temperature determination, the true temperature can not be determined with this experiment. Hence, this is no temperature calibration experiment.

Temperature correction will not be disabled during the experiments. If L23==1 or L23==2 the original values will be retained in the BSVT at the end of the test. If L23==11 odr L23==12 the newly determined correction values will be activated in the BSVT.

If L23==2 or L23==12 the specifications for highest and lowest TE and number of points are ignored and will be taken from the following constants:

CNST 50: lowest TE used for calibration.

CNST 51: highest TE used for calibration.

CNST 52: number of points (TE settings) used for calibration.

If no specifications are given and L23==1 the following default values will be used:

Three experiments with 298, 308, and 318 K are acquired.

LOCNUC is set to off as a default.


## 5.6 Experiments for Fourier Spectrometer (CMR)

As a prerequisite for NMRPT the system must be shimmed and the probe must be locked with the external lock. Experiments are acquired without sample rotation.

Experiments for Fourier Spectrometer are only implemented for TopSpin 4.1.0 and higher.

## 5.6.1 P90 1H pulse calibration (NPT\_1H\_CM\_R\_p90determination\_1h)

**Test Sample:** 1 M Methanol-13C in D2O  
 Z10692  
**Solvent:** D2O  
**Lock parameter:** External Lock  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 100).

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX



## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 1H		SI 2048	
PARMODE 0	Data Dimension	WDW 3	
PULPROG zg		LB 0.750 Hz	
NS 1		SSB 2.000	
DS 0		PH_mod 1	pk
RG 1.000	no optim.	ME_mod 2	LPfc
O1P 3.416 ppm		NCOEF 20	
SWH 230.766 Hz		ABSF1 1000.000 ppm	
AQ 0.650 s		ABSF2 -1000.000 ppm	
FIDRES 1.538 Hz		F1P 5.496 ppm	
D 1 15.318 s	AQ+D1=const	F2P 5.096 ppm	
P 1 11.0 us	90deg NUC1	CY 11.000 cm	
PLW 1 1.1 W	Pow@90deg(Specs) NUC1		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.


The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

## 5.6.2 P90 13C pulse calibration (NPT\_13C\_CMV\_p90determination\_13c)

**Test Sample:** 1 M Methanol-13C in D2O  
 Z10692  
**Solvent:** D2O  
**Lock parameter:** External Lock  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 100).

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000 Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters	F1 PROC	Parameters
NUC1 13C		SI 8192	
NUC2 1H		WDW 1	
PARMODE 0	Data Dimension	LB 0.300 Hz	
PULPROG zgpg		SSB 0.000	
NS 1		PH_mod 1	pk
DS 0		ME_mod 0	LPfc
RG 1.000	no optim.	NCOEF 0	
O1P 49.435 ppm		ABSF1 240.000 ppm	
O2P 3.590 ppm		ABSF2 -10.000 ppm	
SWH 8196.722 Hz		F1P 180.000 ppm	
AQ 0.500 s		F2P -20.000 ppm	
FIDRES 2.001 Hz		CY 11.000 cm	
D 1 29.500 s	AQ+D1=const		
P 1 17.2 us	90deg NUC1		
PCPD 2 120.0 W	PCPD NUC2		
PLW 1 4.8 W	Pow@90deg(Specs) NUC1		
PLW 12 0.018 W	Pow@CPD NUC2		
PLW 13 0.009 W	Pow@CPD NOE NUC2		
TE 298.000 K	default		

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.


The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skiped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

## 5.6.3 1H lineshape (NPT\_1H\_CMV\_lineshape)

**Test Sample:** 20% Chloroform in Acetone-D6  
Z10689  
**Solvent:** Acetone  
**Lock parameter:** External Lock  
**Sample State:** Rotation off



### Example Printout

Proton line shape spectrum with offset calibration to O1P as overview spectrum. Printing is from +125.0 Hz to -125.0 Hz. The expansion plot to the left shows chloroform signal with higher resolution (+7.0 Hz, -7.0 Hz).

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU				Parameters	F1 PROC				Parameters
NUC1	1H				SI	16384			
PULPROG	zg30				WDW	0			
NS	1				LB	0.000	Hz		
DS	0				PC	1.000			
RG	1.000		no optim.		F1P	8.640	ppm		
O1P	7.410	ppm			F2P	7.440	ppm		
SWH	1000.000	Hz			CY	1000.000	cm		
TD	32768				<b>NMRPT</b>				Parameters
AQ	16.384	s			CNST 50	0.200			Scaling factor for CY
FIDRES	0.061	Hz							
D 1	9.116	s	AQ+D1=const						
P 1	11.0	us	90deg Pulse						
PLW 1	1.1	W	Pow@90deg(Specs)						


## Experiment Description

Before acquiring the final spectrum, the exact signal position is determined with NS=1. The carrier position is afterwards set to the position optimized  $O1 = \text{peak frequency [Hz]} - (SWH/4)$

This procedure allows the application of a strip transformation with the parameters STSR=0, STSI=SI/2 resulting in the signal position on-resonance and in most cases resulting in an optimal automatic phase correction.

## 5.6.4 1H sensitivity (NPT\_1H\_CMV\_sensitivity)

**Test Sample:** 1% Ethylbenzene (EB) in Chloroform-D  
Z10690  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** External Lock  
**Sample State:** Rotation off



### Example Printout

Bottom: 1H overview spectrum of ethylbenzene.

Top right: Expanded region showing the methylene signal used for evaluation.

### Control Option for Acquisition (L23)

1 default


## Parameters

F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	1H			SI	16384		
PULPROG	zg			LB	1.000	Hz	1.0
NS	1			SIGF1	3.500	ppm	
DS	0			SIGF2	1.900	ppm	
RG	1.000		no optim.	NOISF1	20.000	ppm	
O1P	4.000	ppm		NOISF2	8.000	ppm	
SW	36.753	ppm		CY	11.000	cm	
TD	16384						
AQ	0.557	s	field dependent				
FIDRES	1.795	Hz	field dependent				
D 1	113.574	s					
P 1	11.0	us	90deg Pulse				
PLW 1	1.1	W	Pow@90deg(Specs)				
TE	298.000	K	default				

## Experiment Description

## 5.6.5 1H background with sample (NPT\_1H\_CMV\_background)

**Test Sample:** 1% Ethylbenzene (EB) in Chloroform-D  
Z10690  
**Solvent:** CDCl<sub>3</sub>  
**Lock parameter:** External Lock  
**Sample State:** Rotation off



### Example Printout

Top: 1H Background signal spectrum with sample. Sharp signal arises from sample, broad signal could arise from solid compound in the probe.

Bottom: Same spectrum scaled to see baseline distortions

### Control Option for Acquisition (L23)

1 default



## Parameters

<b>F1 ACQU</b>			Parameters	<b>CNST 10</b>	45.000		Flip angle for P90
NUC1	1H			<b>F1 PROC</b>			Parameters
PULPROG	npt_zg0			SI	16384		
NS	32			WDW	1		
DS	0			LB	1.000	Hz	
RG	1.000		no optim.	CY	11.000	cm	
O1P	4.000	ppm		<b>NMRPT</b>			Parameters
SWH	7812.500	Hz		CNST 50	10.000		Scaling factor for CY
TD	16384						
AQ	1.049	s					
FIDRES	0.954	Hz					
D 1	3.116	s	AQ+D1=const				
P 0	5.5	us	P 1 * CNST 10 / 90				
P 1	11.0	us	90deg Pulse				
PLW 1	1.1	W	Pow@90deg(Specs)				
TE	298.000	K	default				

## Experiment Description


Background signal measurements are executed using a small flip angle to compensate for usually long T1 relaxation time of solid signals.

Full spectrum is normally printed without any baseline correction.

## 5.6.6 <sup>13</sup>C sensitivity with 1H decoupling (NPT\_13C\_CMRSensitivity\_dec1h)

---

**Test Sample:** 100% Ethylbenzene (EB)  
Z10694  
**Solvent:** None  
**Lock parameter:** External Lock  
**Sample State:** Rotation off



### Example Printout

Carbon-13 sensitivity test with 1H decoupling.

### Control Option for Acquisition (L23)

1 default

## Parameters


F1 ACQU			Parameters	F1 PROC			Parameters
NUC1	13C			SI	262144		
PULPROG	zgpg			WDW	1		
NS	1			LB	0.300	Hz	
DS	0			PC	1.400		
RG	1.000		no optim.	F1P	180.000	ppm	
O1P	80.000	ppm		F2P	-20.000	ppm	
O2P	19.970	ppm		CY	11.000	cm	
SW	198.766	ppm					
TD	262144						
AQ	6.554	s	field dependent				
FIDRES	0.153	Hz	field dependent				
D 1	823.446	s	AQ+D1=const				
P 1	17.2	us	90deg NUC1				
PCPD 2	120.0	W	PCPD NUC2				
PLW 1	4.8	W	Pow@90deg(Specs) NUC1				
PLW 12	0.018	W	Pow@CPD NUC2				
PLW 13	0.009	W	Pow@CPD NOE NUC2				
CPDPRG2	waltz64		decoupl. sequence				
DIGMOD	3		baseopt				
TE	298.000	K	default				

## Experiment Description

Carbon-13 sensitivity test with 1H decoupling. Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from 140.0 to 124.0 ppm, while the best 40 ppm noise region is determined over the range from 124.0 to 80.0 ppm.

## 5.6.7 P90 31P pulse calibration (NPT\_31P\_CM\_R\_p90determination\_31p)

**Test Sample:** 0.485 M Triphenyl Phosphate (TPP, [C6H5]3PO4) in Acetone-D6  
**CMR\_TPP**  
**Solvent:** Acetone  
**Lock parameter:** External Lock  
**Sample State:** Rotation off



### Example Printout

P90 pulse determination using 1D method by varying the excitation pulse. The result of a CONVTO1D routine shows six experiments from 90 to 270 deg (PROCNO 100).

### Control Option for Acquisition (L23)

- 1 default
- 2 skip SINO check on PROCNO 11
- 11 ignore specifications (optimize power for pulse length from prosol)
- 12 ignore specifications and skip SINO check on PROCNO 11
- 21 ignore specifications and optimize pulse length for power from prosol
- 22 ignore specs, skip SINO check, and optimize pulse length for power from prosol
- 100 Same as xx but skip automatic phase correction and apply manually set values.
- +XX
- 1000Same as xxx but skip automatic O1P determination
- +XXX

## Parameters

F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	31P			SI	32768		
NUC2	1H			WDW	1		
PARMODE	0		Data Dimension	LB	2.000	Hz	
PULPROG	zgig			SSB	0.000		
NS	1			PH_mod	1		pk
DS	0			ME_mod	0		LPfc
RG	1.000		no optim.	NCOEF	0		
O1P	0.000	ppm		ABSF1	-17.332	ppm	
O2P	5.000	ppm		ABSF2	-17.860	ppm	
SWH	8196.722	Hz		F1P	-17.357	ppm	
AQ	0.393	s		F2P	-17.835	ppm	
FIDRES	2.542	Hz		CY	11.000	cm	
D 1	17.757	s	AQ+D1=const				
P 1	10.5	us	90deg NUC1				
PCPD 2	120.0	W	PCPD NUC2				
PLW 1	5.0	W	Pow@90deg(Specs) NUC1				
PLW 12	0.016	W	Pow@CPD NUC2				
TE	298.000	K	default				

## Experiment Description

This experiment depends on the entry of the 90 degree pulse in the PROSOL table.

As usual for pulse determinations the exact signal offset (O1P) is first determined. The resulting spectrum is stored in PROCNO 10 in magnitude mode, the FID is discarded.


The next step is the determination of the correct phase of the signal using the corrected O1P. The result is stored in PROCNO 11, the FID is discarded. After phase correction SINO will be executed. The experiment will be aborted, if a minimal SINO of 10 is not achieved. SINO check can be skipped with L23=2, 12, or 22

If the evaluation shows a deviation for the 90 degree pulse which is not in the interval [97%, 100%], a new power value will be calculated based on the specified pulse length, followed by the update of the PROSOL table and a new execution of the measurement with the newly set power. This cycle can be repeated up to six times and its results are stored under PROCNO 100+n and 200, 201.

A protocol is shown on the lower right side of the printout, containing a table with PLW90, P90, P90[det], Deviation (%).

## 5.6.8 $^{31}\text{P}$ sensitivity with $^1\text{H}$ decoupling (NPT\_31P\_CMRSensitivity\_dec1h)

**Test Sample:** 0.485 M Triphenyl Phosphate (TPP,  $[\text{C}_6\text{H}_5]_3\text{PO}_4$ ) in Acetone-D6  
CMR\_TPP  
**Solvent:** Acetone  
**Lock parameter:** External Lock  
**Sample State:** Rotation off



### Example Printout

Phosphorous-31 sensitivity test with  $^1\text{H}$  decoupling.

### Control Option for Acquisition (L23)

1 default

## Parameters

F1 ACQU	Parameters			F1 PROC	Parameters		
NUC1	31P			SI	32768		
NUC2	1H			WDW	1		
PULPROG	zgig			LB	0.500	Hz	
NS	1			PC	1.400		
DS	0			F1P	-17.357	ppm	
RG	1.000		no optim.	F2P	-17.835	ppm	
O1P	-17.000	ppm		CY	11.000	cm	
O2P	5.000	ppm					
SW	50.606	ppm					
TD	6450						
AQ	0.393	s	field dependent				
FIDRES	2.542	Hz	field dependent				
D 1	120.607	s	AQ+D1=const				
P 1	10.5	us	90deg NUC1				
PCPD 2	120.0	W	PCPD NUC2				
PLW 1	5.0	W	Pow@90deg(Specs) NUC1				
PLW 12	0.016	W	Pow@CPD NUC2				
CPDPRG2	waltz64		decoupl. sequence				
DIGMOD	1		baseopt				
TE	298.000	K	default				

## Experiment Description

Phosphorous-31 sensitivity test with 1H decoupling. Processing is using line broadening (LB) and baseline correction (ABS). Evaluation is carried out by the AU `sinocal`. The signal is searched over the range from -15.0 to -21.0 ppm, while the best 5 ppm noise region is determined over the range from 0.0 to -36.0 ppm.

## 6 Appendix

### 6.1 Sample List

Sample ID	Diameter	Description
CD3CN	n.a.	Acetonitrile-D <sub>3</sub>
CH3CN_D2O	n.a.	Acetonitrile/D <sub>2</sub> O
CMR_TPP	5.0	0.485 M Triphenyl Phosphate (TPP, [C <sub>6</sub> H <sub>5</sub> ] <sub>3</sub> PO <sub>4</sub> ) in Acetone-D <sub>6</sub>
H177072	5.0	Pseudo Honey
H5798	n.a.	5 ug/ul of 1,3,5-Trimethoxybenzene (C <sub>6</sub> H <sub>3</sub> (OCH <sub>3</sub> ) <sub>3</sub> ) in Acetonitrile/D <sub>2</sub> O 70/30
H5799-D2O	n.a.	Methyl-, Ethyl-, Propyl- and Butyl-Ester of para-Hydroxybenzoic Acid, 6.57 mM each in Acetonitrile/D <sub>2</sub> O 30/70
H5799-SPE	n.a.	Methyl-, Ethyl-, Propyl- and Butyl-Ester of para-Hydroxybenzoic Acid, 6.57 mM each in Acetonitrile/D <sub>2</sub> O 30/70
H5799-SPE-1.7	1.7	Methyl-, Ethyl-, Propyl- and Butyl-Ester of para-Hydroxybenzoic Acid, 6.57 mM each in Acetonitrile/D <sub>2</sub> O 30/70
H5799-SPE-3	3.0	Methyl-, Ethyl-, Propyl- and Butyl-Ester of para-Hydroxybenzoic Acid, 6.57 mM each in Acetonitrile/D <sub>2</sub> O 30/70
H5799-SPE-5	5.0	Methyl-, Ethyl-, Propyl- and Butyl-Ester of para-Hydroxybenzoic Acid, 6.57 mM each in Acetonitrile/D <sub>2</sub> O 30/70
H7284	n.a.	0.5% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
H7284-01	n.a.	1% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
H7284-02	n.a.	3% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
H7285	n.a.	2nM Sucrose in D <sub>2</sub> O
H9630	n.a.	800 ng of 1,3,5-Trimethoxybenzene (C <sub>6</sub> H <sub>3</sub> (OCH <sub>3</sub> ) <sub>3</sub> ) in 30 ul Acetonitrile/D <sub>2</sub> O 50/50
H9630-01	n.a.	800 ng of 1,3,5-Trimethoxybenzene (C <sub>6</sub> H <sub>3</sub> (OCH <sub>3</sub> ) <sub>3</sub> ) in 60 ul Acetonitrile/D <sub>2</sub> O 50/50
H9630-02	n.a.	800 ng of 1,3,5-Trimethoxybenzene (C <sub>6</sub> H <sub>3</sub> (OCH <sub>3</sub> ) <sub>3</sub> ) in 120 ul Acetonitrile/D <sub>2</sub> O 50/50
H9630-06	n.a.	800 ng of 1,3,5-Trimethoxybenzene (C <sub>6</sub> H <sub>3</sub> (OCH <sub>3</sub> ) <sub>3</sub> ) in 10 ul Acetonitrile/D <sub>2</sub> O 50/50
MeOD	n.a.	Methanol-D <sub>4</sub>
Z10029	3.0	0.3% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z10030	3.0	1% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z10031	3.0	3% Chloroform (CHCl <sub>3</sub> ), 0.2% TMS in Acetone-D <sub>6</sub>
Z10033	3.0	0.1% Ethylbenzene (EB) in Chloroform-D
Z10034	3.0	10% Ethylbenzene (EB) in Chloroform-D
Z10035	3.0	40% Dioxane in Benzene-D <sub>6</sub> (ASTM)
Z10036	3.0	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z10037	3.0	100 mM Urea- <sup>15</sup> N ([ <sup>15</sup> NH <sub>2</sub> ] <sub>2</sub> CO), 100 mM Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH) in Dimethyl Sulfoxide-D <sub>6</sub>



Sample ID	Diameter	Description
Z100372	2.5	0.5 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z100375	3.0	0.25 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z100376	3.0	0.5 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z100377	3.0	1 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z10038	3.0	0.0485 M Triphenyl Phosphate (TPP, [C <sub>6</sub> H <sub>5</sub> ] <sub>3</sub> PO <sub>4</sub> ) in Acetone-D <sub>6</sub>
Z100384	5.0	0.05 M Sodium Chloride (NaCl), 2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z100385	5.0	0.1 M Sodium Chloride (NaCl), 2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z100386	5.0	0.15 M Sodium Chloride (NaCl), 2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z100387	5.0	0.2 M Sodium Chloride (NaCl), 2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z100388	5.0	0.25 M Sodium Chloride (NaCl), 2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z100389	5.0	0.5 M Sodium Chloride (NaCl), 2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z10039	3.0	90% Formamide (HCONH <sub>2</sub> ) in Dimethyl Sulfoxide-D <sub>6</sub>
Z10040	3.0	0.05% Trifluorotoluene (TFT, a,a,a-CF <sub>3</sub> C <sub>6</sub> H <sub>5</sub> ) in Chloroform-D
Z10046	3.0	0.1 mg/ml Gadolinium Chloride (GdCl <sub>3</sub> ), 0.1% Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH), 1% H <sub>2</sub> O in D <sub>2</sub> O
Z10053	3.0	Temperature Calibration - 99.8% Methanol-D <sub>4</sub>
Z10056	3.0	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 99% H <sub>2</sub> O + 1% D <sub>2</sub> O
Z10075	5.0	1 M Potassium Chloride (KCl) in D <sub>2</sub> O
Z10076	10.0	1 M Potassium Chloride (KCl) in D <sub>2</sub> O
Z10078	5.0	45% Chloroform-D (CDCl <sub>3</sub> ), 45% Chloroform (CHCl <sub>3</sub> ) in 10% Hexafluorobenzene (C <sub>6</sub> F <sub>6</sub> )
Z10079	10.0	45% Chloroform-D (CDCl <sub>3</sub> ), 45% Chloroform (CHCl <sub>3</sub> ) in 10% Hexafluorobenzene (C <sub>6</sub> F <sub>6</sub> )
Z10082	2.5	0.1 mg/ml Gadolinium Chloride (GdCl <sub>3</sub> ), 0.1% Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH), 1% H <sub>2</sub> O in D <sub>2</sub> O
Z10083	5.0	0.1 mg/ml Gadolinium Chloride (GdCl <sub>3</sub> ), 0.1% Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH), 1% H <sub>2</sub> O in D <sub>2</sub> O
Z10084	8.0	0.1 mg/ml Gadolinium Chloride (GdCl <sub>3</sub> ), 0.1% Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH), 1% H <sub>2</sub> O in D <sub>2</sub> O
Z10085	10.0	0.1 mg/ml Gadolinium Chloride (GdCl <sub>3</sub> ), 0.1% Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH), 1% H <sub>2</sub> O in D <sub>2</sub> O
Z100926	1.0	3% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z100927	1.0	0.1% Ethylbenzene (EB) in Chloroform-D
Z100929	1.0	40% Dioxane in Benzene-D <sub>6</sub> (ASTM)
Z100930	1.0	10 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O

Sample ID	Diameter	Description
Z100932	1.0	100 mM Urea- <sup>15</sup> N ([ <sup>15</sup> NH <sub>2</sub> ] <sub>2</sub> CO), 100 mM Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH) in Dimethyl Sulfoxide-D <sub>6</sub>
Z100933	1.0	0.1 mg/ml Gadolinium Chloride (GdCl <sub>3</sub> ), 0.1% Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH), 1% H <sub>2</sub> O in D <sub>2</sub> O
Z100934	1.0	0.485 M Triphenyl Phosphate (TPP, [C <sub>6</sub> H <sub>5</sub> ] <sub>3</sub> PO <sub>4</sub> ) in Acetone-D <sub>6</sub>
Z100937	1.0	0.05% Trifluorotoluene (TFT, a,a,a-CF <sub>3</sub> C <sub>6</sub> H <sub>5</sub> ) in Chloroform-D
Z10120	5.0	0.1% Ethylbenzene (EB) in Chloroform-D
Z10121	10.0	0.1% Ethylbenzene (EB) in Chloroform-D
Z10136	2.5	90% Formamide (HCONH <sub>2</sub> ) in Dimethyl Sulfoxide-D <sub>6</sub>
Z10153	5.0	10% Ethylbenzene (EB) in Chloroform-D
Z10154	10.0	10% Ethylbenzene (EB) in Chloroform-D
Z10163	5.0	40% Dioxane in Benzene-D <sub>6</sub> (ASTM)
Z10164	10.0	40% Dioxane in Benzene-D <sub>6</sub> (ASTM)
Z101710	8.0	0.25 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z101712	10.0	0.5 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z101714	10.0	1 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z101715	1.0	0.25 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z101716	1.0	0.5 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z101717	1.0	1 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z10187	5.0	90% Formamide (HCONH <sub>2</sub> ) in Dimethyl Sulfoxide-D <sub>6</sub>
Z10188	10.0	90% Formamide (HCONH <sub>2</sub> ) in Dimethyl Sulfoxide-D <sub>6</sub>
Z10191	5.0	1 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z10192	2.5	0.25 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z10201	5.0	0.0485 M Triphenyl Phosphate (TPP, [C <sub>6</sub> H <sub>5</sub> ] <sub>3</sub> PO <sub>4</sub> ) in Acetone-D <sub>6</sub>
Z10202	10.0	0.0485 M Triphenyl Phosphate (TPP, [C <sub>6</sub> H <sub>5</sub> ] <sub>3</sub> PO <sub>4</sub> ) in Acetone-D <sub>6</sub>
Z10209	5.0	85% Hexamethyldisiloxane (HMDSO, [[CH <sub>3</sub> ] <sub>3</sub> Si] <sub>2</sub> O) in Benzene-D <sub>6</sub>
Z10210	10.0	85% Hexamethyldisiloxane (HMDSO, [[CH <sub>3</sub> ] <sub>3</sub> Si] <sub>2</sub> O) in Benzene-D <sub>6</sub>
Z10230	5.0	3% Chloroform (CHCl <sub>3</sub> ), 0.2% TMS in Acetone-D <sub>6</sub>
Z10234	5.0	0.05% Trifluorotoluene (TFT, a,a,a-CF <sub>3</sub> C <sub>6</sub> H <sub>5</sub> ) in Chloroform-D
Z10235	10.0	0.05% Trifluorotoluene (TFT, a,a,a-CF <sub>3</sub> C <sub>6</sub> H <sub>5</sub> ) in Chloroform-D
Z10241	5.0	2 mM Lysozyme in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z10242	2.5	40% Dioxane in Benzene-D <sub>6</sub> (ASTM)
Z10246	5.0	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O (60 mm filling height)
Z10247	8.0	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z10248	5.0	1% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z10249	8.0	1% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z10250	10.0	1% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z10253	8.0	10% Ethylbenzene (EB) in Chloroform-D
Z10255	8.0	40% Dioxane in Benzene-D <sub>6</sub> (ASTM)

Sample ID	Diameter	Description
Z10256	8.0	90% Formamide (HCONH <sub>2</sub> ) in Dimethyl Sulfoxide-D <sub>6</sub>
Z10257	8.0	0.0485 M Triphenyl Phosphate (TPP, [C <sub>6</sub> H <sub>5</sub> ] <sub>3</sub> PO <sub>4</sub> ) in Acetone-D <sub>6</sub>
Z10260	8.0	3% Chloroform (CHCl <sub>3</sub> ), 0.2% TMS in Acetone-D <sub>6</sub>
Z10263	5.0	100 mM Urea- <sup>15</sup> N ([ <sup>15</sup> NH <sub>2</sub> ] <sub>2</sub> CO), 100 mM Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH) in Dimethyl Sulfoxide-D <sub>6</sub>
Z10264	8.0	100 mM Urea- <sup>15</sup> N ([ <sup>15</sup> NH <sub>2</sub> ] <sub>2</sub> CO), 100 mM Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH) in Dimethyl Sulfoxide-D <sub>6</sub>
Z10265	10.0	100 mM Urea- <sup>15</sup> N ([ <sup>15</sup> NH <sub>2</sub> ] <sub>2</sub> CO), 100 mM Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH) in Dimethyl Sulfoxide-D <sub>6</sub>
Z10267	2.5	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z10268	10.0	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z10270	2.5	0.1% Ethylbenzene (EB) in Chloroform-D
Z10272	2.5	1% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z10274	2.5	100 mM Urea- <sup>15</sup> N ([ <sup>15</sup> NH <sub>2</sub> ] <sub>2</sub> CO), 100 mM Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH) in Dimethyl Sulfoxide-D <sub>6</sub>
Z10275	2.5	3% Chloroform (CHCl <sub>3</sub> ), 0.2% TMS in Acetone-D <sub>6</sub>
Z10276	2.5	0.0485 M Triphenyl Phosphate (TPP, [C <sub>6</sub> H <sub>5</sub> ] <sub>3</sub> PO <sub>4</sub> ) in Acetone-D <sub>6</sub>
Z10284	5.0	0.25 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z10285	10.0	0.25 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z10288	5.0	0.5 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z10292	2.5	10% Ethylbenzene (EB) in Chloroform-D
Z10610	10.0	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 99% H <sub>2</sub> O + 1% D <sub>2</sub> O
Z10627	5.0	Temperature Calibration - 99.8% Methanol-D <sub>4</sub>
Z10628	10.0	Temperature Calibration - 99.8% Methanol-D <sub>4</sub>
Z10650	5.0	Diffusion Dry Glycerol
Z10688	5.0	5% H <sub>2</sub> O, 0.6 mM CuSO <sub>4</sub> in D <sub>2</sub> O
Z10689	5.0	20% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z10690	5.0	1% Ethylbenzene (EB) in Chloroform-D
Z10692	5.0	1 M Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH) in D <sub>2</sub> O
Z10694	5.0	100% Ethylbenzene (EB)
Z10701	5.0	0.3% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z10702	10.0	0.3% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z107150	5.0	0.15 M Sodium Chloride (NaCl), 2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O (shaped tube)
Z107151	5.0	0.25 M Sodium Chloride (NaCl), 2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O (shaped tube)
Z107152	5.0	0.5 M Sodium Chloride (NaCl), 2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O (shaped tube)
Z10717	1.7	1% Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub>
Z10718	1.7	0.1% Ethylbenzene (EB) in Chloroform-D
Z10719	1.7	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O

Sample ID	Diameter	Description
Z10720	1.7	10 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O
Z10721	1.7	100 mM Urea- <sup>15</sup> N ([ <sup>15</sup> NH <sub>2</sub> ] <sub>2</sub> CO), 100 mM Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH) in Dimethyl Sulfoxide-D <sub>6</sub>
Z10722	1.7	0.0485 M Triphenyl Phosphate (TPP, [C <sub>6</sub> H <sub>5</sub> ] <sub>3</sub> PO <sub>4</sub> ) in Acetone-D <sub>6</sub>
Z10723	1.7	10% Ethylbenzene (EB) in Chloroform-D
Z10724	1.7	40% Dioxane in Benzene-D <sub>6</sub> (ASTM)
Z10727	1.7	0.1 mg/ml Gadolinium Chloride (GdCl <sub>3</sub> ), 0.1% Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH), 1% H <sub>2</sub> O in D <sub>2</sub> O
Z10728	1.7	0.05% Trifluorotoluene (TFT, a,a,a-CF <sub>3</sub> C <sub>6</sub> H <sub>5</sub> ) in Chloroform-D
Z10729	1.7	0.25 M Sodium Chloride (NaCl) in Deuterium Oxide (D <sub>2</sub> O)
Z10730	1.7	0.5 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z10731	1.7	1 M Sodium Chloride (NaCl) in D <sub>2</sub> O
Z10734	1.7	Temperature Calibration - 99.8% Methanol-D <sub>4</sub>
Z10902	5.0	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O (40 mm filling height)
Z10908	5.0	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 99% H <sub>2</sub> O + 1% D <sub>2</sub> O
Z142220	4.0	Chloroform (CHCl <sub>3</sub> ) in Acetone-D <sub>6</sub> (50 ul)
Z142221	4.0	0.1% Ethylbenzene (EB) in Chloroform-D (50 ul)
Z142222	4.0	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O (50 ul)
Z142223	4.0	100 mM Urea- <sup>15</sup> N ([ <sup>15</sup> NH <sub>2</sub> ] <sub>2</sub> CO), 100 mM Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH) in Dimethyl Sulfoxide-D <sub>6</sub> (50 ul)
Z142224	4.0	40% Dioxane in Benzene-D <sub>6</sub> (ASTM, 50 ul)
Z142226	4.0	0.0485 M Triphenyl Phosphate (TPP, [C <sub>6</sub> H <sub>5</sub> ] <sub>3</sub> PO <sub>4</sub> ) in Acetone-D <sub>6</sub> (50 ul)
Z142227	4.0	90% Formamide (HCONH <sub>2</sub> ) in Dimethyl Sulfoxide-D <sub>6</sub> (50 ul)
Z142228	4.0	0.05% Trifluorotoluene (TFT, a,a,a-CF <sub>3</sub> C <sub>6</sub> H <sub>5</sub> ) in Chloroform-D
Z142229	4.0	85% Hexamethyldisiloxane (HMDSO, [[CH <sub>3</sub> ] <sub>3</sub> Si] <sub>2</sub> O) in Benzene-D <sub>6</sub> (50 ul)
Z142231	4.0	0.1 mg/ml Gadolinium Chloride (GdCl <sub>3</sub> ), 0.1% Methanol- <sup>13</sup> C ( <sup>13</sup> CH <sub>3</sub> OH), 1% H <sub>2</sub> O in D <sub>2</sub> O (50 ul)
Z151210	7.0	Potassium Bromide (KBr, 234 ul)
Z151211	7.0	Adamantane (234 ul)
Z151212	7.0	Alpha-glycine (234 ul)
Z151213	7.0	2- <sup>13</sup> C, <sup>15</sup> N alpha-glycine (234 ul)
Z151214	7.0	Ammonium Dihydrogenphosphate (NH <sub>4</sub> H <sub>2</sub> PO <sub>4</sub> , 80 ul)
Z151216	7.0	Ammonium Trifluoroacetate (CF <sub>3</sub> CO <sub>2</sub> NH <sub>4</sub> , 80 ul)
Z151220	4.0	Potassium Bromide (KBr, 80 ul)
Z151221	4.0	Adamantane (50 ul)
Z151222	4.0	Alpha-glycine (50 ul)
Z151223	4.0	2- <sup>13</sup> C, <sup>15</sup> N alpha-glycine (50 ul)

Sample ID	Diameter	Description
Z151224	4.0	Ammonium Dihydrogenphosphate (NH <sub>4</sub> H <sub>2</sub> PO <sub>4</sub> , 50 ul)
Z151226	4.0	Ammonium Trifluoroacetate (CF <sub>3</sub> CO <sub>2</sub> NH <sub>4</sub> , 50 ul)
Z151230	3.2	Potassium Bromide (KBr, 34 ul)
Z151231	3.2	Adamantane (34 ul)
Z151232	3.2	Alpha-glycine (34 ul)
Z151233	3.2	2- <sup>13</sup> C, <sup>15</sup> N alpha-glycine (34 ul)
Z151234	3.2	Ammonium Dihydrogenphosphate (NH <sub>4</sub> H <sub>2</sub> PO <sub>4</sub> , 34 ul)
Z151236	3.2	Ammonium Trifluoroacetate (CF <sub>3</sub> CO <sub>2</sub> NH <sub>4</sub> , 34 ul)
Z151240	3.2	Thinwalled Potassium Bromide (KBr, 45.5 ul)
Z151241	3.2	Thinwalled Adamantane (45.5 ul)
Z151242	3.2	Thinwalled Alpha-glycine (45.5 ul)
Z151243	3.2	Thinwalled 2- <sup>13</sup> C, <sup>15</sup> N alpha-glycine (45.5 ul)
Z151244	3.2	Thinwalled Ammonium Dihydrogenphosphate (NH <sub>4</sub> H <sub>2</sub> PO <sub>4</sub> , 45.5 ul)
Z151246	3.2	Thinwalled Ammonium Trifluoroacetate (CF <sub>3</sub> CO <sub>2</sub> NH <sub>4</sub> , 45.5 ul)
Z151250	2.5	Potassium Bromide (KBr, 13.6 ul)
Z151251	2.5	Adamantane (13.6 ul)
Z151252	2.5	Alpha-glycine (12 mg, 13.6 ul)
Z151253	2.5	2- <sup>13</sup> C, <sup>15</sup> N alpha-glycine (12 mg, 13.6 ul)
Z151254	2.5	Ammonium Dihydrogenphosphate (NH <sub>4</sub> H <sub>2</sub> PO <sub>4</sub> , 13.6 ul)
Z151256	2.5	Ammonium Trifluoroacetate (CF <sub>3</sub> CO <sub>2</sub> NH <sub>4</sub> , 13.6 ul)
Z151260	1.9	Potassium Bromide (KBr, 13.1 ul)
Z151261	1.9	Adamantane (13.1 ul)
Z151262	1.9	Alpha-glycine (10 mg, 13.1 ul)
Z151263	1.9	2- <sup>13</sup> C, <sup>15</sup> N alpha-glycine (10 mg, 13.1 ul)
Z151264	1.9	Ammonium Dihydrogenphosphate (NH <sub>4</sub> H <sub>2</sub> PO <sub>4</sub> , 13.1 ul)
Z151266	1.9	Ammonium Trifluoroacetate (CF <sub>3</sub> CO <sub>2</sub> NH <sub>4</sub> , 13.1 ul)
Z151270	1.3	Potassium Bromide (KBr, 3.0 ul)
Z151271	1.3	Adamantane (3.0 ul)
Z151272	1.3	Alpha-glycine (2 mg, 3.0 ul)
Z151273	1.3	2- <sup>13</sup> C, <sup>15</sup> N alpha-glycine (2 mg, 3.0 ul)
Z151274	1.3	Ammonium Dihydrogenphosphate (NH <sub>4</sub> H <sub>2</sub> PO <sub>4</sub> , 3.0 ul)
Z151276	1.3	Ammonium Trifluoroacetate (CF <sub>3</sub> CO <sub>2</sub> NH <sub>4</sub> , 3.0 ul)
Z163271	0.7	Potassium Bromide (KBr, 0.5 ul)
Z163274	0.7	Adamantane (0.5 ul)
Z163275	0.7	Alpha-glycine 0.5 ul)
Z163276	0.7	2- <sup>13</sup> C, <sup>15</sup> N alpha-glycine (0.5 ul)
Z163277	0.7	Ammonium Dihydrogenphosphate (NH <sub>4</sub> H <sub>2</sub> PO <sub>4</sub> , 0.5 ul)
Z163278	0.7	Ammonium Trifluoroacetate (CF <sub>3</sub> CO <sub>2</sub> NH <sub>4</sub> , 0.5 ul)
Z180181	5.0	2 mM Sucrose, 0.5 mM DSS, 2 mM NaN <sub>3</sub> in 90% H <sub>2</sub> O + 10% D <sub>2</sub> O (40 mm filling height, AvCo Sample Autocal)

Sample ID	Diameter	Description
Z183103	3.2	Potassium Bromide
Z183104	3.2	Adamantane
Z183105	3.2	Alpha-glycine
Z183106	3.2	2- <sup>13</sup> C, <sup>15</sup> N alpha-glycine



## 7 Contact

### 7.1 Manufacturer

---

Bruker BioSpin GmbH  
Silberstreifen 4  
D-76287 Rheinstetten  
Germany

*<http://www.bruker.com>*

WEEE DE43181702

### 7.2 NMR Hotlines

---

Contact our NMR service centers.

Bruker BioSpin NMR provides dedicated hotlines and service centers, so that our specialists can respond as quickly as possible to all your service requests, applications questions, software or technical needs.



Please select the NMR service center or hotline you wish to contact from our list available at: *<https://www.bruker.com/service/information-communication/helpdesk.html>*

Phone: +49 721-5161-6155

E-mail: *[nmr-support@bruker.com](mailto:nmr-support@bruker.com)*

---





---

● **Bruker Corporation**

[info@bruker.com](mailto:info@bruker.com)  
[www.bruker.com](http://www.bruker.com)